

## E5071C OPT TDR Network Analyzer Firmware Revision History

Rev. 20150323

---

**Note:**

The purpose of this document is to provide an overview of important changes that could affect a majority of customers. If you have any concerns about a specific issue, contact your local Agilent representative.

---

**Firmware for Windows 7**

*Revision B.xx.xx can be installed onto the E5071C with Windows 7 only.*

**Revision B.02.02** ... Released October 2014

1. Rebranding to Keysight

**Revision B.02.00** ... Released January 2014

2. Supports the E5071C with Windows 7.

**Firmware for Windows XP**

*Revision A.xx.xx can be installed onto the E5071C with Windows XP only.*

**Revision A.01.57** ... Released January 2014

1. Added the function to avoid this firmware from being installed onto the E5071C with Windows 7.

**Revision A.01.56** ... Released June 2012

1. Fixed Minor Bug..

**Revision A.01.53** ... Released January 2012

1. Fixed Minor Bug.

**Revision A.01.52** ... Released December 2011

1. Supports Windows 7.

**Revision A.01.50** ... Released September 2011

1. Added Advanced Waveform Analysis.
2. Added Hot TDR option.
3. Added manual scale for Eye Diagram.
4. Added Cal status indicator.

**Revision A.01.03** ... Released August 2011

1. Fixed problem on eye diagram when setting negative value for zero level.
2. Fixed minor bugs.

**Revision A.01.02** ... Released July 2011

1. Height (dB) is deleted from the result.
2. TDR Mask Test - It is required to define the values of the Logic 1 Lv and Logic 0 Lv in the scale tab for Mask.