

External Oscilloscope Time Correlation and Data Display

[Online Help](#)

Notices

© Keysight Technologies 2013-2014, 2015

No part of this manual may be reproduced in any form or by any means (including electronic storage and retrieval or translation into a foreign language) without prior agreement and written consent from Keysight Technologies as governed by United States and international copyright laws.

Trademarks

Adobe®, Acrobat®, and the Acrobat Logo® are trademarks of Adobe Systems Incorporated

Manual Part Number

Version 06.10.0000

Edition

May 2015

Available in electronic format only

Keysight Technologies
1900 Garden of the Gods Road
Colorado Springs, CO 80907 USA

Warranty

THE MATERIAL CONTAINED IN THIS DOCUMENT IS PROVIDED "AS IS," AND IS SUBJECT TO BEING CHANGED, WITHOUT NOTICE, IN FUTURE EDITIONS. FURTHER, TO THE MAXIMUM EXTENT PERMITTED BY APPLICABLE LAW, KEYSIGHT DISCLAIMS ALL WARRANTIES, EITHER EXPRESS OR IMPLIED WITH REGARD TO THIS MANUAL AND ANY INFORMATION CONTAINED HEREIN, INCLUDING BUT NOT LIMITED TO THE IMPLIED WARRANTIES OF MERCHANTABILITY AND FITNESS FOR A PARTICULAR PURPOSE. KEYSIGHT SHALL NOT BE LIABLE FOR ERRORS OR FOR INCIDENTAL OR CONSEQUENTIAL DAMAGES IN CONNECTION WITH THE FURNISHING, USE, OR PERFORMANCE OF THIS DOCUMENT OR ANY INFORMATION CONTAINED HEREIN. SHOULD KEYSIGHT AND THE USER HAVE A SEPARATE WRITTEN AGREEMENT WITH WARRANTY TERMS COVERING THE MATERIAL IN THIS DOCUMENT THAT CONFLICT WITH THESE

TERMS, THE WARRANTY TERMS IN THE SEPARATE AGREEMENT WILL CONTROL.

Technology Licenses

The hardware and/or software described in this document are furnished under a license and may be used or copied only in accordance with the terms of such license.

U.S. Government Rights

The Software is "commercial computer software," as defined by Federal Acquisition Regulation ("FAR") 2.101. Pursuant to FAR 12.212 and 27.405-3 and Department of Defense FAR Supplement ("DFARS") 227.7202, the U.S. government acquires commercial computer software under the same terms by which the software is customarily provided to the public. Accordingly, Keysight provides the Software to U.S. government customers under its standard commercial license, which is embodied in its End User License Agreement (EULA), a copy of which can be found at <http://www.keysight.com/find/sweula>. The license set forth in the EULA represents the exclusive authority by which the U.S. government may use, modify, distribute, or disclose the Software. The EULA and the license set forth therein, does not require or permit, among other things, that Keysight: (1) Furnish technical information related to commercial computer software or commercial computer software documentation that is not customarily provided to the public; or (2) Relinquish to, or otherwise provide, the government rights in excess of these rights customarily provided to the public to use, modify, reproduce, release, perform, display, or disclose commercial computer software or commercial computer software documentation. No additional government requirements beyond those set forth in the EULA shall apply, except to the extent that those terms, rights, or licenses are explicitly required from all providers of commercial computer software pursuant to the FAR and the DFARS and are set forth specifically in writing elsewhere in the EULA. Keysight shall be under no obligation to update, revise or otherwise modify the Software. With respect to any technical data as defined by FAR 2.101, pursuant to FAR 12.211 and 27.404.2 and DFARS 227.7102, the U.S. government acquires no

greater than Limited Rights as defined in FAR 27.401 or DFAR 227.7103-5 (c), as applicable in any technical data.

Safety Notices

CAUTION

A CAUTION notice denotes a hazard. It calls attention to an operating procedure, practice, or the like that, if not correctly performed or adhered to, could result in damage to the product or loss of important data. Do not proceed beyond a CAUTION notice until the indicated conditions are fully understood and met.

WARNING

A WARNING notice denotes a hazard. It calls attention to an operating procedure, practice, or the like that, if not correctly performed or adhered to, could result in personal injury or death. Do not proceed beyond a WARNING notice until the indicated conditions are fully understood and met.

Safety Summary

The following general safety precautions must be observed during all phases of operation of this instrument. Failure to comply with these precautions or with specific warnings or operating instructions in the product manuals violates safety standards of design, manufacture, and intended use of the instrument. Keysight Technologies assumes no liability for the customer's failure to comply with these requirements. Product manuals are provided with your instrument on CD-ROM and/or in printed form. Printed manuals are an option for many products. Manuals may also be available on the Web. Go to www.keysight.com and type in your product number in the Search field at the top of the page.

General	Do not use this product in any manner not specified by the manufacturer. The protective features of this product may be impaired if it is used in a manner not specified in the operation instructions.
Before Applying Power	Verify that all safety precautions are taken. Make all connections to the unit before applying power. Note the instrument's external markings described in "Safety Symbols".
Ground the Instrument	If your product is provided with a grounding type power plug, the instrument chassis and cover must be connected to an electrical ground to minimize shock hazard. The ground pin must be firmly connected to an electrical ground (safety ground) terminal at the power outlet. Any interruption of the protective (grounding) conductor or disconnection of the protective earth terminal will cause a potential shock hazard that could result in personal injury.
Fuses	See the user's guide or operator's manual for information about line-fuse replacement. Some instruments contain an internal fuse, which is not user accessible.
Do Not Operate in an Explosive Atmosphere	Do not operate the instrument in the presence of flammable gases or fumes.
Do Not Remove the Instrument Cover	Only qualified, service-trained personnel who are aware of the hazards involved should remove instrument covers. Always disconnect the power cable and any external circuits before removing the instrument cover.
Cleaning	Clean the outside of the instrument with a soft, lint-free, slightly dampened cloth. Do not use detergent or chemical solvents.
Do Not Modify the Instrument	Do not install substitute parts or perform any unauthorized modification to the product. Return the product to an Keysight Sales and Service Office for service and repair to ensure that safety features are maintained.
In Case of Damage	Instruments that appear damaged or defective should be made inoperative and secured against unintended operation until they can be repaired by qualified service personnel.

CAUTION

A CAUTION notice denotes a hazard. It calls attention to an operating procedure, practice, or the like that, if not correctly performed or adhered to, could result in damage to the product or loss of important data. Do not proceed beyond a CAUTION notice until the indicated conditions are fully understood and met.

WARNING

A WARNING notice denotes a hazard. It calls attention to an operating procedure, practice, or the like that, if not correctly performed or adhered to, could result in personal injury or death. Do not proceed beyond a WARNING notice until the indicated conditions are fully understood and met.

Safety Symbols

Table 1 Safety Symbol

Symbol	Description
	Direct current
	Alternating current
	Both direct and alternating current
	Three phase alternating current
	Three phase alternating current
	Earth ground terminal
	Protective earth ground terminal
	Frame or chassis ground terminal
	Terminal is at earth potential
	Equipotentiality
N	Neutral conductor on permanently installed equipment
L	Line conductor on permanently installed equipment
	On (mains supply)
	Off (mains supply)
	Stand by (mains supply). The instrument is not completely disconnected from the mains supply when the power switch is in the stand by position
	In position of a bi-stable push switch

Symbol	Description
	Out position of a bi-stable push switch
	Equipment protected throughout by DOUBLE INSULATION or REINFORCED INSULATION
	Caution, refer to accompanying documentation
	Caution, risk of electric shock
	Do not apply around or remove from HAZARDOUS LIVE conductors
	Application around and removal from HAZARDOUS LIVE conductors is permitted
	Caution, hot surface
	Ionizing radiation
CAT I	IEC Measurement Category I
CAT II	Measurement Category II
CAT III	Measurement Category III
CAT IV	Measurement Category IV

Compliance and Environmental Information

Table 2 Compliance and Environmental Information

Safety Symbol	Description
	CSA is the Canadian certification mark to demonstrate compliance with the Safety requirements.
	The C-tick mark is a registered trademark of the Spectrum Management Agency of Australia. This signifies compliance with the Australia EMC Framework regulations under the terms of the Radio Communication Act of 1992.
	CE compliance marking to the EU Safety and EMC Directives. ISM GRP-1A classification according to the international EMC standard. ICES/NMB-001 compliance marking to the Canadian EMC standard.

External Oscilloscope Time Correlation and Data Display—At a Glance

The Add External Oscilloscope wizard automates and simplifies the process of setting up an external oscilloscope and a logic analyzer / a protocol analyzer to cross-trigger and make time-correlated measurements (using markers). You can also display data captured by the oscilloscope in the Logic and Protocol Analyzer GUI's display windows. This external oscilloscope correlation and data display feature is sometimes called View Scope.

External oscilloscope correlation works with the following oscilloscopes, set up for remote control via LAN:

- Keysight InfiniiVision 2000 X-Series DSO and MSO oscilloscopes
- Keysight InfiniiVision 3000 X-Series DSO and MSO oscilloscopes
- Keysight InfiniiVision 4000 X-Series DSO and MSO oscilloscopes (New in 5.51 Release)
- Keysight InfiniiVision 5000 Series oscilloscopes
- Keysight InfiniiVision 6000 Series DSO and MSO oscilloscopes
- Keysight InfiniiVision 7000 Series DSO and MSO oscilloscopes
- Keysight Infiniium 548xx Series oscilloscopes
- Keysight Infiniium 8000 Series DSO and MSO oscilloscopes
- Keysight Infiniium 80000 Series oscilloscopes
- Keysight Infiniium 9000 Series DSO and MSO oscilloscopes
- Keysight Infiniium 90000 A-Series oscilloscopes
- Keysight Infiniium 90000 X-Series oscilloscopes
- Keysight Infiniium 90000 Q-Series oscilloscopes
- Keysight Infiniium 90000 X-Series MSO oscilloscopes

The Add External Oscilloscope wizard can provide default (typical) deskew values for the type of logic analyzer and oscilloscope being used. You can also use the E5850A time correlation fixture to run a calibration procedure that provides the most accurate deskew values for the actual instruments being used. For more information about obtaining the E5850A time correlation fixture, contact your nearest Keysight Technologies Sales Office.

TIP

The instructions on the E5850A time correlation fixture box are for 16700-series logic analysis systems only. For 16900-series logic analysis systems or 1680/90-series logic analyzers, follow the instructions in this online help.

In This Book

- Getting Started with External Oscilloscope Time Correlation and Data Display (see [page 13](#))
- Step 1. Network the oscilloscope and logic analyzer (see [page 14](#))
- Step 2. Run the Add External Oscilloscope wizard (see [page 16](#))
- Step 3. Make a measurement (see [page 19](#))
- Working with Markers (see [page 21](#))
- Displaying Imported Oscilloscope Data (see [page 23](#))
- Controlling the External Oscilloscope via the Web (see [page 27](#))
- Loading and Saving Configurations (see [page 31](#))
- Changing External Oscilloscope Setup Options (see [page 33](#))
- Deleting the External Oscilloscope (see [page 35](#))
- Troubleshooting (see [page 37](#))
- Finding the External Oscilloscope IP Address (see [page 38](#))
- LAN Problems (see [page 39](#))

- [General Triggering Problems \(see page 40\)](#)
- [Unexpected or Inconsistent Results \(see page 41\)](#)
- [Correlation Drift Over Long Acquisitions \(see page 42\)](#)
- [Reference \(see page 43\)](#)
- [External Oscilloscope Setup Dialog \(see page 44\)](#)
- [Concepts \(see page 49\)](#)
- [Trigger Deskew \(see page 50\)](#)
- [How the Calibration Procedure Works \(see page 51\)](#)
- [Uncertainty \(see page 53\)](#)
- [What Do the Skew Numbers Mean? \(see page 54\)](#)
- [External Oscilloscope Module Control, COM Automation \(see page 55\)](#)
- [External Oscilloscope Module Setup, XML Format \(see page 57\)](#)

Contents

Safety Summary	3
Compliance and Environmental Information	6
External Oscilloscope Time Correlation and Data Display—At a Glance	7
1 Getting Started with External Oscilloscope Time Correlation and Data Display	
Step 1. Network the oscilloscope and logic analyzer	14
Using a Point-to-Point Network	14
Step 2. Run the Add External Oscilloscope wizard	16
Additional Connection Information	16
Step 3. Make a measurement	19
2 Working with Markers	
3 Displaying Imported Oscilloscope Data	
4 Controlling the External Oscilloscope via the Web	
5 Loading and Saving Configurations	
6 Changing External Oscilloscope Setup Options	
7 Deleting the External Oscilloscope	
8 External Oscilloscope Troubleshooting	
Finding the External Oscilloscope IP Address	38
LAN Problems	39
General Triggering Problems	40
Unexpected or Inconsistent Results	41
Correlation Drift Over Long Acquisitions	42

9 External Oscilloscope Reference

External Oscilloscope Setup Dialog	44
Channels Tab	44
Trigger Tab	45
Connection Tab	46

10 External Oscilloscope Concepts

Trigger Deskew	50
How the Calibration Procedure Works	51
Uncertainty	53
What Do the Skew Numbers Mean?	54

11 External Oscilloscope Module Control, COM Automation

12 External Oscilloscope Module Setup, XML Format

<AnalogDataSource> Element	58
<BusSignal> Element	59
<BusSignals> Element	60
<BusSignalSetup> Element	61
<Channel1/2/3/4> Element	62
<Channel1/2/3/4DataTable> Element	63
<Channels> Element	64
<ChannelsSetup> Element	65
<ConnectionSetup> Element	66
<Data> Element	67
<DigitalChannels> Element	69
<DigitalChannelsDataTable> Element	70
<DigitalDataSource> Element	71
<Folder> Element	72
<Markers> Element	73
<Name> Element	74
<Names> Element	75
<Pod1/2> Element	76
<Programming> Element	77

<SampleNumberDataSource> Element	79
<SetupInfo> Element (for AnalogDataSource)	80
<SetupInfo> Element (for DigitalDataSource)	81
<SetupInfo> Element (for SampleNumberDataSource)	82
<SetupInfo> Element (for TimeDataSource)	83
<TimeDataSource> Element	84
<TriggerSetup> Element	85
<WaveformData> Element	86

Index

1 Getting Started with External Oscilloscope Time Correlation and Data Display

Software for external oscilloscope time correlation and data display is built-in to the *Keysight Logic Analyzer* application.

After the external oscilloscope and logic analyzer are networked, the Add External Oscilloscope wizard takes you through the process of connecting the logic analyzer to the oscilloscope. Once that is done, you are ready to make a measurement.

- Step 1. Network the oscilloscope and logic analyzer (see [page 14](#))
- Step 2. Run the Add External Oscilloscope wizard (see [page 16](#))
- Step 3. Make a measurement (see [page 19](#))

Step 1. Network the oscilloscope and logic analyzer

NOTE

If your logic analyzer and external oscilloscope are both already connected to a LAN, you can skip this step.

NOTE

When using a 6000-series oscilloscope, the LAN port must be selected as the I/O Controller. (Press the **Utility** button; then, press the **I/O** softkey; use the Entry knob to select **LAN**; then, press the **Controller** softkey.)

If your logic analyzer and external oscilloscope are not already connected to a LAN, a simple point-to-point connection may be the most convenient connection between the instruments (see Using a Point-to-Point Network (see [page 14](#))).

If you do not want to use a point-to-point connection, refer to the logic analyzer or external oscilloscope operating system instructions for setting up the instrument on the network.

Using a Point-to-Point Network

To set up a point-to-point connection between the logic analyzer and the oscilloscope, use the cross-over (blue) cable (provided with the E5850A time correlation fixture) to connect the instruments.

- 1 Connect one end of the cross-over cable to the network interface on the logic analyzer and the other end of the cable to the network interface on the oscilloscope.
- 2 Configure the networking software on the oscilloscope (see Oscilloscope Point-to-Point Network Set Up (see [page 14](#))).
- 3 Configure the networking software on the logic analyzer (see Logic Analyzer Point-to-Point Network Set Up (see [page 14](#))).


Oscilloscope Point-to-Point Network Set Up

- 1 Make sure networking is enabled on the oscilloscope.
- 2 Set the oscilloscope's IP Address to 192.0.2.231.
- 3 Set the Subnet Mask to 255.255.255.0.
- 4 If there is a WINS resolution option, disable it.
- 5 Remove any installed gateways.
- 6 If there is a DNS option, disable it.
- 7 On a 6000-series oscilloscope, select the LAN port as the I/O Controller:
 - a Press the Utility button.
 - b Press the I/O softkey.
 - c Use the Entry knob to select LAN.
 - d Press the Controller softkey.


See Also • Logic Analyzer Point-to-Point Network Set Up (see [page 14](#))

Logic Analyzer Point-to-Point Network Set Up

- 1 From the Start menu, select Setting>Network and Dial-up Connection.
- 2 Right-click on Local Area Connection and select Properties.
- 3 Ensure that the TCP/IP component is installed, then select Properties.



- 4 In the Properties dialog, set the IP Address and Subnet Mask.
 - a Set the IP Address to 192.0.2.232.
 - b Set the Subnet Mask to 255.255.255.0.




- 5 Select the Advanced tab.
 - a In the IP Settings tab, remove any installed gateways.
 - b In the DNS tab, remove all DNS addresses from the list.
 - c In the WINS tab, remove all WINS addresses from the list.
- 6 Close the dialogs and click Yes if prompted to continue.

- See Also**
- Oscilloscope Point-to-Point Network Set Up (see [page 14](#))
 - Step 2. Run the Add External Oscilloscope wizard (see [page 16](#))

Step 2. Run the Add External Oscilloscope wizard

For step-by-step instructions on setting up the connection between the logic analyzer and the oscilloscope, run the Add External Oscilloscope wizard:

- 1 From the menu bar, select Setup>Add External Scope....



- 2 Follow the instructions in the wizard.

- See Also**
- Additional Connection Information (see [page 16](#))
 - Deleting the External Oscilloscope (see [page 35](#))
 - Finding the External Oscilloscope IP Address (see [page 38](#))

Additional Connection Information

- Connecting the LAN (see [page 16](#))
- Connecting the BNC Cables (see [page 16](#))
- Connecting the Oscilloscope Probe to the Time Correlation Fixture (see [page 17](#))
- Connecting the Logic Analyzer to the Time Correlation Fixture (see [page 17](#))
- Connecting the Power Supply to the Time Correlation Fixture (see [page 17](#))

Connecting the LAN

An external oscilloscope module uses a network connection between the oscilloscope and the 1680/90-series logic analyzer or the 16900-series logic analysis system to produce time-correlated measurements and to import oscilloscope data. The instruments must be connected on a local area network (LAN) or connected point-to-point with a cross-over cable (which is supplied with the E5850A time correlation fixture).

NOTE

When using a 6000-series oscilloscope, the LAN port must be selected as the I/O Controller. (Press the **Utility** button; then, press the **I/O** softkey; use the Entry knob to select **LAN**; then, press the **Controller** softkey.)

You can connect only one oscilloscope to the logic analysis system for time-correlated measurements. The logic analysis system does not support the connection of two or more oscilloscopes at once.

Connecting the BNC Cables

BNC cable connections depend on which model external oscilloscope you have. For instructions for connecting BNC see the Add External Oscilloscope wizard.

Connecting the Oscilloscope Probe to the Time Correlation Fixture

NOTE

It is important to use the same probe for the calibration procedure as you use to make measurements on your device under test. This provides the best possible deskew accuracy. All connections must include a ground connection.

Connect the external oscilloscope channel 1 probe to any point labeled CAL on the front of the E5850A time correlation fixture. Any point labeled CAL is equivalent to any other point labeled CAL.

- If you are using a differential probe, you *must* connect the (+) probe terminal to the CAL terminal and the (-) probe terminal to the GND connection on the time correlation fixture.
- If you are using a grabber probe, connect the grabber probe to the CAL terminal on the front of the time correlation fixture. Then, connect the ground lead.
- If you are using a twin lead probe, connect the twin lead probe to the time correlation fixture. Then, connect the ground lead.
- If you are using a sharp-tip probe, connect the sharp tip probe to the CAL terminal on the front of the time correlation fixture.

Connecting the Logic Analyzer to the Time Correlation Fixture

NOTE

It is important to use the same probes for the calibration procedure as you use to make measurements in the device under test. This provides the best possible deskew accuracy. All connections must include a ground connection.

There are three ways to connect to the CAL signals on the time correlation fixture:

- Using flying leads.
- Using high-density connectors (MICTOR or Samtec).
- Using grabber probes.

Flying Leads (Socket Probe)

Connect pod 2, channel 0 of the logic analyzer you are using to any point labeled CAL on the front of the time correlation fixture. If you are using multiple card logic analyzers, use pod 2 of the master card. Include a ground connection.

High-Density Connectors

If you are using a MICTOR connector on your device under test, use the MICTOR receptacle on the top of the time correlation fixture, which includes the CAL connection. Be sure to connect pod 2 of the logic analyzer to the even receptacle of the MICTOR high-density adapter cable. You can use a terminated or an unterminated high-density adapter cable.

If you are using Samtec connectors on your device under test, use the Samtec receptacle on the top of the time correlation fixture, which includes the CAL connection.

Grabber Probe

If you are using a grabber probe on your device under test, connect the grabber probe to the terminal labeled CAL. Be sure to include the ground connection.

Connecting the Power Supply to the Time Correlation Fixture

Connect the power supply to the time correlation fixture as shown in the Add External Oscilloscope wizard.

WARNING

Shock hazard. Use only the 0950-2546 power supply and cord.

CAUTION

Equipment damage. Use only the 0950-2546 power supply and cord.

The BNC cables and the power supply must remain connected to the time correlation fixture when you are performing a time correlated measurement. The probes only need to be connected during the calibration procedure.

NOTE

Remember that the skew is affected by cable lengths, probe types, and the instruments. If you change any of these, you will need to run the calibration procedure again. To re-run the calibration procedure, select **Setup>(External Oscilloscope)>Trigger...**; then, in the Trigger tab of the External Oscilloscope Setup dialog, click **Run Auto-Deskew....**

Step 3. Make a measurement

When the calibration is complete you may proceed with your measurements as follows:

- 1 If you used the E5850A time correlation fixture to calibrate deskew values, disconnect the external oscilloscope probe and the logic analyzer leads from the time correlation fixture.
- 2 Connect the oscilloscope probe and the logic analyzer leads to your device under test. Do not disconnect the BNC cables.
- 3 Set up the trigger condition.
- 4 Press the Run button on the logic analyzer front panel.
- 5 Select the Waveform tab to view your measurement results.

Logic Analyzer Triggers External Oscilloscope

When the software is configured for Logic Analyzer triggers External Oscilloscope, you can set any trigger condition you desire in the logic analyzer. However, do not change the oscilloscope's trigger setup.

External Oscilloscope Triggers Logic Analyzer

When the software is configured for External Oscilloscope triggers Logic Analyzer, you may set any trigger condition you desire in the oscilloscope. However, do not change the logic analyzer's trigger setup.

See Also

- Working with Markers (see [page 21](#))
- General Triggering Problems (see [page 40](#))

2 Working with Markers

Marker Moves

Marker moves are coordinated between the logic analyzer and the oscilloscope. Markers are coordinated in time, relative to the trigger, taking the measured skew into account.

On the logic analyzer, the markers, called Ax(External Scope) and Bx(External Scope), are time correlated among all display windows that are displaying measurements from a given run. On the oscilloscope, there are two X-Axis (time) markers called Ax and Bx. Marker moves are coordinated between Ax(External Scope) and Ax and between Bx(External Scope) and Bx. The option to coordinate between Ax(External Scope) and Bx or Bx(External Scope) and Ax is not available.

Coordinated marker moves are useful for quickly identifying a given point in the trace on the separate logic analyzer and oscilloscope displays. The marker properties for the Ax(External Scope) and Bx(External Scope) markers can be accessed through the Markers menu.

Oscilloscope Markers

Ax Follows the Logic Analyzer's Ax(External Scope)

When you move the Ax(External Scope) marker in the logic analyzer waveform display, a command is sent to the oscilloscope to position the Ax marker at the same point in time.

Bx Follows the Logic Analyzer's Bx(External Scope)

When you move the Bx(External Scope) marker in the logic analyzer waveform display, a command is sent to the oscilloscope to position the Bx marker at the same point in time.

Logic Analyzer Markers

Ax(External Scope) Follows Oscilloscope's Ax

When you move the Ax marker on the oscilloscope, the Ax(External Scope) marker in the logic analyzer waveform display is moved to the same point in time.

Bx(External Scope) Follows Oscilloscope's Bx

When you move the Bx marker on the oscilloscope, the Bx(External Scope) marker in the logic analyzer waveform display is moved to the same point in time.


3 Displaying Imported Oscilloscope Data

After running the Add External Oscilloscope wizard, the external oscilloscope module appears in the Overview window with a connection to a new Waveform window.


To display logic analyzer data along with imported oscilloscope data

You can insert rows of data to display captured logic analyzer data in the same Waveform window as imported oscilloscope data:


- 1 Right-click on a row in the Waveform window, and choose **Insert Row Before...** or **Insert Row After...**



- 2 In the Insert dialog, select the logic analyzer data you would like to display with the imported oscilloscope data, and click **OK**.



After inserting the row, the captured logic analyzer data appears in the Waveform window along with the imported oscilloscope data. For example:



Note that, in the Waveform window, you can group imported digital channels into buses (see "To group signals into a bus" (in the online help)).

To change imported oscilloscope data options

- 1 Click the external oscilloscope module's  options button.
- 2 In the External Oscilloscope Setup dialog's Channels tab (see [page 44](#)), you can:
 - Select the oscilloscope channels whose data you want to import, including digital channels and buses from a mixed-signal oscilloscope.
 - Change oscilloscope analog channel signal names.

- Change oscilloscope channel signal colors.
 - Specify other external oscilloscope setup options.
- 3 When you are done changing imported data options, click OK to close the External Oscilloscope Setup dialog.

4 Controlling the External Oscilloscope via the Web

The External Scope Web Control VbaView window is an embedded Internet Explorer window pointed at the web interface of an external oscilloscope module.

NOTE

Oscilloscope performance slows down when its web control interface is being used; we recommend that you only connect to the oscilloscope's web control interface temporarily, to set up oscilloscope triggers, etc.; then, use the data import feature of the external oscilloscope module (see [Displaying Imported Oscilloscope Data](#) (see [page 23](#))) to view the waveforms captured by the oscilloscope.

Make sure the oscilloscope's web control interface is disconnected when running the external oscilloscope module's automatic deskew calibration.


To use the External Scope Web Control window:

- 1 Make sure the web control feature is enabled on the oscilloscope.
- 2 After you have added the external oscilloscope (see [Step 2](#). Run the Add External Oscilloscope wizard (see [page 16](#))), choose Window>New VbaView>External Scope Web Control....
- 3 In the Add New Window After dialog, select the external oscilloscope module, and click OK.
- 4 In the External Scope Web Control window, click the Web Control navigation bar to control the oscilloscope via its web interface.


NOTE

If you see a blank screen (for example, if you installed a newer version of Internet Explorer or a Windows Service Pack that has removed the Java Virtual Machine (JVM)), you need to download a JVM to enable the web control feature. For example, you can download Sun's Java Virtual Machine freely from "www.java.com".

Depending on the particular oscilloscope, the External Scope Web Control window looks something like:



In this case, clicking the Remote Front Panel link results in a window like:



5 Loading and Saving Configurations

When saving a configuration that includes an external oscilloscope module, the last skew run is saved along with the host name of the external oscilloscope. The oscilloscope setup (and data if you choose) will also be saved in the logic analyzer's configuration file.

When the configuration is loaded at a later time, the logic analysis system will automatically connect to the external oscilloscope and restore the setup that was saved. If the oscilloscope is not available on the network when the configuration is loaded this automatic connection and setup will fail.

If the external oscilloscope is a shared resource and is likely to be moved to another location, this option can have undesired effects. If the oscilloscope is on the network with the same Host Name or IP Address, the saved configuration will be loaded into that oscilloscope, possibly changing a setup that someone else has created.

NOTE


If two logic analyzers attempt to connect to a single oscilloscope, an error message will appear stating that the oscilloscope is locked by another user.

See Also

- "To save a configuration file" (in the online help)
- "To open a configuration file" (in the online help)
- "ALA vs. XML, When to Use Each Format" (in the online help)

6 Changing External Oscilloscope Setup Options

- 1 Choose the one of the commands from the Setup>(External Oscilloscope) menu.



The External Oscilloscope Setup dialog is opened with one of these tabs selected:


Channels tab (see page 44)	Lets you select the oscilloscope channels to import data from and lets you specify other external oscilloscope setup options.
Trigger tab (see page 45)	Used for setting the trigger direction and running the trigger deskew.
Connection tab (see page 46)	Shows the oscilloscope name or IP address, model, and serial number.

- 2 Make the desired changes.
- 3 Click OK.

7 Deleting the External Oscilloscope

To delete the external oscilloscope:

- 1 Choose Setup>(External Oscilloscope)>Delete from the menu bar.



See Also · Getting Started, Step 2. Run the Add External Oscilloscope wizard (see [page 16](#))

8 External Oscilloscope Troubleshooting

General Troubleshooting

- Finding the External Oscilloscope IP Address (see [page 38](#))
- LAN Problems (see [page 39](#))
- General Triggering Problems (see [page 40](#))
- Unexpected or Inconsistent Results (see [page 41](#))
- Correlation Drift Over Long Acquisitions (see [page 42](#))

Finding the External Oscilloscope IP Address

The external oscilloscope time correlation and data display software requires the external oscilloscope's IP address. To find the external oscilloscope IP address:

- On an Keysight Infiniium oscilloscope, select Utilities->GPIB Setup....
The IP address is shown in the LAN Interface portion of the window. Ignore the brackets and text, and use only the numbers (such as 130.29.68.220). Use that string in the Add External Oscilloscope wizard.
- On an Keysight 6000 Series oscilloscope, press the front panel Utilities button; then, press the I/O softkey.
The IP address is shown in the LAN portion of the resulting I/O configuration information box. Use that IP address string in the Add External Oscilloscope wizard.

If this string is not entered correctly or if someone changes the external oscilloscope's IP address, the logic analyzer will not be able to communicate with the oscilloscope. If the string happens to correspond to another oscilloscope on the network, the logic analyzer may find and connect to that oscilloscope, giving the appearance that you are correctly connected when in fact you are not.

If you enter a string that does not correspond to any device on your network, the software may appear to hang for several minutes while it searches for the device. Only one connection from a logic analyzer to a given oscilloscope can be supported.

LAN Problems

If you see a LAN error message:

- If you are using a 6000-series oscilloscope, make sure the LAN port is selected as the I/O Controller. (Press the Utility button; then, press the I/O softkey; use the Entry knob to select LAN; then, press the Controller softkey.)
- Check the LAN connections to the external oscilloscope and the logic analysis system.
- Check with your LAN administrator to ensure that the LAN is currently functioning. In the Connection tab of the External Oscilloscope Setup dialog, make sure the correct IP address is entered.
- If errors are reported, ask your LAN administrator to help you establish networking on the logic analyzer and oscilloscope. Cycle power on both instruments.
- If problems persist, contact Keysight Technologies for technical support.

Only one logic analysis system may connect to the external oscilloscope via the LAN

If two logic analysis systems try to connect to the same external oscilloscope, a message will appear stating that the oscilloscope is locked by another user. Connection to multiple external oscilloscopes from one logic analysis system is not supported.

See Also

- Finding the External Oscilloscope IP Address (see [page 38](#))

General Triggering Problems

- Logic analyzer triggers external oscilloscope (see [page 40](#))
- External oscilloscope triggers logic analyzer (see [page 40](#))
- Why isn't the red trigger marker where I expect? (see [page 40](#))
- Check cable connections (see [page 40](#))
- Check probe connections (see [page 40](#))
- Use grounds (see [page 40](#))
- External oscilloscope must be in the correct mode (see [page 40](#))
- Ensure no one else is using the external oscilloscope (see [page 40](#))

Logic analyzer triggers external oscilloscope

When Logic Analyzer triggers External Oscilloscope is selected, the oscilloscope is configured to trigger on a rising edge of Aux Trig In at 1.0 V. If you manually change the oscilloscope trigger, the logic analysis system will automatically restore the oscilloscope's trigger condition to trigger on the rising edge of Aux Trig In at 1.0 V the next time you run the logic analyzer. You can set any trigger condition you desire in the logic analyzer because you are using it to trigger the oscilloscope.

External oscilloscope triggers logic analyzer

When External Oscilloscope triggers Logic Analyzer is selected, the logic analyzer is configured to trigger on an external Trigger In. If you manually change the logic analyzer trigger, you will not be able to correctly time-correlate the instruments. You may set any trigger condition you desire in the oscilloscope because you are using it to trigger the logic analyzer.

Why isn't the red trigger marker where I expect?

Whenever you are using the external oscilloscope module, you are performing a multiple instrument measurement. Measurements from each instrument are time correlated, but each instrument retains its actual trigger point.

Check cable connections

The BNC cables must remain connected between the logic analyzer, the time correlation fixture, and the oscilloscope whenever you are using the time correlation feature.

Check probe connections

Be sure you are using the correct logic analyzer cables. Trace the cables from the logic analyzer to the E5850A time correlation fixture or device under test.

Be sure you are using the correct oscilloscope probe. Follow it from the instrument to the E5850A time correlation fixture or the device under test.

Use grounds

Your connections to the front of the time correlation fixture must include grounds. For example, you should connect the oscilloscope probe ground clip to the ground terminal on the E5850A time correlation fixture. Ground leads for each logic analyzer pod, and preferably each logic analyzer channel should be connected to a ground terminal on the E5850A fixture too.

External oscilloscope must be in the correct mode

Note that only Single Sweep, Real Time acquisition can be meaningfully correlated with a logic analyzer. Meaningful correlation may not be possible if the oscilloscope is set to sample in Equivalent Time, or if the oscilloscope sweep is set to Sweep or Triggered Sweep.

Ensure no one else is using the external oscilloscope

If someone else is using the external oscilloscope, either from the front panel or via the LAN, the logic analysis system will not be able to correctly control the oscilloscope.

Unexpected or Inconsistent Results

If you are experiencing unexpected or inconsistent deskew results, check the following items.

- Check cable connections (see [page 40](#))
- Check probe connection (see [page 40](#))
- External oscilloscope must be in the correct mode (see [page 40](#))
- Ensure no one else is using the external oscilloscope (see [page 40](#))

Correlation Drift Over Long Acquisitions

If you see correlation drift in long acquisitions as markers get farther away from the trigger, you can fix the problem by using the logic analysis system's 10 MHz CLOCK IN input (if present) to keep the oscilloscope and logic analyzer in sync.

The logic analysis system's CLOCK IN input is auto-sensing at power up; therefore, to use this input:

- 1 Make sure the logic analysis system is powered off.
- 2 Set up the 10 MHz clock source, 5.5 V max pk, either as an output from the oscilloscope or from another source that is fed into both the oscilloscope and logic analysis system.
- 3 Connect the clock source to the logic analysis system's CLOCK IN input BNC.
- 4 Power on the logic analysis system.

9 External Oscilloscope Reference


- External Oscilloscope Setup Dialog (see [page 44](#))

External Oscilloscope Setup Dialog

- Channels Tab (see [page 44](#))
- Trigger Tab (see [page 45](#))
- Connection Tab (see [page 46](#))

Channels Tab

The Channels tab lets you select the oscilloscope channels to import data from and lets you specify other external oscilloscope setup options.




Import Channels	Lets you choose oscilloscope channels to import data from, rename oscilloscope channel signals, and change the color associated with channel data. The color associated with a channel is used when the analog signal is inserted into a Waveform display window. Changing a channel's color in this dialog does not affect the color of waveforms currently displayed.
Current Probe Sense Resistor	This section provides a checkbox and an input box for each of the listed oscilloscope channels. <ul style="list-style-type: none"> Select the checkbox to indicate that you are using a current probe when importing data from the oscilloscope channel with which this checkbox is displayed. In the input box, specify the resistance used at the current probe point to allow scope power analysis from the scope voltage and supplied resistance. The default value used for resistance is 50 mOhm. <p>Note: If you select the Current probe checkbox for an oscilloscope channel, then Amperage is used as the Base unit for displaying this channel's data in the Waveform Viewer and Listing Viewer of the Logic and Protocol Analyzer GUI. If the Current probe checkbox is not selected, then Voltage is used as the Base unit for displaying the channel's data in the Waveform Viewer and Listing Viewer of the Logic and Protocol Analyzer GUI. Refer to the help books - <i>Analyzing Waveform Data</i> and <i>Analyzing Listing Data</i> in the <i>Logic and Protocol Analyzer Online Help</i> to know more about the display of analog signals in these viewers.</p>
Import Digital Channels and Buses	Lets you import digital channels and defined buses from a mixed-signal oscilloscope and change the color associated with digital channel data. The color associated with digital channels and buses is used when the signals are inserted into a Waveform display window. Changing the color in this dialog does not affect the color of waveforms currently displayed. Note that, in the Waveform window, you can group imported digital channels into buses (see "To group signals into a bus" (in the online help)).
Save External Oscilloscope Setup with Configuration	Specifies whether the external oscilloscope's configuration settings will be saved along with the time correlation settings when saving the logic analyzer's configuration file. When enabled and the logic analyzer configuration file is loaded, the logic analysis system will attempt to reload the external oscilloscope settings saved earlier. Note: The <i>Keysight Logic Analyzer</i> application will not load external oscilloscope settings from one oscilloscope model into a different oscilloscope model. For example, settings for a Infiniium 54810A oscilloscope will not load into a Infiniium 54831B oscilloscope.
Limit Import to N Samples	Used to limit the number of oscilloscope data samples to import.

- See Also**
- Connection Tab (see [page 46](#))
 - Trigger Tab (see [page 45](#))
 - Loading and Saving Configurations (see [page 31](#))

Trigger Tab

The Trigger tab is used for setting the trigger direction and running the trigger deskew.




Trigger Direction	<p>The Logic Analyzer triggers External Oscilloscope option sets up the logic analyzer and oscilloscope so that when the logic analyzer generates the trigger signal, the oscilloscope receives the signal through its Aux Trig In/Ext Trig channel and begins capturing data. You must manually set up the logic analyzer trigger.</p> <p>The External Oscilloscope triggers Logic Analyzer option sets up the logic analyzer and oscilloscope so that when the oscilloscope generates the trigger signal, the logic analyzer receives the signal through its Trigger In port and begins capturing data. You must manually set up the oscilloscope trigger. Note that the oscilloscope will not trigger the second half of a split logic analyzer.</p> <p>The Analyzer selection lets you choose the appropriate logic analyzer module for time correlation with the external oscilloscope.</p>
Trigger Deskew	<p>The Automatic option specifies that either default deskew values or the deskew values determined by the calibration procedure are used. Automatic deskew gives two separate deskew values, one for each trigger direction. To re-run the automatic deskew, click Run Auto-Deskew...</p> <p>Note: Make sure the external oscilloscope's web control interface is disconnected before running the automatic deskew calibration procedure.</p> <p>Note: Deskew calibration values differ depending upon the direction of the trigger.</p> <p>The Manual options lets you enter a deskew value in the numeric field provided. The deskew value entered is used for both trigger directions (that is, manual deskew mode does not maintain separate values for each trigger direction).</p> <p>Note: When Automatic deskew is selected, the deskew value is reset to the value obtained during the last calibration procedure.</p>

- See Also**
- Connection Tab (see [page 46](#))
 - Channels Tab (see [page 44](#))

Connection Tab

The Connection tab shows the external oscilloscope name or IP address, model, and serial number. If the oscilloscope becomes disconnected, you can reconnect through the Connection tab by clicking Reconnect.



- See Also**
- Channels Tab (see [page 44](#))
 - Trigger Tab (see [page 45](#))
 - Finding the External Oscilloscope IP Address (see [page 38](#))

10 External Oscilloscope Concepts

The following information helps explain how the external oscilloscope connection works.

- [Trigger Deskew \(see page 50\)](#)
- [How the Calibration Procedure Works \(see page 51\)](#)
- [Uncertainty \(see page 53\)](#)
- [What Do the Skew Numbers Mean? \(see page 54\)](#)

Trigger Deskew

When adding an external oscilloscope or running automatic deskew, trigger deskew values can be obtained by:

- Using default calibration values – In this case, default (typical) values for the logic analyzer and oscilloscope are used. This is sufficient for most customers.
- Using previous calibration values – This is appropriate when you have previously performed the calibration procedure and the hardware setup (instruments, probing, and cabling) has not changed. The previous values are used.
- Performing the calibration procedure (which requires the E5850A time correlation fixture) – To provide the minimum skew possible between the two instruments, the skew is measured by the calibration procedure.

After you have run the calibration procedure, you can remove the channel connections from the front of the E5850A time correlation fixture. When calibration is complete, all of your oscilloscope and logic analysis channels are available for probing your device under test.

Remember that the skew is affected by cable lengths, probe types, and the instruments themselves. If any of these change, you need to perform the calibration procedure again.

How the Calibration Procedure Works

- Propagation Delays (see [page 51](#))
- Trigger Points are Different (see [page 51](#))
- The Objective (see [page 51](#))
- Setup Matters (see [page 51](#))
- The CAL Signal (see [page 51](#))
- Logic Analyzer Triggers External Oscilloscope (see [page 52](#))
- Measuring the Skew (see [page 52](#))
- An Average is Used (see [page 52](#))
- External Oscilloscope Triggers Logic Analyzer (see [page 52](#))
- The Setup is Ready (see [page 52](#))

Propagation Delays

The logic analyzer and the external oscilloscope are able to drive a signal out of their chassis when a trigger condition occurs. These signals are Trigger Out on the logic analyzer and Aux Out on the oscilloscope.

The logic analyzer and the oscilloscope are also able to arm and/or trigger on the receipt of an external signal. These signals are Trigger In on the logic analyzer and Aux Trig In on the oscilloscope.

On both instruments, there is a propagation time between when a trigger event occurs and when its output signal is driven, and a propagation delay between when an input signal occurs and when the instrument actually triggers. There will also be propagation time in the cable between the instruments. We call the sum of all these propagation times the *skew* between the instruments.

Trigger Points are Different

For example, you can:

- 1 Connect a 50 ohm cable from the logic analyzer Trigger Out to the oscilloscope Aux Trig In.
- 2 Set a trigger condition of your choice on the logic analyzer.
- 3 Set the oscilloscope to Edge trigger on Aux Trig In at 1.0 V.
- 4 Run the oscilloscope.
- 5 Run the logic analyzer.

After the logic analyzer triggers, the oscilloscope will trigger. Both instruments will identify their trigger point. Unfortunately, because of the skew, these trigger points will not represent the same point in time. Unless the signal(s) being probed have unique identifying patterns, it may be very difficult or impossible to tell which edge(s) on one instrument correspond to which edge(s) on the other instrument.

The Objective

The point of the calibration procedure is to measure the skew between the instruments, remember the results of those measurements, and take those results into account when displaying subsequent logic analyzer and oscilloscope measurements so that time correlated marker moves can be created.

Setup Matters

It is important to be aware that the skew measurement is specific to the setup used for the measurement. The skew will change for different models of logic analyzers and oscilloscopes. Skew will vary based on the type and length of oscilloscope probe used. Skew will vary among the various probing options available for logic analysis modules such as flying leads, MICTOR connectors, Samtec connectors, etc. For the best possible time correlation, you must perform the calibration procedure with the same probing you will later use for measurements on your device under test.

The CAL Signal

The skew is measured by probing the CAL signal of the E5850A time correlation fixture with both Pod 2, Channel 0 of the logic analysis module and Channel 1 of the oscilloscope.

The CAL signal is a low frequency oscillator (approximately 62 kHz, with a 16 usec period). Because the period of this oscillator is at least two decimal orders of magnitude greater than the skew between the instruments, we know that the edge nearest the trigger is the same edge on both instruments.

It is critically important that the instrument that is to receive the external trigger signal is running, armed, and waiting before the driving instrument sends the trigger signal. If that is not the case, the instruments could very well be looking at different edges of the oscillator.

During the calibration procedure, and during subsequent regular single or repetitive runs, the external oscilloscope software ensures that the instruments are always armed and run in the correct sequence.

Analyzer Triggers External Oscilloscope

To measure skew in the Logic Analyzer triggers External Oscilloscope direction, the external oscilloscope software sets up both the logic analyzer and the oscilloscope. The logic analyzer is placed in Timing Mode at its highest sample rate (shortest sample period). A signal name is defined for the "DeskewEdge" (CAL signal) provided by the E5850A time correlation fixture. The logic analyzer is configured to trigger on the rising edge of the "DeskewEdge" signal. The oscilloscope is configured to trigger on its Aux Trig In line, and Channel 1 is turned on and configured to measure the "DeskewEdge" signal from the time correlation fixture. It is set up for Single Sweep, Real Time Sampling Mode at its maximum Sampling Rate and maximum Memory Depth.

Measuring the Skew

The skew is measured by first running the oscilloscope. After the logic analyzer is armed, it is run. The logic analyzer triggers on the next edge of "DeskewEdge", drives Trigger Out to the oscilloscope, which then triggers. The software retrieves the captured Channel 1 waveform from the oscilloscope and measures the time from the trigger point to the rising edge on that waveform. The rising edge will occur earlier in time than the trigger. This time is the skew.

An Average is Used

The measurement is repeated several times and the average is taken to be the skew. The reason for taking the average of multiple measurements is because the logic analyzer is sampling the "DeskewEdge" signal at a regular interval. The edge may actually occur near the beginning or near the end of the logic analyzer's sample period. The oscilloscope is also sampling at a regular interval, generally at a much higher frequency (shorter period). Because of the different and asynchronous sampling between the two instruments, a skew of +/- one analyzer sample period can occur in any single measurement.

External Oscilloscope Triggers Logic Analyzer

Measuring skew in the External Oscilloscope triggers Logic Analyzer direction is very similar to the Logic Analyzer triggers External Oscilloscope direction. The oscilloscope is set up to trigger on a rising edge on Channel 1 and drive Aux Out (TTL Trig Out) to the logic analysis system Trigger In. The logic analysis module is configured to trigger on Trigger In. When the measurement is run, the analysis module is run first so it is armed and waiting for the Aux Out/Trigger In signal from the oscilloscope. When the oscilloscope is run, it triggers and drives the trigger signal to the logic analysis system which then triggers and fills analysis memory. (Note that oscilloscopes that have a "TTL Trig Out" port can drive the logic analyzer's Trigger In input directly, while oscilloscopes that have "Aux Out" ports must be routed through the E5850A time correlation fixture for buffering.) The skew is then measured as the time between when the rising edge occurred and the trigger occurred in the logic analysis system. The measurement is run several times and an average is taken.

The Setup is Ready

The analyzer pod and oscilloscope probe may now be disconnected from the time correlation fixture and used to probe the device under test.

Remember that only Single Sweep, Real Time acquisition can be meaningfully correlated with a logic analyzer.

Uncertainty

Because the logic analyzer samples signals at intervals of one duration and the oscilloscope will usually sample at intervals of an entirely different duration, and because signals can actually transition at any point in time with any possible relation to the sample period(s), there will always be some uncertainty about when the signal actually transitioned versus when the analyzer noticed the transition.

Another source of uncertainty is the circuit used by the intermodule bus (IMB) to cross-correlate the times of trigger events from different modules. This circuitry has a resolution of 2 ns.

The uncertainty will always be at least +/- one timing analyzer sample period plus 2 ns of intermodule bus uncertainty.

During the calibration procedure, the maximum and minimum skew are measured. If the maximum measured skew is outside the expected range, it can indicate that one or more of the connections to the time correlation fixture were not made or were made incorrectly. A warning dialog will appear, suggesting that you check all the connections and run the calibration procedure again.

What Do the Skew Numbers Mean?

The skew (which is displayed within the Trigger tab of the External Oscilloscope Setup dialog) is the average trigger delay that is introduced external to the logic analyzer. The portion of the total skew that is internal to the logic analysis system is compensated for internally on a per-run basis. This internally measured skew is added to the constant skew that is displayed to compute the total skew. The total skew is added to the trigger time for the external oscilloscope, thereby time-correlating its trigger position with the trigger position of the analyzer.

The skew number will be positive when the analyzer triggers the external oscilloscope. When the analyzer triggers the oscilloscope, a delay occurs between the time when the analyzer sees the trigger condition and when the Trigger Out signal arrives at the oscilloscope Aux Trig In.

The skew number will be negative when the external oscilloscope triggers the analyzer. When the oscilloscope triggers the analyzer, a delay occurs between the time when the oscilloscope sees the trigger condition and when the Aux Out signal arrives at the logic analyzer Trigger In.

In the Trigger tab of the External Oscilloscope Setup dialog, you can use the Manual Trigger Deskew to change the fixed part of the total skew computation. This feature can be used to remove constant skew that was not measured during the Auto Deskew. If you use different probing accessories when you probe your device under test than when you ran the Auto Deskew, you will be causing additional constant skew. This is one example of where Manual Trigger Deskew can be useful.

11 External Oscilloscope Module Control, COM Automation

The *Keysight Logic Analyzer* application includes the COM Automation Server. This software lets you write programs that control the *Keysight Logic Analyzer* application from remote computers on the Local Area Network (LAN).

In a COM automation program, you can configure a module by:

- Loading a configuration file (which configures the complete logic analyzer setup).
- Using the "Module" (in the online help) object's "DoCommands" (in the online help) method with an XML-format string parameter (see Module Setup, XML Format (see [page 57](#))).

You can get information about a module's configuration using the Module object's "QueryCommand" (in the online help) method. Queries supported by the External Oscilloscope module are listed below.

For more information about logic analyzer COM automation and module objects in general, see "COM Automation" (in the online help).

XML-Based Queries Supported

The External Oscilloscope module supports the following XML-based queries (made with the "Module" (in the online help) object's "QueryCommand" (in the online help) method).

Query	Description
GetAllSetup	Returns the current setup, using the full tag set, used for writing generic configuration files (see the XML format Module Setup, XML Format (see page 57)).

See Also

- "COM Automation" (in the online help)
- Module Setup, XML Format (see [page 57](#))

12 External Oscilloscope Module Setup, XML Format

When you save logic analyzer configurations to XML format files, setup information for the External Oscilloscope module is included.

This XML format setup information is also used when writing COM automation programs to control the logic analyzer from a remote computer.

XML elements for the External Oscilloscope module have the following hierarchy:

```
<BusSignalSetup> (see page 61)
  <BusSignals> (see page 60)
    <Folder> (see page 72)
      <BusSignal> (see page 59)
        <Channels> (see page 64)
          <Names> (see page 75)
            <Name> (see page 74)
<ChannelsSetup> (see page 65)
<TriggerSetup> (see page 85)
<ConnectionSetup> (see page 66)
<Markers> (see page 73)
<Programming> (see page 77)
<Data> (see page 67)
  <Channel1/2/3/4> (see page 62)
    <Channel1/2/3/4DataTable> (see page 63)
      <AnalogDataSource> (see page 58)
        <SetupInfo> (see page 80)
      <TimeDataSource> (see page 84)
        <SetupInfo> (see page 83)
      <SampleNumberDataSource> (see page 79)
        <SetupInfo> (see page 82)
    <DigitalChannels> (see page 69)
      <WaveformData> (see page 86)
        <Pod1/2> (see page 76)
      <DigitalChannelsDataTable> (see page 70)
        <DigitalDataSource> (see page 71)
          <SetupInfo> (see page 81)
        <TimeDataSource> (see page 84)
          <SetupInfo> (see page 83)
        <SampleNumberDataSource> (see page 79)
          <SetupInfo> (see page 82)
```

- See Also**
- "XML Format" (in the online help)
 - Module Control, COM Automation (see [page 55](#))

<AnalogDataSource> Element

The <AnalogDataSource> element contains information about the external oscilloscope channel's voltage column in the binary data file.

Attributes

Name	Description
Type	'Page'

Children This element can have the following children: <SetupInfo> (for AnalogDataSource) (see [page 80](#)).

Parents This element can have the following parents: <Channel1/2/3/4DataTable> (see [page 63](#)).

Example

```
<AnalogDataSource Type='Page'>
  <SetupInfo RowCount='1048000' PageSize='49152' YInc='0.15625'
    YOrg='-8.55' YRef='0' FileOffset='0'/>
</AnalogDataSource>
```

<BusSignal> Element

The <BusSignal> element contains an oscilloscope channel signal definition.

Attributes

Name	Description
Color	'hex_RGB_value' (the color associated with this channel's data)
DefaultLabel	'F' (false) or 'T' (true)
Enabled	'F' (false) or 'T' (true)
Name	'string'

Children This element can have the following children: <Channels> (see [page 64](#)), <Names> (see [page 75](#)).

Parents This element can have the following parents: <Folder> (see [page 72](#)).

Examples

```
<BusSignal Name='My Scope C1' Enabled='T' Color='hFFFF00'>
  <Channels>1</Channels>
</BusSignal>
```

```
<BusSignal Name='My Scope D15-0' DefaultLabel='T'>
  <Channels>D[15:0]</Channels>
  <Names>
    <Name Index='0'>D0</Name>
    <Name Index='1'>D1</Name>
    <Name Index='2'>D2</Name>
    <Name Index='3'>D3</Name>
    <Name Index='4'>D4</Name>
    <Name Index='5'>D5</Name>
    <Name Index='6'>D6</Name>
    <Name Index='7'>D7</Name>
    <Name Index='8'>D8</Name>
    <Name Index='9'>D9</Name>
    <Name Index='10'>D10</Name>
    <Name Index='11'>D11</Name>
    <Name Index='12'>D12</Name>
    <Name Index='13'>D13</Name>
    <Name Index='14'>D14</Name>
    <Name Index='15'>D15</Name>
  </Names>
</BusSignal>
```

<BusSignals> Element

The <BusSignals> element contains oscilloscope channel signal definitions.

Children This element can have the following children: <Folder> (see [page 72](#)).

Parents This element can have the following parents: <BusSignalSetup> (see [page 61](#)).

Example

```
<BusSignals>
  <Clear />
  <Folder Name='Channel 1'>
    <BusSignal Name='My Scope C1' Enabled='T' Color='hFFFF00'>
      <Channels>C1</Channels>
    </BusSignal>
  </Folder>
  <Folder Name='Channel 2'>
    <BusSignal Name='My Scope C2' Enabled='T' Color='h00FF00'>
      <Channels>C2</Channels>
    </BusSignal>
  </Folder>
  <Folder Name='Digital Channels' Enabled='T' Color='h00FFFF'>
    <BusSignal Name='My Scope D15-0' DefaultLabel='T'>
      <Channels>D[15:0]</Channels>
      <Names>
        <Name Index='0'>D0</Name>
        <Name Index='1'>D1</Name>
        <Name Index='2'>D2</Name>
        <Name Index='3'>D3</Name>
        <Name Index='4'>D4</Name>
        <Name Index='5'>D5</Name>
        <Name Index='6'>D6</Name>
        <Name Index='7'>D7</Name>
        <Name Index='8'>D8</Name>
        <Name Index='9'>D9</Name>
        <Name Index='10'>D10</Name>
        <Name Index='11'>D11</Name>
        <Name Index='12'>D12</Name>
        <Name Index='13'>D13</Name>
        <Name Index='14'>D14</Name>
        <Name Index='15'>D15</Name>
      </Names>
    </BusSignal>
  </Folder>
</BusSignals>
```

<BusSignalSetup> Element

The <BusSignalSetup> element contains a <BusSignals> element.

Children This element can have the following children: <BusSignals> (see [page 60](#)).

Parents This element can have the following parents: "<Module>" (in the [online help](#)).

When used in COM automation, this element is returned by the "QueryCommand method" (in the [online help](#))'s GetAllSetup query. You can also use this element string as an XMLCommand with the "DoCommands method" (in the [online help](#)) to configure the External Oscilloscope module.

Example

```
<BusSignalSetup>
  <BusSignals>
    <Clear />
    <Folder Name='Channel 1'>
      <BusSignal Name='My Scope C1' Enabled='T' Color='hFFFFFF00'>
        <Channels>C1</Channels>
      </BusSignal>
    </Folder>
    <Folder Name='Channel 2'>
      <BusSignal Name='My Scope C2' Enabled='T' Color='h00FF00'>
        <Channels>C2</Channels>
      </BusSignal>
    </Folder>
    <Folder Name='Digital Channels' Enabled='T' Color='h00FFFF'>
      <BusSignal Name='My Scope D15-0' DefaultLabel='T'>
        <Channels>D[15:0]</Channels>
        <Names>
          <Name Index='0'>D0</Name>
          <Name Index='1'>D1</Name>
          <Name Index='2'>D2</Name>
          <Name Index='3'>D3</Name>
          <Name Index='4'>D4</Name>
          <Name Index='5'>D5</Name>
          <Name Index='6'>D6</Name>
          <Name Index='7'>D7</Name>
          <Name Index='8'>D8</Name>
          <Name Index='9'>D9</Name>
          <Name Index='10'>D10</Name>
          <Name Index='11'>D11</Name>
          <Name Index='12'>D12</Name>
          <Name Index='13'>D13</Name>
          <Name Index='14'>D14</Name>
          <Name Index='15'>D15</Name>
        </Names>
      </BusSignal>
    </Folder>
  </BusSignals>
</BusSignalSetup>
```

<Channel1/2/3/4> Element

The <Channel1/2/3/4> element describes the part of the binary data file that contains a particular oscilloscope channel's data.

Attributes

Name	Description
MaxVoltage	'floating_point_number' (maximum voltage for the full acquisition of this channel, represented in volts)
MinVoltage	'floating_point_number' (minimum voltage for the full acquisition of this channel, represented in volts)
NumBytesPerSample	'number' (number of bytes per sample in the oscilloscope data)
NumSamples	'number' (number of "points" or samples in the oscilloscope data)
Valid	'F' (false) or 'T' (true, when data is valid)
XInc	'floating_point_number' (the X coordinate increment value in seconds)
XOrg	'floating_point_number' (the X origin's value in seconds)
XRef	'number' (the X coordinate value associated with the X origin)
YInc	'floating_point_number' (the Y coordinate increment value in volts)
YOrg	'floating_point_number' (the Y origin's value in volts)
YRef	'number' (the Y coordinate value associated with the Y origin)

Children This element can have the following children: <Channel1/2/3/4DataTable> (see [page 63](#)).

Parents This element can have the following parents: <Data> (see [page 67](#)).

Example

```
<Channel1 Valid='T' XInc='5.000000e-010' XOrg='-2.619995e-004' XRef='0'
  YInc='1.562500e-001' YOrg='-8.550000e+000' YRef='0'
  MinVoltage='-3.080000e+000' MaxVoltage='5.670000e+000'
  NumBytesPerSample='1' NumSamples='1048000'>
  <Channel1DataTable RowCount='1048000' TriggerIndex='523999'
    OffsetTime='99.9666666666667 ns' TimePrecision='1 ps'
    FileName='ext_scope_data02of02.mfb'>
    <AnalogDataSource Type='Page'>
      <SetupInfo RowCount='1048000' PageSize='49152' YInc='0.15625'
        YOrg='-8.55' YRef='0' FileOffset='0'/>
    </AnalogDataSource>
    <TimeDataSource Type='Periodic'>
      <SetupInfo RowCount='1048000' TriggerIndex='523999'
        SamplePeriod='500 ps'/>
    </TimeDataSource>
    <SampleNumberDataSource Type='Periodic'>
      <SetupInfo RowCount='1048000' TriggerIndex='523999'
        SamplePeriod='1'/>
    </SampleNumberDataSource>
  </Channel1DataTable>
</Channel1>
```

<Channel1/2/3/4DataTable> Element

The <Channel1/2/3/4DataTable> element describes the part of the binary data file that contains a particular oscilloscope channel's data.

Attributes

Name	Description
FileName	'relative_path_file_name'
OffsetTime	'number "time_unit" (in the online help)'
RowCount	'number'
TimePrecision	'number "time_unit" (in the online help)'
TriggerIndex	'number'

Children This element can have the following children: <AnalogDataSource> (see [page 58](#)), <TimeDataSource> (see [page 84](#)), <SampleNumberDataSource> (see [page 79](#)).

Parents This element can have the following parents: <Data> (see [page 67](#)).

Example

```
<Channel1DataTable RowCount='1048000' TriggerIndex='523999'
  OffsetTime='99.966666666667 ns' TimePrecision='1 ps'
  FileName='ext_scope_data02of02.mfb'>
  <AnalogDataSource Type='Page'>
    <SetupInfo RowCount='1048000' PageSize='49152' YInc='0.15625'
      YOrg='-8.55' YRef='0' FileOffset='0' />
  </AnalogDataSource>
  <TimeDataSource Type='Periodic'>
    <SetupInfo RowCount='1048000' TriggerIndex='523999'
      SamplePeriod='500 ps' />
  </TimeDataSource>
  <SampleNumberDataSource Type='Periodic'>
    <SetupInfo RowCount='1048000' TriggerIndex='523999'
      SamplePeriod='1' />
  </SampleNumberDataSource>
</Channel1DataTable>
```

<Channels> Element

The `<Channels>` element specifies the oscilloscope channel assigned to the signal name.

Data This element's data is a string representing the oscilloscope channel for the signal.

Parents This element can have the following parents: `<BusSignal>` (see [page 59](#)).

Example `<Channels>1</Channels>`

<ChannelsSetup> Element

The <ChannelsSetup> element describes the external oscilloscope channels setup options.

Attributes

Name	Description
MemoryDepth	' number ' (the number of oscilloscope data samples to import)
NumChannels	' number ' (the number of oscilloscope channels)
SaveScopeSetup	' F ' (false) or ' T ' (true)

Parents

This element can have the following parents: "<Module>" (in the online help).

When used in COM automation, this element is returned by the "QueryCommand method" (in the online help)'s GetAllSetup query. You can also use this element string as an XMLCommand with the "DoCommands method" (in the online help) to configure the External Oscilloscope module.

Example

```
<ChannelsSetup NumChannels='2' SaveScopeSetup='T'
  MemoryDepth='1048576' />
```

<ConnectionSetup> Element

The <ConnectionSetup> element describes the oscilloscope connection options.

Attributes

Name	Description
LAProbe	'LAProbeGrabber', 'LAProbeMictor', 'LAProbeSamtec', or 'LAProbeSocket' (the probe connection from the analyzer to the deskewing signal)
Name	'string' (hostname of the oscilloscope to connect to)
ScopeProbe	'100-pin probe', '38-pin probe', 'Flying Lead probe', or 'ScopeProbeGrabber' (the probe connection from the oscilloscope to the deskewing signal)

Parents This element can have the following parents: "<Module>" (in the online help).

When used in COM automation, this element is returned by the "QueryCommand method" (in the online help)'s GetAllSetup query. You can also use this element string as an XMLCommand with the "DoCommands method" (in the online help) to configure the External Oscilloscope module.

Example

```
<ConnectionSetup
  Name='130.29.68.220'
  LAProbe='LAProbeSocket'
  ScopeProbe='ScopeProbeGrabber'
/>
```

<Data> Element

When data and setup information are saved to XML format configuration files, the <Data> element contains information about the binary data files that are saved.

Children This element can have the following children: <Channel1/2/3/4> (see [page 62](#)), <DigitalChannels> (see [page 62](#)).

Parents This element can have the following parents: "<Module>" (in the online help).

Example

```
<Data>
  <Channel1 Valid='T' XInc='5.000000e-010' XOrg='-2.668685e-004'
    XRef='0' YInc='6.250000e-002' YOrg='-4.984000e+000' YRef='0'
    MinVoltage='-1.047000e+000' MaxVoltage='2.016000e+000'
    NumBytesPerSample='1' NumSamples='1048600'>
    <Channel1DataTable RowCount='1048600' TriggerIndex='533737'
      OffsetTime='54 ns' TimePrecision='1 ps'
      FileName='ViewScope2_data03of04.mfb'>
      <AnalogDataSource Type='Page'>
        <SetupInfo RowCount='1048600' PageSize='49152'
          YInc='0.0625' YOrg='-4.984' YRef='0' YMin='-1.047'
          YMax='2.016' FileOffset='0' />
      </AnalogDataSource>
      <TimeDataSource Type='Periodic'>
        <SetupInfo RowCount='1048600' TriggerIndex='533737'
          SamplePeriod='500 ps' />
      </TimeDataSource>
      <SampleNumberDataSource Type='Periodic'>
        <SetupInfo RowCount='1048600' TriggerIndex='533737'
          SamplePeriod='1' />
      </SampleNumberDataSource>
    </Channel1DataTable>
  </Channel1>
  <Channel2 Valid='T' XInc='5.000000e-010' XOrg='-2.668680e-004'
    XRef='0' YInc='1.562500e-001' YOrg='2.000000e-001' YRef='0'
    MinVoltage='-2.460000e+000' MaxVoltage='5.360000e+000'
    NumBytesPerSample='1' NumSamples='1048600'>
    <Channel2DataTable RowCount='1048600' TriggerIndex='533736'
      OffsetTime='0 s' TimePrecision='10 ps'
      FileName='ViewScope2_data03of04.mfb'>
      <AnalogDataSource Type='Page'>
        <SetupInfo RowCount='1048600' PageSize='49152'
          YInc='0.15625' YOrg='0.2' YRef='0' YMin='-2.46'
          YMax='5.36' FileOffset='2097200' />
      </AnalogDataSource>
      <TimeDataSource Type='Periodic'>
        <SetupInfo RowCount='1048600' TriggerIndex='533736'
          SamplePeriod='500 ps' />
      </TimeDataSource>
      <SampleNumberDataSource Type='Periodic'>
        <SetupInfo RowCount='1048600' TriggerIndex='533736'
          SamplePeriod='1' />
      </SampleNumberDataSource>
    </Channel2DataTable>
  </Channel2>
</DigitalChannels>
```

```

<WaveformData>
  <Pod1 Valid='T' XInc='1.000000e-009' XOrg='-2.668490e-004'
    XRef='0' YInc='1.000000e+000' YOrg='0.000000e+000'
    YRef='0' MinVoltage='0.000000e+000'
    MaxVoltage='0.000000e+000' NumBytesPerSample='1'
    NumSamples='524300' />
  <Pod2 Valid='T' XInc='1.000000e-009' XOrg='-2.668490e-004'
    XRef='0' YInc='1.000000e+000' YOrg='0.000000e+000'
    YRef='0' MinVoltage='0.000000e+000'
    MaxVoltage='0.000000e+000' NumBytesPerSample='1'
    NumSamples='524300' />
</WaveformData>
<DigitalChannelsDataTable RowCount='524300' TriggerIndex='266849'
  OffsetTime='0 s' TimePrecision='10 ps'
  FileName='ViewScope2_data03of04.mfb'>
  <DigitalDataSource Type='Page'>
    <SetupInfo RowCount='524300' BitsPerSample='16'
      PageSize='49152' BitsPerBonusBit='16'
      FileOffset='4194400' />
  </DigitalDataSource>
  <TimeDataSource Type='Periodic'>
    <SetupInfo RowCount='524300' TriggerIndex='266849'
      SamplePeriod='1 ns' />
  </TimeDataSource>
  <SampleNumberDataSource Type='Periodic'>
    <SetupInfo RowCount='524300' TriggerIndex='266849'
      SamplePeriod='1' />
  </SampleNumberDataSource>
</DigitalChannelsDataTable>
</DigitalChannels>
</Data>

```

Binary Data File Format

For the external oscilloscope module, there is a single binary data file that contains the module's sample number, voltage, and time data. Each channel's data is stored contiguously, with the XML elements under <Data> pointing to the byte offsets within the file for each channel. Data is dumped simply by sample/row, starting with the first sample of data, all the way through the last.

If the XML with data file is read back into the system, the paging information in the XML file is used to load the data by pages to increase performance. There is no checksumming of any kind on the data, so data integrity is assumed to be valid whenever these files are loaded. If the file is altered in any way, it should be kept consistent with the format described.

If you are interested in processing the binary data files saved with XML format configuration files, please contact Keysight Technologies ("<http://www.keysight.com/find/contactus>") for sample parsing code.

<DigitalChannels> Element

The <DigitalChannels> element describes the part of the binary data file that contains a particular oscilloscope's digital channels data.

Children This element can have the following children: <WaveformData> (see [page 86](#)), <DigitalChannelsDataTable> (see [page 70](#)).

Parents This element can have the following parents: <Data> (see [page 67](#)).

Example

```
<DigitalChannels>
  <WaveformData>
    <Pod1 Valid='T' XInc='1.000000e-009' XOrg='-2.668490e-004'
      XRef='0' YInc='1.000000e+000' YOrg='0.000000e+000'
      YRef='0' MinVoltage='0.000000e+000'
      MaxVoltage='0.000000e+000' NumBytesPerSample='1'
      NumSamples='524300' />
    <Pod2 Valid='T' XInc='1.000000e-009' XOrg='-2.668490e-004'
      XRef='0' YInc='1.000000e+000' YOrg='0.000000e+000'
      YRef='0' MinVoltage='0.000000e+000'
      MaxVoltage='0.000000e+000' NumBytesPerSample='1'
      NumSamples='524300' />
  </WaveformData>
  <DigitalChannelsDataTable RowCount='524300' TriggerIndex='266849'
    OffsetTime='0 s' TimePrecision='10 ps'
    FileName='ViewScope2_data03of04.mfb'>
    <DigitalDataSource Type='Page'>
      <SetupInfo RowCount='524300' BitsPerSample='16'
        PageSize='49152' BitsPerBonusBit='16'
        FileOffset='4194400' />
    </DigitalDataSource>
    <TimeDataSource Type='Periodic'>
      <SetupInfo RowCount='524300' TriggerIndex='266849'
        SamplePeriod='1 ns' />
    </TimeDataSource>
    <SampleNumberDataSource Type='Periodic'>
      <SetupInfo RowCount='524300' TriggerIndex='266849'
        SamplePeriod='1' />
    </SampleNumberDataSource>
  </DigitalChannelsDataTable>
</DigitalChannels>
```

<DigitalChannelsDataTable> Element

The <DigitalChannelsDataTable> element describes the part of the binary data file that contains a particular oscilloscope's digital channels data.

Attributes

Name	Description
FileName	'relative_path_file_name'
OffsetTime	'number "time_unit" (in the online help)'
RowCount	'number'
TimePrecision	'number "time_unit" (in the online help)'
TriggerIndex	'number'

Children This element can have the following children: <DigitalDataSource> (see [page 71](#)), <TimeDataSource> (see [page 84](#)), <SampleNumberDataSource> (see [page 79](#)).

Parents This element can have the following parents: <DigitalChannels> (see [page 69](#)).

Example

```
<DigitalChannelsDataTable RowCount='524300' TriggerIndex='266849'
  OffsetTime='0 s' TimePrecision='10 ps'
  FileName='ViewScope2_data03of04.mfb'>
  <DigitalDataSource Type='Page'>
    <SetupInfo RowCount='524300' BitsPerSample='16'
      PageSize='49152' BitsPerBonusBit='16'
      FileOffset='4194400' />
  </DigitalDataSource>
  <TimeDataSource Type='Periodic'>
    <SetupInfo RowCount='524300' TriggerIndex='266849'
      SamplePeriod='1 ns' />
  </TimeDataSource>
  <SampleNumberDataSource Type='Periodic'>
    <SetupInfo RowCount='524300' TriggerIndex='266849'
      SamplePeriod='1' />
  </SampleNumberDataSource>
</DigitalChannelsDataTable>
```

<DigitalDataSource> Element

The <DigitalDataSource> element contains information about the external oscilloscope's digital data column in the binary data file.

Attributes

Name	Description
Type	'Page'

Children This element can have the following children: <SetupInfo> (for DigitalDataSource) (see [page 81](#)).

Parents This element can have the following parents: <DigitalChannelsDataTable> (see [page 70](#)).

Example

```
<DigitalDataSource Type='Page'>
  <SetupInfo RowCount='524300' BitsPerSample='16'
    PageSize='49152' BitsPerBonusBit='16'
    FileOffset='4194400' />
</DigitalDataSource>
```

<Folder> Element

The <Folder> element contains the bus/signal definition for an oscilloscope channel.

Attributes

Name	Description
Name	'string'

Children This element can have the following children: <BusSignal> (see [page 59](#)).

Parents This element can have the following parents: <BusSignals> (see [page 60](#)).

Example

```
<Folder Name='Channel 1'>
  <BusSignal Name='My Scope C1' Enabled='T' Color='hFFFF00'>
    <Channels>1</Channels>
  </BusSignal>
</Folder>
```

<Markers> Element

The <Markers> element describes the Ax and Bx oscilloscope marker times relative to the system trigger.

Attributes

Name	Description
MarkerAxTime	'floating_point_number' (relative to system trigger)
MarkerBxTime	'floating_point_number' (relative to system trigger)

Parents

This element can have the following parents: "<Module>" (in the online help).

When used in COM automation, this element is returned by the "QueryCommand method" (in the online help)'s GetAllSetup query. You can also use this element string as an XMLCommand with the "DoCommands method" (in the online help) to configure the External Oscilloscope module.

Example

```
<Markers
  MarkerAxTime= '-0.000000005313333333333328'
  MarkerBxTime= '0.0000001816666666666672'
/>
```

<Name> Element

The <Name> element gives a name to an individual digital channel imported from an external oscilloscope.

Attributes

Name	Description
Index	'number'

Parents This element can have the following parents: <Names> (see [page 75](#)).

Example `<Name Index='0'>D0</Name>`

<Names> Element

The <Names> element contains the names of the digital channels imported from an external oscilloscope.

Children This element can have the following children: <Name> (see [page 74](#)).

Parents This element can have the following parents: <BusSignal> (see [page 60](#)).

Example

```
<Names>
  <Name Index=' 0 ' >D0</Name>
  <Name Index=' 1 ' >D1</Name>
  <Name Index=' 2 ' >D2</Name>
  <Name Index=' 3 ' >D3</Name>
  <Name Index=' 4 ' >D4</Name>
  <Name Index=' 5 ' >D5</Name>
  <Name Index=' 6 ' >D6</Name>
  <Name Index=' 7 ' >D7</Name>
  <Name Index=' 8 ' >D8</Name>
  <Name Index=' 9 ' >D9</Name>
  <Name Index='10 ' >D10</Name>
  <Name Index='11 ' >D11</Name>
  <Name Index='12 ' >D12</Name>
  <Name Index='13 ' >D13</Name>
  <Name Index='14 ' >D14</Name>
  <Name Index='15 ' >D15</Name>
</Names>
```

<Pod1/2> Element

The <Pod1/2> element describes the part of the binary data file that contains a particular oscilloscope's digital data.

Attributes

Name	Description
MaxVoltage	'floating_point_number' (maximum voltage for the full acquisition of this pod, represented in volts)
MinVoltage	'floating_point_number' (minimum voltage for the full acquisition of this pod, represented in volts)
NumBytesPerSample	'number' (number of bytes per sample in the oscilloscope data)
NumSamples	'number' (number of "points" or samples in the oscilloscope data)
Valid	'F' (false) or 'T' (true, when data is valid)
XInc	'floating_point_number' (the X coordinate increment value in seconds)
XOrg	'floating_point_number' (the X origin's value in seconds)
XRef	'number' (the X coordinate value associated with the X origin)
YInc	'floating_point_number' (the Y coordinate increment value in volts)
YOrg	'floating_point_number' (the Y origin's value in volts)
YRef	'number' (the Y coordinate value associated with the Y origin)

Parents This element can have the following parents: <Data> (see [page 67](#)).

Example

```
<Pod1 Valid='T' XInc='1.000000e-009' XOrg='-2.668490e-004'
  XRef='0' YInc='1.000000e+000' YOrg='0.000000e+000'
  YRef='0' MinVoltage='0.000000e+000'
  MaxVoltage='0.000000e+000' NumBytesPerSample='1'
  NumSamples='524300' />
<Pod2 Valid='T' XInc='1.000000e-009' XOrg='-2.668490e-004'
  XRef='0' YInc='1.000000e+000' YOrg='0.000000e+000'
  YRef='0' MinVoltage='0.000000e+000'
  MaxVoltage='0.000000e+000' NumBytesPerSample='1'
  NumSamples='524300' />
```

<Programming> Element

The <Programming> element describes oscilloscope and logic analyzer setup information.

Attributes

Name	Description
LATrigState	'string' (the XML trigger specification for the logic analyzer to do auto-deskew in state mode)
LATrigTiming	'string' (the XML trigger specification for the logic analyzer to do auto-deskew in timing mode)
ScopeEdgeSource	'Channel1', 'Channel2', 'Channel3', or 'Channel4', (the edge source for the deskewing signal)
ScopeTrigMode	'TriggerModeAdvanced', 'TriggerModeEdge', or 'TriggerModeGlitch' (the oscilloscope's trigger mode for deskewing)
ScopeTrigSweep	'TriggerSweepAuto', 'TriggerSweepSingle', or 'TriggerSweepTriggered' (the oscilloscope's trigger sweep setting for deskewing)
ScopeAuxExtEdgeSlope	'SlopeNegative' or 'SlopePositive' (the slope for the oscilloscope's deskewing signal)
ScopeAuxExtTrigLevel	'floating_point_number'

Parents This element can have the following parents: "<Module>" (in the online help).

When used in COM automation, this element is returned by the "QueryCommand method" (in the online help)'s GetAllSetup query. You can also use this element string as an XMLCommand with the "DoCommands method" (in the online help) to configure the External Oscilloscope module.

Example

```
<Programming
  ScopeTrigSweep='TriggerSweepSingle'
  ScopeTrigMode='TriggerModeEdge'
  ScopeEdgeSource='Channel1'
  ScopeAuxExtEdgeSlope='SlopePositive'
  ScopeAuxExtTrigLevel='0.000000e+000'
  LATrigTiming='
    &lt;Module&gt;
      &lt;Trigger Mode=&quot;Timing&quot;&gt;
        &lt;Sequence Level=&quot;1&quot;&gt;
          &lt;If&gt;
            &lt;Event&gt;
              &lt;Anything/&gt;
            &lt;/Event&gt;
            &lt;Occurrence Value=&quot;1&quot;/&gt;
            &lt;Action&gt;
              &lt;TriggerAction Operator=&quot;Fill Memory&quot;/&gt;
            &lt;/Action&gt;
          &lt;/If&gt;
        &lt;/Sequence&gt;
      &lt;/Trigger&gt;
    &lt;/Module&gt;
  '
```

```

    LATrigState='
&lt;Module&gt;
  &lt;Trigger Mode=&quot;State&quot;&gt;
    &lt;StoreQual&gt;
      &lt;Event&gt;
        &lt;Anything/&gt;
      &lt;/Event&gt;
    &lt;/StoreQual&gt;
    &lt;Sequence Level=&quot;1&quot;&gt;
      &lt;If&gt;
        &lt;Event&gt;
          &lt;Anything/&gt;
        &lt;/Event&gt;
        &lt;Occurrence Value=&quot;1&quot;/&gt;
        &lt;Action&gt;
          &lt;TriggerAction Operator=&quot;Fill Memory&quot;&gt;
            &lt;StoreQual&gt;
              &lt;Event&gt;
                &lt;Anything/&gt;
              &lt;/Event&gt;
            &lt;/StoreQual&gt;
          &lt;/TriggerAction&gt;
        &lt;/Action&gt;
      &lt;/If&gt;
    &lt;/Sequence&gt;
  &lt;/Trigger&gt;
&lt;/Module&gt;
'
/>

```

<SampleNumberDataSource> Element

The <SampleNumberDataSource> element contains information about the external oscilloscope channel's sample number column in the binary data file.

Attributes

Name	Description
Type	'Periodic'

Children This element can have the following children: <SetupInfo> (for SampleNumberDataSource) (see [page 82](#)).

Parents This element can have the following parents: <Channel1/2/3/4DataTable> (see [page 63](#)).

Example

```
<SampleNumberDataSource Type='Periodic'>
  <SetupInfo RowCount='1048000' TriggerIndex='523999' SamplePeriod='1'/
>
</SampleNumberDataSource>
```

<SetupInfo> Element (for AnalogDataSource)

The <SetupInfo> element contains additional setup information for the <AnalogDataSource> (see [page 58](#)) element.

Attributes

Name	Description
FileOffset	'number' (bytes within the binary data file where the data begins)
PageSize	'number' (of bits per page of data)
RowCount	'number' (of samples)
YInc	'floating_point_number' (the Y coordinate increment value in volts)
YOrg	'floating_point_number' (the Y origin's value in volts)
YRef	'number' (the Y coordinate value associated with the Y origin)

Parents This element can have the following parents: <AnalogDataSource> (see [page 58](#)).

Example

```
<SetupInfo RowCount='1048000' PageSize='49152' YInc='0.15625'
  YOrg='-8.55' YRef='0' FileOffset='0' />
```

<SetupInfo> Element (for DigitalDataSource)

The <SetupInfo> element contains additional setup information for the <DigitalDataSource> (see [page 71](#)) element.

Attributes

Name	Description
BitsPerBonusBit	'number'
BitsPerSample	'number'
FileOffset	'number' (bytes within the binary data file where the data begins)
PageSize	'number' (of bits per page of data)
RowCount	'number' (of samples)

Parents This element can have the following parents: <DigitalDataSource> (see [page 71](#)).

Example

```
<SetupInfo RowCount='524300' BitsPerSample='16'
  PageSize='49152' BitsPerBonusBit='16'
  FileOffset='4194400' />
```

<SetupInfo> Element (for SampleNumberDataSource)

The <SetupInfo> element contains additional setup information for the <SampleNumberDataSource> (see [page 79](#)) element.

Attributes

Name	Description
RowCount	'number' (of samples)
SamplePeriod	'1'
TriggerIndex	'number' (zero-based)

Parents This element can have the following parents: <SampleNumberDataSource> (see [page 79](#)).

Example <SetupInfo RowCount='1048000' TriggerIndex='523999' SamplePeriod='1' />

<SetupInfo> Element (for TimeDataSource)

The <SetupInfo> element contains additional setup information for the <TimeDataSource> (see [page 84](#)) element.

Attributes

Name	Description
RowCount	'number' (of samples)
SamplePeriod	'number "time_unit" (in the online help) ' (sampling period for the timing data)
TriggerIndex	'number' (zero-based)

Parents

This element can have the following parents: <TimeDataSource> (see [page 84](#)).

Examples

```
<SetupInfo RowCount='1048000' TriggerIndex='523999' SamplePeriod='500 ps' />
```

<TimeDataSource> Element

The <TimeDataSource> element contains information about the external oscilloscope channel's time column in the binary data file.

Attributes

Name	Description
Type	'Periodic'

Children This element can have the following children: <SetupInfo> (for TimeDataSource) (see [page 83](#)).

Parents This element can have the following parents: <Channel1/2/3/4DataTable> (see [page 63](#)).

Examples

```
<TimeDataSource Type='Periodic'>
  <SetupInfo RowCount='1048000' TriggerIndex='523999' SamplePeriod='500
ps' />
</TimeDataSource>
```

<TriggerSetup> Element

The <TriggerSetup> element describes the external oscilloscope trigger options.

Attributes

Name	Description
Analyzer	'string' (name of logic analyzer module the oscilloscope is associated with)
AutoDeskew	'F' (false) or 'T' (true)
AutoDeskewExtIntTime	'floating_point_number' (the deskew time for the oscilloscope triggering the logic analyzer)
AutoDeskewIntExtTime	'floating_point_number' (the deskew time for the logic analyzer triggering the oscilloscope)
ManualDeskewExtIntTime	'floating_point_number' (the deskew time for the oscilloscope triggering the logic analyzer)
ManualDeskewIntExtTime	'floating_point_number' (the deskew time for the logic analyzer triggering the oscilloscope)
TriggerDirection	'LATrigScope' or 'ScopeTrigLA'

Parents This element can have the following parents: "<Module>" (in the online help).

When used in COM automation, this element is returned by the "QueryCommand method" (in the online help)'s GetAllSetup query. You can also use this element string as an XMLCommand with the "DoCommands method" (in the online help) to configure the External Oscilloscope module.

Example

```

<TriggerSetup
  TriggerDirection='LATrigScope'
  AutoDeskew='T'
  Analyzer='My 1682D-1'
  AutoDeskewIntExtTime='9.996667e-008'
  AutoDeskewExtIntTime='-1.150000e-008'
  ManualDeskewIntExtTime='9.996667e-008'
  ManualDeskewExtIntTime='-1.150000e-008'
/>

```

<WaveformData> Element

The <WaveformData> element contains information about the data associated with the oscilloscope's digital channels.

Children This element can have the following children: <Pod1/2> (see [page 76](#)).

Parents This element can have the following parents: <DigitalChannels> (see [page 69](#)).

Example

```
<WaveformData>
  <Pod1 Valid='T' XInc='1.000000e-009' XOrg='-2.668490e-004'
    XRef='0' YInc='1.000000e+000' YOrg='0.000000e+000'
    YRef='0' MinVoltage='0.000000e+000'
    MaxVoltage='0.000000e+000' NumBytesPerSample='1'
    NumSamples='524300' />
  <Pod2 Valid='T' XInc='1.000000e-009' XOrg='-2.668490e-004'
    XRef='0' YInc='1.000000e+000' YOrg='0.000000e+000'
    YRef='0' MinVoltage='0.000000e+000'
    MaxVoltage='0.000000e+000' NumBytesPerSample='1'
    NumSamples='524300' />
</WaveformData>
```

Index

Numerics

10 MHz CLOCK IN input, [42](#)

A

Add External Oscilloscope wizard, [16](#)
AnalogDataSource element, [58](#)
automatic trigger deskew, [45](#)

B

buses, external oscilloscope, [44](#)
BusSignal element, [59](#)
BusSignals element, [60](#)
BusSignalSetup element, [61](#)

C

calibration procedure, [51](#)
calibration procedure, external
oscilloscope time correlation, [50](#)
Channel1/2/3/4 element, [62](#)
Channel1/2/3/4DataTable element, [63](#)
Channels element, [64](#)
channels, external oscilloscope, [44](#)
ChannelsSetup, XML element (External
Oscilloscope), [65](#)
CLOCK IN input, [42](#)
COM Automation, external
oscilloscope, [55](#)
concepts, external oscilloscope time
correlation, [49](#)
configuration files, external
oscilloscope, [31](#)
connection information, external
oscilloscope, [16](#)
connection, external oscilloscope, [46](#)
ConnectionSetup, XML element
(External Oscilloscope), [66](#)

D

data display, external oscilloscope, [7](#)
Data element, [67](#)
default calibration values, external
oscilloscope time correlation, [50](#)
deskew results, unexpected or
inconsistent external
oscilloscope, [41](#)
deskew, trigger, [45](#)

digital channels, external
oscilloscope, [44](#)
DigitalChannels element, [69](#)
DigitalChannelsDataTable element, [70](#)
DigitalDataSource element, [71](#)
drift, correlation, [42](#)

E

E5850A Time Correlation, [7](#)
external oscilloscope data,
displaying, [23](#)
external oscilloscope IP address, [38](#)
External Oscilloscope Setup dialog, [44](#)
external oscilloscope setup options, [33](#)
external oscilloscope setup, saving, [44](#)
external oscilloscope time correlation
and data display, [7](#)
external oscilloscope triggers logic
analyzer, [45](#)
external oscilloscope, connection
information, [16](#)
external oscilloscope, deleting, [35](#)
external oscilloscope, split analyzer
and, [16, 45](#)
External Scope Web Control VbaView
window, [27](#)

F

Folder element, [72](#)

G

getting started, external oscilloscope
time correlation and data
display, [13](#)

L

LAN problems, external
oscilloscope, [39](#)
limit import to N samples, external
oscilloscope, [44](#)
logic analyzer triggers external
oscilloscope, [45](#)

M

manual trigger deskew, [45](#)
markers, external oscilloscope, [21](#)

Markers, XML element (External
Oscilloscope), [73](#)
measurement, external
oscilloscope, [19](#)

N

Name element, [74](#)
Names element, [75](#)
network external oscilloscope and
logic analyzer, [14](#)

O

options, external oscilloscope, [44](#)
oscilloscope (external) time
correlation, [7](#)
oscilloscope data, displaying, [23](#)

P

Pod1/2 element, [76](#)
point-to-point network set up, external
oscilloscope, [14](#)
point-to-point network set up, logic
analyzer, [14](#)
point-to-point network, external
oscilloscope and logic analyzer, [14](#)
Programming, XML element (External
Oscilloscope), [77](#)

R

reference, external oscilloscope, [43](#)

S

SampleNumberDataSource
element, [79](#)
SetupInfo element, [80, 81, 82, 83](#)
skew, external oscilloscope time
correlation, [54](#)
split analyzer, external oscilloscope
and, [16, 45](#)

T

TimeDataSource element, [84](#)
trigger deskew, [45](#)
trigger deskew, external oscilloscope
time correlation, [50](#)

- trigger direction, [45](#)
- triggering problems, external
 - oscilloscope, [40](#)
- TriggerSetup, XML element (External Oscilloscope), [85](#)
- troubleshooting, external
 - oscilloscope, [37](#)

U

- uncertainty in external oscilloscope
 - time correlation, [53](#)

V

- VbaView window, External Scope Web Control, [27](#)
- View Scope, [7](#)

W

- waveform display, external
 - oscilloscope, [7](#)
- WaveformData element, [86](#)
- web control, external oscilloscope, [27](#)
- wizard, Add External Oscilloscope, [16](#)

X

- XML format, external oscilloscope, [57](#)