

E5053A Security Features

Rev. 2.0



October 2014
Copyright 2008-2014 Keysight Technologies

Contacting Keysight Sales and Service Offices

Assistance with test and measurements needs and information on finding a local Keysight office is available on the internet at, <http://www.keysight.com/find/assist>. If you do not have access to the internet, please contact your field engineer.

Note: In any correspondence or telephone conversation, refer to the signal generator by its model number and full serial number. With this information, the Keysight representative can determine whether your unit is still within its warranty period.

Product Declassification and Security

Model Number(s): E5053A

Product Name: Microwave Downconverter

Product Family Name: Signal Source Analyzer

This document describes instrument security features and the steps to declassify an instrument through memory sanitization or removal. For additional information [please go to www.keysight.com/find/ad](http://www.keysight.com/find/ad) and click on the security instrument tab.

Memory Implementation

This section contains information on the types of memory available in your instrument. It explains the size of memory, how it is used, its location, volatility, and the sanitization procedure.

User Accessible

There are not any User Accessible memories in this model.

System Use Only – Not User Accessible

1) Flash memory (non-volatile) 32 kBytes

Used to store factory default system and calibration data.

It cannot be written to this non-volatile flash memory by an external user. There are no memory cleaning Procedure.

Other issues

There are no BIOS in this model.