

## E5092A Security Features

Rev. 2.0



October 2014  
Copyright 2008-2014 Keysight Technologies

## **Contacting Keysight Sales and Service Offices**

---

Assistance with test and measurements needs and information on finding a local Keysight office is available on the internet at, <http://www.keysight.com/find/assist>. If you do not have access to the internet, please contact your field engineer.

Note: In any correspondence or telephone conversation, with the information of its model number and full serial number, the Keysight representative can determine whether your unit is still within its warranty period.

## **Product Declassification and Security**

---

Model Number(s): E5092A

Product Name: Configurable Multiport Test Set

Product Family Name: Network Analyzer

---

This document describes instrument security features and the steps to declassify an instrument through memory sanitization or removal. For additional information [please go to www.keysight.com/find/ad](http://www.keysight.com/find/ad) and click on the security instrument tab.

## **Memory Implementation**

---

This section contains information on the types of memory available in your instrument. It explains the size of memory, how it is used, its location, volatility, and the sanitization procedure.

### **User Accessible**

**There are not any User Accessible memories in this model.**

### **System Use Only – Not User Accessible**

**1) EEPROM (non-volatile) 64 k bytes**

Used to store factory default system and calibration data.

It cannot be written to this non-volatile EEPROM by an external user. There is no memory cleaning procedure.

### ***Other issues***

There are no BIOS in this model.