

Keysight Technologies E8257D and E8663D Option C09

Product Note

The Keysight E8257D and E8663D Option C09 is a standard instrument that has been modified by moving all of the BNC connectors from the front panel to the rear panel with one exception; the RF Output connector will remain on the front panel.

For information concerning the operation and connections, reference the standard Keysight PSG User's and Service Guide.

In all other respects this instrument is identical to the standard instrument specifications.

Inspect the shipment. Keep the shipping container and packaging material until you have inspected the contents of the shipment for completeness and have verified the instrument mechanically and electrically. If there is physical damage refer to [“Contacting Keysight” on page 2](#). Keep the damaged shipping materials (if any) for inspection by the carrier and a Keysight Technologies representative.

Contacting Keysight

Assistance with test and measurement needs, and information on finding a local Keysight office are available on the Internet at:

<http://www.keysight.com/find/assist>

You can also purchase accessories or documentation items on the Internet at:

<http://www.keysight.com/find>

If you do not have access to the Internet, contact your field engineer.

NOTE In any correspondence or telephone conversation, refer to the product by its model number and full serial number. With this information, the Keysight representative can determine the warranty status of your unit.

