

Keysight Calibration Kit Definitions > 8510 WR-90 A.1

## WR-90 A.1 Calibration Kit

[Download CK\\_WR90A1](#)

### Standard Definitions

Std. No.	Standard Type	C0/L0 Term Z	C1/L1	C2/L2	C3/L3	Load Type	Offset Delay	Offset Z0	Offset Loss	Min Freq (GHz)	Max Freq (GHz)	Medium	Standard Label
#1	Short	0	0	0	0	Fixed	0	1	0	6.555	13.111	Waveguide	SHORT
#2	Short	0	0	0	0	Fixed	0	1	0	0	999	Waveguide	
#3	Short	0	0	0	0	Fixed	32.6332	1	0	6.555	13.111	Waveguide	1/4 OFFS
#4	Short	0	0	0	0	Fixed	0	1	0	0	999	Waveguide	
#5	Short	0	0	0	0	Fixed	0	1	0	0	999	Waveguide	
#6	Short	0	0	0	0	Fixed	0	1	0	0	999	Waveguide	
#7	Short	0	0	0	0	Fixed	0	1	0	0	999	Waveguide	
#8	Short	0	0	0	0	Fixed	0	1	0	0	999	Waveguide	
#9	Load					Fixed	0	1	0	6.555	13.111	Waveguide	FIXED
#10	Load					Sliding	0	1	0	6.555	13.111	Waveguide	SLIDING
#11	Thru					Fixed	0	1	0	6.555	13.111	Waveguide	THRU
#12	Short	0	0	0	0	Fixed	0	1	0	0	999	Waveguide	
#13	Short	0	0	0	0	Fixed	0	1	0	0	999	Waveguide	
#14	Thru					Fixed	0	1	0	6.555	13.111	Waveguide	THRU
#15	Thru					Fixed	32.6332	1	0	6.555	13.111	Waveguide	1/4 DELAY
#16	Short	0	0	0	0	Fixed	0	1	0	0	999	Waveguide	
#17	Short	0	0	0	0	Fixed	0	1	0	0	999	Waveguide	
#18	Short	0	0	0	0	Fixed	0	1	0	0	999	Waveguide	
#19	Short	0	0	0	0	Fixed	0	1	0	0	999	Waveguide	
#20	Load					Offset	32.6332	1	0	6.555	13.111	Waveguide	OFFSET
#21	Short	0	0	0	0	Fixed	0	1	0	0	999	Waveguide	

### Class Assignments

	A	B	C	D	E	Standard Class Label
S11A	1					SHORT
S11B	3					OFFSET
S11C	9	20				LOADS
S22A	1					SHORT
S22B	3					OFFSET
S22C	9	20				LOADS
Forward Transmission	11					THRU
Reverse Transmission	11					THRU
Forward Match	11					THRU
Reverse Match	11					THRU
Forward Isolation	9					ISOL'N STD
Reverse Isolation	9					ISOL'N STD
Frequency Response	1	11				RESPONSE
TRL Thru or Line	14					THRU
TRL Reflect	1					SHORT
TRL Line or Match	15					LINE
Adapter	13					undefined

VNA = Agilent 8510C

SYSTEM Z0 = 50 ohm

TRL Set Ref = Thru

TRL Cal Z0 = Line Z0

TRL Lowband Frequency = 2 GHz

© Keysight Technologies 2000-2021

[Privacy](#)

[Terms](#)

[Feedback](#)

[Contact Keysight](#)