



**Manufacturer's Name:** Agilent Technologies, Inc.  
**Manufacturer's Address:** 1900 Garden of the Gods Road  
**Supplier's Address:** Colorado Springs, CO 80907  
USA

**Declares under sole responsibility that the product as originally delivered**

**Product Name:** Digital and Mixed-Signal Oscilloscope (1,2.5G,4 GHz bandwidths)  
**Model Number:** DS09104A, DS09254A, DS09404A, MS09104A, MS09254A, MS09404A  
**Product Options:** This declaration covers all options of the above product(s)

**and conforms with the following product standards:**

EMC	Standard	Limit
	IEC 61326-1:2005 / EN 61326-1:2006	
	CISPR 11 / EN 55011	Group 1 Class A
	IEC 61000-4-2 / EN 61000-4-2	4 kV CD, 8 kV AD
	IEC 61000-4-3 / EN 61000-4-3	3 V/m, 80-1000 MHz & 1.4-2 GHz, 1V/m 2-2.7 GHz
	IEC 61000-4-4 / EN 61000-4-4	0.5 kV signal lines, 1 kV power lines
	IEC 61000-4-5 / EN 61000-4-5	0.5 kV line-line, 1 kV line-ground
	IEC 61000-4-6 / EN 61000-4-6	3 V, 0.15-80 MHz
	IEC 61000-4-11 / EN 61000-4-11	1 cycle, >95%
	Canada: ICES-001:2004	
	Australia/New Zealand: AS/NZS CISPR 11:2004	

The product was tested in a typical configuration with Agilent Technologies test systems.

**Safety** IEC 61010-1:2001 / EN 61010-1:2001  
Canada: CAN/CSA-C22.2 No. 61010-1-04  
USA: ANSI/UL 61010-1:2004  
UL94

**Supplementary Information: Supplementary Information:**

**This DoC applies to above-listed products placed on the EU market after:**

**June 15, 2009**

Date

**Merlin Loblick**  
Quality Manager

For further information, please contact your local Agilent Technologies sales office, agent or distributor, or Agilent Technologies Deutschland GmbH, Herrenberger Straße 130, D 71034 Böblingen, Germany.

## **Product Regulations**

### **EMC**

IEC 61326-1:2004 / EN 61326-1 :2006

### **Product Safety**

IEC 61010-1:2001 / EN 61010-1:2001

Canada: CAN/CSA-C22.2 No. 61010-1-04

USA: ANSI/UL 61010-1:2004

UL94

### **Sound Pressure Level**

Less than 60 dBA

### **Regulatory Information for Canada**

Canada: ICES-001:2004

This ISM device complies with Canadian ICES-001.

Cet appareil ISM est conforme à la norme NMB-001 du Canada.

### **Regulatory Information for Australia/New Zealand**

This ISM device complies with Australian/New Zealand AS/NZS CISPR 11:2004



Manual Part Number: 54904-92000

Printed in Malaysia

