



# Agilent E3668B

## Feedthru Panel Kit

**For:**

***E7590A— 1.3 meter Instrument Rack***

***E3661B— 1.6 meter Instrument Rack***

***E3662B— 2.0 meter Instrument Rack***

## Installation Guide



**Agilent Technologies**

November 2001  
Printed in U.S.A. E1101



E3668-90002

# Notice

The information contained in this document is subject to change without notice.

**AGILENT TECHNOLOGIES MAKES NO WARRANTY OF ANY KIND WITH REGARD TO THIS MATERIAL, INCLUDING, BUT NOT LIMITED TO, THE IMPLIED WARRANTIES OF MERCHANTABILITY AND FITNESS FOR A PARTICULAR PURPOSE.**

Agilent Technologies shall not be liable for errors contained herein or for incidental or consequential damages in connection with the furnishing, performance, or use of this material. This document contains proprietary information, which is protected by copyright. All rights are reserved. No part of this document may be photocopied, reproduced, or translated into another language without the prior written consent of Agilent Technologies.

**Copyright © 1998, 2001 Agilent Technologies, Inc.**

## Contents

The E3668B Feedthru Panel Kit contains the following items.

Item	Quantity	Description	Part Number
1	1	Feedthru Panel	E3668-60002
2	2	Sheet Metal Nuts (10-32)	0590-0804
3	2	Screw, special, 10-32 x 0.5"	0570-1366
4	1	Installation Instruction Sheet	E3668-90002

## Tools Required

To install the Feedthru Panel Kit, you will need a Phillips screwdriver

## Installation Procedure

Refer to the illustration on page 4. The feedthru panel occupies 1 EIA unit (44 mm). The panel is designed to attach to the center hole within the EIA envelope.

1. Locate the appropriate holes on the front columns of the rack.
2. Position the two sheet metal nuts over the appropriate holes in the front columns.
3. Attach the feedthru panel to the front columns of the rack using two screws.
4. Route the cables through the openings in the feedthru panel.

This completes installation of the Feedthru Panel Kit.

Place this *Note* with your rack documentation.

