DECLARATION OF CONFORMITY

According to ISO/IEC Guide 22 and CEN/CENELEC EN 45014



Manufacturer's Name:Agilent Technologies, Inc.Manufacturer's Address:1900 Garden of the Gods RoadSupplier's Address:Colorado Springs, CO 80907

USA

Declares under sole responsibility that the product as originally delivered

Product Name:Active High Voltage Differential ProbeModel Number:N2791A, N2792A, N2793A and N2891A

Product Options: This declaration covers all options of the above product(s)

and conforms with the following product standards:

| ЕМС | Standard | Limit |
|-----|------------------------------------|--|
| | IEC 61326-1:2005 / EN 61326-1:2006 | |
| | CISPR 11 / EN 55011 | Group 1 Class A |
| | IEC 61000-4-2 / EN 61000-4-2 | 4 kV CD, 8 kV AD |
| | IEC 61000-4-3 / EN 61000-4-3 | 3 V/m, 80-1000 MHz & 1.4-2 GHz, 1V/m 2-2.7 GHz |
| | IEC 61000-4-4 / EN 61000-4-4 | 0.5 kV signal lines, 1 kV power lines |
| | IEC 61000-4-5 / EN 61000-4-5 | 0.5 kV line-line, 1 kV line-ground |
| | IEC 61000-4-6 / EN 61000-4-6 | 3 V, 0.15-80 MHz |
| | IEC 61000-4-11 / EN 61000-4-11 | 1 cycle, >95% |
| | | |

Canada: ICES-001:2004

Australia/New Zealand: AS/NZS CISPR 11:2004

The product was tested in a typical configuration with Agilent Technologies test systems.

Safety IEC 61010-1:2001 / EN 61010-1:2001

Canada: CAN/CSA-C22.2 No. 61010-1-04

USA: ANSI/UL 61010-1:2004

UL94

Supplementary Information:

This DoC applies to above-listed products placed on the EU market after:

| April 28, 2009 | Merlin Jobelesk |
|----------------|-----------------|
| Date | Merlin Loblick |
| | Quality Manager |

For further information, please contact your local Agilent Technologies sales office, agent or distributor, or Agilent Technologies Deutschland GmbH, Herrenberger Straße 130, D 71034 Böblingen, Germany.

Product Regulations

EMC

IEC 61326-1:2004 / EN 61326-1:2006

Product Safety

IEC 61010-1:2001 / EN 61010-1:2001 Canada: CAN/CSA-C22.2 No. 61010-1-04 USA: ANSI/UL 61010-1:2004 UL94

Sound Pressure Level

Less than 60 dBA

Regulatory Information for Canada

Canada: ICES-001:2004 This ISM device complies with Canadian ICES-001. Cet appareil ISM est confomre à la norme NMB-001 du Canada.

Regulatory Information for Australia/New Zealand

This ISM device complies with Australian/New Zealand AS/NZS CISPR 11:2004



Manual Part Number: N2793-92001, February 2010



Printed in Malaysia