




Using the Agilent N5442A BNC Adapter

The Agilent N5442A BNC Adapter can be used to connect any 50 Ω Autoprobe device (such as an InfiniiMax I/II probe amplifier or BNC cable) to a 90000-X Series oscilloscope.



Connect one side of the adapter (**A**) to one of the input clutches on a 90000-X Series oscilloscope and connect the other side of the adapter (**B**) to the 50 Ω device.



Agilent Technologies

Agilent Technologies, Inc.

Printed in Malaysia

January 2010

Document Part Number: N5442-97000



Agilent Technologies