



Agilent X-Series Signal Analyzer

**This manual provides documentation for the
following X-Series Analyzers:**

**PXA Signal Analyzer N9030A
MXA Signal Analyzer N9020A
EXA Signal Analyzer N9010A**

N9062A SCPI Language Compatibility Guide



Agilent Technologies

Notices

© Agilent Technologies, Inc. 2010

No part of this manual may be reproduced in any form or by any means (including electronic storage and retrieval or translation into a foreign language) without prior agreement and written consent from Agilent Technologies, Inc. as governed by United States and international copyright laws.

Trademark Acknowledgements

Microsoft® is a U.S. registered trademark of Microsoft Corporation.

Windows® and MS Windows® are U.S. registered trademarks of Microsoft Corporation.

Adobe Reader® is a U.S. registered trademark of Adobe System Incorporated.

Java™ is a U.S. trademark of Sun Microsystems, Inc.

MATLAB® is a U.S. registered trademark of Math Works, Inc.

Norton Ghost™ is a U.S. trademark of Symantec Corporation.

Manual Part Number

N9062-90001

Print Date

July 2010

Printed in USA

Agilent Technologies, Inc.
1400 Fountaingrove Parkway
Santa Rosa, CA 95403

Warranty

The material contained in this document is provided “as is,” and is subject to being changed, without notice, in future editions. Further, to the maximum extent permitted by applicable law, Agilent disclaims all warranties, either express or implied, with regard to this manual and any information contained herein, including but not limited to the implied warranties of merchantability and fitness for a particular purpose. Agilent shall not be liable for errors or for incidental or consequential damages in connection with the furnishing, use, or performance of this document or of any information contained herein. Should Agilent and the user have a separate written agreement with warranty terms covering the material in this document that conflict with these terms, the warranty terms in the separate agreement shall control.

Technology Licenses

The hardware and/or software described in this document are furnished under a license and may be used or copied only in accordance with the terms of such license.

Restricted Rights Legend

If software is for use in the performance of a U.S. Government prime contract or subcontract, Software is delivered and

licensed as “Commercial computer software” as defined in DFAR 252.227-7014 (June 1995), or as a “commercial item” as defined in FAR 2.101(a) or as “Restricted computer software” as defined in FAR 52.227-19 (June 1987) or any equivalent agency regulation or contract clause. Use, duplication or disclosure of Software is subject to Agilent Technologies’ standard commercial license terms, and non-DOD Departments and Agencies of the U.S. Government will receive no greater than Restricted Rights as defined in FAR 52.227-19(c)(1-2) (June 1987). U.S. Government users will receive no greater than Limited Rights as defined in FAR 52.227-14 (June 1987) or DFAR 252.227-7015 (b)(2) (November 1995), as applicable in any technical data.

Safety Notices

CAUTION

A CAUTION notice denotes a hazard. It calls attention to an operating procedure, practice, or the like that, if not correctly performed or adhered to, could result in damage to the product or loss of important data. Do not proceed beyond a CAUTION notice until the indicated conditions are fully understood and met.

WARNING

A WARNING notice denotes a hazard. It calls attention to an operating procedure, practice, or the like that, if not correctly performed or adhered to, could result in personal injury or death. Do not proceed beyond a WARNING notice until the indicated conditions are fully understood and met.

Warranty

This Agilent technologies instrument product is warranted against defects in material and workmanship for a period of one year from the date of shipment. During the warranty period, Agilent Technologies will, at its option, either repair or replace products that prove to be defective.

For warranty service or repair, this product must be returned to a service facility designated by Agilent Technologies. Buyer shall prepay shipping charges to Agilent Technologies and Agilent Technologies shall pay shipping charges to return the product to Buyer. However, Buyer shall pay all shipping charges, duties, and taxes for products returned to Agilent Technologies from another country.

Where to Find the Latest Information

Documentation is updated periodically. For the latest information about this analyzer, including firmware upgrades, application information, and product information, see the following URLs:

<http://www.agilent.com/find/pxa>

<http://www.agilent.com/find/mxa>

<http://www.agilent.com/find/exa>

To receive the latest updates by email, subscribe to Agilent Email Updates:

<http://www.agilent.com/find/emailupdates>

Information on preventing analyzer damage can be found at:

<http://www.agilent.com/find/tips>

1. Getting Started

What Does the Agilent N9062A SCPI Language Compatibility Application Do?	14
Documentation for the N9062A application.	15
Signal Analyzers with N9062A.	15
Signal Analyzer Updates.	15
N9062A General Rules and Limitations.	16
AC/DC Coupling.	16
Couplings	16
Data Formats	16
Markers	16
Numeric Ranges	16
Remote Control	16
Returning Data	16
Units	17
Status System & SRQ	17
Supported Commands and Queries.	18
Unsupported Commands and Queries.	18
Hardware and Firmware Requirements for N9062A	19
Installing the N9062A application	20
License for N9062A	20
Installation.	20
Licensing	20
License Installation Procedure over USB	21
Verify the Installation	21
Running Software that Requires SCPI Commands	22
Service and Calibration.	23

2. Front Panel Operations

Selecting SCPI Language Compatibility Mode	26
Mode Setup	27
Selecting Compatibility Model	27
R&S FSE.	28
R&S FSP.	28
R&S FSU.	28
Selecting Mode IDN Response	28
Compatibility Model	30
System IDN Response.	30
User IDN Response.	30
Restore Mode Defaults	30
Send Commands	31

3. Compatible SCPI Commands

SCPI Summary Table	34
Agilent N9062A Supported IEEE SCPI Commands	49
Supported SCPI Commands	51
Mode Selection	51
Mode Setup	51
Compatibility Mode	51

R&S FSE	52
R&S FSP	52
R&S FSU	52
Mode IDN Response	52
Compatibility Model	53
System IDN Response	53
User IDN Response	53
ABORt Subsystem	53
ABORt	54
CALCulate:DELTAmarker Subsystem	54
CALCulate<1 2>:DELTAmarker<1...4>[:STATE] ON OFF	54
CALCulate<1 2>:DELTAmarker<1...4>:MODE ABSolute RELative	54
CALCulate<1 2>:DELTAmarker<1...4>:AOFF	55
CALCulate<1 2>:DELTAmarker<1...4>:TRACe <1...3>	55
CALCulate<1 2>:DELTAmarker<1...4>:X<0...MAX>(frequency sweep time)	55
CALCulate<1 2>:DELTAmarker<1...4>:X:RELative?	55
CALCulate<1 2>:DELTAmarker<1...4>:Y?	56
CALCulate<1 2>:DELTAmarker<1...4>:MAXimum[:PEAK]	56
CALCulate<1 2>:DELTAmarker<1...4>:MAXimum:LEFT	56
CALCulate<1 2>:DELTAmarker<1...4>:MAXimum:NEXT	56
CALCulate<1 2>:DELTAmarker<1...4>:MAXimum:RIGHT	57
CALCulate<1 2>:DELTAmarker<1...4>:MINimum[:PEAK]	57
CALCulate<1 2>:DELTAmarker<1...4>:MINimum:LEFT	57
CALCulate<1 2>:DELTAmarker<1...4>:MINimum:NEXT	57
CALCulate<1 2>:DELTAmarker<1...4>:MINimum:RIGHT	57
CALCulate<1 2>:DELTAmarker<1...4>:LINK ON OFF	58
CALCulate<1 2>:DELTAmarker<1...4>:FUNCTioN:FIXed[:STATE] ON OFF	58
CALCulate<1 2>:DELTAmarker<1...4>:FUNCTioN:FIXed:RPOint:MAXimum[:PEAK] <numeric_value>	59
CALCulate<1 2>:DELTAmarker<1...4>:FUNCTioN:FIXed:RPOint:Y <numeric_value>	59
CALCulate<1 2>:DELTAmarker<1...4>:FUNCTioN:FIXed:RPOint:Y:OFFSet <numeric_value>	59
CALCulate<1 2>:DELTAmarker<1...4>:FUNCTioN:FIXed:RPOint:X <numeric_value>	60
CALCulate<1 2>:DELTAmarker<1...4>:FUNCTioN:PNOise[:STATE] ON OFF	60
CALCulate<1 2>:DELTAmarker<1...4>:FUNCTioN:PNOise:AUTO ON OFF	60
CALCulate<1 2>:DELTAmarker<1...4>:FUNCTioN:PNOise:RESult?	61
CALCulate:LIMit Subsystem	61
CALCulate<1 2>:LIMit<1...6>:ACTive?	61
CALCulate<1 2>:LIMit<1...6>:TRACe <1...3>	62
CALCulate<1 2>:LIMit<1...6>:STATe ON OFF	62
CALCulate<1 2>:LIMit<1...6>:UNIT DBM DBPW WATT DBUV DBMV VOLT DBUA AMPere DB DEG RAD S HZ PCT UNITLESS	62
CALCulate<1 2>:LIMit<1...6>:FAIL?	63
CALCulate<1 2>:LIMit<1...6>:CLEar[:IMMediate]	63
CALCulate<1 2>:LIMit<1...6>:COMMeNt <string>	63
CALCulate<1 2>:LIMit<1...6>:COPY 1 to 6 <name>	64
CALCulate<1 2>:LIMit<1...6>:NAME <name of limit line>	64
CALCulate<1 2>:LIMit<1...6>:DELeTe	65
CALCulate:LIMit:ESpectrum Subsystem	65
CALCulate<1 2>:LIMit<1...6>:ESpectrum:MODE AUTO MANual USER	65

CALCulate<1 2>:LIMit<1...6>:ESpectrum:VALue <numeric_value>	65
CALCulate<1 2>:LIMit<1...6>:ESpectrum:REStore	66
CALCulate:LIMit:CONTRol Subsystem	66
CALCulate<1 2>:LIMit<1...6>:CONTRol[:DATA] <numeric_value>,<numeric_value>..	66
CALCulate<1 2>:LIMit<1...6>:CONTRol:DOMain FREQUency TIME	66
CALCulate<1 2>:LIMit<1...6>:CONTRol:OFFSet <numeric_value>	67
CALCulate<1 2>:LIMit<1...6>:CONTRol:MODE RELAtive ABSolute.	67
CALCulate<1 2>:LIMit<1...6>:CONTRol:SHIFt <numeric_value>	67
CALCulate<1 2>:LIMit<1...6>:CONTRol:SPACing LINear LOGarithmic	68
CALCulate:LIMit:LOWer Subsystem	68
CALCulate<1 2>:LIMit<1...6>:LOWer[:DATA] <numeric_value>,<numeric_value>..	68
CALCulate<1 2>:LIMit<1...6>:LOWer:STATe ON OFF	69
CALCulate<1 2>:LIMit<1...6>:LOWer:OFFSet <numeric_value>	69
CALCulate<1 2>:LIMit<1...6>:LOWer:MARGin <numeric_value>	69
CALCulate<1 2>:LIMit<1...6>:LOWer:MODE RELAtive ABSolute	70
CALCulate<1 2>:LIMit<1...6>:LOWer:SHIFt <numeric_value>	70
CALCulate<1 2>:LIMit<1...6>:LOWer:SPACing LINear LOGarithmic	71
CALCulate<1 2>:LIMit<1...6>:LOWer:THReshold <numeric_value>	71
CALCulate:LIMit:UPPer Subsystem	71
CALCulate<1 2>:LIMit<1...6>:UPPer[:DATA] <numeric_value>,<numeric_value>..	72
CALCulate<1 2>:LIMit<1...6>:UPPer:STATe ON OFF.	72
CALCulate<1 2>:LIMit<1...6>:UPPer:OFFSet <numeric_value>.	72
CALCulate<1 2>:LIMit<1...6>:UPPer:MARGin <numeric_value>	73
CALCulate<1 2>:LIMit<1...6>:UPPer:MODE RELAtive ABSolute	73
CALCulate<1 2>:LIMit<1...6>:UPPer:SHIFt <numeric_value>	73
CALCulate<1 2>:LIMit<1...6>:UPPer:SPACing LINear LOGarithmic.	74
CALCulate<1 2>:LIMit<1...6>:UPPer:THReshold <numeric_value>.	74
CALCulate:MARKer Subsystem	75
CALCulate<1 2>:MARKer<1...4>[:STATe] ON OFF	75
CALCulate<1 2>:MARKer<1...4>:AOFF	75
CALCulate<1 2>:MARKer<1...4>:TRACe 1 to 3	75
CALCulate<1 2>:MARKer<1...4>:X <0...MAX> (frequency sweep time)	76
CALCulate<1 2>:MARKer<1...4>:X:SLIMits[:STATe] ON OFF	76
CALCulate<1 2>:MARKer<1...4>:X:SLIMits:LEFT <0...MAX> (frequency sweep time).	76
CALCulate<1 2>:MARKer<1...4>:X:SLIMits:RIGHT <0...MAX> (frequency sweep time)	77
CALCulate<1 2>:MARKer<1...4>:COUNt ON OFF.	77
CALCulate<1 2>:MARKer<1...4>:COUNt:RESolution 0.1 1 10 100 1000 10000 Hz	77
CALCulate<1 2>:MARKer<1...4>:COUNt:FREQUency?	78
CALCulate<1 2>:MARKer<1...4>:LOEXclude ON OFF	78
CALCulate<1 2>:MARKer<1...4>:Y?	78
CALCulate<1 2>:MARKer<1...4>:Y:PERCent	78
CALCulate<1 2>:MARKer<1...4>:MAXimum[:PEAK]	79
CALCulate<1 2>:MARKer<1...4>:MAXimum:NEXT	79
CALCulate<1 2>:MARKer<1...4>:MAXimum:LEFT	79
CALCulate<1 2>:MARKer<1...4>:MAXimum:RIGHT	79
CALCulate<1 2>:MARKer<1...4>:MAXimum:AUTO ON OFF.	80
CALCulate<1 2>:MARKer<1...4>:MINimum:AUTO ON OFF	80
CALCulate<1 2>:MARKer<1...4>:MINimum[:PEAK].	81
CALCulate<1 2>:MARKer<1...4>:MINimum:NEXT	81

CALCulate<1 2>:MARKer<1...4>:MINimum:RIGHT	81
CALCulate<1 2>:MARKer<1...4>:MINimum:LEFT	81
CALCulate<1 2>:MARKer<1...4>:PEXCursion <numeric_value>	81
CALCulate:MARKer:FUNCTion Subsystem	82
CALCulate<1 2>:MARKer<1...4>:FUNCTion:NOISe[:STATe]	82
CALCulate<1 2>:MARKer<1...4>:FUNCTion:NOISe:RESult?	82
CALCulate<1 2>:MARKer<1...4>:FUNCTion:REFerence?	82
CALCulate:MARKer:FUNCTion:POWer Subsystem	83
CALCulate<1 2>:MARKer<1...4>:FUNCTion:POWer:SElect ACPower CPOWer MCACpower	
OBANdwidth OBWidth CN CN0	83
CALCulate<1 2>:MARKer<1...4>:FUNCTion:POWer:RESult? ACPower CPOWer MCACpower	
OBANdwidth OBWidth CN CN0	83
CALCulate<1 2>:MARKer<1...4>:FUNCTion:POWer:RESult:PHZ ON OFF	83
CALCulate:STATistics Subsystem	84
CALCulate:STATistics:APD[:STATe] ON OFF	84
CALCulate:STATistics:CCDF[:STATe] ON OFF	84
CALCulate:STATistics:CCDF:X<1...3>? P0_01 P0_1 P1 P10	85
CALCulate:STATistics:NSAMples <100...1E9>	85
CALCulate:STATistics:SCALE:AUTO ONCE	85
CALCulate:STATistics:SCALE:X:RLEVel -130dBm to 30dBm	85
CALCulate:STATistics:SCALE:X:RANGe 10dB to 200dB	86
CALCulate:STATistics:SCALE:Y:UPPer 1E-8 to 1.0	86
CALCulate:STATistics:SCALE:Y:LOWer 1E-9 to 0.1	87
CALCulate:STATistics:PRESet	87
CALCulate:STATistics:RESult<1...3>? MEAN PEAK CFACtor ALL	87
CALCulate:THReshold Subsystem	87
CALCulate<1 2>:DLINe<1 2> MINimum...MAXimum	88
CALCulate<1 2>:DLINe<1 2>:STATe ON OFF	88
CALCulate<1 2>:THReshold MINimum to MAXimum (depending on current unit)	88
CALCulate<1 2>:THReshold:STATe ON OFF	89
CALCulate<1 2>:FLINe<1 2> 0...fmax	89
CALCulate<1 2>:FLINe<1 2>:STATe ON OFF	89
CALCulate<1 2>:TLINe<1 2> 0...1000s	89
CALCulate<1 2>:TLINe<1 2>:STATe ON OFF	90
CALCulate:UNIT Subsystem	90
CALCulate<1 2>:UNIT:POWer DBM V A W DB PCT UNITLESS DBPW WATT DBUV	
DBMV VOLT DBUA AMPere	90
CALibration Subsystem	91
CALibration[:ALL]?	91
CALibration:STATe ON OFF	92
DISPlay Subsystem	92
DISPlay:FORMat SINGLE SPLit	92
DISPlay:ANNotation:FREQuency ON OFF	92
DISPlay:LOGO ON OFF	93
DISPlay:PSAVE[:STATe] ON OFF	93
DISPlay:PSAVE:HOLDoff 1 to 60	93
DISPlay:CMAp<1 to 26>:DEFault<1 2>	94
DISPlay:CMAp<1 to 26>:HSL <hue>,<sat>,<lum>	94
DISPlay:CMAp<1 to 26>:PDEFined BLACK BLUE BROWn GREen CYAN RED MAGenta	

YELLow WHITe DGRAY LGRAY LBLUe LGREen LCYan LRED LMAGenta	94
DISPlay[:WINDow<1 2>]:SElect	95
DISPlay[:WINDow<1 2>]:SIZE LARGe SMALl	95
DISPlay[:WINDow<1 2>]:TEXT[:DATA] <string>	96
DISPlay[:WINDow<1 2>]:TEXT:STATe ON OFF	96
DISPlay[:WINDow<1 2>]:TIME ON OFF	96
DISPlay[:WINDow<1 2>]:TRACe<1 to 3>:Y[:SCALe] 10dB to 200dB	97
DISPlay[:WINDow<1 2>]:TRACe<1 to 3>:Y[:SCALe]:MODE ABSolute RELativ	97
DISPlay[:WINDow<1 2>]:TRACe<1 to 3>:Y[:SCALe]:RLEVel	97
DISPlay[:WINDow<1 2>]:TRACe<1 to 3>:Y[:SCALe]:RLEVel:OFFSet	98
DISPlay[:WINDow<1 2>]:TRACe<1 to 3>:Y[:SCALe]:PDIVision	98
DISPlay[:WINDow<1 2>]:TRACe<1 to 3>:Y.SPACing LInear LOGarithmic LDB	99
DISPlay[:WINDow<1 2>]:TRACe<1...3>:MODE WRITe VIEW AVERAge MAXHold MINHold 99	
DISPlay[:WINDow<1 2>]:TRACe<1 to 3>[:STATe] ON OFF	100
FORMat Subsystem	100
FORMat[:DATA].	100
HCOPy Subsystem	100
HCOPy:ABORt.	100
HCOPy:CMAP<1 to 26>:DEFault1 2 3	101
HCOPy:CMAP<1 to 26>:HSL <hue>,<sat>,<lum>	101
HCOPy:CMAP<1 to 26>:PDEFined BLACK BLUE BROWn GREen CYAN RED MAGenta YELLow WHITe DGRAY LGRAY LBLUe LGREen LCYan LRED LMAGenta	101
HCOPy:DESTination<1 2> <string>	102
HCOPy:DEVice:COLor ON OFF	102
HCOPy:DEVice:LANGuage<1 2> GDI WMF EWMF BMP	102
HCOPy[:IMMediate<1 2>].	102
HCOPy:ITEM:ALL	103
HCOPy:ITEM:WINDow<1 2>:TABLe:STATe ON OFF	103
HCOPy:ITEM:WINDow<1 2>:TEXT <string>	103
HCOPy:ITEM:WINDow<1 2>:TRACe:STATe ON OFF	104
HCOPy:PAGE:ORientation<1 2> LANDscape PORTRait	104
INITiate Subsystem	104
INITiate<1 2>:CONTinuous ON OFF	104
INITiate<1 2>:CONMeas	105
INITiate<1 2>[:IMMediate]	105
INPut Subsystem	105
INPut<1 2>:ATTenuation	105
INPut<1 2>:ATTenuation:AUTO	105
INPut<1 2>:EATT	106
INPut<1 2>:EATT:AUTO	106
INPut<1 2>:EATT:STATe	107
INPut<1 2>:IMPedance	107
INPut<1 2>:GAIN:STATe	107
INSTrument Subsystem	108
INSTrument[:SElect] SANalyzer DDEMod ADEMod MGSM WCDPower BWCDpower MWCDpower BC2K BDO	108
INSTrument:NSElect <numeric value>.	108
MMEMory Subsystem	109

MMEMory:CATalog? <path>	110
MMEMory:CDIRectory <directory_name>	110
MMEMory:COpy <file_source>,<file_destination>	110
MMEMory:DATA <file_name>[,<block_data>]	110
MMEMory:DElete <file_name>	111
MMEMory:INItialize <msus>	111
MMEMory:LOAD:STATe 1,<file_name>	111
MMEMory:LOAD:AUTO 1,<file_name>	111
MMEMory:MDIRectory <directory_name>	112
MMEMory:MOVE <file_source>,<file_destination>	112
MMEMory:MSIS <device>	112
MMEMory:NAME <file_name>	112
MMEMory:RDIRectory <directory_name>	113
MMEMory:STORE<1 2>:STATe 1,<file_name>	113
MMEMory:STORE<1 2>:TRACe <1..6>,<file_name>	113
MMEMory:CLear:STATe 1,<file_name>	113
MMEMory:CLear:ALL	113
MMEMory:SElect[:ITEM]:HWSettings ON OFF	114
MMEMory:SElect[:ITEM]:TRACe[:ACTive] ON OFF	114
MMEMory:SElect[:ITEM]:LINes:ALL ON OFF	114
MMEMory:SElect[:ITEM]:SCData ON OFF	115
MMEMory:SElect[:ITEM]:TRANsducer:ALL ON OFF	115
MMEMory:SElect[:ITEM]:ALL	115
MMEMory:SElect[:ITEM]:NONE	115
MMEMory:SElect[:ITEM]:DEFault	116
MMEMory:COMMeNt <string>	116
SENSE:AVErAge Subsystem	116
[SENSe<1 2>:]AVErAge:COUNt 0 to 32767	116
[SENSe<1 2>:]AVErAge[:STATe<1 to 3>] ON OFF	117
[SENSe<1 2>:]AVErAge:TYpe VIDEo LINear	117
SENSE:BANdwidth Subsystem	117
[SENSe<1 2>:]BANdwidth BWIDth[:RESolution] <numeric_value>	117
[SENSe<1 2>:]BANdwidth BWIDth[:RESolution]:AUTO ON OFF	118
[SENSe<1 2>:]BANdwidth BWIDth[:RESolution]:RATio 0.0001 to 1	118
[SENSe<1 2>:]BANdwidth BWIDth[:RESolution]:TYpe NORMal FFT CFILter RRC	119
[SENSe<1 2>:]BANdwidth BWIDth:VIDeo 1Hz to 10MHz	119
[SENSe<1 2>:]BANdwidth BWIDth:VIDeo:AUTO ON OFF	119
[SENSe<1 2>:]BANdwidth BWIDth:VIDeo:RATio 0.01 to 1000	120
SENSE:CORRection Subsystem	120
[SENSe<1 2>:]CORRection:EGAI:n:INPut[:MAGNitude] -200...200dB	120
SENSE:DETEctor Subsystem	121
[SENSe<1 2>:]DETEctor<1..3>[:FUNction] APEak NEGative POSitive SAMple RMS AVErAge QPEak	121
[SENSe<1 2>:]DETEctor<1 to 3>[:FUNction]:AUTO ON OFF	121
SENSE:FREQuency Subsystem	122
[SENSe<1 2>:]FREQuency:CENTer 0 to fmax	122
[SENSe<1 2>:]FREQuency:CENTer:STEP 0 to fmax	122
[SENSe<1 2>:]FREQuency:CENTer:STEP:LINK SPAN RBW OFF	122
[SENSe<1 2>:]FREQuency:CENTer:STEP:LINK:FACTor 1 to 100 PCT	123

[SENSe<1 2>]:FREQUency:SPAN 0 to fmax	123
[SENSe<1 2>]:FREQUency:SPAN:FULL	124
[SENSe<1 2>]:FREQUency:STARt 0 to fmax	124
[SENSe<1 2>]:FREQUency:STOP 0 to fmax	124
[SENSe<1 2>]:FREQUency:MODE CW FIXEd SWEep	125
[SENSe<1 2>]:FREQUency:OFFSet	125
SENSe:POWer Subsystem	125
[SENSe<1 2>]:POWer:ACHannel:SPACing:CHANnel 100 Hz to 2000 MHz	125
[SENSe<1 2>]:POWer:ACHannel:SPACing:ACHannel 100 Hz to 2000 MHz	126
[SENSe<1 2>]:POWer:ACHannel:SPACing:ALTErnate<1 2> 100 Hz to 2000 MHz	126
[SENSe<1 2>]:POWer:ACHannel:TXCHannel:COUNt 1 2 3 4	126
[SENSe<1 2>]:POWer:ACHannel:ACPairs 0 1 2 3	127
[SENSe<1 2>]:POWer:ACHannel:BANDwidth BWIDth[:CHANnel] 100 Hz to 1000 MHz	127
[SENSe<1 2>]:POWer:ACHannel:BANDwidth BWIDth:ACHannel 100Hz–1000MHz	128
[SENSe<1 2>]:POWer:ACHannel:BANDwidth BWIDth:ALTErnate<1 2> 100 Hz to 1000 MHz	128
[SENSe<1 2>]:POWer:ACHannel:MODE ABSolute RELative	128
[SENSe<1 2>]:POWer:ACHannel:REFerence:AUTO ONCE	129
[SENSe<1 2>]:POWer:ACHannel:REFerence:TXCHannel:AUTO MINimum MAXimum LHIGhest 129	
[SENSe<1 2>]:POWer:ACHannel:REFerence:TXCHannel:MANual 1 2 3 4	129
[SENSe<1 2>]:POWer:ACHannel:PRESet ACPower CPOWer MCACpower OBANDwidth OBWidth CN CN0	130
[SENSe<1 2>]:POWer:ACHannel:PRESet:RLEVel	130
[SENSe<1 2>]:POWer:BANDwidth BWIDth 10 to 99.9PCT	130
[SENSe<1 2>]:POWer:HSPeed ON OFF	131
[SENSe<1 2>]:POWer:NCORrection ON OFF	131
[SENSe<1 2>]:POWer:TRACe 1 to 3	131
SENSe:ROSCillator Subsystem	132
[SENSe<1 2>]:ROSCillator:SOURce INTernal EXTErnal	132
SENSe:SWEep Subsystem	132
[SENSe<1 2>]:SWEep:TIME 2.5ms to 16000s (frequency domain) 1us to 16000s (time domain)	132
[SENSe<1 2>]:SWEep:TIME:AUTO ON OFF	133
[SENSe<1 2>]:SWEep:COUNt 0 to 32767	133
[SENSe<1 2>]:SWEep:EGATE ON OFF	133
[SENSe<1 2>]:SWEep:EGATE:TYPE LEVel EDGE	134
[SENSe<1 2>]:SWEep:EGATE:POLarity POSitive NEGative	134
[SENSe<1 2>]:SWEep:EGATE:HOLDoff 125 ns to 100 s	134
[SENSe<1 2>]:SWEep:EGATE:LENGth 0 to 100 s	135
[SENSe<1 2>]:SWEep:EGATE:SOURce EXTErnal IFPower RFPower	135
[SENSe<1 2>]:SWEep:POINts 125 to 8001	135
STATus Subsystem	136
STATus:OPERation[:EVENT]?	136
STATus:OPERation:CONDition?	136
STATus:OPERation:ENABle	136
STATus:OPERation:PTRansition	137
STATus:OPERation:NTRansition	137
STATus:PRESet	137
STATus:QUEStionable[:EVENT]	138
STATus:QUEStionable:CONDition	138

STATus:QUEStionable:ENABle	138
STATus:QUEStionable:PTRansition	139
STATus:QUEStionable:NTRansition	139
STATus:QUEStionable:FREQuency[:EVENT].	139
STATus:QUEStionable:FREQuency:CONDition	139
STATus:QUEStionable:FREQuency:ENABle	140
STATus:QUEStionable:FREQuency:PTRansition	140
STATus:QUEStionable:FREQuency:NTRansition	140
STATus:QUEue[:NEXT]?	141
SYSTem Subsystem	141
SYSTem:COMMunicate:GPIB[:SELF]:ADDRess	141
SYSTem:DATE	142
SYSTem:DISPlay:FPANel ON OFF	142
SYSTem:DISPlay:UPDate	142
SYSTem:ERRor?	142
SYSTem:ERRor:LIST?	143
SYSTem:ERRor:CLEar:ALL	143
SYSTem:PRESet	143
SYSTem:TIME	143
SYSTem:VERSion	143
TRACe Subsystem	144
TRACe<1 2>[:DATA] TRACE1 TRACE2 TRACE3 SPURious ABITstream PWCDp CTABLE, <block> <numeric_value>	144
TRACe:IQ Subsystem	144
TRACe<1 2>:IQ:DATA	144
TRACe<1 2>:IQ:SET <filter type>,<rbw>,<sample rate>,<trigger source>,<trigger slope>,<pretrigger samples>,<# of samples>	145
TRACe:IQ:SRATe 15.625kHz to 32MHz	145
TRACe:IQ:STATe ON OFF	145
TRIGger Subsystem	145
TRIGger<1 2>[:SEQUence]:SOURce	146
TRIGger<1 2>[:SEQUence]:LEVel:IFPower -30 to -10DBM	146
TRIGger<1 2>[:SEQUence]:LEVel:RFPower	146
TRIGger<1 2>[:SEQUence]:LEVel:VIDeo 0 to 100PCT	147
TRIGger<1 2>[:SEQUence]:HOLDoff	147
TRIGger<1 2>[:SEQUence]:SLOPe	147
4. A Brief Introduction to the SCPI Language	
Overview	150
SCPI Language Basics	150
Command Syntax	151
Command Statement Rules Overview	152
Creating Valid Commands	152
Special Characters in Commands	153
Parameters in Commands	153
Variable Parameters	155
Block Program Data	155
Putting Multiple Commands on the Same Line	156
SCPI Termination and Separator Syntax	156

This chapter gives a quick overview of the N9062A application on Agilent X-series analyzers.

What Does the Agilent N9062A SCPI Language Compatibility Application Do?

The Agilent N9062A SCPI Language Compatibility application is designed to allow X-Series signal analyzers (including EXA, MXA and PXA) to be controlled using the SCPI programming commands (as known as SCPI commands, or SCPI) from the following Rohde & Schwarz (R&S) instruments:

- R&S FSE
- R&S FSP
- R&S FSU

The X-Series analyzer with the N9062A application installed is able to emulate the FSE/FSP/FSU measurement functions as closely as possible in many automated systems with minimal or no modification to the currently used measurement software.

The N9062A supports over 278 SCPI commands from the FSE/FSP/FSU instruments and 16 of IEEE488.2 standard SCPI commands.

NOTE

The N9062A supports commands from the Spectrum Analyzer (default) mode. Commands from other options or modes are not supported.

The N9062A does not support the same front panel operations, user interface and measurement screen display as the FSE/FSP/FSU instruments.

The N9062A does not support the Split Screen feature.

The N9062A partly supports Channel Power measurements, OBW measurements, ACP measurements and IQ traces.

Documentation for the N9062A application

Signal Analyzers with N9062A

When you purchase your X-Series signal analyzer with the N9062A SCPI Language Compatibility, this manual - the *SCPI Language Compatibility Guide* (N9062-90001) is included on the documentation CD and is installed on the analyzer in the online help.

For information on PXA series analyzers and other related documentation, refer to the PXA web site at <http://www.agilent.com/find/pxa>.

For information on MXA series analyzers and other related documentation, refer to the MXA web site at <http://www.agilent.com/find/mxa>.

For information on EXA series analyzers and other related documentation, refer to the EXA web site at <http://www.agilent.com/find/exa>.

This *SCPI Language Compatibility Guide* is not designed to be a comprehensive guide to all SCPI commands. It gives brief descriptions of the supported commands, and highlights important functional or behavioral differences that you should be aware of when transferring your existing code to your X-Series analyzer. For a fuller description of these SCPI commands, refer to the manuals supplied with your original analyzer.

Signal Analyzer Updates

For the latest information about this instrument, including software upgrades, application information, and product information, please visit the URL above.

N9062A General Rules and Limitations

The N9062A application has been designed to emulate the remote operation of the R&S FSE/FSP/FSU instruments for the most commonly used commands needed for basic spectrum analyzer functions. It is not an absolute direct replacement for these instruments in all cases, and does not support instrument modes beyond the core spectrum analyzer. It also does not emulate the local front-panel interface (buttons) and display as used for manual operation.

AC/DC Coupling

The N9062A application will default to DC coupling on preset.

Couplings

The N9062A application uses the auto coupling features of the X-Series analyzers, and may be slightly different from the FSE/FSP/FSU. To eliminate the possibilities of "Meas Uncal" errors between auto and manual values, values will default to the X-Series auto settings where applicable (for example resolution bandwidth).

The parameter's default value is adjusted to match that of the FSE/FSP/FSU unless otherwise stated.

Data Formats

N9062A application supports data formats defined for the X-Series analyzers. For more information, see *SA Programmer's Guide*.

Markers

The N9062A application emulates the behavior of FSE/FSP/FSU instruments.

Numeric Ranges

Numeric ranges are limited to that of the X-Series analyzers unless otherwise stated. For more information about X-Series numeric ranges, see *SA Programmer's Guide*.

Remote Control

The N9062A application in X-Series signal analyzers supports remote operation through the following interfaces: GPIB, LAN, USB or Telnet.

Returning Data

Data returned in response to a command or query is defined in the following [Chapter 3, "Compatible SCPI Commands," on page 33](#), unless otherwise stated. For more information about X-Series data returning, see *SA Programmer's Guide*.

The data returned from queries, especially from a trace query, match the FSE/FSP/FSU for the most common data formats (ASCII, Binary 32).

Units

The N9062A application supports the following standard units. Any units supported by FSE/FSP/FSU commands that are not supported by the N9062A are noted in the table and details found in [Chapter 3](#), “Compatible SCPI Commands,” on page 33.

Frequency Units: HZ, KHZ, MHZ, GHZ

Amplitude Units: DB, DBM, DBUV, DBMV, DBUA, V, A, W

Time Units: KS, S, MS, US, NS

Status System & SRQ

The STATus subsystem is a remote-only command structure that allows a controller to determine the operating status of the instrument. The system can also allow the instrument to request service from the controller on an interrupt basis. The system is a complex hierarchy, where specific details of an instrument’s condition are summarized upward to higher level registers. The use of this system is optional, and your programs may vary widely in the depth to which you use it.

The topmost Status Byte Register, Standard Event Status Register, and their SRQ behavior, are specified by IEEE 488.2 standard. The N9062A application adopts the Status system of the Agilent X-Series analyzers, which conforms to the IEEE 488.2 standard and behaves like that of any IEEE488.2-compliant instrument (including the FSE, FSP, or FSU). These status conditions, events, or commands should emulate any compliant instrument, for example:

- operation complete, *OPC and *OPC?
- Status Byte bits, including Error/Event reporting, plus *CLS, *STB
- Standard Event Status Register bits, plus *ESE, *ESR, *PSC
- SRQ behavior and *SRE mask
- most error conditions (top level), plus STATus:QUEue? and SYSTem:ERRor? queries
- use of STATus:OPERation and STATus:QUEStionable commands, although specific bits may be mapped differently

At lower levels of details, different vendors have implemented their status systems differently. Therefore, if a program (written for FSE/FSP/FSU) makes deep or sophisticated use of the status subsystem, it may not run correctly with N9062A and so require modification. The top differences include (but are not limited to) the following:

- RF & IF Overload condition
- parallel polling, IST-flag (Individual Status), and PPE (Parallel Poll Enable) mask
- limit or limit margin mask failure reporting
- sweep break feature (of a multi-sweep measurement)
- specific error message numbers, specific error message text
- status conditions associated with Screen B (specifically)

For more information about the X-Series Status system, see *SA Programmer’s Guide*.

Supported Commands and Queries

Only a subset of the FSE/FSP/FSU series commands is supported in the N9062A application. The list of supported commands was determined by feedback from our customers combined with technical considerations and constraints.

The supported FSE/FSP/FSU commands are summarized with short notes in [Table 3-1 on page 34](#). In some cases, a command is accepted without error, but no action taken, for example, a display line.

Unsupported Commands and Queries

If a command or query is valid for FSE/FSP/FSU products but not supported by the N9062A application, the X-Series analyzer may handle it in the following way:

- Generate an error message. This kind of SCPI commands is not covered in this manual.

NOTE	If your use case is not fully supported, contact your local Agilent Technologies Sales and Service Office, or if in the United States, call 1-800-829-4444.
-------------	---

Hardware and Firmware Requirements for N9062A

One of the following Agilent signal analyzers is required to run the N9062A application.

Table 1-1 Compatible Agilent PXA Series Signal Analyzers

Analyzer Model Number	Upper Frequency Limit	Firmware
N9030A-503	3.6 GHz	Rev A.06.00 or later
N9030A-508	8.4 GHz	Rev A.06.00 or later
N9030A-513	13.6 GHz	Rev A.06.00 or later
N9030A-526	26.5 GHz	Rev A.06.00 or later

Table 1-2 Compatible Agilent MXA Series Signal Analyzers

Analyzer Model Number	Upper Frequency Limit	Firmware
N9020A-503	3.6 GHz	Rev A.06.00 or later
N9020A-508	8.4 GHz	Rev A.06.00 or later
N9020A-513	13.6 GHz	Rev A.06.00 or later
N9020A-526	26.5 GHz	Rev A.06.00 or later

Table 1-3 Compatible Agilent EXA Series Signal Analyzers

Analyzer Model Number	Upper Frequency Limit	Firmware
N9010A-503	3.6 GHz	Rev A.06.00 or later
N9010A-507	7 GHz	Rev A.06.00 or later
N9010A-513	13.6 GHz	Rev A.06.00 or later
N9010A-526	26.5 GHz	Rev A.06.00 or later

For maximum compatibility, you should select an X-Series analyzer that equals or exceeds the frequency range of the FSE/FSP/FSU analyzer you are replacing.

Installing the N9062A application

N9062A SCPI Language Compatibility for the R&S FSE/FSP/FSU analyzers is a licensed application on the X-Series analyzers. The application must be installed and licensed on the X-Series analyzer, PXA, MXA and EXA, for it to work correctly.

License for N9062A

License N9062A-2FP is required to provide SCPI language compatibility on Agilent X-Series analyzers.

Installation

The license is installed on the X-Series analyzer in one of the following ways:

- If you purchased a new X-Series analyzer with the N9062A application then the product is installed and licensed and ready to use.
- If you have an X-Series analyzer and have subsequently purchased the N9062A application then you can download the N9062A application from the Agilent website. The N9062A application is installed as part of a software upgrade. See the link below for Signal Analyzers software upgrade site. After upgrading your software you should then use your entitlement certificate to license the product (see Licensing below).

The latest revision of the software may be downloaded from:

http://www.agilent.com/find/pxa_software

http://www.agilent.com/find/mxa_software

http://www.agilent.com/find/exa_software

NOTE No calibration is required after the N9062A application is installed.

Licensing

When you order a licensed product, you receive an entitlement certificate. Instructions are provided on your entitlement certificate to direct you to the Web site to redeem your certificate for a license key. You need to provide your instrument product number and serial number, and the entitlement certificate number.

Required Information:	Front Panel Key Path:
Model #: (Ex. N9020A)	
Instrument Serial Number: _____	System > Show > System

The license is downloaded from the license website onto a USB storage device so that it can be loaded

into the instrument.

A license key is usually for one instrument model/serial number combination. The license key can only be installed on that instrument.

License Installation Procedure over USB

1. Redeem the Option Upgrade Entitlement Certificate by following the instructions on the Certificate.
2. After redeeming your Option Upgrade Entitlement Certificate you will receive an e-mail with an attached License File.
3. Locate a USB storage device and save the .lic file to the root directory of the USB storage device.
4. Connect the USB storage device to one of the signal analyzer USB ports. Windows detects the new hardware and may display the configuration menu. This menu may be configured according to your preferences.
5. The signal analyzer automatically consumes the license file. (This may take a few minutes) When the license is consumed the Agilent License Manager displays a “Successful License Installation” message.
6. Alternatively the license file can be manually installed over USB or LAN by placing the license file in the following folder on the signal analyzer.
C:\Program Files\Agilent\licensing

Verify the Installation

1. Press **System > Show > System** to display the list of installed applications.
2. Verify that the new application appears in the list.

If you require further assistance, please contact the Agilent support team.

Online assistance: <http://www.agilent.com/find/assist>

If you do not have access to the Internet, contact your local Agilent Technologies Sales and Service Office, or if in the United States, call 1-800-829-4444.

Running Software that Requires SCPI Commands

You only have access to a subset of SCPI commands from FSE/FSP/FSU instruments. The SCPI commands supported by X-Series analyzers are listed in [Table 3-1 on page 34](#).

If you are not familiar with the SCPI language, [Chapter 4](#), “A Brief Introduction to the SCPI Language,” on [page 149](#) contains some useful information.

Service and Calibration

Since the Performance Verification and Adjustment Software uses the SCPI command language, you need to exit the N9062A (SCPI Language Compatibility) application and change to the Spectrum Analyzer application prior to calibration or service of your Agilent X-Series analyzer.


This chapter describes the available front panel operations that are supported by the N9062A application on the Agilent X-series analyzers including EXA, MXA and PXA.

Selecting SCPI Language Compatibility Mode

NOTE Option N9062A is required to use the SCPI Language Compatibility mode.

To select the SCPI Language Compatibility mode, press the **Mode** hardkey on the X-Series analyzer front panel and then select the **SCPI Language Compatibility** mode. If there are more than six modes on the signal analyzer, then use the More button to find the **SCPI Language Compatibility** mode selection. [Figure 2-1](#) is an example mode menu map showing the N9062A application installed.

Figure 2-1 Example Mode Menu Map for X-Series Analyzers



Alternatively, you can select this mode using following SCPI commands.

Key Path:	Front-panel key
Remote Command	:INST SCPIRLC :INST:NSElect 270
Initial S/W Revision:	A.06.00 or later

Mode Setup

NOTE License N9062A-2FP is required for emulating FSE, FSP or FSU.


Allows you to adjust the mode setup configuration for current mode.

Key Path:	Front-panel key
Initial S/W Revision:	A.06.00 or later

Selecting Compatibility Model

To select the compatibility model you wish to emulate, press the **Mode Setup** hardkey on the front panel and then press **Compatibility Model**. [Figure 2-2](#) shows the menu map that allows you to select the Compatibility Mode from FSE, FSP and FSU.

Figure 2-2 Mode Setup > Compatibility Mode Selection Menu Map



NOTE When the **Mode IDN Response** selects Compatibility Model, it does not affect the response to the command “ID?” which always returns the model number of your X-Series analyzer, like “N9020A”.

The selected compatibility model determines the response to the “IDN?” command. Setting the Compatibility Model to ‘R&S FSE’, ‘R&S FSP’ or ‘R&S FSU’ affects the response to the SCPI command ‘*IDN?’ This command will return the model number and firmware version number of the corresponding instrument.

Alternatively, you can select the desired compatibility mode using the following SCPI command.

Key Path:	Mode Setup
-----------	-------------------

Front Panel Operations
Mode Setup

Remote Command	[:SENSE] :RLC:TYPE FSE FSP FSU [:SENSE] :RLC:TYPE?
Example	RLC:TYPE FSU RLC:TYPE?
Preset	FSP
State Saved	Saved in instrument state.
Range	FSE FSP FSU
Notes	If you don't install the N9062A-2FP license, SCPI Language Compatibility mode is not available.
Initial S/W Revision:	A.06.00 or later

R&S FSE

Selects emulation of FSE.

Key Path:	Mode Setup, Compatibility Model
Example:	RLC:TYPE FSE
Initial S/W Revision:	A.06.00 or later

R&S FSP

Selects emulation of FSP.

Key Path:	Mode Setup, Compatibility Model
Example:	RLC:TYPE FSP
Initial S/W Revision:	A.06.00 or later

R&S FSU

Selects emulation of FSU.

Key Path:	Mode Setup, Compatibility Model
Example:	RLC:TYPE FSU
Initial S/W Revision:	A.06.00 or later


Selecting Mode IDN Response

NOTE This selection is active only within the SCPI Language Compatibility mode.

To select the response to the command '*IDN?' as what you desire, press the **Mode Setup** hardkey on

the front panel and then press **Mode IDN Response**. Figure 2-3 shows the menu map that allows you to select the response type from the list.

Figure 2-3 Mode Setup > Mode IDN Response Selection Menu Map



The available selections are as follows:

Compatibility

Model Sets the response to the command `*IDN?` to the model number and firmware version number that you select from the Compatibility Model list, like `Rohde&Schwarz , FSP-26 , 123456/789 , 04 . 00`.

System IDN

Response Sets the response to the command `*IDN?` to the model number and firmware version number of your X-Series analyzer, like `Agilent Technologies , N9020A , US04000119 , A . 06 . 00 _ R0008`.

User

Sets the response to the command `*IDN?` to the name that you want. To press **User**, enter a name for your case and press **Done**, you will get the name when you read from the command `*IDN?`, like `My Analyzer`.

By default, the response is set the same as **System IDN Response**. If you don't enter a name, the response is also set the same as **System IDN Response**.

NOTE The **System IDN Response** selection refers back to the system-level IDN response as selected under the **System** key (either **Factory** or **User**). These selections apply to the entire instrument, across all modes.

Alternatively, you can select the desired response using the following SCPI command.

Key Path:	Mode Setup
Remote Command:	<code>[:SENSE] :RLC :IDN :TYPE SYSTem COMPatibility USER</code> <code>[:SENSE] :RLC :IDN :TYPE ?</code>

Front Panel Operations
Mode Setup

Example:	SENS:RLC:IDN:TYPE SYSTem RLC:IDN:TYPE?
State Saved:	Saved in instrument state.
Range:	SYSTem COMPatibility USER
Initial S/W Revision:	A.06.00 or later

Compatibility Model

Sets the response to *IDN command as the emulated instrument, like Rohde&Schwarz, FSP-26, 123456/789, 04.00.

Key Path:	Mode Setup, Mode IDN Response
Example:	RLC:IDN:TYPE COMP
Initial S/W Revision:	A.06.00 or later

System IDN Response

Sets the response to *IDN command as the analyzer system, like “N9020A”.

Key Path:	Mode Setup, Mode IDN Response
Example:	RLC:IDN:TYPE SYST
Initial S/W Revision:	A.06.00 or later

User IDN Response

Allows you to set or get the user-defined mode IDN response.

If you don't enter anything, the system IDN response will be used instead.

Key Path:	Mode Setup
Remote Command:	[:SENSE] :RLC:IDN:USER <string> [:SENSE] :RLC:IDN:USER?
Example:	RLC:IDN:USER “abc” RLC:IDN:USER?
State Saved:	Saved in instrument state.
Initial S/W Revision:	A.06.00 or later

Restore Mode Defaults

Resets the state for the currently active mode by resetting the mode persistent setting to their default values and by performing a mode preset. This function will never cause a mode switch. This function performs a full preset on the active mode.

Send Commands

After you finish setting up the N9062A application, you can send FSE/FSP/FSU commands to X-Series analyzers.

It is recommended to use Agilent 82357B USB/GPIB interface and free instrument to PC connection tool Agilent IO Libraries Suite when you connect the instrument to a PC. For more details, visit:

<http://www.agilent.com/find/82357B>

<http://www.agilent.com/find/iolib>

NOTE The commands can be sent via a GPIB, LAN, USB, or Telnet connection that your X-Series analyzer can support.

This chapter lists the IEEE standard SCPI commands and the FSE/FSP/FSU SCPI commands that are supported by the N9062A application on the Agilent X-series analyzers including EXA, MXA and PXA and gives brief details on the parameters of the commands.

It's assumed that you have previous experience in programming SCPI. This chapter is not intended to teach you everything about the SCPI programming language.

If you are not familiar with the SCPI language, [Chapter 4](#), "A Brief Introduction to the SCPI Language," on page 149 contains some useful information.

If you are using other optional programming compatibility modes, you should refer to the manual that came with the option.

NOTE

Most FSE/FSP/FSU SCPI commands have a numeric suffix at the first node, corresponding to the Split Screen A/B feature, for example, CALCulate<1|2>, INITiate<1|2> and so on. However, the N9062A does not support suffix "2". That is, Full Screen is assumed, and all commands are mapped as if to parameter <1>. Besides, the numeric suffix of "1" can be omitted, for example, CALCulate1 is equal to CALCulate.

If you send a command with the numeric suffix of "2", an error message will be displayed in the X-Series message area.

SCPI Summary Table

This table is a brief summary of the FSE/FSP/FSU SCPI commands covered by the N9062A SCPI Language Compatibility application in the X-Series analyzers, with short notes on operation. For more details, see “Supported SCPI Commands” on page 51.

Table 3-1 SCPI Commands Summary Table

SCPI Command	Corresponding Model	Support	Notes
Abort	FSE/FSP/FSU	Yes	
CALCulate<1 2>:DELTamarker<1...4>[:STATe] ON OFF	FSE/FSP/FSU	Yes	All DELTmarker functions are not supported in CCDF and APD measurements.
CALCulate<1 2>:DELTamarker<1...4>:MODE ABSolute RELative	FSE/FSP/FSU	Yes	
CALCulate<1 2>:DELTamarker<1...4>:AOFF	FSE/FSP/FSU	Yes	
CALCulate<1 2>:DELTamarker<1...4>:TRACe 1 to 3	FSE/FSP/FSU	Yes	
CALCulate<1 2>:DELTamarker<1...4>:X 0 to MAX (frequency sweep time)	FSE/FSP/FSU	Yes	
CALCulate<1 2>:DELTamarker<1...4>:X:RELati ve?	FSE/FSP/FSU	Yes	
CALCulate<1 2>:DELTamarker<1...4>:Y?	FSE/FSP/FSU	Yes	The only unit supported is dB.
CALCulate<1 2>:DELTamarker<1...4>:MAXimu m[:PEAK]	FSE/FSP/FSU	Yes	
CALCulate<1 2>:DELTamarker<1...4>:MAXimu m:LEFT	FSE/FSP/FSU	Yes	
CALCulate<1 2>:DELTamarker<1...4>:MAXimu m:NEXT	FSE/FSP/FSU	Yes	
CALCulate<1 2>:DELTamarker<1...4>:MAXimu m:RIGHT	FSE/FSP/FSU	Yes	
CALCulate<1 2>:DELTamarker<1...4>:MINimu m[:PEAK]	FSE/FSP/FSU	Yes	
CALCulate<1 2>:DELTamarker<1...4>:MINimu m:LEFT	FSE/FSP/FSU	Yes	
CALCulate<1 2>:DELTamarker<1...4>:MINimu m:NEXT	FSE/FSP/FSU	Yes	

Table 3-1 SCPI Commands Summary Table

SCPI Command	Corresponding Model	Support	Notes
CALCulate<1 2>:DELTamarker<1...4>:MINimum:RIGHT	FSE/FSP/FSU	Yes	
CALCulate<1 2>:DELTamarker<1...4>:LINK ON OFF	FSU	Yes	
CALCulate<1 2>:DELTamarker<1...4>:FUNCTION:FIXed[:STATe] ON OFF	FSE/FSP/FSU	Yes	
CALCulate<1 2>:DELTamarker<1...4>:FUNCTION:FIXed:RPOint:MAXimum[:PEAK] <numeric_value>	FSP/FSU	Yes	
CALCulate<1 2>:DELTamarker<1...4>:FUNCTION:FIXed:RPOint:Y <numeric_value>	FSE/FSP/FSU	Yes	
CALCulate<1 2>:DELTamarker<1...4>:FUNCTION:FIXed:RPOint:Y:OFFSet <numeric_value>	FSE/FSP/FSU	Yes	
CALCulate<1 2>:DELTamarker<1...4>:FUNCTION:FIXed:RPOint:X <numeric_value>	FSE/FSP/FSU	Yes	
CALCulate<1 2>:DELTamarker<1...4>:FUNCTION:PNOise[:STATe]	FSE/FSP/FSU	Yes	
CALCulate<1 2>:DELTamarker<1...4>:FUNCTION:PNOise:AUTO ON OFF	FSU	Yes	
CALCulate<1 2>:DELTamarker<1...4>:FUNCTION:PNOise:RESult?	FSE/FSP/FSU	Yes	
CALCulate<1 2>:LIMit<1...8>:ACTive?	FSE/FSU	Yes ^a	All LIMit functions are not supported in CCDF and APD measurements.
CALCulate<1 2>:LIMit<1...8>:TRACe 1 to 3	FSE/FSP/FSU	Yes	
CALCulate<1 2>:LIMit<1...8>:STATe ON OFF	FSE/FSP/FSU	Yes	
CALCulate<1 2>:LIMit<1...8>:UNIT DBM DBPW WATT DBUV DBMV VOLT DBUA AMPere DB DEG RAD S HZ PCT UNITLESS	FSE/FSP/FSU	Yes	The units supported are DBM, DBPW, WATT, DBUV, DBMV, VOLT, DBUA, and DB. All limit lines use the same unit.
CALCulate<1 2>:LIMit<1...8>:FAIL?	FSE/FSP/FSU	Yes	
CALCulate<1 2>:LIMit<1...8>:CLEar[:IMMedia te]	FSE/FSP/FSU	Yes	
CALCulate<1 2>:LIMit<1...8>:COMMENT <string>	FSE/FSP/FSU	Yes	

Table 3-1 SCPI Commands Summary Table

SCPI Command	Corresponding Model	Support	Notes
CALCulate<1 2>:LIMit<1...8>:COPY 1 to 8 <name>	FSE/FSP/FSU	Yes	Do not support the parameter option <name>
CALCulate<1 2>:LIMit<1...8>:NAME <name of limit line>	FSE/FSP/FSU	Yes	
CALCulate<1 2>:LIMit<1...8>:DELete	FSE/FSP/FSU	Yes	
CALCulate<1 2>:LIMit<1...8>:ESPectrum:MODE AUTO MANual USER	FSP	Accepted ^b	
CALCulate<1 2>:LIMit<1...8>:ESPectrum:VALue <numeric_value>	FSP	Accepted	
CALCulate<1 2>:LIMit<1...8>:ESPectrum:RESTore	FSP	Accepted	
CALCulate<1 2>:LIMit<1...8>:CONTRol[:DATA] <numeric_value>,<numeric_value>...	FSE/FSP/FSU	Yes	
CALCulate<1 2>:LIMit<1...8>:CONTRol:DOMai n FREQuency TIME	FSE/FSP/FSU	Yes	
CALCulate<1 2>:LIMit<1...8>:CONTRol:OFFSet <numeric_value>	FSE/FSP/FSU	Yes	
CALCulate<1 2>:LIMit<1...8>:CONTRol:MODE RELative ABSolute	FSE/FSP/FSU	Yes	
CALCulate<1 2>:LIMit<1...8>:CONTRol:SHIFt <numeric_value>	FSE/FSP/FSU	Yes	
CALCulate<1 2>:LIMit<1...8>:CONTRol:SPACin g LINear LOGarithmic	FSE/FSP/FSU	Yes	
CALCulate<1 2>:LIMit<1...8>:LOWer[:DATA] <numeric_value>,<numeric_value>...	FSE/FSP/FSU	Yes	The unit supported are DBM and DB.
CALCulate<1 2>:LIMit<1...8>:LOWer:STATe ON OFF	FSE/FSP/FSU	Yes	
CALCulate<1 2>:LIMit<1...8>:LOWer:OFFSet <numeric_value>	FSE/FSP/FSU	Yes	The unit supported are DBM and DB.
CALCulate<1 2>:LIMit<1...8>:LOWer:MARGIN <numeric_value>	FSE/FSP/FSU	Yes	The unit supported are DBM and DB.
CALCulate<1 2>:LIMit<1...8>:LOWer:MODE RELative ABSolute	FSE/FSP/FSU	Yes	
CALCulate<1 2>:LIMit<1...8>:LOWer:SHIFt <numeric_value>	FSE/FSP/FSU	Yes	The unit supported are DBM and DB.

Table 3-1 SCPI Commands Summary Table

SCPI Command	Corresponding Model	Support	Notes
CALCulate<1 2>:LIMit<1...8>:LOWer:SPACing LINear LOGarithmic	FSE/FSP/FSU	Yes	
CALCulate<1 2>:LIMit<1...8>:LOWer:THReshold <numeric_value>	FSP/FSU	Yes	The only unit supported is DBM.
CALCulate<1 2>:LIMit<1...8>:UPPer[:DATA] <numeric_value>,<numeric_value>...	FSE/FSP/FSU	Yes	The unit supported are DBM and DB.
CALCulate<1 2>:LIMit<1...8>:UPPer:STATe ON OFF	FSE/FSP/FSU	Yes	
CALCulate<1 2>:LIMit<1...8>:UPPer:OFFSet <numeric_value>	FSE/FSP/FSU	Yes	The unit supported are DBM and DB.
CALCulate<1 2>:LIMit<1...8>:UPPer:MARGin <numeric_value>	FSE/FSP/FSU	Yes	The unit supported are DBM and DB.
CALCulate<1 2>:LIMit<1...8>:UPPer:MODE RELative ABSolute	FSE/FSP/FSU	Yes	
CALCulate<1 2>:LIMit<1...8>:UPPer:SHIFt <numeric_value>	FSE/FSP/FSU	Yes	The unit supported are DBM and DB.
CALCulate<1 2>:LIMit<1...8>:UPPer:SPACing LINear LOGarithmic	FSE/FSP/FSU	Yes	
CALCulate<1 2>:LIMit<1...8>:UPPer:THReshold <numeric_value>	FSP/FSU	Yes	The only unit supported is DBM.
CALCulate<1 2>:MARKer<1...4>[:STATe] ON OFF	FSE/FSP/FSU	Yes	
CALCulate<1 2>:MARKer<1...4>:AOFF	FSE/FSP/FSU	Yes	
CALCulate<1 2>:MARKer<1...4>:TRACe 1 to 3	FSE/FSP/FSU	Yes	
CALCulate<1 2>:MARKer<1...4>:X 0 to MAX (frequency sweep time)	FSE/FSP/FSU	Yes	
CALCulate<1 2>:MARKer<1...4>:X:SLIMits[:STATe] ON OFF	FSE/FSP/FSU	Yes	
CALCulate<1 2>:MARKer<1...4>:X:SLIMits:LEFT 0 to MAX (frequency sweep time)	FSP/FSU	Yes	
CALCulate<1 2>:MARKer<1...4>:X:SLIMits:RIGHT 0 to MAX (frequency sweep time)	FSP/FSU	Yes	
CALCulate<1 2>:MARKer<1...4>:COUNT ON OFF	FSE/FSP/FSU	Yes	

Table 3-1 SCPI Commands Summary Table

SCPI Command	Corresponding Model	Support	Notes
CALCulate<1 2>:MARKer<1...4>:COUNT:RESolution 0.1 1 10 100 1000 10000 Hz	FSE/FSP/FSU	Yes	Support the parameter option: 1 10 100 1000 10000 100000 1000000 Hz
CALCulate<1 2>:MARKer<1...4>:COUNT:FREQuency?	FSE/FSP/FSU	Yes	
CALCulate<1 2>:MARKer<1...4>:LOEXclude ON OFF	FSE/FSP/FSU	Yes	
CALCulate<1 2>:MARKer<1...4>:Y?	FSE/FSP/FSU	Yes	
CALCulate<1 2>:MARKer<1...4>:Y:PERCent	FSP/FSU	Accepted	
CALCulate<1 2>:MARKer<1...4>:MAXimum[:PEAK]	FSE/FSP/FSU	Yes	
CALCulate<1 2>:MARKer<1...4>:MAXimum:NEXT	FSE/FSP/FSU	Yes	
CALCulate<1 2>:MARKer<1...4>:MAXimum:LEFT	FSE/FSP/FSU	Yes	
CALCulate<1 2>:MARKer<1...4>:MAXimum:RIGHT	FSE/FSP/FSU	Yes	
CALCulate<1 2>:MARKer<1...4>:MAXimum:AUTO ON OFF	FSU	Yes	
CALCulate<1 2>:MARKer<1...4>:MINimum:AUTO ON OFF	FSU	Yes	
CALCulate<1 2>:MARKer<1...4>:MINimum[:PEAK]	FSE/FSP/FSU	Yes	
CALCulate<1 2>:MARKer<1...4>:MINimum:NEXT	FSE/FSP/FSU	Yes	
CALCulate<1 2>:MARKer<1...4>:MINimum:RIGHT	FSE/FSP/FSU	Yes	
CALCulate<1 2>:MARKer<1...4>:MINimum:LEFT	FSE/FSP/FSU	Yes	
CALCulate<1 2>:MARKer<1...4>:PEXCursion <numeric_value>	FSE/FSP/FSU	Yes	The only unit supported is DB.
CALCulate<1 2>:MARKer<1...4>:FUNCTION:Noise[:STATE] ON OFF	FSE/FSP/FSU	Yes	
CALCulate<1 2>:MARKer<1...4>:FUNCTION:Noise:RESULT?	FSE/FSP/FSU	Yes	

Table 3-1 SCPI Commands Summary Table

SCPI Command	Corresponding Model	Support	Notes
CALCulate<1 2>:MARKer<1...4>:FUNction:POWer:SElect ACPower CPOWer MCACpower OBANdwidth OBWidth CN CN0	FSE/FSP/FSU	Yes	Support ACPower, MCACPower, CPOWer, OBANdwidth and OBWidth.
CALCulate<1 2>:MARKer<1...4>:FUNction:POWer:RESult? ACPower CPOWer MCACpower OBANdwidth OBWidth CN CN0	FSE/FSP/FSU	Yes	Support ACPower, MCACPower, CPOWer, OBANdwidth and OBWidth.
CALCulate:STATistics:APD[:STATe] ON OFF	FSP/FSU	Accepted	
CALCulate:STATistics:CCDF[:STATe] ON OFF	FSP/FSU	Yes	
CALCulate:STATistics:CCDF:X<1 to 3>? P0_01 P0_1 P1 P10	FSU	Yes	
CALCulate:STATistics:NSAMples 100 to 1E9	FSP/FSU	Yes	Min = 1000
CALCulate:STATistics:SCALE:AUTO ONCE	FSP/FSU	Accepted	
CALCulate:STATistics:SCALE:X:RLEVEL -130dBm to 30dBm	FSP/FSU	Accepted	
CALCulate:STATistics:SCALE:X:RANGE 10dB to 200dB	FSP/FSU	Accepted	
CALCulate:STATistics:SCALE:Y:UPPer 1E-8 to 1.0	FSP/FSU	Accepted	
CALCulate:STATistics:SCALE:Y:LOWer 1E-9 to 0.1	FSP/FSU	Accepted	
CALCulate:STATistics:PRESet	FSP/FSU	Accepted	
CALCulate:STATistics:RESult<1 to 3>? MEAN PEAK CFACtor ALL	FSP/FSU	Yes	
CALCulate<1 2>:DLINe<1 2> MINimum...MAXimum	FSE/FSP/FSU	Accepted	
CALCulate<1 2>:DLINe<1 2>:STATe ON OFF	FSE/FSP/FSU	Yes	
CALCulate<1 2>:THReshold MINimum to MAXimum (depending on current unit)	FSE/FSP/FSU	Yes	
CALCulate<1 2>:THReshold:STATe ON OFF	FSE/FSP/FSU	Yes	
CALCulate<1 2>:FLINe<1 2> 0...fmax	FSE/FSP/FSU	Accepted	
CALCulate<1 2>:FLINe<1 2>:STATe ON OFF	FSE/FSP/FSU	Accepted	
CALCulate<1 2>:TLINe<1 2> 0...1000s	FSE/FSP/FSU	Accepted	

Table 3-1 SCPI Commands Summary Table

SCPI Command	Corresponding Model	Support	Notes
CALCulate<1 2>:TLINe<1 2>:STATe ON OFF	FSE/FSP/FSU	Accepted	
CALCulate<1 2>:UNIT:POWer DBM V A W DB PCT UNITLESS DBPW WATT DBUV DBMV VOLT DBUA AMPere	FSE/FSP/FSU	Yes	Support DBM DBUV DBMV DBUA V A W.
CALibration[:ALL]?	FSE/FSP/FSU	Yes	
CALibration:STATe ON OFF	FSE/FSP/FSU	Yes	
DISPlay:FORMat SINGLE SPLit	FSE/FSP/FSU	Accepted	Most of screen display functions are not supported. Most of the DISPlay related commands are accepted, but take no action, unless otherwise stated.
DISPlay:ANNotation:FREQuency ON OFF	FSE/FSP/FSU	Accepted	
DISPlay:LOGO ON OFF	FSE/FSP/FSU	Accepted	
DISPlay:PSAVe[:STATe] ON OFF	FSE/FSP/FSU	Accepted	
DISPlay:PSAVe:HOLDoff 1 to 60	FSE/FSP/FSU	Accepted	
DISPlay:CMAP<1 to 26>:DEFault<1 2>	FSE/FSP/FSU	Accepted	
DISPlay:CMAP<1 to 26>:HSL <hue>,<sat>,<lum>	FSE/FSP/FSU	Accepted	
DISPlay:CMAP<1 to 26>:PDEFined BLACK BLUE BROWn GREen CYAN RED MAGenta YELLow WHITE DGRAY LGRAY LBLUe LGREen LCYan LRED LMAGenta	FSE/FSP/FSU	Accepted	
DISPlay[:WINDow<1 2>]:SElect	FSE/FSP/FSU	Accepted	
DISPlay[:WINDow<1 2>]:SIZE LARGe SMALl	FSP/FSU	Accepted	
DISPlay[:WINDow<1 2>]:TEXT[:DATA] <string>	FSE/FSP/FSU	Accepted	
DISPlay[:WINDow<1 2>]:TEXT:STATe ON OFF	FSE/FSP/FSU	Accepted	
DISPlay[:WINDow<1 2>]:TIME ON OFF	FSE/FSP/FSU	Accepted	
DISPlay[:WINDow<1 2>]:TRACe<1 to 3>:Y[:SCALe] 10dB to 200dB	FSE/FSP/FSU	Yes	

Table 3-1 SCPI Commands Summary Table

SCPI Command	Corresponding Model	Support	Notes
DISPlay[:WINDow<1 2>]:TRACe<1 to 3>:Y[:SCALe]:MODE ABSolute RELativ	FSE/FSP/FSU	Accepted	
DISPlay[:WINDow<1 2>]:TRACe<1 to 3>:Y[:SCALe]:RLEVel	FSE/FSP	Yes	
DISPlay[:WINDow<1 2>]:TRACe<1 to 3>:Y[:SCALe]:RLEVel:OFFSet	FSE/FSP/FSU	Yes	
DISPlay[:WINDow<1 2>]:TRACe<1 to 3>:Y[:SCALe]:PDIVision	FSE/FSP/FSU	Yes	The only unit supported is DB.
DISPlay[:WINDow<1 2>]:TRACe<1 to 3>:Y:SPACing LINear LOGarithmic LDB	FSE/FSP/FSU	Yes	Do not support the parameter option LDB.
DISPlay[:WINDow<1 2>]:TRACe<1 to 3>:MODE WRITe VIEW AVERAge MAXHold MINHold	FSE/FSP/FSU	Yes	
DISPlay[:WINDow<1 2>]:TRACe<1 to 3>[:STATe] ON OFF	FSE/FSP/FSU	Yes	
FORMat[:DATA]	FSE/FSP/FSU	Yes	The supported formats are X-Series specific: ASC,8 REAL,32 REAL,64 and INT,32
HCOPy:ABORt	FSE/FSP/FSU	Yes	HCOPy functions are not supported, except HCOPY:ABORT and HCOPY[:IMMEDIATE].
HCOPy:CMAP<1 to 26>:DEFault1 2 3	FSP/FSU	Accepted	All the printer setting commands are accepted, but take no action.
HCOPy:CMAP<1 to 26>:HSL <hue>,<sat>,<lum>	FSP/FSU	Accepted	
HCOPy:CMAP<1 to 26>:PDEFined BLACK BLUE BROWn GREen CYAN RED MAGenta YELLow WHITe DGRAY LGRAY LBLUe LGREen LCYan LRED LMAGenta	FSP/FSU	Accepted	
HCOPy:DESTination<1 2> <string>	FSE/FSP/FSU	Accepted	
HCOPy:DEVice:COLor ON OFF	FSE/FSP/FSU	Accepted	
HCOPy:DEVice:LANGUage<1 2> GDI WMF EWmf BMP	FSE/FSP/FSU	Accepted	

Table 3-1 SCPI Commands Summary Table

SCPI Command	Corresponding Model	Support	Notes
HCOPY[:IMMEDIATE<1 2>]	FSE/FSP/FSU	Yes	HCOPY functions are not supported, except HCOPIY:ABORT and HCOPIY[:IMMEDIATE].
HCOPY:ITEM:ALL	FSE/FSP/FSU	Accepted	
HCOPY:ITEM:WINDOW<1 2>:TABLE:STATE ON OFF	FSE/FSP/FSU	Accepted	
HCOPY:ITEM:WINDOW<1 2>:TEXT <string>	FSE/FSP/FSU	Accepted	
HCOPY:ITEM:WINDOW<1 2>:TRACE:STATE ON OFF	FSE/FSP/FSU	Accepted	
HCOPY:PAGE:ORIENTATION<1 2> LANDSCAPE PORTRAIT	FSE/FSP/FSU	Accepted	
INITIATE<1 2>:CONTINUOUS ON OFF	FSE/FSP/FSU	Yes	
INITIATE<1 2>:CONMEAS	FSE/FSP/FSU	Yes	
INITIATE<1 2>[:IMMEDIATE]	FSE/FSP/FSU	Yes	
INPUT<1 2>:ATTENUATION	FSE/FSP/FSU	Yes	Only support 2 dB steps.
INPUT<1 2>:ATTENUATION:AUTO	FSE/FSP/FSU	Yes	
INPUT<1 2>:EATT	FSP/FSU	Yes	Max = 24 dB
INPUT<1 2>:EATT:AUTO	FSP/FSU	Yes	
INPUT<1 2>:EATT:STATE	FSP/FSU	Yes	
INPUT<1 2>:IMPEDANCE	FSE/FSP/FSU	Yes	
INPUT<1 2>:GAIN:STATE	FSP/FSU	Yes	
INSTRUMENT[:SELECT] SANALYZER DDEMOD ADEMOD MGSM WCDPOWER BWCDPOWER MWCDPOWER BC2K BDO	FSE/FSP/FSU	Yes	Agilent supports its own list with more options.
INSTRUMENT:NSELECT <numeric value>	FSE/FSP/FSU	Yes	Agilent supports its own list with more options.
MMEMORY:CATALOG? <path>	FSE/FSP/FSU	Yes	
MMEMORY:CDIRECTORY <directory_name>	FSE/FSP/FSU	Yes	
MMEMORY:COPY <file_source>,<file_destination>	FSE/FSP/FSU	Yes	
MMEMORY:DATA <file_name>[,<block data>]	FSE/FSP/FSU	Yes	
MMEMORY:DELETE <file_name>	FSE/FSP/FSU	Yes	

Table 3-1 SCPI Commands Summary Table

SCPI Command	Corresponding Model	Support	Notes
MMEMory:INITialize <msus>	FSE/FSP/FSU	Accepted	
MMEMory:LOAD:STATe 1,<file_name>	FSE/FSP/FSU	Yes	
MMEMory:LOAD:AUTO 1,<file_name>	FSE/FSP/FSU	Yes	
MMEMory:MDIRectory <directory_name>	FSE/FSP/FSU	Yes	
MMEMory:MOVE <file_source>,<file_destination>	FSE/FSP/FSU	Yes	
MMEMory:MSIS <device>	FSE/FSP/FSU	Accepted	
MMEMory:NAME <file_name>	FSE/FSP/FSU	Yes	Only support "*.png" files.
MMEMory:RDIRectory <directory_name>	FSE/FSP/FSU	Yes	
MMEMory:STORe<1 2>:STATe 1,<file_name>	FSE/FSP/FSU	Yes	
MMEMory:STORe<1 2>:TRACe 1 to 3,<file_name>	FSE/FSP/FSU	Yes	Agilent supports TRACE1 TRACE2 TRACE3 TRACE4 TRACE5 TRACE6 ALL. Do not support Split Screen (screen A or screen B).
MMEMory:CLEar:STATe 1,<file_name>	FSE/FSP/FSU	Accepted	
MMEMory:CLEar:ALL	FSE/FSP/FSU	Accepted	
MMEMory:SElect[:ITEM]:HWSettings ON OFF	FSE/FSP/FSU	Accepted	The following MMEM:SEL commands related to printer settings are accepted, but take no action.
MMEMory:SElect[:ITEM]:TRACe[:ACTive] ON OFF	FSE/FSP/FSU	Accepted	
MMEMory:SElect[:ITEM]:LINes:ALL ON OFF	FSE/FSP/FSU	Accepted	
MMEMory:SElect[:ITEM]:SCData ON OFF	FSE/FSP/FSU	Accepted	
MMEMory:SElect[:ITEM]:TRANsducer:ALL ON OFF	FSE/FSP/FSU	Accepted	
MMEMory:SElect[:ITEM]:ALL	FSE/FSP/FSU	Accepted	
MMEMory:SElect[:ITEM]:NONE	FSE/FSP/FSU	Accepted	
MMEMory:SElect[:ITEM]:DEFault	FSE/FSP/FSU	Accepted	
MMEMory:COMMeNt <string>	FSE/FSP/FSU	Accepted	

Table 3-1 SCPI Commands Summary Table

SCPI Command	Corresponding Model	Support	Notes
[SENSe<1 2>:]AVERAge:COUNT 0 to 32767	FSE/FSP/FSU	Yes	Max = 10000
[SENSe<1 2>:]AVERAge[:STATe<1 to 3>] ON OFF	FSE/FSP/FSU	Yes	
[SENSe<1 2>:]AVERAge:TYPE VIDEO LINear	FSP/FSU	Yes	
[SENSe<1 2>:]BANDwidth BWIDth[:RESolutio n] <numeric_value>	FSE/FSP/FSU	Yes	Min = 1 Hz Max = 8 MHz
[SENSe<1 2>:]BANDwidth BWIDth[:RESolutio n]:AUTO ON OFF	FSE/FSP/FSU	Yes	
[SENSe<1 2>:]BANDwidth BWIDth[:RESolutio n]:RATio 0.0001 to 1	FSE/FSP/FSU	Yes	
[SENSe<1 2>:]BANDwidth BWIDth[:RESolutio n]:TYPE NORMAl FFT CFILter RRC	FSP/FSU	Yes	
[SENSe<1 2>:]BANDwidth BWIDth:VIDeo 1Hz to 10MHz	FSE/FSP/FSU	Yes	Max = 50 MHz
[SENSe<1 2>:]BANDwidth BWIDth:VIDeo:AUTO ON OFF	FSE/FSP/FSU	Yes	
[SENSe<1 2>:]BANDwidth BWIDth:VIDeo:RATio 0.01 to 1000	FSP/FSU	Yes	
[SENSe<1 2>:]CORRection:EGAIN:INPut[:MAGNitude] -200...200dB	FSP/FSU	Yes	Min = -81.90 dB Max = 81.90 dB
[SENSe<1 2>:]DETEctor<1..3>[:FUNctio n] APEak NEGative POSitive SAMPlE RMS AVERAge QPEak	FSE/FSP/FSU	Yes	
[SENSe<1 2>:]DETEctor<1 to 3>[:FUNctio n]:AUTO ON OFF	FSE/FSP/FSU	Yes	
[SENSe<1 2>:]FREQUency:CENTer 0 to fmax	FSE/FSP/FSU	Yes	The value of fmax depends on the analyzer hardware.
[SENSe<1 2>:]FREQUency:CENTer:STEP 0 to fmax	FSE/FSP/FSU	Yes	
[SENSe<1 2>:]FREQUency:CENTer:STEP:LINK SPAN RBW OFF	FSE/FSP/FSU	Yes	
[SENSe<1 2>:]FREQUency:CENTer:STEP:LINK :FACTor 1 to 100 PCT	FSE/FSP/FSU	Yes	
[SENSe<1 2>:]FREQUency:SPAN 0 to fmax	FSE/FSP/FSU	Yes	
[SENSe<1 2>:]FREQUency:SPAN:FULL	FSE/FSP/FSU	Yes	

Table 3-1 SCPI Commands Summary Table

SCPI Command	Corresponding Model	Support	Notes
[SENSe<1 2>:]FREQuency:STARt 0 to fmax	FSE/FSP/FSU	Yes	
[SENSe<1 2>:]FREQuency:STOP 0 to fmax	FSE/FSP/FSU	Yes	
[SENSe<1 2>:]FREQuency:MODE CW FIXEd SWEEp	FSE/FSP/FSU	Yes	
[SENSe<1 2>:]FREQuency:OFFSet	FSE/FSP/FSU	Yes	
[SENSe<1 2>:]POWer:ACHannel:SPACing:CHANnel 100 Hz to 2000 MHz	FSE/FSP/FSU	Yes	
[SENSe<1 2>:]POWer:ACHannel:SPACing:ACHannel 100 Hz to 2000 MHz	FSE/FSP/FSU	Yes	
[SENSe<1 2>:]POWer:ACHannel:SPACing:ALTErnate<1 2> 100 Hz to 2000 MHz	FSE/FSP/FSU	Yes	
[SENSe<1 2>:]POWer:ACHannel:TXCHannel:COUNt 1 2 3 4	FSP/FSU	Yes	
[SENSe<1 2>:]POWer:ACHannel:ACPairs 0 1 2 3	FSE/FSP/FSU	Yes	
[SENSe<1 2>:]POWer:ACHannel:BANDwidth BWIDTH[:CHANnel] 100 Hz to 1000 MHz	FSE/FSP/FSU	Yes	
[SENSe<1 2>:]POWer:ACHannel:BANDwidth BWIDTH:ACHannel 100Hz–1000MHz	FSE/FSP/FSU	Yes	
[SENSe<1 2>:]POWer:ACHannel:BANDwidth BWIDTH:ALTErnate<1 2> 100 Hz to 1000 MHz	FSP/FSU	Yes	
[SENSe<1 2>:]POWer:ACHannel:MODE ABSolute RELAtive	FSE/FSP/FSU	Yes	
[SENSe<1 2>:]POWer:ACHannel:REFerence:AUTO ONCE	FSE/FSP/FSU	Accepted	
SENSe<1 2>:]POWer:ACHannel:REFerence:TXCHannel:AUTO MINimum MAXimum LHIGhest	FSP/FSU	Yes	
[SENSe<1 2>:]POWer:ACHannel:REFerence:TXCHannel:MANual 1 2 3 4	FSP/FSU	Yes	
[SENSe<1 2>:]POWer:ACHannel:PRESet ACPower CPOWer MCACpower OBANDwidth OBWidth CN CN0	FSE/FSP/FSU	Yes	
[SENSe<1 2>:]POWer:ACHannel:PRESet:RLEVel	FSP/FSU	Accepted	

Table 3-1 SCPI Commands Summary Table

SCPI Command	Corresponding Model	Support	Notes
[SENSe<1 2>:]POWer:BAWidth BWIDTH 10 to 99.9PCT	FSP/FSU	Yes	
[SENSe<1 2>:]POWer:HSPeed ON OFF	FSP/FSU	Yes	
[SENSe<1 2>:]POWer:NCORrection ON OFF	FSP/FSU	Yes	
[SENSe<1 2>:]POWer:TRACe 1 to 3	FSP/FSU	Accepted	
[SENSe<1 2>:]ROSCillator:SOURce INTernal EXTernal	FSE/FSP/FSU	Yes	
[SENSe<1 2>:]SWEep:TIME 2.5ms to 16000s (frequency domain) 1us to 16000s (time domain)	FSE/FSP/FSU	Yes	
[SENSe<1 2>:]SWEep:TIME:AUTO ON OFF	FSE/FSP/FSU	Yes	
[SENSe<1 2>:]SWEep:COUNt 0 to 32767	FSE/FSP/FSU	Yes	Max = 10000
[SENSe<1 2>:]SWEep:EGATe ON OFF	FSE/FSP/FSU	Yes	
[SENSe<1 2>:]SWEep:EGATe:TYPE LEVel EDGE	FSE/FSP/FSU	Yes	
[SENSe<1 2>:]SWEep:EGATe:POLarity POSitive NEGative	FSE/FSP/FSU	Yes	
[SENSe<1 2>:]SWEep:EGATe:HOLDoff 125 ns to 100 s	FSE/FSP/FSU	Yes	
[SENSe<1 2>:]SWEep:EGATe:LENGth 0 to 100 s	FSE/FSP/FSU	Yes	
[SENSe<1 2>:]SWEep:EGATe:SOURce EXTernal IFPower RFPower	FSE/FSP/FSU	Yes	
[SENSe<1 2>:]SWEep:POINts 125 to 8001	FSP/FSU	Yes	
STATus:OPERation[:EVENT]?	FSE/FSP/FSU	Yes	
STATus:OPERation:CONDition?	FSE/FSP/FSU	Yes	
STATus:OPERation:ENABle	FSE/FSP/FSU	Yes	Max = 32767
STATus:OPERation:PTRansition	FSE/FSP/FSU	Yes	Max = 32767
STATus:OPERation:NTRansition	FSE/FSP/FSU	Yes	Max = 32767
STATus:PRESet	FSE/FSP/FSU	Yes	
STATus:QUEStionable[:EVENT]	FSE/FSP/FSU	Yes	
STATus:QUEStionable:CONDition	FSE/FSP/FSU	Yes	
STATus:QUEStionable:ENABle	FSE/FSP/FSU	Yes	Max = 32767

Table 3-1 SCPI Commands Summary Table

SCPI Command	Corresponding Model	Support	Notes
STATus:QUEStionable:PTRansition	FSE/FSP/FSU	Yes	Max = 32767
STATus:QUEStionable:NTRansition	FSE/FSP/FSU	Yes	Max = 32767
STATus:QUEStionable:FREQuency[:EVENT]	FSE/FSP/FSU	Yes	
STATus:QUEStionable:FREQuency:CONDition	FSE/FSP/FSU	Yes	
STATus:QUEStionable:FREQuency:ENABle	FSE/FSP/FSU	Yes	Max = 32767
STATus:QUEStionable:FREQuency:PTRansition	FSE/FSP/FSU	Yes	Max = 32767
STATus:QUEStionable:FREQuency:NTRansition	FSE/FSP/FSU	Yes	Max = 32767
STATus:QUEue[:NEXT]?	FSE/FSP/FSU	Yes	Agilent supports its own Error messages. See <i>Instrument Messages</i> for more details.
SYSTem:COMMunicate:GPIB[:SELF]:ADDRes s	FSE/FSP/FSU	Yes	
SYSTem:DATE	FSE/FSP/FSU	Yes	Date display format is, for example, 2010,5,10
SYSTem:DISPlay:FPANel ON OFF	FSP/FSU	Yes	
SYSTem:DISPlay:UPDate	FSE/FSP/FSU	Accepted	
SYSTem:ERRor?	FSE/FSP/FSU	Yes	
SYSTem:ERRor:LIST?	FSP/FSU	Accepted	
SYSTem:ERRor:CLEar:ALL	FSP/FSU	Accepted	
SYSTem:PRESet	FSE/FSP/FSU	Yes	
SYSTem:TIME	FSE/FSP/FSU	Yes	
SYSTem:VERSion	FSE/FSP/FSU	Yes	
TRACe<1 2>[:DATA] TRACE1 TRACE2 TRACE3 SPURious ABITstream PWCDp CTABLE, <block> <numeric_value>	FSE/FSP/FSU	Yes	Support the parameter list as: TRACe[1][:DATA] TRACE1 TRACE2 TRACE3 TRACE4 TRACE5 TRACE6, <data> where <data> can be ASCII, REAL or INTeGer.
TRACe<1 2>:IQ:DATA	FSP/FSU	Yes	

Table 3-1 SCPI Commands Summary Table

SCPI Command	Corresponding Model	Support	Notes
TRACe<1 2>:IQ:SET <filter type>,<rbw>,<sample rate>,<trigger source>,<trigger slope>,<pretrigger samples>,<# of samples>	FSP/FSU	Yes	
TRACe<1 2>:IQ:SRATe 15.625kHz to 32MHz	FSP/FSU	Yes	
TRACe:IQ:STATe ON OFF	FSP/FSU	Yes	
TRIGger<1 2>[:SEQuence]:SOURce	FSE/FSP/FSU	Yes	The following source selections are not supported: IFPower, TV, AF, FM, AM and PM.
TRIGger<1 2>[:SEQuence]:LEVel:IFPower -30 to -10DBM	FSP/FSU	Accepted	
TRIGger<1 2>[:SEQuence]:LEVel:RFPower <real>	FSP/FSU	Yes	Min = -200 dBm Max = 100 dBm
TRIGger<1 2>[:SEQuence]:LEVel:VIDeo 0 to 100PCT	FSE/FSP/FSU	Yes	
TRIGger<1 2>[:SEQuence]:HOLDoff <time>	FSE/FSP/FSU	Yes	Min = 0 Max = 500 ms
TRIGger<1 2>[:SEQuence]:SLOPe POSitive NEGative	FSE/FSP/FSU	Yes	

- a. Agilent supports up to 6 limit lines.
- b. "Accepted" means this command is accepted, but takes no action and reports no error.

Agilent N9062A Supported IEEE SCPI Commands

The Agilent N9062A application also supports the following IEEE 488.2 SCPI commands:

Table 3-2 IEEE 488.2 SCPI Commands

SCPI Commands	Description
*CAL?	Performs a full alignment and returns a number indicating the success of the alignment. A zero is returned if the alignment is successful. A one is returned if any part of the alignment fails.
*CLS	Clears the status byte. It does this by emptying the error queue and clearing all bits in all of the event registers. The status byte registers summarize the states of the other registers. It is also responsible for generating service requests.
*ESE <number> *ESE?	Selects the desired bits from the standard event status enable register. This register monitors I/O errors and synchronization conditions such as operation complete, request control, query error, device dependent error, execution error, command error and power on. The selected bits are OR'd to become a summary bit (bit 5) in the status byte register which can be queried. The query returns the state of the standard event status enable register.
*ESR?	Queries and clears the standard event status event register. (This is a destructive read.)
ID?	This command does not follow any of the SCPI rules. It only exists so that production can get a universal response to the ID? command for every instrument.
*IDN?	Returns an instrument identification information. The string will contain the model number, serial number and firmware revision. The response is organized into four fields separated by commas. The field definitions are as follows: 1.Manufacturer 2.Model 3.Serial number 4.Firmware version
*OPC *OPC?	The *OPC command sets bit 0 in the standard event status register to "1" when pending operations have finished. It does not hold off subsequent operations. The *OPC? query stops new commands from being processed until the current processing is complete. Then it returns a "1", and the program continues. This query can be used to synchronize events of other instruments on the external bus.
*OPT?	Returns a string of all the installed instrument options. It is a comma separated list such as: "BAC,BAH". There are a few options that include more than one mode. An instrument with one of these options will report the option number once for each mode.
*RCL <register>	This command recalls the instrument state from the specified instrument memory register.

Table 3-2 IEEE 488.2 SCPI Commands

SCPI Commands	Description
*RST	<p>This command presets the instrument to a factory defined condition that is appropriate for remote programming operation. *RST is equivalent to performing the commands:</p> <ol style="list-style-type: none"> 1.:SYSTem:PRESet, with preset type set to MODE. 2.*CLS which clears the STATus bits and error queue 3.:INITiate:CONTinuous OFF which selects single sweep/measurement <p>*RST does not change the mode and only resets the parameters for the current mode.</p>
*SAV <register>	This command saves the instrument state to the specified instrument memory register.
*SRE <integer> *SRE?	<p>This command sets the value of the service request enable register.</p> <p>The query returns the value of the register.</p>
*STB?	Returns the value of the status byte register without erasing its contents.
*TRG	This command triggers the instrument. Use the command :TRIGger[:SEQuence]:SOURce to select the trigger source.
*TST?	<p>This query performs a self test and returns a number indicating the success of the testing. A zero is returned if the test is successful.</p>
*WAI	<p>This command causes the instrument to wait until all pending commands/processes are completed before executing any additional commands.</p> <p>There is no query form for the command.</p>

Supported SCPI Commands

Mode Selection

NOTE When Option N9062A–2FP is licensed, the SCPI Language Compatibility mode is available.

To invoke SCPI Language Compatibility mode, enter either of the following commands¹:

:INST SCILC

:INST:NSElect 270

Key Path	Front-panel key
Remote Command	:INSTrument[:SElect] SCILC :INSTrument:NSElect 270

Mode Setup

Allows you to adjust the Mode Setup settings.

Key Path	Front-panel key
Initial S/W Revision	A.06.00

Compatibility Mode

The command allows you to set the model you want to emulate.

The query returns the model currently being emulated.

This setting affects the response string of “*IDN?”, and the entire SCPI tree, according to the current selection.

Key Path	Mode Setup
Remote Command	[:SENSe]:RLC:TYPE FSE FSP FSU [:SENSe]:RLC:TYPE?
Example	SENS:RLC:TYPE FSU RLC:TYPE?
State Saved	Saved in instrument state.

1. After changing into or out of SCPI Language Compatibility mode, allow seconds of delay before sending subsequent commands.

Compatible SCPI Commands
Supported SCPI Commands

Range	FSE FSP FSU
Initial S/W Revision	A.06.00 or later

R&S FSE

Selects emulation of the R&S FSE.

Key Path	Mode Setup, Compatibility Mode
Example	RLC:TYPE FSE
Initial S/W Revision	A.06.00 or later

R&S FSP

Selects emulation of the R&S FSP.

Key Path	Mode Setup, Compatibility Mode
Example	RLC:TYPE FSP
Initial S/W Revision	A.06.00 or later

R&S FSU

Selects emulation of the R&S FSU.

Key Path	Mode Setup, Compatibility Mode
Example	RLC:TYPE FSU
Initial S/W Revision	A.06.00 or later

Mode IDN Response

The command allows you to set or query the Mode IDN Response type.

This setting affects the response string of “*IDN?”, and the whole SCPI tree, according to the current selection.

The query returns the current state of this setting.

Key Path	Mode Setup
Remote Command	[:SENSe] :RLC:IDN:TYPE SYSTem COMPAtibility USER [:SENSe] :RLC:IDN:TYPE?
Example	SENS:RLC:IDN:TYPE SYSTem RLC:IDN:TYPE?
State Saved	Saved in instrument state.
Range	SYSTem COMPAtibility USER

Initial S/W Revision	A.06.00 or later
----------------------	------------------

Compatibility Model

Sets the response to the *IDN command to be a string describing the emulated instrument, such as “FSP”.

Key Path	Mode Setup, Mode IDN Response
Example	RLC:IDN:TYPE COMP
Initial S/W Revision	A.06.00

System IDN Response

Sets the response to the *IDN command to be a string describing the analyzer model, such as “N9020A”.

Key Path	Mode Setup, Mode IDN Response
Example	RLC:IDN:TYPE SYST
Initial S/W Revision	A.06.00

User IDN Response

Allows you to set or query the user-defined mode IDN response.

This setting affects the response string of “*IDN?”, and the whole SCPI tree, according to the current selection.

The query returns the current User IDN Response string, unless the string is empty, in which case the System IDN Response is returned instead.

Key Path	Mode Setup
Remote Command	[:SENSe] :RLC:IDN:USER <string> [:SENSe] :RLC:IDN:USER?
Example	SENS:RLC:IDN:USER “abc” RLC:IDN:USER?
State Saved	Saved in instrument state.
Initial S/W Revision	A.06.00

ABORt Subsystem

The ABORt subsystem contains commands for aborting triggered actions. An action may be triggered again immediately after being aborted. All commands trigger events, and therefore have no *RST value.

ABORt

Aborts any current measurement and resets the trigger system.

Remote Command	:ABORt
Example	:ABOR
Remote Command Notes	FSE, FSP, FSU

CALCulate:DELTamarker Subsystem

The CALCulate:DELTamarker subsystem controls the delta-marker functions.

NOTE All delta marker functions are not supported in the CCDF and APD measurements.

CALCulate<1|2>:DELTamarker<1...4>[:STATe] ON | OFF

When delta marker 1 is selected, this command switches the delta marker on and off.

When any one of delta markers 2 through 4 is selected, that marker becomes the delta marker. If the marker was not activated, it becomes activated, and is placed at the maximum of the measurement curve.

If the numeric suffix is omitted, delta marker 1 is automatically selected.

The query returns the current state of this setting.

Remote Command	:CALCulate [1] :DELTamarker [1] 2 3 4 [:STATe] ON OFF 1 0 :CALCulate [1] :DELTamarker [1] 2 3 4 [:STATe] ?
Example	CALC:DELT ON
Remote Command Notes	FSE, FSP, FSU
Preset	OFF
State Saved	Saved in instrument state.

CALCulate<1|2>:DELTamarker<1...4>:MODE ABSolute|RELative

The command switches the frequency input of the delta marker between Absolute and Relative modes.

The query returns the current state of this setting.

Remote Command	:CALCulate [1] :DELTamarker [1] 2 3 4 :MODE ABSolute RELative :CALCulate [1] :DELTamarker [1] 2 3 4 :MODE?
Example	CALC:DELT:MODE ABS CALC:DELT:MODE REL CALC:DELT:MODE?

Remote Command Notes	FSE, FSP, FSU
Preset	RELative
State Saved	Saved in instrument state.
Range	ABSolute RELative

CALCulate<1|2>:DELTamarker<1...4>:AOFF

Switches off all active delta markers.

Remote Command	:CALCulate [1] :DELTamarker [1] 2 3 4 :AOFF
Example	CALC:DELT:AOFF
Remote Command Notes	FSE, FSP, FSU

CALCulate<1|2>:DELTamarker<1...4>:TRACe <1...3>

Assigns the specified delta marker to the specified trace.

Remote Command	:CALCulate [1] :DELTamarker [1] 2 3 4 :TRACe 1 2 3
Example	CALC:DELT:TRAC 1
Remote Command Notes	FSE, FSP, FSU
Preset	1
State Saved	Saved in instrument state.
Min	1
Max	3

CALCulate<1|2>:DELTamarker<1...4>:X<0...MAX>(frequency | sweep time)

The command positions the specified delta marker at the indicated frequency (span > 0) or time (span = 0).

The query returns the current value of this parameter.

Remote Command	:CALCulate [1] :DELTamarker [1] 2 3 4 :X <real> :CALCulate [1] :DELTamarker [1] 2 3 4 :X?
Example	CALC:DELT:X 10.7MHz CALC:DELT:X?
Remote Command Notes	FSE, FSP, FSU

CALCulate<1|2>:DELTamarker<1...4>:X:RELative?

Queries the frequency or time value of the specified delta marker, relative to marker 1, or to the reference position (for CALCulate:DELTamarker:FUNCTION:FIXed:STATe ON). If required, the corresponding

delta marker is activated.

This command is not supported in the CCDF and APD measurements.

Remote Command	:CALCulate [1] :DELTamarker [1] 2 3 4 :X:RELative?
Example	CALC:DELT:X:REL?
Remote Command Notes	FSE, FSP, FSU

CALCulate<1|2>:DELTamarker<1...4>:Y?

Queries the measured value of the specified delta marker. If required, the specified delta marker is activated. The returned result is a value relative to marker 1, or to the reference position if the reference is fixed.

The only unit supported is dB.

Remote Command	:CALCulate [1] :DELTamarker [1] 2 3 4 :Y?
Example	CALC:DELT:Y?
Remote Command Notes	FSE, FSP, FSU

CALCulate<1|2>:DELTamarker<1...4>:MAXimum[:PEAK]

Places the delta marker at the current maximum value of the trace. If required, the specified delta marker is first activated.

Remote Command	:CALCulate [1] :DELTamarker [1] 2 3 4 :MAXimum [:PEAK]
Example	CALC:DELT:MAX
Remote Command Notes	FSE, FSP, FSU

CALCulate<1|2>:DELTamarker<1...4>:MAXimum:LEFT

Places the delta marker at the next smaller maximum to the left of the current value (that is, in order of decreasing X values). If required, the specified delta marker is first activated.

Remote Command	:CALCulate [1] :DELTamarker [1] 2 3 4 :MAXimum:LEFT
Example	CALC:DELT:MAX:LEFT
Remote Command Notes	FSE, FSP, FSU

CALCulate<1|2>:DELTamarker<1...4>:MAXimum:NEXT

Places the delta marker at the next smaller maximum on the trace. If required, the specified delta marker is first activated.

Remote Command	:CALCulate [1] :DELTamarker [1] 2 3 4 :MAXimum:NEXT
Example	CALC:DELT:MAX:NEXT

Remote Command Notes	FSE, FSP, FSU
----------------------	---------------

CALCulate<1|2>:DELTamarker<1...4>:MAXimum:RIGHT

Places the delta marker at the next smaller maximum to the right of the current value (that is, in order of increasing X values). If required, the specified delta marker is first activated.

Remote Command	:CALCulate [1] :DELTamarker [1] 2 3 4 :MAXimum:RIGHT
Example	CALC:DELT:MAX:RIGH
Remote Command Notes	FSE, FSP, FSU

CALCulate<1|2>:DELTamarker<1...4>:MINimum[:PEAK]

Places the delta marker at the current minimum on the trace. If required, the specified delta marker is first activated.

Remote Command	:CALCulate [1] :DELTamarker [1] 2 3 4 :MINimum[:PEAK]
Example	CALC:DELT:MIN
Remote Command Notes	FSE, FSP, FSU

CALCulate<1|2>:DELTamarker<1...4>:MINimum:LEFT

Places the delta marker at the next higher minimum to the left of the current value (that is, in order of decreasing X values). If required, the specified delta marker is first activated.

Remote Command	:CALCulate [1] :DELTamarker [1] 2 3 4 :MINimum:LEFT
Example	CALC:DELT:MIN:LEFT
Remote Command Notes	FSE, FSP, FSU

CALCulate<1|2>:DELTamarker<1...4>:MINimum:NEXT

Places the delta marker at the next higher minimum of the trace. If required, the specified delta marker is first activated.

Remote Command	:CALCulate [1] :DELTamarker [1] 2 3 4 :MINimum:NEXT
Example	CALC:DELT:MIN:NEXT
Remote Command Notes	FSE, FSP, FSU

CALCulate<1|2>:DELTamarker<1...4>:MINimum:RIGHT

Places the delta marker at the next higher minimum to the right of the current value (that is, in order of increasing X values). If required, the specified delta marker is first activated.

Remote Command	:CALCulate [1] :DELTamarker [1] 2 3 4 :MINimum:RIGHT
-----------------------	--

Example	CALC:DELT:MIN:RIGH
Remote Command Notes	FSE, FSP, FSU

CALCulate<1|2>:DELTamarker<1...4>:LINK ON | OFF

The command switches the link between delta marker 1 and marker 1 on or off.

If the link is “on”, and the X-axis value of marker 1 is changed, then delta marker 1 will change by the same amount.

The numeric suffix of DELTmarker may only be 1, or be omitted, because this functionality is only available for marker 1 and delta marker 1.

The query returns the current state of this setting.

Remote Command	:CALCulate [1] :DELTamarker [1] 2 3 4 :LINK ON OFF 1 0 :CALCulate [1] :DELTamarker [1] 2 3 4 :LINK?
Example	CALC:DELT:LINK ON
Remote Command Notes	FSU
Preset	OFF
State Saved	Saved in instrument state.

CALCulate<1|2>:DELTamarker<1...4>:FUNCTION:FIXed[:STATe] ON | OFF

The command switches between relative measurement, and measurement relative to a fixed reference.

If required, marker 1 is activated and a peak search is performed.

If marker 1 is activated, its position becomes the reference point for the measurement. The reference point can then be changed, using commands CALCulate:DELTamarker:FUNCTION:FIXed:RPOint:X and :RPOint:Y, independently of the position of marker 1 and of any trace.

The query returns the current state of this setting.

Remote Command	:CALCulate [1] :DELTamarker [1] 2 3 4 :FUNCTION:FIXed[:STATe] ON OFF 1 0 :CALCulate [1] :DELTamarker [1] 2 3 4 :FUNCTION:FIXed[:STATe] ?
Example	CALC:DELT:FUNC:FIX ON
Remote Command Notes	FSE, FSP, FSU
Preset	OFF
State Saved	Saved in instrument state.

**CALCulate<1|2>:DELTamarker<1...4>:FUNCTION:FIXed:RPOint:MAXimum[:PEAK]
<numeric_value>**

Sets the reference point for all delta markers when you are making measurements using a fixed reference point (enabled by the command CALC:DELT:FUNC:FIX:STAT ON). You can specify a value using the <numeric_value> parameter, or specify the peak of the selected trace [:PEAK].

For phase-noise measurements (selected via CALCulate:DELTamarker:FUNCTION:PNOise:STATe ON), the command specifies a new reference point for delta marker 2.

Remote Command	:CALCulate [1] :DELTamarker [1] 2 3 4 :FUNCTION:FIXed:RPOint:MAXimum[:PEAK]
Example	CALC:DELT:FUNC:FIX:RPO:MAX
Remote Command Notes	FSP, FSU

CALCulate<1|2>:DELTamarker<1...4>:FUNCTION:FIXed:RPOint:Y <numeric_value>

Sets or queries a Y-axis reference point for all delta markers when you are making measurements using a fixed reference point (enabled by CALCulate:DELTamarker:FUNCTION:FIXed:STATe ON).

For phase-noise measurements (selected via CALCulate:DELTamarker:FUNCTION:PNOise:STATe ON), the command specifies a new reference point level for delta marker 2.

Remote Command	:CALCulate [1] :DELTamarker [1] 2 3 4 :FUNCTION:FIXed:RPOint:Y <amplitude> :CALCulate [1] :DELTamarker [1] 2 3 4 :FUNCTION:FIXed:RPOint:Y?
Example	CALC:DELT2:FUNC:FIX:RPO:Y -10DBM CALC:DELT2:FUNC:FIX:RPO:Y?
Remote Command Notes	FSE, FSP, FSU

CALCulate<1|2>:DELTamarker<1...4>:FUNCTION:FIXed:RPOint:Y:OFFSet <numeric_value>

Sets or queries the additional offset for all delta markers when you are making measurements using a fixed reference value (enabled by CALCulate:DELTamarker:FUNCTION:FIXed:STATe ON).

For phase-noise measurements (selected via CALCulate:DELTamarker:FUNCTION:PNOise:STATe ON), specifies an additional offset that is included in the display of delta marker 2.

Remote Command	:CALCulate [1] :DELTamarker [1] 2 3 4 :FUNCTION:FIXed:RPOint:Y:OFFSet <rel_ampl> :CALCulate [1] :DELTamarker [1] 2 3 4 :FUNCTION:FIXed:RPOint:Y:OFFSet?
Example	CALC:DELT2:FUNC:FIX:RPO:Y:OFFS 10DB CALC:DELT2:FUNC:FIX:RPO:Y:OFFS?
Remote Command Notes	FSE, FSP, FSU

Preset	0
--------	---

CALCulate<1|2>:DELTamarker<1...4>:FUNCTION:FIXed:RPOint:X <numeric_value>

Sets or queries the reference frequency or time for all delta markers when you are making measurements using a fixed reference value (enabled by CALCulate:DELTamarker:FUNCTION:FIXed:STATe ON).

For phase-noise measurements (selected via CALCulate:DELTamarker:FUNCTION:PNOise:STATe ON), specifies a reference frequency or time for delta marker 2.

Remote Command	:CALCulate [1] :DELTamarker [1] 2 3 4 :FUNCTION:FIXed:RPOint:X <real> :CALCulate [1] :DELTamarker [1] 2 3 4 :FUNCTION:FIXed:RPOint:X?
Example	CALC:DELT2:FUNC:FIX:RPO:X 10MHz CALC1:DELT:FUNC:FIX:RPO:X?
Remote Command Notes	FSE, FSP, FSU

CALCulate<1|2>:DELTamarker<1...4>:FUNCTION:PNOise[:STATe] ON|OFF

The command switches phase-noise measurement on or off for all active delta markers. All correction values are included.

If required, marker 1 is activated, and a peak search is performed. If marker 1 is activated, its position becomes the measurement's reference point.

The measurement's reference point can be modified with commands CALCulate:DELTamarker:FUNCTION:FIXed:RPOint:X and ...:RPOint:Y, independently of the position of marker 1 and of any trace.

The DELTmarker numeric suffix <1 to 4> is not required for this command.

The query returns the current state of this setting.

Remote Command	:CALCulate [1] :DELTamarker [1] 2 3 4 :FUNCTION:PNOise[:STATe] ON OFF 1 0 :CALCulate [1] :DELTamarker [1] 2 3 4 :FUNCTION:PNOise[:STATe] ?
Example	CALC:DELT:FUNC:PNO ON
Remote Command Notes	FSE, FSP, FSU
Preset	OFF
State Saved	Saved in instrument state.

CALCulate<1|2>:DELTamarker<1...4>:FUNCTION:PNOise:AUTO ON | OFF

The command switches on or off automatic peak search action for fixed reference marker 1. When activated, this search occurs at the end of each sweep.

You can use this functionality to track a drifting source during a phase noise measurement. Delta marker 2, which shows the phase noise measurement result, retains the delta frequency value. Therefore the results of phase noise measurement with a specified offset remain valid.

When marker 2 reaches the limit of the span, the delta marker value is automatically adjusted to be within the span. In such cases, select a wider span.

The query returns the current state of this setting.

Remote Command	:CALCulate [1] :DELTamarker [1] 2 3 4 :FUNction:PNOise:AUTO ON OFF 1 0 :CALCulate [1] :DELTamarker [1] 2 3 4 :FUNction:PNOise:AUTO ?
Example	CALC:DELT:FUNC:PNO:AUTO ON
Remote Command Notes	FSU
Preset	OFF
State Saved	Saved in instrument state.

CALCulate<1|2>:DELTamarker<1...4>:FUNction:PNOise:RESult?

Queries the result of a phase noise measurement. If required, the measurement is switched on.

Remote Command	:CALCulate [1] :DELTamarker [1] 2 3 4 :FUNction:PNOise:RESult?
Example	CALC:DELT:FUNC:PNO:RES?
Remote Command Notes	FSE, FSP, FSU

CALCulate:LIMit Subsystem

The CALCulate:LIMit subsystem includes the limit lines and limit checks. Limit lines may be defined as upper or lower. The limit line Y-values correspond to the X-axis values. The number of x and y values must be identical.

X Series Analyzers support up to 6 active limit lines at the same time.

CALCulate<1|2>:LIMit<1...6>:ACTive?

Queries the names of all active limit lines. For this command, all the numeric suffixes are ignored.

The return values are sorted alphabetically. If there are no active limit lines, an empty string is returned.

Remote Command	:CALCulate [1] :LIMit [1] 2 3 4 5 6 :ACTive?
Example	CALC:LIM:ACT?
Remote Command Notes	FSE, FSU
Preset	""

State Saved	Saved in instrument state.
-------------	----------------------------

CALCulate<1|2>:LIMit<1...6>:TRACe <1...3>

The command assigns a limit line to a trace.

The query returns the current assignment.

Remote Command	:CALCulate [1] :LIMit [1] 2 3 4 5 6 :TRACe 1 2 3 :CALCulate [1] :LIMit [1] 2 3 4 5 6 :TRACe?
Example	CALC:LIM:TRAC 1
Remote Command Notes	FSE, FSP, FSU
Preset	1
State Saved	Saved in instrument state.
Min	1
Max	3

CALCulate<1|2>:LIMit<1...6>:STATe ON | OFF

The command switches the limit check for the specified limit line on or off.

The limit check result may be queried using CALCulate:LIMit<1...6>:FAIL?.

The query returns the current state of this setting.

Remote Command	:CALCulate [1] :LIMit [1] 2 3 4 5 6 :STATe ON OFF 1 0 :CALCulate [1] :LIMit [1] 2 3 4 5 6 :STATe?
Example	CALC:LIM:STAT ON
Remote Command Notes	FSE, FSP, FSU
Preset	OFF
State Saved	Saved in instrument state.

CALCulate<1|2>:LIMit<1...6>:UNIT DBM | DBPW | WATT | DBUV | DBMV | VOLT | DBUA | AMPere | DB | DEG | RAD | S | HZ | PCT | UNITLESS

The command specifies the unit for all limit lines.

This mode supports only the following amplitude units: DBM, DBPW, WATT, DBUV, DBMV, VOLT, DBUA and DB.

The command ignores the numeric suffix of LIMit, and sets all limit lines to the same unit.

The query returns the current state of this setting.

Remote Command	:CALCulate [1] :LIMit [1] 2 3 4 5 6 :UNIT DBM DBUV DBMV DBUA A V W :CALCulate [1] :LIMit [1] 2 3 4 5 6 :UNIT?
Example	CALC:LIM:UNIT DBM CALC:LIM:UNIT?
Remote Command Notes	FSE, FSP, FSU
Preset	DBM
State Saved	Saved in instrument state.
Range	DBM DBUV DBMV DBUA A V W

CALCulate<1|2>:LIMit<1...6>:FAIL?

Returns the limit check result for the specified limit line.

To obtain a valid result, you must perform a complete sweep.

The numeric value returned has the following meaning:

0	PASS
1	FAIL
2	MARGIN

Remote Command	:CALCulate [1] :LIMit [1] 2 3 4 5 6 :FAIL?
Example	CALC:LIM:FAIL?
Remote Command Notes	FSE, FSP, FSU
Min	0
Max	2

CALCulate<1|2>:LIMit<1...6>:CLEar[:IMMediate]

Deletes the current limit check result for all limit lines.

Remote Command	:CALCulate [1] :LIMit [1] 2 3 4 5 6 :CLEar [:IMMediate]
Example	CALC:LIM:CLE
Remote Command Notes	FSE, FSP, FSU

CALCulate<1|2>:LIMit<1...6>:COMMeNt <string>

The command defines a comment for the specified limit line. The comment string may have a maximum length of 40 characters.

The query returns the current value of this string.

Remote Command	:CALCulate [1] :LIMit [1] 2 3 4 5 6 :COMMeNt <string> :CALCulate [1] :LIMit [1] 2 3 4 5 6 :COMMeNt?
Example	CALC:LIM:COMM 'GSM2' CALC:LIM:COMM?
Remote Command Notes	FSE, FSP, FSU
Preset	""
State Saved	Saved in instrument state.

CALCulate<1|2>:LIMit<1...6>:COPY 1 to 6 | <name>

Copies the limit line specified by the numeric suffix of LIMit onto the limit line specified by the command's parameter.

This mode does not support the parameter option <name>.

Remote Command	:CALCulate [1] :LIMit [1] 2 3 4 5 6 :COpy <int>
Example	CALC:LIM:COpy 2
Remote Command Notes	FSE, FSP, FSU
Min	1
Max	6

CALCulate<1|2>:LIMit<1...6>:NAME <name of limit line>

The command assigns a name to the limit line specified by the numeric suffix of LIMit. A limit line with this name is created, if it does not already exist.

The query returns the current value of this string.

Remote Command	:CALCulate [1] :LIMit [1] 2 3 4 5 6 :NAME <string> :CALCulate [1] :LIMit [1] 2 3 4 5 6 :NAME?
Example	CALC:LIM:NAME 'test' CALC:LIM:NAME?
Remote Command Notes	FSE, FSP, FSU
Preset	'REM1' to 'REM6' for lines 1 to 6
State Saved	Saved in instrument state.

CALCulate<1|2>:LIMit<1...6>:DELeTe

Deletes the specified limit line.

Remote Command	:CALCulate [1] :LIMit [1] 2 3 4 5 6 :DELeTe
Example	CALC:LIM:DEL
Remote Command Notes	FSE, FSP, FSU

CALCulate:LIMit:ESpectrum Subsystem

The CALCulate:LIMit:SPECTrum subsystem defines the limit check for spectrum emission measurements.

The following SCPI commands of this subsystem are accepted by the analyzer, but result in no action, and report no error.

CALCulate<1|2>:LIMit<1...6>:ESpectrum:MODE AUTO | MANual | USER

This command is accepted, but takes no action and reports no error.

Remote Command	:CALCulate [1] :LIMit [1] 2 3 4 5 6 :ESpectrum:MODE AUTO MANual USER :CALCulate [1] :LIMit [1] 2 3 4 5 6 :ESpectrum:MODE?
Example	CALC:LIM:ESP:MODE AUTO CALC:LIM:ESP:MODE
Remote Command Notes	FSP
Preset	AUTO
State Saved	Saved in instrument state.
Range	AUTO MANual USER

CALCulate<1|2>:LIMit<1...6>:ESpectrum:VALue <numeric_value>

Sets manual limit line selection. The limit line is selected by specifying the expected dBm power.

This command is accepted, but takes no action and reports no error.

Remote Command	:CALCulate [1] :LIMit [1] 2 3 4 5 6 :ESpectrum:VALue <integer> :CALCulate [1] :LIMit [1] 2 3 4 5 6 :ESpectrum:VALue?
Example	CALC:LIM:ESP:VAL 39 CALC:LIM:ESP:VAL?
Remote Command Notes	FSP
Preset	0

CALCulate<1|2>:LIMit<1...6>:ESpectrum:REStore

Restores the standard limit lines for the spectrum emission mask measurement. Any modifications that were made to the standard limit lines are lost, and the default setting is restored.

This command is accepted, but takes no action and reports no error.

Remote Command	:CALCulate [1] :LIMit [1] 2 3 4 5 6 :ESpectrum:REStore
Example	CALC:LIM:ESP:REST
Remote Command Notes	FSP

CALCulate:LIMit:CONTRol Subsystem

The CALCulate:LIMit:CONTRol subsystem controls the X-axis.

NOTE All limit functions are not supported in the CCDF and APD measurements.

CALCulate<1|2>:LIMit<1...6>:CONTRol[:DATA] <numeric_value>,<numeric_value>..

The command specifies the X-axis values (frequencies or times) of the upper or lower limit lines.

This mode does not support any unit for this command.

The query returns the current values of these settings.

Remote Command	:CALCulate [1] :LIMit [1] 2 3 4 5 6 :CONTRol [:DATA] <real>,<real>.. :CALCulate [1] :LIMit [1] 2 3 4 5 6 :CONTRol [:DATA] ?
Example	CALC:LIM:CONT:DATA 1.0e7,2.0e7,1.0e8,3.0e8,1.0e9 CALC:LIM:CONT:DATA?
Remote Command Notes	FSE, FSP, FSU
State Saved	Saved in instrument state.

CALCulate<1|2>:LIMit<1...6>:CONTRol:DOMain FREQuency | TIME

The command switches between frequency or time domain for the X-axis.

The query returns the current state of this setting.

Remote Command	:CALCulate [1] :LIMit [1] 2 3 4 5 6 :CONTRol:DOMain FREQuency TIME :CALCulate [1] :LIMit [1] 2 3 4 5 6 :CONTRol:DOMain?
Example	CALC:LIM:CONT:DOM FREQ CALC:LIM:CONT:DOM?

Remote Command Notes	FSE, FSP, FSU
Preset	FREQuency
State Saved	Saved in instrument state.
Range	FREQuency TIME

CALCulate<1|2>:LIMit<1...6>:CONTrol:OFFSet <numeric_value>

The command defines an offset for the X-axis value of the specified relative limit line, in the frequency or time domain.

The query returns the current value of this setting.

Remote Command	:CALCulate [1] :LIMit [1] 2 3 4 5 6 :CONTrol:OFFSet <real> :CALCulate [1] :LIMit [1] 2 3 4 5 6 :CONTrol:OFFSet?
Example	CALC:LIM:CONT:OFFS 1 CALC:LIM:CONT:OFFS?
Remote Command Notes	FSE, FSP, FSU
Preset	0

CALCulate<1|2>:LIMit<1...6>:CONTrol:MODE RELative | ABSolute

The command switches between relative or absolute scaling for the X-axis of the specified limit line.

The query returns the current state of this setting.

Remote Command	:CALCulate [1] :LIMit [1] 2 3 4 5 6 :CONTrol:MODE RELative ABSolute :CALCulate [1] :LIMit [1] 2 3 4 5 6 :CONTrol:MODE?
Example	CALC:LIM:CONT:MODE ABS CALC:LIM:CONT:MODE
Remote Command Notes	FSE, FSP, FSU
Preset	ABS
State Saved	Saved in instrument state.
Range	RELative ABSolute

CALCulate<1|2>:LIMit<1...6>:CONTrol:SHIFt <numeric_value>

The command shifts the specified limit line in the X-direction, by the specified value.

The query returns the current value of this setting for the specified limit line.

Remote Command	:CALCulate [1] :LIMit [1] 2 3 4 5 6 :CONTrol :SHIFt <real> :CALCulate [1] :LIMit [1] 2 3 4 5 6 :CONTrol :SHIFt?
Preset	0
SCPI Command Notes SCPI Command Notes	FSE, FSP, FSU
SCPI Example[DEF]SCPI Example	CALC:LIM:CONT:SHIF 1 CALC:LIM:CONT:SHIF?

CALCulate<1|2>:LIMit<1...6>:CONTrol:SPACing LINear | LOGarithmic

The command switches between linear or logarithmic interpolation for calculation of limit lines from frequency points.

The query returns the current state of this setting.

Remote Command	:CALCulate [1] :LIMit [1] 2 3 4 5 6 :CONTrol :SPACing LINear LOGarithmic :CALCulate [1] :LIMit [1] 2 3 4 5 6 :CONTrol :SPACing?
Example	CALC:LIM:CONT:SPAC LOG CALC:LIM:CONT:SPAC
Remote Command Notes	FSE, FSP, FSU
Preset	LIN
State Saved	Saved in instrument state.
Range	LINear LOGarithmic

CALCulate:LIMit:LOWer Subsystem

The CALCulate:LIMit:LOWer subsystem controls the lower limit line.

NOTE All limit functions are not supported in the CCDF and APD measurements.

CALCulate<1|2>:LIMit<1...6>:LOWer[:DATA] <numeric_value>,<numeric_value>..

The command defines a set of values for the specified lower limit line.

The query returns the current value set.

This mode does not support any unit for this command.

Remote Command	CALCulate [1] :LIMit [1] 2 3 4 5 6 :LOWer [:DATA] <real>, <real>... CALCulate [1] :LIMit [1] 2 3 4 5 6 :LOWer [:DATA] ?
Example	CALC:LIM:LOW:DATA 1,20,100,300,1 CALC:LIM:LOW:DATA?
Remote Command Notes	FSE, FSP, FSU
State Saved	Saved in instrument state.

CALCulate<1|2>:LIMit<1...6>:LOWer:STATe ON | OFF

The command activates or deactivates the specified lower limit line. Limit check is activated separately, using CALC:LIM:STAT ON.

The query returns the current state of this setting.

Remote Command	:CALCulate [1] :LIMit [1] 2 3 4 5 6 :LOWer:STATe ON OFF 1 0 :CALCulate [1] :LIMit [1] 2 3 4 5 6 :LOWer:STATe?
Example	CALC:LIM:LOW:STAT OFF CALC:LIM:LOW:STAT?
Remote Command Notes	FSE, FSP, FSU
Preset	OFF
State Saved	Saved in instrument state.

CALCulate<1|2>:LIMit<1...6>:LOWer:OFFSet <numeric_value>

The command defines an offset for the Y axis of the specified relative lower limit line.

This mode does not support any unit for this command.

The query returns the current value of this setting.

Remote Command	:CALCulate [1] :LIMit [1] 2 3 4 5 6 :LOWer:OFFSet <real> :CALCulate [1] :LIMit [1] 2 3 4 5 6 :LOWer:OFFSet?
Example	CALC:LIM:LOW:OFFS 1 CALC:LIM:LOW:OFFS?
Remote Command Notes	FSE, FSP, FSU
Preset	0

CALCulate<1|2>:LIMit<1...6>:LOWer:MARGin <numeric_value>

The command defines a margin for the specified lower limit line. At the margin level, out-of-limit values

are signaled (if the limit check is active), but are not treated as violations of the limit value.

This mode does not support any unit for this command.

The query returns the current value of this setting.

Remote Command	:CALCulate [1] :LIMit [1] 2 3 4 5 6 :LOWer:MARGin <real> :CALCulate [1] :LIMit [1] 2 3 4 5 6 :LOWer:MARGin?
Example	CALC:LIM:LOW:MARG 10 CALC:LIM:LOW:MARG?
Remote Command Notes	FSE, FSP, FSU
Preset	0

CALCulate<1|2>:LIMit<1...6>:LOWer:MODE RELative | ABSolute

The command switches between relative or absolute scaling for the Y-axis of the specified lower limit line.

Selecting RELative causes the unit to be switched to dB.

The query returns the current state of this setting.

Remote Command	:CALCulate [1] :LIMit [1] 2 3 4 5 6 :LOWer:MODE RELative ABSolute :CALCulate [1] :LIMit [1] 2 3 4 5 6 :LOWer:MODE?
Example	CALC:LIM:LOW:MODE ABS CALC:LIM:LOW:MODE?
Remote Command Notes	FSE, FSP, FSU
Preset	ABS
State Saved	Saved in instrument state.
Range	RELative ABSolute

CALCulate<1|2>:LIMit<1...6>:LOWer:SHIFt <numeric_value>

The command shifts the specified limit line in the Y-direction, by the specified amount.

This mode does not support any unit for this command.

The query returns the current value of this setting.

Remote Command	:CALCulate [1] :LIMit [1] 2 3 4 5 6 :LOWer:SHIFt <real> :CALCulate [1] :LIMit [1] 2 3 4 5 6 :LOWer:SHIFt?
Example	CALC:LIM:LOW:SHIF 1 CALC:LIM:LOW:SHIF?

Remote Command Notes	FSE, FSP, FSU
Preset	0

CALCulate<1|2>:LIMit<1...6>:LOWer:SPACing LINear | LOGarithmic

The command switches between linear or logarithmic interpolation for the lower limit line.

The query returns the current state of this setting.

Remote Command	:CALCulate [1] :LIMit [1] 2 3 4 5 6 :LOWer:SPACing LINear LOGarithmic :CALCulate [1] :LIMit [1] 2 3 4 5 6 :LOWer:SPACing?
Example	CALC:LIM:LOW:SPAC LIN CALC:LIM:LOW:SPAC?
Remote Command Notes	FSE, FSP, FSU
Preset	LIN
State Saved	Saved in instrument state.
Range	LINear LOGarithmic

CALCulate<1|2>:LIMit<1...6>:LOWer:THReshold <numeric_value>

The command defines an absolute threshold value for limit lines with relative Y-axis scaling.

The absolute threshold value is used for the limit check when it exceeds the relative limit value.

This mode supports only dBm units.

The query returns the current value of this setting.

Remote Command	CALCulate [1] :LIMit [1] 2 3 4 5 6 :LOWer:THReshold <ampl> CALCulate [1] :LIMit [1] 2 3 4 5 6 :LOWer:THReshold?
Example	CALC:LIM:LOW:THR 1DBM CALC:LIM:LOW:THR?
Remote Command Notes	FSP, FSU
Preset	-200dBm

CALCulate:LIMit:UPPer Subsystem

The CALCulate:LIMit:UPPer subsystem controls the upper limit line.

NOTE All limit functions are not supported in the CCDF and APD measurements.

CALCulate<1|2>:LIMit<1...6>:UPPer[:DATA] <numeric_value>,<numeric_value>..

The command defines a set of values for the specified upper limit line.

This mode does not support any unit for this command.

The query returns the current value set.

Remote Command	:CALCulate [1] :LIMit [1] 2 3 4 5 6 :UPPer [:DATA] <real> , <real>... :CALCulate [1] :LIMit [1] 2 3 4 5 6 :UPPer [:DATA] ?
Example	CALC:LIM:UPP:DATA -10,0,0,-10,-5 CALC:LIM:UPP:DATA?
Remote Command Notes	FSE, FSP, FSU
State Saved	Saved in instrument state.

CALCulate<1|2>:LIMit<1...6>:UPPer:STATe ON | OFF

The command activates or deactivates the specified upper limit line. Limit check is activated separately, using CALC:LIM:STAT ON.

The query returns the current state of this setting.

Remote Command	:CALCulate [1] :LIMit [1] 2 3 4 5 6 :UPPer :STATe ON OFF 1 0 :CALCulate [1] :LIMit [1] 2 3 4 5 6 :UPPer :STATe?
Example	CALC:LIM:UPP:STAT OFF CALC:LIM:UPP:STAT?
Remote Command Notes	FSE, FSP, FSU
Preset	OFF
State Saved	Saved in instrument state.

CALCulate<1|2>:LIMit<1...6>:UPPer:OFFSet <numeric_value>

The command defines an offset for the Y-axis of the specified relative upper limit line.

This mode does not support any unit for this command.

The query returns the current value of this setting.

Remote Command	CALCulate [1] :LIMit [1] 2 3 4 5 6 :UPPer :OFFSet <real> CALCulate [1] :LIMit [1] 2 3 4 5 6 :UPPer :OFFSet?
Example	CALC:LIM:UPP:OFFS 1 CALC:LIM:UPP:OFFS?
Remote Command Notes	FSE, FSP, FSU

Preset	0
--------	---

CALCulate<1|2>:LIMit<1...6>:UPPer:MARGin <numeric_value>

The command defines a margin for the specified upper limit line. At the margin level, out-of-limit values are signaled (if the limit check is active), but are not treated as violations of the limit value.

This mode does not support any unit for this command.

The query returns the current value of this setting.

Remote Command	:CALCulate [1] :LIMit [1] 2 3 4 5 6 :UPPer:MARGin <real> :CALCulate [1] :LIMit [1] 2 3 4 5 6 :UPPer:MARGin?
Example	CALC:LIM:UPP:MARG 10 CALC:LIM:UPP:MARG?
Remote Command Notes	FSE, FSP, FSU
Preset	0

CALCulate<1|2>:LIMit<1...6>:UPPer:MODE RELative | ABSolute

The command switches between relative or absolute scaling for the Y-axis of the specified upper limit line.

Selecting RELative causes the unit to be switched to dB.

The query returns the current state of this setting.

Remote Command	:CALCulate [1] :LIMit [1] 2 3 4 5 6 :UPPer:MODE RELative ABSolute :CALCulate [1] :LIMit [1] 2 3 4 5 6 :UPPer:MODE?
Example	CALC:LIM:UPP:MODE ABS CALC:LIM:UPP:MODE?
Remote Command Notes	FSE, FSP, FSU
Preset	ABS
State Saved	Saved in instrument state.
Range	RELative ABSolute

CALCulate<1|2>:LIMit<1...6>:UPPer:SHIFt <numeric_value>

The command shifts the specified limit line in the Y-direction by the specified amount.

This mode does not support any unit for this command.

The query returns the current value of this setting.

Remote Command	CALCulate[1]:LIMit[1] 2 3 4 5 6:UPPer:SHIFt <real> CALCulate[1]:LIMit[1] 2 3 4 5 6:UPPer:SHIFt?
Example	CALC:LIM:UPP:SHIF 1 CALC:LIM:UPP:SHIF?
Remote Command Notes	FSE, FSP, FSU
Preset	0

CALCulate<1|2>:LIMit<1...6>:UPPer:SPACing LINear | LOGarithmic

The command switches between linear or logarithmic interpolation for the upper limit line.

The query returns the current state of this setting.

Remote Command	:CALCulate[1]:LIMit[1] 2 3 4 5 6:UPPer:SPACing LINear LOGarithmic :CALCulate[1]:LIMit[1] 2 3 4 5 6:UPPer:SPACing?
Example	CALC:LIM:UPP:SPAC LIN CALC:LIM:UPP:SPAC?
Remote Command Notes	FSE, FSP, FSU
Preset	LIN
State Saved	Saved in instrument state.
Range	LINear LOGarithmic

CALCulate<1|2>:LIMit<1...6>:UPPer:THReshold <numeric_value>

The command specifies an absolute threshold value for limit lines with relative Y-axis scaling. The absolute threshold value is used for the limit check when it exceeds the relative limit value.

This mode supports only dBm units.

The query returns the current value of this setting.

Remote Command	:CALCulate[1]:LIMit[1] 2 3 4 5 6:UPPer:THReshold <ampl> :CALCulate[1]:LIMit[1] 2 3 4 5 6:UPPer:THReshold?
Example	CALC:LIM:UPP:THR 1DBM CALC:LIM:UPP:THR?
Remote Command Notes	FSP, FSU
Preset	-200dBm

CALCulate:MARKer Subsystem

The CALCulate:MARKer subsystem controls marker functions in the instrument.

NOTE All marker functions are not supported in the CCDF and APD measurements.

CALCulate<1|2>:MARKer<1...4>[:STATe] ON | OFF

The command switches the specified marker on or off.

If no numeric suffix is supplied, marker 1 is selected automatically.

If marker 2, 3 or 4 is selected and used as a delta marker, it is switched to marker mode.

The query returns the current state of this setting.

Remote Command	:CALCulate [1] :MARKer [1] 2 3 4 :STATe ON OFF 1 0 :CALCulate [1] :MARKer [1] 2 3 4 :STATe?
Example	CALC:MARK3:STAT ON
Remote Command Notes	FSE, FSP, FSU
Preset	OFF
State Saved	Saved in instrument state.

CALCulate<1|2>:MARKer<1...4>:AOFF

Turns off all active markers.

All delta markers are turned off.

All active marker and delta marker measurement functions are turned off.

Remote Command	:CALCulate [1] :MARKer [1] 2 3 4 :AOFF
Example	CALC:MARK:AOFF
Remote Command Notes	FSE, FSP, FSU

CALCulate<1|2>:MARKer<1...4>:TRACe 1 to 3

The command assigns the specified marker to the trace defined by the command parameter.

If required, the corresponding marker is turned on.

The query returns the current assignment.

Remote Command	:CALCulate [1] :MARKer [1] 2 3 4 :TRACe 1 2 3 :CALCulate [1] :MARKer [1] 2 3 4 :TRACe?
Example	CALC:MARK:TRAC 2

Remote Command Notes	FSE, FSP, FSU
Preset	1
Min	1
Max	3

CALCulate<1|2>:MARKer<1...4>:X <0...MAX> (frequency | sweep time)

The command places the selected marker at the indicated frequency, time, or level (for APD or CCDF measurements).

If marker 2, 3 or 4 is selected and used as delta marker, it is switched to marker mode.

The query returns the current value of this setting.

Remote Command	:CALCulate [1] :MARKer [1] 2 3 4 :X <real> :CALCulate [1] :MARKer [1] 2 3 4 :X?
Example	CALC:MARK:X 1000000000 CALC:MARK:X?
Remote Command Notes	FSE, FSP, FSU

CALCulate<1|2>:MARKer<1...4>:X:SLIMits[:STATe] ON | OFF

The command switches the search range limits on or off.

For this command, the numeric suffix <1...4> is ignored.

If time domain power measurement is active, this command limits the evaluation range on the trace.

The query returns the current state of this setting.

Remote Command	:CALCulate [1] :MARKer [1] 2 3 4 :X:SLIMits [:STATe] ON OFF 1 0 :CALCulate [1] :MARKer [1] 2 3 4 :X:SLIMits [:STATe] ?
Example	CALC:MARK:X:SLIM:STAT ON
Remote Command Notes	FSE, FSP, FSU
Preset	OFF
State Saved	Saved in instrument state.

CALCulate<1|2>:MARKer<1...4>:X:SLIMits:LEFT <0...MAX> (frequency | sweep time)

The command sets the left limit of the search range for markers and delta markers.

The query returns the current value of this setting.

Remote Command	:CALCulate[1]:MARKer[1] 2 3 4:X:SLIMits:LEFT <real> :CALCulate[1]:MARKer[1] 2 3 4:X:SLIMits:LEFT?
Example	CALC:MARK:X:SLIM:LEFT 1MHZ CALC:MARK:X:SLIM:LEFT?
Remote Command Notes	FSP, FSU

CALCulate<1|2>:MARKer<1...4>:X:SLIMits:RIGHT <0...MAX> (frequency | sweep time)

The command sets the right limit of the search range for markers and delta markers.

The query returns the current value of this setting.

Remote Command	:CALCulate[1]:MARKer[1] 2 3 4:X:SLIMits:RIGHT <real> :CALCulate[1]:MARKer[1] 2 3 4:X:SLIMits:RIGHT?
Example	CALC:MARK:X:SLIM:RIGH 1MHZ CALC:MARK:X:SLIM:RIGH?
Remote Command Notes	FSP, FSU

CALCulate<1|2>:MARKer<1...4>:COUNT ON | OFF

The command switches on or off operation of the frequency counter at the marker position.

The query returns the current state of this setting.

Remote Command	:CALCulate[1]:MARKer[1] 2 3 4:COUNT ON OFF 1 0 :CALCulate[1]:MARKer[1] 2 3 4:COUNT?
Example	CALC:MARK:COUN ON
Remote Command Notes	FSE, FSP, FSU
Preset	OFF

CALCulate<1|2>:MARKer<1...4>:COUNT:RESolution 0.1 | 1 | 10 | 100 | 1000 | 10000 Hz

The command specifies the resolution of the frequency counter.

This mode supports the following resolutions: 1 | 10 | 100 | 1000| 10000 | 100000 | 1000000 Hz.

The query returns the current value of this setting.

Remote Command	:CALCulate[1]:MARKer[1] 2 3 4:COUNT:RESolution <freq> :CALCulate[1]:MARKer[1] 2 3 4:COUNT:RESolution?
Example	CALC:MARK:COUN:RES 1kHz CALC:MARK:COUN:RES?

Remote Command Notes	FSE, FSP, FSU
Preset	1KHZ
Min	1
Max	10KHz

CALCulate<1|2>:MARKer<1...4>:COUNT:FREQUency?

Queries the frequency counter result for the specified marker.

Remote Command	:CALCulate [1] :MARKer [1] 2 3 4 :COUNT :FREQUency?
Example	CALC:MARK:COUN:FREQ?
Remote Command Notes	FSE, FSP, FSU

CALCulate<1|2>:MARKer<1...4>:LOEXclude ON | OFF

The command switches local oscillator suppression for peak search on or off.

This setting is valid for all markers and delta markers.

The query returns the current state of this setting.

Remote Command	:CALCulate [1] :MARKer [1] 2 3 4 :LOEXclude ON OFF 1 0 :CALCulate [1] :MARKer [1] 2 3 4 :LOEXclude?
Example	CALC:MARK:LOEX ON
Remote Command Notes	FSE, FSP, FSU
Preset	ON
State Saved	Saved in instrument state.

CALCulate<1|2>:MARKer<1...4>:Y?

Queries the measured value of the specified marker.

If required, the marker is activated, or switched to marker mode.

Remote Command	:CALCulate [1] :MARKer [1] 2 3 4 :Y?
Example	CALC:MARK2:Y 1
Remote Command Notes	FSE, FSP, FSU

CALCulate<1|2>:MARKer<1...4>:Y:PERCent

This command positions the selected marker in the selected window to the given probability. If marker 2, 3 or 4 is selected and used as a delta marker, it is switched to marker mode.

This command is accepted, but takes no action and reports no error.

Remote Command	:CALCulate [1] :MARKer [1] 2 3 4 :Y:PERCent <float> :CALCulate [1] :MARKer [1] 2 3 4 :Y:PERCent?
Example	CALC:MARK2:Y:PERC 95PCT
Remote Command Notes	FSP, FSU
Min	0
Max	100

CALCulate<1|2>:MARKer<1...4>:MAXimum[:PEAK]

Places the delta marker at the current maximum on the measured curve. If required, the delta marker is first activated.

This command is an event, and therefore has no *RST value and no query.

Remote Command	:CALCulate [1] :MARKer [1] 2 3 4 :MAXimum [:PEAK]
Example	CALC:MARK:MAX
Remote Command Notes	FSE, FSP, FSU

CALCulate<1|2>:MARKer<1...4>:MAXimum:NEXT

Places the marker at the next smaller maximum value of the corresponding trace.

Remote Command	:CALCulate [1] :MARKer [1] 2 3 4 :MAXimum:NEXT
Example	CALC:MARK:MAX:NEXT
Remote Command Notes	FSE, FSP, FSU

CALCulate<1|2>:MARKer<1...4>:MAXimum:LEFT

Places the marker at the next smaller maximum to the left of the current value (that is, in order of decreasing X values) on the trace.

Remote Command	:CALCulate [1] :MARKer [1] 2 3 4 :MAXimum:LEFT
Example	CALC:MARK:MAX:LEFT
Remote Command Notes	FSE, FSP, FSU

CALCulate<1|2>:MARKer<1...4>:MAXimum:RIGHT

Places the delta marker at the next smaller maximum to the right of the current value (that is, in order of increasing X values). If required, the corresponding delta marker is first activated.

This command is an event, and therefore has no *RST value and no query.

Remote Command	CALCulate [1] :MARKer [1] 2 3 4 :MINimum:RIGHT
Example	CALC:MARK:MIN:RIGH
Remote Command Notes	FSE, FSP, FSU

CALCulate<1|2>:MARKer<1...4>:MAXimum:AUTO ON | OFF

For marker 1, the command switches on or off operation of automatic maximum peak search at the end of each sweep.

The marker search limit settings are taken into account.

For this command, the numeric suffix <1 to 4> is ignored.

The query returns the current state of this setting.

Remote Command	:CALCulate [1] :MARKer [1] 2 3 4 :MAXimum:AUTO ON OFF 1 0 :CALCulate [1] : MARKer [1] 2 3 4 :MAXimum:AUTO?
Example	CALC:MARK:MAX:AUTO ON CALC:MARK:MAX:AUTO?
Remote Command Notes	FSU
Preset	OFF
State Saved	Saved in instrument state.

CALCulate<1|2>:MARKer<1...4>:MINimum:AUTO ON | OFF

For marker 1, the command switches on or off operation of automatic minimum peak search at the end of each sweep.

The marker search limit settings are taken into account.

For this command, the numeric suffix <1 to 4> is ignored.

The query returns the current state of this setting.

Remote Command	:CALCulate [1] :MARKer [1] 2 3 4 :MINimum:AUTO ON OFF 1 0 :CALCulate [1] : MARKer [1] 2 3 4 :MINimum:AUTO?
Example	CALC:MARK:MIN:AUTO ON CALC:MARK:MIN:AUTO?
Remote Command Notes	FSU
Preset	OFF
State Saved	Saved in instrument state.

CALCulate<1|2>:MARKer<1...4>:MINimum[:PEAK]

Places the delta marker at the current maximum on the trace.

If required, the corresponding delta marker is first activated.

This command is an event, and therefore has no *RST value and no query.

Remote Command	:CALCulate [1] :MARKer [1] 2 3 4 :MINimum [:PEAK]
Example	CALC:MARK:MIN
Remote Command Notes	FSE, FSP, FSU

CALCulate<1|2>:MARKer<1...4>:MINimum:NEXT

Places the marker at the next smaller maximum of the corresponding trace.

Remote Command	:CALCulate [1] :MARKer [1] 2 3 4 :MINimum:NEXT
Example	CALC:MARK:MIN:NEXT
Remote Command Notes	FSE, FSP, FSU

CALCulate<1|2>:MARKer<1...4>:MINimum:RIGHT

Places the delta marker at the next smaller maximum to the right of the current value (that is, in order of increasing X values).

If required, the corresponding delta marker is first activated.

This command is an event, and therefore has no *RST value and no query.

Remote Command	CALCulate [1] :MARKer [1] 2 3 4 :MAXimum:RIGHT
Example	CALC:MARK:MAX:RIGH
Remote Command Notes	FSE, FSP, FSU

CALCulate<1|2>:MARKer<1...4>:MINimum:LEFT

Places the marker at the next smaller maximum to the left of the current value (that is, in order of decreasing X values) on the trace.

Remote Command	:CALCulate [1] :MARKer [1] 2 3 4 :MINimum:LEFT
Example	CALC:MARK:MIN:LEFT
Remote Command Notes	FSE, FSP, FSU

CALCulate<1|2>:MARKer<1...4>:PEXCursion <numeric_value>

The command specifies the peak excursion.

This mode supports only dB units.

The query returns the current value of this setting.

Remote Command	:CALCulate [1] :MARKer [1] 2 3 4 :PEXCursion <rel_amp1> :CALCulate [1] :MARKer [1] 2 3 4 :PEXCursion?
Example	CALC:MARK:PEXC 10dB
Remote Command Notes	FSE, FSP, FSU
Preset	6dB
State Saved	Saved in instrument state.

CALCulate:MARKer:FUNCTION Subsystem

CALCulate<1|2>:MARKer<1...4>:FUNCTION:NOISe[:STATe]

The command switches noise measurement on or off for all markers.

The noise power density is measured at the marker positions. The result may be queried using CALCulate:MARKer:FUNCTION:NOISe:RESult?.

The query returns the current state of this setting.

Remote Command	:CALCulate [1] :MARKer [1] 2 3 4 :FUNCTION:NOISe [:STATe] ON OFF 1 0 :CALCulate [1] :DELTaMarker [1] 2 3 4 :FUNCTION:NOISe [:STATe] e] ?
Example	CALC:MARK:FUNC:NOIS ON
Remote Command Notes	FSE, FSP, FSU
Preset	OFF
State Saved	Saved in instrument state.

CALCulate<1|2>:MARKer<1...4>:FUNCTION:NOISe:RESult?

Queries the noise measurement result.

To obtain a valid result, a complete sweep must be performed after switching on the function, but before querying the measured value. This is possible only in Single sweep mode.

Remote Command	:CALCulate [1] :MARKer [1] 2 3 4 :FUNCTION:NOISe:RESult?
Example	CALC:MARK:FUNC:NOIS:RES?
Remote Command Notes	FSE, FSP, FSU

CALCulate<1|2>:MARKer<1...4>:FUNCTION:REFerence?

Sets the reference level to the power measured by the specified marker.

If marker 2, 3 or 4 is selected, and used as delta marker, it is switched to marker mode.

This command is an event, and therefore has neither *RST value nor query.

The value of the numeric suffix <1 to 4> is ignored.

Remote Command	:CALCulate [1] :MARKer [1] 2 3 4 :FUNction:REFerence
Example	CALC:MARK:FUNC:REF
Remote Command Notes	FSE, FSP, FSU

CALCulate:MARKer:FUNction:POWer Subsystem

The CALCulate:MARKer:FUNction:POWer subsystem includes commands for control of power measurement.

CALCulate<1|2>:MARKer<1...4>:FUNction:POWer:SElect ACPower | CPOWer | MCACpower | OBANdwidth | OBWidth | CN | CN0

Selects and activates the specified power measurement type.

This mode supports: ACPower, MCACpower, CPOWer, OBANdwidth and OBWidth.

Remote Command	CALCulate [1] :MARKer [1] 2 3 4 :FUNction:POWer:SElect ACPower CPOWer MCACpower OBANdwidth OBWidth CN CN0
Example	CALC:MARK:FUNC:POW:SEL OBW
Remote Command Notes	FSE, FSP, FSU

CALCulate<1|2>:MARKer<1...4>:FUNction:POWer:RESult? ACPower | CPOWer | MCACpower | OBANdwidth | OBWidth | CN | CN0

Queries the result for the specified power measurement type.

This mode supports ACPower, CPOWer, MCACpower, OBANdwidth and OBWidth.

Remote Command	CALCulate [1] :MARKer [1] 2 3 4 :FUNction:POWer:RESult? ACPower CPOWer MCACpower OBANdwidth OBWidth CN CN0
Example	CALC:MARK:FUNC:POW:RES? OBW
Remote Command Notes	FSE, FSP, FSU

CALCulate<1|2>:MARKer<1...4>:FUNction:POWer:RESult:PHZ ON | OFF

The command switches the query response for power measurement results between absolute values (OFF) and output referred to measurement bandwidth (ON).

The query returns the current state of this setting.

Remote Command	:CALCulate [1] :MARKer [1] 2 3 4 :FUNction:POWer:RESult:PHZ ON OFF 1 0 :CALCulate [1] :MARKer [1] 2 3 4 :FUNction:POWer:RESult:PHZ ?
Example	CALC:MARK:FUNC:POW:RES:PHZ OFF CALC:MARK:FUNC:POW:RES:PHZ?
Remote Command Notes	FSP, FSU
Preset	OFF
State Saved	Saved in instrument state.

CALCulate:STATistics Subsystem

The CALCulate:STATistics subsystem controls the statistical measurement functionality.

CALCulate:STATistics:APD[:STATe] ON | OFF

This command is accepted, but takes no action and reports no error.

Remote Command	:CALCulate:STATistics:APD:STATe ON OFF 1 0 :CALCulate:STATistics:APD:STATe?
Example	CALC:STAT:APD:STAT ON
Remote Command Notes	FSP, FSU
Preset	OFF
State Saved	Saved in instrument state.

CALCulate:STATistics:CCDF[:STATe] ON | OFF

The command activates or deactivates measurement of the complementary cumulative distribution function (CCDF).

When this function is activated, the APD measurement is turned off.

The query returns the current state of this setting.

Remote Command	:CALCulate:STATistics:CCDF:STATe ON OFF 1 0 :CALCulate:STATistics:CCDF:STATe?
Example	CALC:STAT:CCDF:STAT ON
Remote Command Notes	FSP, FSU
Preset	OFF
State Saved	Saved in instrument state.

CALCulate:STATistics:CCDF:X<1...3>? P0_01 | P0_1 | P1 | P10

Queries the level values for the probabilities 0.01%, 0.1%, 1% and 10%.

Remote Command	CALCulate:STATistics:CCDF:X[1] 2 3? P0_01 P0_1 P1 P10
Example	CALC:STAT:CCDF:X? P10
Remote Command Notes	FSU

CALCulate:STATistics:NSAMples <100...1E9>

The command specifies the number of measurement points to be acquired for the statistical measurement functions.

The minimum value supported by this mode is 1000.

The query returns the current value of this setting.

Remote Command	:CALCulate[1]:STATistics:NSAMples <int> :CALCulate[1]:STATistics:NSAMples?
Example	CALC:STAT:NSAM 100000 CALC:STAT:NSAM?
Remote Command Notes	FSP, FSU
Preset	100000
State Saved	Saved in instrument state.
Min	1000
Max	1E9

CALCulate:STATistics:SCALE:AUTO ONCE

This command is accepted, but takes no action and reports no error.

Remote Command	:CALCulate[1]:STATistics:SCALE:AUTO ONCE
Example	CALC:STAT:SCALE:AUTO ONCE
Remote Command Notes	FSP, FSU

CALCulate:STATistics:SCALE:X:RLEVEL -130dBm to 30dBm

Specifies the reference level for the X-axis of the measurement display. The setting is identical to the reference level setting using the command DISPLAY:WINDOW:TRACE:Y:RLEVEL.

This command is accepted, but takes no action and reports no error.

Remote Command	:CALCulate[1]:STATistics:SCALE:X:RLEVEL <ampl> :CALCulate[1]:STATistics:SCALE:X:RLEVEL?
-----------------------	--

Example	CALC:STAT:SCAL:X:RLEV -60dBm CALC:STAT:SCAL:X:RLEV?
Remote Command Notes	FSP, FSU
Preset	-20dBm
State Saved	Saved in instrument state.
Min	-130
Max	30

CALCulate:STATistics:SCALE:X:RANGe 10dB to 200dB

Specifies the level range for the X-axis of the measurement display. The setting is identical to the level range setting defined with the command DISPLAY:WINDOW:TRACE:Y:SCALE.

This command is accepted, but takes no action and reports no error.

Remote Command	:CALCulate [1] :STATistics:SCALE:X:RANGe <rel_ampl> :CALCulate [1] :STATistics:SCALE:X:RANGe?
Example	CALC:STAT:SCAL:X:RANG 20dB CALC:STAT:SCAL:X:RANG?
Remote Command Notes	FSP, FSU
Preset	100dB
State Saved	Saved in instrument state.
Min	10
Max	200

CALCulate:STATistics:SCALE:Y:UPPer 1E-8 to 1.0

Specifies the upper limit for the Y-axis of the display, in statistical measurements. Since probabilities are specified on the Y axis, the entered numerical values are dimensionless.

This command is accepted, but takes no action and reports no error.

Remote Command	:CALCulate [1] :STATistics:SCALE:Y:UPPer <real> :CALCulate [1] :STATistics:SCALE:Y:UPPer?
Example	CALC:STAT:SCAL:Y:UPP 0.01 CALC:STAT:SCAL:Y:UPP?
Remote Command Notes	FSP, FSU
Preset	1.0
State Saved	Saved in instrument state.

Min	1e-8
Max	1.0

CALCulate:STATistics:SCALE:Y:LOWer 1E-9 to 0.1

Specifies the lower limit for the Y-axis of the display, in statistical measurements. Since probabilities are specified on the Y axis, the entered numerical values are dimensionless.

This command is accepted, but takes no action and reports no error.

Remote Command	:CALCulate[1]:STATistics:SCALE:Y:LOWer <real> :CALCulate[1]:STATistics:SCALE:Y:LOWer?
Example	CALC:STAT:SCAL:Y:LOW 0.001 CALC:STAT:SCAL:Y:LOW?
Remote Command Notes	FSP, FSU
Preset	1e-6
State Saved	Saved in instrument state.
Min	1e-9
Max	0.1

CALCulate:STATistics:PRESet

Resets the scaling of the X- and Y-axes in a statistical measurement. The following values are set:

This command is accepted, but takes no action and reports no error.

Remote Command	:CALCulate[1]:STATistics:PRESet
Example	CALC:STAT:PRES
Remote Command Notes	FSP, FSU

CALCulate:STATistics:RESult<1...3>? MEAN | PEAK | CFACTOR | ALL

Queries the results of statistical measurements of a recorded trace.

Currently, only TRACE1 is valid for CCDF measurements.

Remote Command	CALCulate:STATistics:RESult? MEAN PEAK CFACTOR ALL
Example	CALC:STAT:RES? ALL
Remote Command Notes	FSP, FSU

CALCulate:THReshold Subsystem

The CALCulate:THReshold subsystem controls the threshold value for maximum/minimum marker

searches.

CALCulate<1|2>:DLINe<1|2> MINimum...MAXimum

Specifies the position of Display Line 1 or 2. These lines enable the user to mark any levels in the diagram. The unit depends on the setting made with CALC:UNIT.

This command is accepted, but takes no action and reports no error.

Remote Command	:CALCulate [1] :DLINe [1] 2 <ampl> :CALCulate [1] :DLINe [1] 2?
Example	CALC:DLIN -20DBM
Remote Command Notes	FSE, FSP, FSU
State Saved	Saved in instrument state.

CALCulate<1|2>:DLINe<1|2>:STATe ON | OFF

The command turns Display Line 1 or 2 (level lines) on or off.

The query returns the current state of this setting.

Remote Command	:CALCulate [1] :DLINe [1] 2 :STATe ON OFF 1 0 :CALCulate [1] :DLINe [1] 2 :STATe?
Example	CALC:DLIN:STAT ON
Remote Command Notes	FSE, FSP, FSU
Preset	OFF
State Saved	Saved in instrument state.

CALCulate<1|2>:THReshold MINimum to MAXimum (depending on current unit)

The command specifies the threshold value for the maximum/minimum search of markers with marker search functions MAX PEAK, NEXT PEAK, etc.

The associated display line is automatically turned on.

The query returns the current value of this setting.

Remote Command	:CALCulate [1] :THReshold <ampl> :CALCulate [1] :THReshold?
Example	CALC:THR -82DBM
Remote Command Notes	FSE, FSP, FSU
State Saved	Saved in instrument state.

CALCulate<1|2>:THReshold:STATe ON | OFF

The command turns the threshold line on or off.

The unit depends on the setting performed by CALC:UNIT.

The query returns the current state of this setting.

Remote Command	:CALCulate [1] :THReshold:STATe ON OFF 1 0 :CALCulate [1] :THReshold:STATe?
Example	CALC:THR:STAT ON
Remote Command Notes	FSE, FSP, FSU
Preset	OFF
State Saved	Saved in instrument state.

CALCulate<1|2>:FLINe<1|2> 0...fmax

Specifies the position of the frequency lines.

This command is accepted, but takes no action and reports no error.

Remote Command	:CALCulate [1] :FLINe [1] 2 <freq> :CALCulate [1] :FLINe [1] 2?
Example	CALC:FLIN 120MHz
Remote Command Notes	FSE, FSP, FSU
State Saved	Saved in instrument state.

CALCulate<1|2>:FLINe<1|2>:STATe ON | OFF

Switches the frequency line on or off.

This command is accepted, but takes no action and reports no error.

Remote Command	:CALCulate [1] :FLINe [1] 2:STATe ON OFF 1 0 :CALCulate [1] :FLINe [1] 2:STATe?
Example	CALC:FLIN:STAT ON
Remote Command Notes	FSE, FSP, FSU
Preset	OFF
State Saved	Saved in instrument state.

CALCulate<1|2>:TLINe<1|2> 0...1000s

Specifies the position of the time lines.

This command is accepted, but takes no action and reports no error.

Remote Command	:CALCulate [1] :TLINe [1] 2 <time> :CALCulate [1] :TLINe [1] 2?
Example	CALC:TLIN 10ms
Remote Command Notes	FSE, FSP, FSU
State Saved	Saved in instrument state.

CALCulate<1|2>:TLINe<1|2>:STATe ON | OFF

Switches the time line on or off.

This command is accepted, but takes no action and reports no error.

Remote Command	:CALCulate [1] :TLINe [1] 2:STATe ON OFF 1 0 :CALCulate [1] :TLINe [1] 2:STATe?
Example	CALC:TLIN:STAT ON
Remote Command Notes	FSE, FSP, FSU
Preset	OFF
State Saved	Saved in instrument state.

CALCulate:UNIT Subsystem

The CALCulate:Unit subsystem specifies the units for power measurement settings.

CALCulate<1|2>:UNIT:POWer DBM | V | A | W | DB | PCT | UNITLESS |DBPW | WATT | DBUV |DBMV | VOLT | DBUA | AMPere

The command selects the power units.

This mode supports the following unit parameters: DBM | DBUV | DBMV | DBUA | V | A | W.

The query returns the current unit selection.

Remote Command	:CALCulate [1] :UNIT:POWer DBM DBUV DBMV DBUA V A W :CALCulate [1] :UNIT:POWer?
Example	CALC:UNIT:POW DBM
Remote Command Notes	FSE, FSP, FSU
Preset	DBM
State Saved	Saved in instrument state.
Range	DBM DBUV DBMV DBUA V A W
Backward Compatibility	UNIT[1] 2:POWer

dBm

Sets the unit for the selected amplitude scale (log/lin) to dBm.

Remote Command Example:	UNIT:POW DBM
-------------------------	--------------

dBmV

Sets the unit for the selected amplitude scale (log/lin) to dBmV.

Remote Command Example:	UNIT:POW DBMV
-------------------------	---------------

W

Sets the unit for the selected amplitude scale (log/lin) to Watt.

Remote Command Example:	UNIT:POW W
-------------------------	------------

V

Sets the unit for the selected amplitude scale (log/lin) to Volt.

Remote Command Example:	UNIT:POW V
-------------------------	------------

A

Sets the unit for the selected amplitude scale (log/lin) to Ampere.

Remote Command Example:	UNIT:POW A
-------------------------	------------

dB μ V

Sets the unit for the selected amplitude scale (log/lin) to dB μ V.

Remote Command Example:	UNIT:POW DBUV
-------------------------	---------------

dB μ A

Sets the unit for the selected amplitude scale (log/lin) to dB μ A.

Remote Command Example:	UNIT:POW DBUA
-------------------------	---------------

CALibration Subsystem

The CALibration subsystem specifies the data for system error correction.

CALibration[:ALL]?

Initiates the acquisition of system error correction data.

If the acquisition was successful, "0" is returned.

Mode	All
Remote Command	:CALibration[:ALL]?
Example	:CAL?
Remote Command Notes	FSE, FSP, FSU

CALibration:STATe ON | OFF

Specifies whether the current calibration data is to be taken into account by the instrument (ON) or not (OFF).

Remote Command	:CALibration:STATe ON OFF 1 0
Example	CAL:STAT OFF
Remote Command Notes	FSE, FSP, FSU
Preset	ON
State Saved	No

DISPlay Subsystem

The DISPlay subsystem commands control the selection and presentation of textual and graphical information and measurement data.

This mode does not support the screen display function. All DISPlay-related SCPI commands are accepted, but take no action and report no error.

DISPlay:FORMat SINGLE | SPLit

Switches the measurement result display between FULL SCREEN and SPLIT SCREEN.

This command is accepted, but takes no action and reports no error.

Remote Command	:DISPlay:FORMat SINGLE SPLit :DISPlay:FORMat?
Example	DISP:FORM SING
Remote Command Notes	FSE, FSP, FSU
Preset	SINGLE
State Saved	Saved in instrument state.
Range	SINGLE SPLit

DISPlay:ANNotation:FREQuency ON | OFF

Switches X-axis annotation on or off.

This command is accepted, but takes no action and reports no error.

Remote Command	:DISPlay:ANNotation:FREQuency ON OFF 1 0 :DISPlay:ANNotation:FREQuency?
Example	DISP:ANN:FREQ OFF DISP:ANN:FREQ?
Remote Command Notes	FSE, FSP, FSU
Preset	ON
State Saved	Saved in instrument state.

DISPlay:LOGO ON | OFF

Switches the on-screen company logo on or off.

This command is accepted, but takes no action and reports no error.

Remote Command	:DISPlay:LOGO ON OFF 1 0 :DISPlay:LOGO?
Example	DISP:LOGO OFF DISP:LOGO?
Remote Command Notes	FSE, FSP, FSU
Preset	ON
State Saved	Saved in instrument state.

DISPlay:PSAVe[:STATe] ON | OFF

Switches the display power-save mode on or off.

This command is accepted, but takes no action and reports no error.

Remote Command	:DISPlay:PSAVe[:STATe] ON OFF 1 0 :DISPlay:PSAVe[:STATe]?
Example	DISP:PSAV OFF DISP:PSAV?
Remote Command Notes	FSE, FSP, FSU
Preset	OFF
State Saved	Saved in instrument state.

DISPlay:PSAVe:HOLDoff 1 to 60

Specifies the holdoff time for the display's power-save mode.

Compatible SCPI Commands
Supported SCPI Commands

This command is accepted, but takes no action and reports no error.

Remote Command	:DISPlay:PSAVe:HOLDoff <int> :DISPlay:PSAVe:HOLDoff?
Example	DISP:PSAV:HOLD 30 DISP:PSAV:HOLD?
Remote Command Notes	FSE, FSP, FSU
Preset	15
Min	1
Max	60

DISPlay:CMAP<1 to 26>:DEFault<1|2>

Resets the display's screen colors to their defaults.

This command is accepted, but takes no action and reports no error.

Remote Command	:DISPlay:CMAP [1] 2 3 4 5 6 7 8 9 10 11 12 13 14 15 16 17 18 19 20 21 22 23 24 25 26 :DEFault [1] 2
Example	DISP:CMAP:DEF
Remote Command Notes	FSE, FSP, FSU

DISPlay:CMAP<1 to 26>:HSL <hue>,<sat>,<lum>

Defines the instrument's color map.

This command is accepted, but takes no action and reports no error.

Remote Command	:DISPlay:CMAP [1] 2 3 4 5 6 7 8 9 10 11 12 13 14 15 16 17 18 19 20 21 22 23 24 25 26 :HSL <real>,<real>,<real> :DISPlay:CMAP [1] 2 3 4 5 6 7 8 9 10 11 12 13 14 15 16 17 18 19 20 21 22 23 24 25 26 :HSL?
Example	DISP:CMAP:HSL 0.3,0.8,1.0 DISP:CMAP:HSL?
Remote Command Notes	FSE, FSP, FSU
State Saved	Saved in instrument state.

DISPlay:CMAP<1 to 26>:PDEFined BLACK | BLUE | BROWn | GREen | CYAN | RED | MAGenta | YELLow | WHITe | DGRAY | LGRAY | LBLUe | LGREen | LCYan | LRED | LMAGenta

Specifies predefined color values for the instrument's color map.

This command is accepted, but takes no action and reports no error.

Remote Command	:DISPlay:CMAP [1] 2 3 4 5 6 7 8 9 10 11 12 13 14 15 16 17 18 19 20 21 22 23 24 25 26 :PDEFined BLACk BLUE BROWn GREen CYAN RED MAGenta YELLow WHITE DG RAY LGRAY LBLUe LGREen LCYan LRED LMAGenta :DISPlay:CMAP [1] 2 3 4 5 6 7 8 9 10 11 12 13 14 15 16 17 18 19 20 21 22 23 24 25 26 :PDEFined?
Example	DISP:CMAP:PDEF BLAC DISP:CMAP:PDEF?
Remote Command Notes	FSE, FSP, FSU
State Saved	Saved in instrument state.
Range	BLACk BLUE BROWn GREen CYAN RED MAGenta YELLow WHITE DGRAY LGRAY LBLUe LGREen LCYan LRED LMAGenta

DISPlay[:WINDow<1|2>]:SElect

This command is accepted, but takes no action and reports no error.

Remote Command	:DISPlay:WINDow [1] :SElect
Example	DISP:WIND:SEL
Remote Command Notes	FSE, FSP, FSU
Backwards Compatibility SCPI	:DISPlay:SElect

DISPlay[:WINDow<1|2>]:SIZE LARGe|SMALl

Switches the measurement window between full screen and half screen.

This command is accepted, but takes no action and reports no error.

Remote Command	:DISPlay:WINDow [1] :SIZE LARGe SMALl :DISPlay:WINDow [1] :TRACe [1] 2 3 :MODE?
Example	DISP:WIND:SIZE LARG DISP:WIND:SIZE?
Remote Command Notes	FSP, FSU
Preset	SMALl
State Saved	Saved in instrument state.
Range	LARGe SMALl
Backwards Compatibility SCPI	:DISPlay:SIZE

DISPlay[:WINDow<1|2>]:TEXT[:DATA] <string>

Specifies a comment (max. 20 characters) that can be displayed on the screen.

This command is accepted, but takes no action and reports no error.

Remote Command	:DISPlay:WINDow[1]:TEXT[:DATA] <string> :DISPlay:WINDow[1]:TEXT[:DATA] ?
Example	DISP:WIND:TEXT:DATA "123" DISP:WIND:TEXT:DATA?
Remote Command Notes	FSE, FSP, FSU
Preset	""
State Saved	Saved in instrument state.
Backwards Compatibility SCPI	:DISPlay:TEXT[:DATA]

DISPlay[:WINDow<1|2>]:TEXT:STATe ON | OFF

Switches display of the comment (screen title) on or off.

This command is accepted, but takes no action and reports no error.

Remote Command	:DISPlay:WINDow[1]:TEXT :STATe ON OFF 1 0 :DISPlay:WINDow[1]:TEXT :STATe?
Example	DISP:WIND:TEXT:STAT ON DISP:WIND:TEXT:STAT?
Remote Command Notes	FSE, FSP, FSU
Preset	OFF
State Saved	Saved in instrument state.
Backwards Compatibility SCPI	:DISPlay:TEXT:STATe

DISPlay[:WINDow<1|2>]:TIME ON | OFF

Switches on-screen display of date and time on or off.

This command is accepted, but takes no action and reports no error.

Remote Command	:DISPlay:WINDow[1]:TIME ON OFF 1 0 :DISPlay:WINDow[1]:TIME?
Example	DISP:WIND:TIME ON DISP:WIND:TIME?
Remote Command Notes	FSE, FSP, FSU

Preset	OFF
State Saved	Saved in instrument state.
Backwards Compatibility SCPI	:DISPlay:TIME

DISPlay[:WINDow<1|2>]:TRACe<1 to 3>:Y[:SCALe] 10dB to 200dB

The command specifies the display range of the Y-axis (level axis) with logarithmic scaling (DISP:TRAC:Y:SPAC LOG).

The query returns the current value of this setting.

Remote Command	DISPlay:WINDow[1]:TRACe[1] 2 3:Y[:SCALe] <rel_amp> DISPlay:WINDow[1]:TRACe[1] 2 3:Y[:SCALe]?
Example	DISP:WIND:TRAC:Y 20dB DISP:WIND:TRAC:Y?
Remote Command Notes	FSE, FSP, FSU
Preset	100dB
Min	10dB
Max	200dB
Backwards Compatibility SCPI	:DISPlay[1]:TRACe[1] 2 3:Y[:SCALe]

DISPlay[:WINDow<1|2>]:TRACe<1 to 3>:Y[:SCALe]:MODE ABSolute | RELAtiv

Specifies the scale type of the Y-axis (absolute or relative).

This command is accepted, but takes no action and reports no error.

Remote Command	:DISPlay:WINDow[1]:TRACe:Y[:SCALe]:MODE ABSolute RELAtiv :DISPlay:WINDow[1]:TRACe:Y[:SCALe]:MODE?
Example	DISP:WIND:TRAC:Y:MODE ABS DISP:WIND:TRAC:Y:MODE?
Remote Command Notes	FSE, FSP, FSU
Preset	ABS
State Saved	Saved in instrument state.
Range	ABSolute RELAtiv
Backwards Compatibility SCPI	:DISPlay[1]:TRACe[1] 2 3:Y[:SCALe]:MODE

DISPlay[:WINDow<1|2>]:TRACe<1 to 3>:Y[:SCALe]:RLEVel

The command specifies the reference level.

The query returns the current value of this setting.

Remote Command	:DISPlay:WINDow[1]:TRACe[1] 2 3:Y[:SCALe]:RLEVel <ampl> :DISPlay:WINDow[1]:TRACe[1] 2 3:Y[:SCALe]:RLEVel?
Example	DISP:TRAC:Y:RLEV 20dBm
Remote Command Notes	FSE, FSP, FSU
Preset	-20 dBm
State Saved	Saved in instrument state.
Min	-130 dBm
Max	30 dBm
Backwards Compatibility SCPI	:DISPlay:TRACe[1] 2 3:Y[:SCALe]:RLEVel

DISPlay[:WINDow<1|2>]:TRACe<1 to 3>:Y[:SCALe]:RLEVel:OFFSet

The command specifies the offset of the reference level.

The query returns the current value of this setting.

Remote Command	:DISPlay:WINDow[1]:TRACe[1] 2 3:Y[:SCALe]:RLEVel:OFFSet <rel_ampl> :DISPlay:WINDow[1]:TRACe[1] 2 3:Y[:SCALe]:RLEVel:OFFSet ?
Example	DISP:TRAC:Y:RLEV:OFFS -10dB
Remote Command Notes	FSE, FSP, FSU
Preset	0 dB
State Saved	Saved in instrument state.
Min	-200.0 dB
Max	200 dB
Backwards Compatibility SCPI	:DISPlay:TRACe[1] 2 3:Y[:SCALe]:RLEVel:OFFSet

DISPlay[:WINDow<1|2>]:TRACe<1 to 3>:Y[:SCALe]:PDIVision

The command specifies the scaling of the Y-axis, using the currently-selected unit.

This mode supports only dB units.

The query returns the current value of this setting.

Remote Command	DISPlay:WINDow[1]:TRACe<1 to 3>:Y[:SCALe]:PDIVision <rel_ampl> DISPlay:WINDow[1]:TRACe<1 to 3>:Y[:SCALe]:PDIVision?
-----------------------	---

Example	DISP:WIND:TRAC:Y:PDIV 5 DISP:WIND:TRAC:Y:PDIV?
Remote Command Notes	FSE, FSP, FSU
State Saved	Saved in instrument state.
Backwards Compatibility SCPI	:DISPlay[1]:TRACe[1] 2 3:Y[:SCALE]:PDIVision

DISPlay[:WINDow<1|2>]:TRACe<1 to 3>:Y:SPACing LINear | LOGarithmic| LDB

The command switches between linear and logarithmic display.

This mode does not support the parameter option LDB.

The query returns the current state of this setting.

Remote Command	:DISPlay:WINDow[1]:TRACe[1] 2 3:Y:SPACing LINear LOGarithmic :DISPlay:WINDow[1]:TRACe[1] 2 3:Y:SPACing?
Example	DISP:WIND:TRAC:Y:SPAC LOG DISP:WIND:TRAC:Y:SPAC?
Remote Command Notes	FSE, FSP, FSU
Preset	LOG
State Saved	Saved in instrument state.
Backwards Compatibility SCPI	:DISPlay[1]:TRACe[1] 2 3:Y:SPACing

DISPlay[:WINDow<1|2>]:TRACe<1...3>:MODE WRITe | VIEW | AVERAge | MAXHold | MINHold

The command specifies the display type and evaluation mode for the specified trace.

The query returns the current state of this setting for the specified trace.

Remote Command	:DISPlay:WINDow[1]:TRACe[1] 2 3:MODE WRITe VIEW AVERAge MAXHold MINHold :DISPlay:WINDow[1]:TRACe[1] 2 3:MODE?
Example	DISP:WIND1:TRAC:MODE WRIT
Remote Command Notes	FSE, FSP, FSU
Preset	WRITe
State Saved	Saved in instrument state.
Range	WRITe VIEW AVERAge MAXHold MINHold
Backwards Compatibility SCPI	:DISPlay:TRACe[1] 2 3:MODE

DISPlay[:WINDow<1|2>]:TRACe<1 to 3>[:STATe] ON | OFF

The command switches on or off display of the corresponding trace.

The query returns the current state of this setting.

Remote Command	:DISPlay [1] :WINDow [1] :TRACe [1] 2 3 [:STATe] ON OFF 1 0 :DISPlay [1] :WINDow [1] :TRACe [1] 2 3 [:STATe] ?
Example	DISP:TRAC1 ON
Remote Command Notes	FSE, FSP, FSU
Preset	ON for Trace 1, OFF for the others
State Saved	Saved in instrument state.
Backwards Compatibility SCPI	:DISPlay[1]:TRACe[1] 2 3[:STATe]

FORMat Subsystem

The FORMat subsystem specifies the format of data transmitted from and received by the instrument.

FORMat[:DATA]

The command specifies the format for data transmitted from the instrument to a control PC.

This command is specific to X-Series analyzers.

The query returns the current state of this setting.

Remote Command	:FORMat [:DATA] ASCii INTeGer, 32 REAL, 32 REAL 64 :FORMat [:DATA] ?
Example	FORM REAL,32 FORM ASC
Remote Command Notes	FSE, FSP, FSU
Preset	ASCii

HCOPy Subsystem

The HCOPy subsystem controls the rendering of display information on output devices or to files.

Most of the HCOPy subsystem SCPI commands are accepted, but take no action and report no error. There are two exceptions: HCOPy:ABORt and HCOPy[:IMMediate].

HCOPy:ABORt

Aborts hardcopy output.

Remote Command	:HCOPy:ABORt
-----------------------	--------------

Example	HCOP:ABOR
Remote Command Notes	FSE, FSP, FSU

HCOPy:CMAP<1 to 26>:DEFault1|2|3

This command is accepted, but takes no action and reports no error.

Remote Command	:HCOPy:CMAP [1] 2 3 4 5 6 7 8 9 10 11 12 13 14 15 16 17 18 19 20 21 22 23 24 25 26 :DEFault [1] 2 3
Example	HCOP:CMAP:DEF
Remote Command Notes	FSP, FSU

HCOPy:CMAP<1 to 26>:HSL <hue>,<sat>,<lum>

This command is accepted, but takes no action and reports no error.

Remote Command	:HCOPy:CMAP [1] 2 3 4 5 6 7 8 9 10 11 12 13 14 15 16 17 18 19 20 21 22 23 24 25 26 :HSL <real> , <real> , <real> :HCOPy:CMAP [1] 2 3 4 5 6 7 8 9 10 11 12 13 14 15 16 17 18 19 20 21 22 23 24 25 26 :HSL?
Example	HCOP:CMAP:HSL 0.3,0.8,1.0 HCOP:CMAP:HSL?
Remote Command Notes	FSP, FSU
State Saved	Saved in instrument state.

HCOPy:CMAP<1 to 26>:PDEFined BLACK | BLUE | BROWn | GREen | CYAN | RED | MAGenta | YELLow | WHITE | DGRAY | LGRAY | LBLUe | LGREen | LCYan | LRED | LMAGenta

This command is accepted, but takes no action and reports no error.

Remote Command	:HCOPy:CMAP [1] 2 3 4 5 6 7 8 9 10 11 12 13 14 15 16 17 18 19 20 21 22 23 24 25 26 :PDEFined BLACK BLUE BROWn GREen CYAN RED MAGenta YELLow WHITE DGRAY LGRAY LBLUe LGREen LCYan LRED LMAGenta :HCOPy:CMAP [1] 2 3 4 5 6 7 8 9 10 11 12 13 14 15 16 17 18 19 20 21 22 23 24 25 26 :PDEFined?
Example	HCOP:CMAP:PDEF BLAC HCOP:CMAP:PDEF?
Remote Command Notes	FSP, FSU
State Saved	Saved in instrument state.

Range	BLACK BLUE BROWN GREEN CYAN RED MAGENTA YELLOW WHITE DGRAY LGRAY LBLUE LGREEN LCYAN LRED LMAGENTA
-------	---

HCOPY:DESTINATION<1|2> <string>

This command is accepted, but takes no action and reports no error.

Remote Command	:HCOPY:DESTINATION[1] 2 <string> :HCOPY:DESTINATION[1] 2?
Example	HCOP:DEST 'MMEM' HCOP:DEST?
Remote Command Notes	FSE, FSP, FSU
State Saved	Saved in instrument state.

HCOPY:DEVICE:COLOR ON|OFF

This command is accepted, but takes no action and reports no error.

Remote Command	:HCOPY:DEVICE:COLOR ON OFF 1 0 :HCOPY:DEVICE:COLOR?
Example	HCOP:DEV:COL OFF HCOP:DEV:COL?
Remote Command Notes	FSE, FSP, FSU
Preset	OFF
State Saved	Saved in instrument state.

HCOPY:DEVICE:LANGUAGE<1|2> GDI | WMF | EWMF | BMP

This command is accepted, but takes no action and reports no error.

Remote Command	:HCOPY:DEVICE:LANGUAGE [1] 2 GDI WMF EWMF BMP :HCOPY:DEVICE:LANGUAGE?
Example	HCOP:DEV:LANG GDI HCOP:DEV:LANG?
Remote Command Notes	FSE, FSP, FSU
State Saved	Saved in instrument state.
Range	GDI WMF EWMF BMP

HCOPY[:IMMEDIATE<1|2>]

Initiates hard copy output. The numeric suffix specifies which printer configuration should be used for

the output. If there is no suffix, configuration 1 is selected.

Remote Command	:HCOPY[:IMMEDIATE]
Example	HCOP:IMM
Remote Command Notes	FSE, FSP, FSU

HCOPY:ITEM:ALL

This command is accepted, but takes no action and reports no error.

Remote Command	:HCOPY:ITEM:ALL
Example	HCOP:ITEM:ALL
Remote Command Notes	FSE, FSP, FSU

HCOPY:ITEM:WINDOW<1|2>:TABLE:STATE ON | OFF

This command is accepted, but takes no action and reports no error.

Remote Command	:HCOPY:ITEM:WINDOW[1] 2:TABLE:STATE ON OFF 1 0 :HCOPY:ITEM:WINDOW[1] 2:TABLE:STATE?
Example	HCOP:ITEM:WIND:TABL:STAT OFF HCOP:ITEM:WIND:TABL:STAT?
Remote Command Notes	FSE, FSP, FSU
Preset	OFF
State Saved	Saved in instrument state.

HCOPY:ITEM:WINDOW<1|2>:TEXT <string>

This command is accepted, but takes no action and reports no error.

Remote Command	:HCOPY:ITEM:WINDOW[1] 2:TEXT <string> :HCOPY:ITEM:WINDOW[1] 2:TEXT?
Example	HCOP:ITEM:WIND:TEXT 'comment' HCOP:ITEM:WIND:TEXT?
Remote Command Notes	FSE, FSP, FSU
State Saved	Saved in instrument state.

HCOPy:ITEM:WINDow<1|2>:TRACe:STATe ON | OFF

This command is accepted, but takes no action and reports no error.

Remote Command	:HCOPy:ITEM:WINDow[1] 2:TRACe:STATe ON OFF 1 0 :HCOPy:ITEM:WINDow[1] 2:TRACe:STATe?
Example	HCOP:ITEM:WIND:TRAC:STAT OFF HCOP:ITEM:WIND:TRAC:STAT?
Remote Command Notes	FSE, FSP, FSU
Preset	OFF
State Saved	Saved in instrument state.

HCOPy:PAGE:ORIEntation<1|2> LANDscape | PORTrait

This command is accepted, but takes no action and reports no error.

Remote Command	:HCOPy:PAGE:ORIEntation[1] 2 LANDscape PORTrait :HCOPy:PAGE:ORIEntation[1] 2?
Example	HCOP:PAGE:ORI LAND HCOP:PAGE:ORI?
Remote Command Notes	FSE, FSP, FSU
State Saved	Saved in instrument state.
Range	LANDscape PORTrait

INITiate Subsystem

The INITiate subsystem controls the init-measurement function.

INITiate<1|2>:CONTInuous ON | OFF

The command switches the trigger system between the continuously-initiated mode (Continuous) or single measurement mode (Single).

The query returns the current state of this setting.

Remote Command	:INITiate[1]:CONTInuous ON OFF 1 0 :INITiate[1]:CONTInuous?
Example	:INIT:CONT 0 :INIT:CONT 1
Remote Command Notes	FSE, FSP, FSU
Preset	ON

State Saved	Saved in instrument state.
-------------	----------------------------

INITiate<1|2>:CONMeas

In Single sweep mode, this command resumes a stopped measurement from the current position.

Remote Command	:INITiate [1] :CONMeas
Example	INIT:CONM
Remote Command Notes	FSE, FSP, FSU

INITiate<1|2>[:IMMediate]

Initiates a new sweep.

Remote Command	:INITiate [1] [:IMMediate]
Example	INIT
Remote Command Notes	FSE, FSP, FSU

INPut Subsystem

The INPut subsystem controls the input characteristics of the instrument's RF inputs.

INPut<1|2>:ATTenuation

The command specifies the input attenuator setting.

With the Electronic Attenuator Option, this mode supports 2 dB steps.

The query returns the current value of this setting.

Remote Command	:INPut [1] :ATTenuation <rel_ampl> :INPut [1] :ATTenuation?
Example	INP:ATT 40dB
Remote Command Notes	FSE, FSP, FSU
Preset	6dB
State Saved	Saved in instrument state.
Min	0
Max	70

INPut<1|2>:ATTenuation:AUTO

The command switches between coupling of the input attenuation to the reference level (state ON), and manual input attenuation (state OFF).

The query returns the current state of this setting.

Remote Command	:INPut [1] :ATTenuation:AUTO ON OFF 1 0 :INPut [1] :ATTenuation:AUTO?
Example	INP:ATT:AUTO ON
Remote Command Notes	FSP, FSU
Preset	ON
State Saved	Saved in instrument state.

INPut<1|2>:EATT

The command specifies the setting of the electronic input attenuator.

This mode supports a maximum value of 24 dB.

The query returns the current value of this setting.

Remote Command	:INPut [1] :EATT <rel_amp1> :INPut [1] :EATT?
Example	INP:EATT 15dB
Remote Command Notes	FSP, FSU
Preset	0
State Saved	Saved in instrument state.
Min	0
Max	24

INPut<1|2>:EATT:AUTO

The command switches between coupling of the electronic input attenuation to the reference level (state ON), and manual input attenuation (state OFF).

The query returns the current state of this setting.

Remote Command	:INPut [1] :EATT:AUTO ON OFF 1 0 :INPut [1] :EATT:AUTO?
Example	INP:EATT:AUTO ON
Remote Command Notes	FSP, FSU
Preset	ON
State Saved	Saved in instrument state.

INPut<1|2>:EATT:STATe

The command switches electronic input attenuation on or off.

The query returns the current state of this setting.

Remote Command	: INPut [1] : EATT: STATe ON OFF 1 0 : INPut [1] : EATT: STATe?
Example	INP:EATT:STAT ON
Remote Command Notes	FSP, FSU
Preset	OFF
State Saved	Saved in instrument state.

INPut<1|2>:IMPedance

The command sets the instrument's nominal input impedance.

The query returns the current state of this setting.

Remote Command	: INPut [1] : IMPedance 50 75 : INPut [1] : IMPedance?
Example	INP:IMP 75 INP:IMP?
Remote Command Notes	FSE, FSP, FSU
State Saved	Saved in instrument state.
Range	50 75

INPut<1|2>:GAIN:STATe

The command switches the instrument's preamplifier on or off.

The query returns the current state of this setting.

Remote Command	: INPut [1] : GAIN: STATe ON OFF 1 0 : INPut [1] : GAIN: STATe?
Example	INP:GAIN:STAT ON
Remote Command Notes	FSP, FSU
Preset	OFF
State Saved	Saved in instrument state.

INSTRUMENT Subsystem

The INSTRUMENT subsystem allows you to select the analyzer's operating mode.

**INSTRUMENT[:SElect] SANalyzer | DDEMod | ADEMod | MGSM | WCDPower|BWCDpower
[MWCDpower | BC2K| BDO**

The command selects the analyzer's mode according to the value of a text parameter.

This mode supports a parameter set that differs from that listed above. For details, see the table of available parameter values in the section INSTRUMENT:NSElect <numeric value>.

The query returns the current state of this setting.

Remote Command	:INSTRUMENT[:SElect] <string> :INSTRUMENT[:SElect]?
Example	:INST:SEL SA
Remote Command Notes	FSE, FSP, FSU
State Saved	Saved in instrument state.

INSTRUMENT:NSElect <numeric value>

The command selects the analyzer's mode according to the value of a numeric parameter.

This mode supports the parameter values listed in the table below.

Mode Identification Numbers and Strings for X-Series Analyzers:

Mode	:INSTRUMENT:NSElect <integer>	:INSTRUMENT[:SElect] <parameter>
Spectrum Analyzer	1	SA
GSM	3	GSM
cdmaOne	4	CDMA
NADC	5	NADC
PDC	6	PDC
I/Q Analyzer (Basic)	8	BASIC
WCDMA with HSDPA/HSUPA	9	WCDMA
cdma2000	10	CDMA2K
GSM/EDGE/EDGE Evo	13	EDGE GSM
Phase Noise	14	PNOISE
1xEV-DO	15	CDMA1XEV
Combined WLAN	19	CWLAN

Mode	:INSTrument:NSElect <integer>	:INSTrument[:SElect] <parameter>
802.16 OFDMA (WiMAX/WiBro)	75	WIMAXOFDMA
Combined Fixed WiMAX	81	CWIMAXOFDM
Vector Signal Analyzer (VXA)	100	VSA
89601 VSA	101	VSA89601
LTE	102	LTE
iDEN/WiDEN/MotoTalk	103	IDEN
802.16 OFDM (Fixed WiMAX)	104	WIMAXFIXED
LTE TDD	105	LTETDD
TD-SCDMA with HSPA/8PSK	211	TDSCDMA
Noise Figure	219	NFIGURE
Bluetooth	228	BLUETOOTH
Measuring Receiver	233	MRECEIVE
Analog Demod	234	ADEMODO
DVB-T/H	235	DVB
DTMB	236	DTMB
ISDB-T	239	ISDBT
CMMB	240	CMMB
Remote Language Compatibility	266	RLC
SCPI Language Compatibility	270	SCPILC

Remote Command	:INSTrument:NSElect <integer> :INSTrument:NSElect?
Example	:INST:NSEL 1
Remote Command Notes	FSE, FSP, FSU
State Saved	Saved in instrument state.

MMEMory Subsystem

The MMEMory (mass memory) subsystem allows access to the analyzer's storage media, and permits storage of various settings.

MMEMory:CATalog? <path>

Reads the content of the specified directory.

Remote Command	:MMEMory:CATalog <string> :MMEMory:CATalog?
Example	:MMEM:CAT 'C:' :MMEM:CAT?
Remote Command Notes	FSE, FSP, FSU
State Saved	Saved in instrument state.

MMEMory:CDIRectory <directory_name>

The command changes the current directory to that specified by <directory_name>.

The query returns the current directory setting.

Remote Command	:MMEMory:CDIRectory <string> :MMEMory:CDIRectory?
Example	:MMEM:CDIR 'C:\Temp' :MMEM:CDIR?
Remote Command Notes	FSE, FSP, FSU
State Saved	Saved in instrument state.

MMEMory:COpy <file_source>,<file_destination>

If <file_source> specifies a single file, copies the file to <file_destination>.

If <file_source> lists more than one file, copies the files to the destination directory specified by <file_destination>.

Remote Command	:MMEMory:COpy <string>,<string>
Example	:MMEM:COpy 'C:\test.txt' 'D:'
Remote Command Notes	FSE, FSP, FSU

MMEMory:DATA <file_name>[,<block_data>]

The command writes the data block <block_data> from the control computer into the file specified by <file_name> (in the analyzer).

The query transfers the file specified by <file_name> from the analyzer to the control computer.

Remote Command	:MMEMory:DATA <string>[,<block_data>] :MMEMory:DATA? <string>
-----------------------	--

Example	:MMEM:DATA 'TEST.CFG',#217This is the file :MMEM:DATA? 'TEST.CFG'
Remote Command Notes	FSE, FSP, FSU
State Saved	Saved in instrument state.

MMEMory:DELeTe <file_name>

Deletes the file specified by <file_name>.

Remote Command	:MMEMory:DELeTe <string >
Example	:MMEM:DEL 'C:\test.txt'
Remote Command Notes	FSE, FSP, FSU

MMEMory:INITialize <msus>

This command is accepted, but takes no action and reports no error.

Remote Command	:MMEMory:INITialize <string >
Example	:MMEM:INIT 'A:'
Remote Command Notes	FSE, FSP, FSU

MMEMory:LOAD:STATe 1,<file_name>

Loads device settings from the file specified by <file_name>. The file contents are set as the new analyzer state.

Remote Command	:MMEMory:LOAD:STATe 1,<string>
Example	:MMEM:LOAD:STAT 1,'myState.state'
Remote Command Notes	FSE, FSP, FSU

MMEMory:LOAD:AUTO 1,<file_name>

Specifies a settings file, the contents of which will automatically be loaded (and set as the analyzer state) when the analyzer is switched on.

Remote Command	:MMEMory:LOAD:AUTO 1,<string>
Example	:MMEM:LOAD:AUTO 1,'myState.state'
Remote Command Notes	FSE, FSP, FSU

MMEMory:MDIRectory <directory_name>

Creates the new directory specified by <directory_name>.

Remote Command	:MMEMory:MDIRectory <string >
Example	:MMEM:MDIR 'C:\T1'
Remote Command Notes	FSE, FSP, FSU

MMEMory:MOVE <file_source>,<file_destination>

If <file_destination> contains no path information, then <file_source> is renamed to <file_destination>.

If <file_destination> contains path information, <file_source> is moved to the specified path, and renamed if <file_destination> also includes a file name.

Remote Command	:MMEMory:MOVE <string>,<string>
Example	:MMEM:MOVE 'C:\test.txt' 'D:'
Remote Command Notes	FSE, FSP, FSU

MMEMory:MSIS <device>

This command is accepted, but takes no action and reports no error.

Remote Command	:MMEMory:MSIS <string > :MMEMory:MSIS?
Example	:MMEM:MSIS 'A:'
Remote Command Notes	FSE, FSP, FSU

MMEMory:NAME <file_name>

The command specifies the path and name of the destination file, for use when a print-to-file operation is initiated using the command HCOPY:IMMEDIATE.

The specified destination file type must be PNG.

The query returns the current value of this setting.

Remote Command	:MMEMory:NAME <string > :MMEMory:NAME?
Example	:MMEM:NAME '1.PNG' :MMEM:NAME?
Remote Command Notes	FSE, FSP, FSU

MMEMory:RDIrectory <directory_name>

Deletes the directory specified by <directory_name>.

Remote Command	:MMEMory:RDIrectory <string>
Example	:MMEM:RDIR 'C:\Test'
Remote Command Notes	FSE, FSP, FSU

MMEMory:STORe<1|2>:STATe 1,<file_name>

Stores the current analyzer state settings in a set of files. The files use the specified file name, but have differing extensions..

Remote Command	:MMEMory:STORe:STATe 1,<string>
Example	:MMEM:STORe:STAT 1,'myState.state'
Remote Command Notes	FSE, FSP, FSU

MMEMory:STORe<1|2>:TRACe <1...6>,<file_name>

Stores the specified trace (1...6) in an ASCII-format file specified by <file_name>.

Remote Command	:MMEMory:STORe:TRACe TRACE1 TRACE2 TRACE3 TRACE4 TRACE5 TRACE6 ALL,<string>
Example	:MMEM:STORe:TRACe TRACE1,'Trace1.trace'
Remote Command Notes	FSE, FSP, FSU

MMEMory:CLear:STATe 1,<file_name>

Clears the analyzer setting specified by <file_name>.

This command is accepted, but takes no action and reports no error.

Remote Command	:MMEMory:CLear:STATe 1,<string>
Example	:MMEM:CLE:STAT 1,'Trace1.trace'
Remote Command Notes	FSE, FSP, FSU

MMEMory:CLear:ALL

Clears all analyzer settings in the current directory.

This command is accepted, but takes no action and reports no error.

Remote Command	:MMEMory:CLear:ALL
Example	:MMEM:CLE:ALL
Remote Command Notes	FSE, FSP, FSU

MMEMory:SElect[:ITEM]:HWSettings ON | OFF

This command is accepted, but takes no action and reports no error.

Remote Command	:MMEMory:SElect [:ITEM] :HWSettings ON OFF 1 0 :MMEMory:SElect [:ITEM] :HWSettings?
Example	MMEM:SEL:HWS OFF MMEM:SEL:HWS?
Remote Command Notes	FSE, FSP, FSU
Preset	On
State Saved	Saved in instrument state.

MMEMory:SElect[:ITEM]:TRACe[:ACTive] ON | OFF

This command is accepted, but takes no action and reports no error.

Remote Command	:MMEMory:SElect [:ITEM] :TRACe[:ACTive] ON OFF 1 0 :MMEMory:SElect [:ITEM] :TRACe[:ACTive]?
Example	MMEM:SEL:TRAC OFF MMEM:SEL:TRAC?
Remote Command Notes	FSE, FSP, FSU
Preset	OFF
State Saved	Saved in instrument state.

MMEMory:SElect[:ITEM]:LINes:ALL ON | OFF

This command is accepted, but takes no action and reports no error.

Remote Command	:MMEMory:SElect [:ITEM] :LINes:ALL ON OFF 1 0 :MMEMory:SElect [:ITEM] :LINes:ALL?
Example	MMEM:SEL:LIN:ALL OFF MMEM:SEL:LIN:ALL?
Remote Command Notes	FSE, FSP, FSU
Preset	ON
State Saved	Saved in instrument state.

MMEMory:SElect[:ITEM]:SCData ON | OFF

This command is accepted, but takes no action and reports no error.

Remote Command	:MMEMory:SElect [:ITEM] :SCData ON OFF 1 0 :MMEMory:SElect [:ITEM] :SCData?
Example	MMEM:SEL:SCD OFF MMEM:SEL:SCD?
Remote Command Notes	FSE, FSP, FSU
Preset	OFF
State Saved	Saved in instrument state.

MMEMory:SElect[:ITEM]:TRANsducer:ALL ON | OFF

This command is accepted, but takes no action and reports no error.

Remote Command	:MMEMory:SElect [:ITEM] :TRANsducer:ALL ON OFF 1 0 :MMEMory:SElect [:ITEM] :TRANsducer:ALL?
Example	MMEM:SEL:TRAN:ALL OFF MMEM:SEL:TRAN:ALL?
Remote Command Notes	FSE, FSP, FSU
Preset	ON
State Saved	Saved in instrument state.

MMEMory:SElect[:ITEM]:ALL

This command is accepted, but takes no action and reports no error.

Remote Command	:MMEMory:SElect [:ITEM] :ALL
Example	MMEM:SEL:ALL
Remote Command Notes	FSE, FSP, FSU

MMEMory:SElect[:ITEM]:NONE

This command is accepted, but takes no action and reports no error.

Remote Command	:MMEMory:SElect [:ITEM] :NONE
Example	MMEM:SEL:NONE
Remote Command Notes	FSE, FSP, FSU

MMEMory:SELEct[:ITEM]:DEFault

This command is accepted, but takes no action and reports no error.

Remote Command	:MMEMory:SELEct [:ITEM] :DEFault
Example	MMEM:SEL: DEF
Remote Command Notes	FSE, FSP, FSU

MMEMory:COMMEnt <string>

Specifies a comment (max. 60 characters), to be associated with a stored analyzer setting.

This command is accepted, but takes no action and reports no error.

Remote Command	:MMEMory:COMMEnt <string> :MMEMory:COMMEnt?
Example	MMEM:COMM 'Unknown Trace' MMEM:COMM?
Remote Command Notes	FSE, FSP, FSU
State Saved	Saved in instrument state.

SENSE:AVERAge Subsystem

The SENSE:AVERAge subsystem calculates the acquired data average. After several successive measurements, a new test result is obtained.

[SENSE<1|2>:]AVERAge:COUNT 0 to 32767

Specifies the number of measurements that contribute to the average value.

In the SCPI LC mode, the maximum value is 10000 and not 32767 as in FSx.

Remote Command	[:SENSE] :AVERAge:COUNT <integer> [:SENSE] :AVERAge:COUNT?
Example	SENS:AVER:COUN 10 SENS:AVER:COUN?
Remote Command Notes	FSE, FSP, FSU
Preset	1
State Saved	Saved in instrument state.
Min	0
Max	10000
Backward Compatibility	:SENSE1:AVERAge:COUNT

[SENSe<1|2>:]AVERAge[:STATe<1 to 3>] ON | OFF

Switches on or off the selected trace's average calculation (STATe<1 to 3>) within the measurement window.

Remote Command	[:SENSe] :AVERAge :STATe [1] 2 3 ON OFF 1 0 [:SENSe] :AVERAge :STATe [1] 2 3 ?
Example	AVER:STAT1 OFF SENS:AVER:STAT3 ON
Remote Command Notes	FSE, FSP, FSU
Preset	OFF
State Saved	Yes.
Backwards Compatibility SCPI	:SENSe1:AVERAge[:STATe]

[SENSe<1|2>:]AVERAge:TYPE VIDEo | LINear

Specifies the average function type.

Remote Command	[:SENSe] :AVERAge :TYPE VIDEo LINear [:SENSe] :AVERAge ?
Example	AVER:TYPE LIN AVER:TYPE ?
Remote Command Notes	FSP, FSU
Preset	VIDeo
Range	VIDeo LINear
Backwards Compatibility SCPI	:SENSe1:AVERAge:TYPE

SENSe:BANDwidth Subsystem

The SENSe:BANDwidth subsystem specifies the instrument's filter bandwidths. Both BANDwidth and BWIDth perform the same functions.

[SENSe<1|2>:]BANDwidth|BWIDth[:RESolution] <numeric_value>

Specifies the instrument's resolution bandwidth.

In the SCPI LC mode, the maximum resolution bandwidth is 8 MHz (not 10 MHz as in FSx) and the minimum resolution bandwidth is 1 Hz (not 10 Hz as in FSx).

Remote Command	[:SENSe] :BANDwidth BWIDth [:RESolution] <freq> [:SENSe] :BANDwidth BWIDth [:RESolution] ?
-----------------------	---

Compatible SCPI Commands
Supported SCPI Commands

Example	BAND 1 KHZ BAND?
Remote Command Notes	The setting and querying of values depends on the current bandwidth type.
Remote Command Notes	
Remote Command Notes	FSE, FSP, FSU
Preset	3 MHz
State Saved	Saved in instrument state.
Min	1 Hz
Max	8 MHz
Backwards Compatibility SCPI	:SENSe1:BANDwidth BWIDth[:RESolution]

[SENSe<1|2>:]BANDwidth|BWIDth[:RESolution]:AUTO ON | OFF

Automatically couples the instrument's resolution bandwidth to the span or it cancels the coupling.

Remote Command	[:SENSe] :BANDwidth BWIDth[:RESolution] :AUTO OFF ON 0 1 [:SENSe] :BANDwidth BWIDth[:RESolution] :AUTO?
Example	BWID:AUTO ON BWID:AUTO?
Remote Command Notes	FSE, FSP, FSU
Preset	ON
Backwards Compatibility SCPI	:SENSe1:BANDwidth BWIDth[:RESolution]:AUTO

[SENSe<1|2>:]BANDwidth|BWIDth[:RESolution]:RATio 0.0001 to 1

Specifies the resolution bandwidth (Hz) / span (Hz) ratio.

Remote Command	[:SENSe] :BANDwidth BWIDth[:RESolution] :RATio <real> [:SENSe] :BANDwidth BWIDth[:RESolution] :RATio?
Example	BAND:RAT 0.1
Remote Command Notes	FSE, FSP, FSU
Preset	0.02 with BAND:TYPE NORMAl or RBW > 30 kHz 0.01 with BAND:TYPE FFT for RBW 30 kHz
State Saved	Saved in instrument state.
Min	0.0001
Max	1

Backwards Compatibility SCPI	:SENSe1:BANDwidth BWIDth[:RESolution]:RATio
------------------------------	---

[SENSe<1|2>:]BANDwidth|BWIDth[:RESolution]:TYPE NORMal | FFT | CFILter | RRC

Sets the resolution bandwidth filter type to "normal" analog, FIR in 1, 3, 10 steps, or FFT for bandwidths less than 100 kHz.

Remote Command	[:SENSe] :BANDwidth BWIDth[:RESolution] :TYPE NORMal FFT CFILter RRC [:SENSe] :BANDwidth BWIDth[:RESolution] :TYPE?
Example	BAND:TYPE RRC BAND:TYPE?
Remote Command Notes	FSP, FSU
Preset	NORMal
State Saved	Saved in instrument state.
Range	NORMal FFT CFILter RRC
Backwards Compatibility SCPI	:SENSe1:BANDwidth BWIDth[:RESolution]:TYPE

[SENSe<1|2>:]BANDwidth|BWIDth:VIDeo 1Hz to 10MHz

Sets the instrument's video bandwidth to between 10 Hz and 10 MHz in 1, 3, 10 steps.

In the SCPI LC mode, the maximum video bandwidth is 50 MHz and not 10 MHz as in FSx.

Remote Command	[:SENSe] :BANDwidth BWIDth:VIDeo <freq> [:SENSe] :BANDwidth BWIDth:VIDeo?
Example	BAND:VID 1KHZ BAND:VID?
Remote Command Notes	FSE, FSP, FSU
Couplings	
Preset	10 MHz
State Saved	Saved in instrument state.
Min	1 Hz
Max	50 MHz
Backwards Compatibility SCPI	:SENSe1:BANDwidth BWIDth[:RESolution]:VIDeo

[SENSe<1|2>:]BANDwidth|BWIDth:VIDeo:AUTO ON | OFF

Automatically couples the instrument's video bandwidth to its resolution bandwidth or cancels the

coupling.

Remote Command	[:SENSe] :BANDwidth BWIDth:VIDeo:AUTO ON OFF 1 0 [:SENSe] :BANDwidth BWIDth:VIDeo:AUTO?
Example	BAND:VID:AUTO OFF
Remote Command Notes	FSE, FSP, FSU
Preset	ON
State Saved	Saved in instrument state.
Backwards Compatibility SCPI	:SENSe1:BANDwidth BWIDth[:RESolution]:VIDeo:AUTO

[SENSe<1|2>:]BANDwidth|BWIDth:VIDeo:RATio 0.01 to 1000

Specifies the video bandwidth (Hz) / resolution bandwidth (Hz) ratio.

Remote Command	[:SENSe] :BANDwidth BWIDth:VIDeo:RATio <real> [:SENSe] :BANDwidth BWIDth:VIDeo:RATio?
Example	BAND:VID:RAT 2 BAND:VID:RAT?
Remote Command Notes	FSE, FSP, FSU
Preset	3
State Saved	Saved in instrument state.
Min	0.00001
Max	3000000
Backwards Compatibility SCPI	:SENSe1:BANDwidth BWIDth[:RESolution]:VIDeo:RATio

SENSe:CORRection Subsystem

The SENSe:CORRection subsystem controls calibration and normalization.

[SENSe<1|2>:]CORRection:EGain:INPut[:MAGNitude] -200...200dB

Informs the analyzer of an external gain.

SCPI LC mode supports the range of -81.90 to 81.90 dB, not -200 to 200 dB as FSP.

Remote Command	[:SENSe] :CORRection:EGain:INPut [:MAGNitude] <rel_amp> [:SENSe] :CORRection:EGain:INPut [:MAGNitude] ?
-----------------------	--

Example	CORR:EGA:INP 10DB
Remote Command Notes	FSP
Preset	0 dB
State Saved	Saved in instrument state.
Min	-81.90 dB
Max	81.90 dB
Backwards Compatibility SCPI	:SENSe1:CORRection:EGain:INPut[:MAGNitude]

SENSe:DETECTOR Subsystem

The SENSe:DETECTOR subsystem controls measurement data acquisition via selection of the detector for the corresponding trace.

[SENSe<1|2>:]DETECTOR<1..3>[:FUNCTION] APEak | NEGative | POSitive | SAMPlE | RMS | AVERAge | QPEak

Switches on the data acquisition detector in the selected trace.

Remote Command	[:SENSe] :DETECTOR [1] 2 3 [:FUNCTion] APEak NEGative POSitive SAMPlE RMS AVERAge QPEak [:SENSe] :DETECTOR [1] 2 3 [:FUNCTion] ?
Example	DET POS
Remote Command Notes	
Remote Command Notes	FSE, FSP, FSU
Preset	APEak
State Saved	Saved in instrument state.
Range	APEak NEGative POSitive SAMPlE RMS AVERAge QPEak
Backwards Compatibility SCPI	:SENSe1:DETECTOR[1] 2 3[:FUNCTION]

[SENSe<1|2>:]DETECTOR<1 to 3>[:FUNCTION]:AUTO ON | OFF

Either couples the detector to the current trace setting or turns coupling off.

Remote Command	[:SENSe] :DETECTOR [1] 2 3 [:FUNCTion] :AUTO ON OFF 1 0 [:SENSe] :DETECTOR [1] 2 3 [:FUNCTion] :AUTO?
Example	DET:AUTO OFF
Remote Command Notes	FSE, FSP, FSU
Preset	ON

State Saved	Saved in instrument state.
Backwards Compatibility SCPI	:SENSe1:DETEctor[1] 2 3[:FUNCTion]:AUTO

SENSe:FREQUENCY Subsystem

The SENSe:FREQUENCY subsystem specifies the active display's frequency axis. The frequency axis can either be set via the center frequency and span, or via the start and stop frequency.

[SENSe<1|2>:]FREQUENCY:CENTer 0 to fmax

Specifies the analyzer's center frequency or the measuring frequency for span = 0.

Remote Command	[:SENSe] :FREQUENCY:CENTer <freq> [:SENSe] :FREQUENCY:CENTer?
Example	FREQ:CENT 50MHz FREQ:CENT?
Remote Command Notes	FSE, FSP, FSU
Preset	fmax /2 with fmax = maximum frequency
State Saved	Saved in instrument state.
Min	Depends on instrument minimum frequency
Max	Depends on instrument maximum frequency
Backwards Compatibility SCPI	:SENSe1:FREQUENCY:CENTer

[SENSe<1|2>:]FREQUENCY:CENTer:STEP 0 to fmax

Specifies the center frequency's step width.

Remote Command	[:SENSe] :FREQUENCY:CENTer:STEP <freq> [:SENSe] :FREQUENCY:CENTer:STEP?
Example	FREQ:CENT:STEP 120MHz
Remote Command Notes	FSE, FSP, FSU
Preset	1MHz
State Saved	Saved in instrument state.
Min	1 Hz
Max	27.0 GHz
Backwards Compatibility SCPI	:SENSe1:FREQUENCY:CENTer:STEP

[SENSe<1|2>:]FREQUENCY:CENTer:STEP:LINK SPAN | RBW | OFF

Couples the center frequency's step width to the span (span >0) or to the resolution bandwidth (span =

0), or cancels the coupling.

Remote Command	[:SENSe] :FREQuency:CENTer:STEP:LINK SPAN RBW OFF [:SENSe] :FREQuency:CENTer:STEP:LINK?
Example	FREQ:CENT:STEP:LINK RBW FREQ:CENT:STEP:LINK?
Remote Command Notes	FSE, FSP, FSU
Preset	SPAN
State Saved	Saved in instrument state.
Backwards Compatibility SCPI	:SENSe1:FREQuency:CENTer:STEP:LINK

[SENSe<1|2>]:FREQuency:CENTer:STEP:LINK:FACTor 1 to 100 PCT

Couples the center frequency's step width to a factor of the span (span >0) or to the resolution bandwidth (span = 0).

Remote Command	[:SENSe] :FREQuency:CENTer:STEP:LINK:FACTor <int> [:SENSe] :FREQuency:CENTer:STEP:LINK:FACTor?
Example	FREQ:CENT:STEP:LINK:FACTor 20PCT
Remote Command Notes	FSE, FSP, FSU
State Saved	Saved in instrument state.
Min	1
Max	100
Backwards Compatibility SCPI	:SENSe1:FREQuency:CENTer:STEP:LINK:FACTor

[SENSe<1|2>]:FREQuency:SPAN 0 to fmax

Specifies the frequency span.

Remote Command	[:SENSe] :FREQuency:SPAN <freq> [:SENSe] :FREQuency:SPAN?
Example	FREQ:SPAN 10MHz
Remote Command Notes	FSE, FSP, FSU
Preset	3 GHz
State Saved	Saved in instrument state.
Max	26.5 GHz
Backwards Compatibility SCPI	:SENSe1:FREQuency:SPAN

[SENSe<1|2>:]FREQUENCY:SPAN:FULL

Sets the frequency span to maximum.

Remote Command	[:SENSe] :FREQUENCY :SPAN :FULL
Example	FREQ:SPAN:FULL
Remote Command Notes	FSE, FSP, FSU
Backwards Compatibility SCPI	:SENSe1:FREQUENCY:SPAN:FULL

[SENSe<1|2>:]FREQUENCY:START 0 to fmax

Specifies the analyzer's start frequency. This command is available only in the frequency domain (span > 0).

Remote Command	[:SENSe] :FREQUENCY :START <freq> [:SENSe] :FREQUENCY :START?
Example	FREQ:STAR 200MHz FREQ:STAR?
Remote Command Notes	FSE, FSP, FSU
Preset	0 Hz
State Saved	Saved in instrument state.
Min	0Hz
Max	Depends on the instrument maximum frequency

[SENSe<1|2>:]FREQUENCY:STOP 0 to fmax

Specifies the analyzer's stop frequency. This command is available only in the frequency domain (span > 0).

Remote Command	[:SENSe] :FREQUENCY :STOP <freq> [:SENSe] :FREQUENCY :STOP?
Example	FREQ:STOP 220MHz FREQ:STOP?
Remote Command Notes	FSE, FSP, FSU
Preset	3GHZ
State Saved	Saved in instrument state.
Min	0Hz
Max	27.0 GHz

[SENSe<1|2>]:FREQuency:MODE CW | FIXEd | SWEEp

This command switches between time domain (CW | FIXEd) and frequency domain (SWEEp).

Remote Command	[:SENSe] :FREQuency:MODE CW FIXEd SWEEp [:SENSe] :FREQuency:MODE?
Example	FREQ:MODE CW
Remote Command Notes	FSE, FSP, FSU
Preset	SWEEp
Range	CW FIXEd SWEEp
Backwards Compatibility SCPI	:SENSe1:FREQuency:MODE

[SENSe<1|2>]:FREQuency:OFFSet

Specifies the instrument's frequency offset.

Remote Command	[:SENSe] :FREQuency:OFFSet <freq> [:SENSe] :FREQuency:OFFSet?
Example	FREQ:OFFS 10MHz
Remote Command Notes	FSE, FSP, FSU
Preset	0 Hz
State Saved	Saved in instrument state.
Min	-500 GHz
Max	500 GHz
Backwards Compatibility SCPI	:SENSe1:FREQuency:OFFSet

SENSe:POWer Subsystem

The SENSe:POWer subsystem controls the instrument's channel and adjacent channel power measurements.

[SENSe<1|2>]:POWer:ACHannel:SPACing:CHANnel 100 Hz to 2000 MHz

Specifies the channel spacing for the carrier signals. This command is available only for measurements in the frequency domain (span > 0).

Remote Command	[:SENSe] :POWer:ACHannel:SPACing:CHANnel 1000 [:SENSe] :POWer:ACHannel:SPACing:CHANnel
Example	POW:ACH:SPAC:CHAN 2000 POW:ACH:SPAC:CHAN?

Remote Command Notes	FSE, FSP, FSU
Preset	20 kHz
State Saved	Saved in instrument state.
Min	100 Hz
Max	2 GHz

[SENSe<1|2>:]POWer:ACHannel:SPACing:ACHannel 100 Hz to 2000 MHz

Specifies the channel spacing of the adjacent channel to the TX channel. At the same time, the spacing of alternate adjacent channels 1 and 2 is set to double or triple the entered value.

This command is available only in the frequency domain (span > 0).

Remote Command	[:SENSe] :POWer:ACHannel:SPACing:ACHannel <freq> [:SENSe] :POWer:ACHannel:SPACing:ACHannel?
Example	POW:ACH:SPAC:ACH 33kHz
Remote Command Notes	FSE, FSP, FSU
Preset	20KHz
State Saved	Saved in instrument state.
Min	100Hz
Max	2000MHz

[SENSe<1|2>:]POWer:ACHannel:SPACing:ALternate<1|2> 100 Hz to 2000 MHz

Specifies the spacing between the first (ALternate1) or second alternate adjacent channel (ALternate2) and the TX channel.

Remote Command	[:SENSe] :POWer:ACHannel:SPACing:ALternate [1] 2 <freq> [:SENSe] :POWer:ACHannel:SPACing:ALternate [1] 2?
Example	POW:ACH:SPAC:ALT1 100kHz
Remote Command Notes	FSE, FSP, FSU
Preset	40000 60000
State Saved	Saved in instrument state.
Min	100Hz
Max	2000MHz

[SENSe<1|2>:]POWer:ACHannel:TXChannel:COUNT 1 | 2 | 3 | 4

Specifies the number of carrier signals. This command is only available for multicarrier channel and

adjacent-channel power measurements.

Remote Command	[:SENSe] :POWER:ACHannel:TXChannel:COUNT <int> [:SENSe] :POWER:ACHannel:TXChannel:COUNT?
Example	POW:ACH:TXCH:COUN 1 POW:ACH:TXCH:COUN?
Remote Command Notes	FSP, FSU
Preset	4
State Saved	Saved in instrument state.
Min	1
Max	4

[SENSe<1|2>:]POWER:ACHannel:ACPairs 0 | 1 | 2 | 3

Specifies the number of adjacent channels (upper and lower channel in pairs). The number 0 represents a pure channel power measurement.

Remote Command	[:SENSe] :POWER:ACHannel:ACPairs [:SENSe] :POWER:ACHannel:ACPairs?
Example	POW:ACH:ACP 3
Remote Command Notes	FSE, FSP, FSU
Preset	0
Range	0 1 2 3

[SENSe<1|2>:]POWER:ACHannel:BANDwidth|BWIDth[:CHANnel] 100 Hz to 1000 MHz

Specifies the radio communication system's channel bandwidth.

Remote Command	[:SENSe] :POWER:ACHannel:BANDwidth BWIDth[:CHANnel] <freq> [:SENSe] :POWER:ACHannel:BANDwidth BWIDth[:CHANnel]?
Example	POW:ACH:BWID 30kHz
Remote Command Notes	FSE, FSP, FSU
Preset	14KHz
State Saved	Saved in instrument state.
Min	100Hz
Max	1000MHz

[SENSe<1|2>:]POWer:ACHannel:BANDwidth|BWIDth:ACHannel 100Hz–1000MHz

Specifies the radio transmission system’s adjacent channel bandwidth. If the adjacent channel bandwidth is changed, all alternate adjacent channel’s bandwidths are automatically set to the same value.

Remote Command	[:SENSe] :POWer:ACHannel :BANDwidth BWIDth:ACHannel <freq> [:SENSe] :POWer:ACHannel :BANDwidth BWIDth:ACHannel ?
Example	POW:ACH:BWID:ACH 30kHz
Remote Command Notes	FSE, FSP, FSU
Preset	14KHz
State Saved	Saved in instrument state.
Min	100Hz
Max	1000MHz

[SENSe<1|2>:]POWer:ACHannel:BANDwidth|BWIDth:ALternate<1|2> 100 Hz to 1000 MHz

Specifies the radio transmission system’s channel bandwidth for the first and second alternate adjacent channel. If alternate adjacent channel 1’s bandwidth is changed, then alternate adjacent channel 2’s bandwidth is automatically set to the same value.

Remote Command	[:SENSe] :POWer:ACHannel :BANDwidth BWIDth:ALternate [1] 2 <freq> [:SENSe] :POWer:ACHannel :BANDwidth BWIDth:ALternate [1] 2 ?
Example	POW:ACH:BWID:ALT2 30kHz
Remote Command Notes	FSP, FSU
Preset	14KHz
State Saved	Saved in instrument state.
Min	100Hz
Max	1000MHz

[SENSe<1|2>:]POWer:ACHannel:MODE ABSolute | RELative

Switches between absolute and relative adjacent channel measurements.

Remote Command	[:SENSe] :POWer:ACHannel :MODE ABSolute RELative [:SENSe] :POWer:ACHannel :MODE?
Example	POW:ACH:MODE REL
Remote Command Notes	FSE, FSP, FSU

Preset	ABSolute
State Saved	Saved in instrument state.
Range	ABSolute RELative

[SENSe<1|2>:]POWer:ACHannel:REFerence:AUTO ONCE

Sets the relative measurement's reference value to the currently measured channel power.

This command is accepted, but takes no action and reports no error.

Remote Command	[:SENSe] :POWer:ACHannel:REFerence:AUTO ONCE
Example	POW:ACH:REF:AUTO ONCE
Remote Command Notes	FSE, FSP, FSU
State Saved	Saved in instrument state.

[SENSe<1|2>:]POWer:ACHannel:REFerence:TXCHannel:AUTO MINimum | MAXimum | LHIGHest

Activates the automatic selection of the transmission channel that is used as a reference channel in relative adjacent-channel power measurements.

Remote Command	[:SENSe] :POWer:ACHannel:REFerence:TXCHannel:AUTO MINimum MAXimum LHIGhest [:SENSe] :POWer:ACHannel:REFerence:TXCHannel:AUTO?
Example	SENS:POW:ACH:REF:TXCH:AUTO MIN
Remote Command Notes	FSP, FSU
State Saved	Saved in instrument state.
Range	MINimum MAXimum LHIGHest

[SENSe<1|2>:]POWer:ACHannel:REFerence:TXCHannel:MANual 1 | 2 | 3 | 4

Selects the transmission channel to use as a reference channel in relative adjacent-channel power measurements.

The command is only available for multicarrier and adjacent-channel power measurements.

Remote Command	[:SENSe] :POWer:ACHannel:REFerence:TXCHannel:MANual 1 2 [:SENSe] :POWer:ACHannel:REFerence:TXCHannel:MANual?
Example	POW:ACH:REF:TXCH:MAN 2 POW:ACH:REF:TXCH:MAN?
Remote Command Notes	FSP, FSU

Compatible SCPI Commands
Supported SCPI Commands

Preset	1
State Saved	Saved in instrument state.
Min	1
Max	2

[SENSe<1|2>:]POWer:ACHannel:PRESet ACPower | CPOWer | MCACpower | OBANdwidth | OBWidth | CN | CN0

Adjusts the frequency span, the measurement bandwidths including the required detector for the number of channels, the channel bandwidths, and the channel spacings selected in the active power measurement. If required, switches on the adjacent-channel power measurement prior to the adjustment.

Remote Command	[:SENSe] :POWer:ACHannel:PRESet ACPower CPOWer MCACpower OBANdwidth OBWidth CN CN0
Example	SENS:POW:ACH:PRES ACP
Remote Command Notes	FSE, FSP, FSU
State Saved	No
Range	ACPower CPOWer MCACpower OBANdwidth OBWidth CN CN0

[SENSe<1|2>:]POWer:ACHannel:PRESet:RLEVel

Adjusts the reference level to the measured channel power. If required, switches on the adjacent channel power measurement.

This command is accepted, but takes no action and reports no error.

Remote Command	[:SENSe] :POWer:ACHannel:PRESet:RLEVel
Example	SENS:POW:ACH:PRES:RLEV
Remote Command Notes	FSP, FSU

[SENSe<1|2>:]POWer:BANDwidth|BWIDth 10 to 99.9PCT

Specifies the power percentage with respect to the total power.

Remote Command	[:SENSe] :POWer:BANDwidth BWIDth <real> [:SENSe] :POWer:BANDwidth BWIDth?
Example	POW:BAND 99.0 POW:BAND?
Remote Command Notes	FSP, FSU
Preset	99PCT
State Saved	Saved in instrument state.

Min	10
Max	99.99

[SENSe<1|2>:]POWer:HSPeEd ON | OFF

Switches on or off the high-speed channel or adjacent channel power measurement.

Remote Command	[:SENSe] :POWer:HSPeEd ON OFF 1 0 [:SENSe] :POWer:HSPeEd?
Example	POW:HSP ON POW:HSP?
Remote Command Notes	FSP, FSU
Preset	OFF
State Saved	Saved in instrument state.

[SENSe<1|2>:]POWer:NCORrection ON | OFF

Switches on or off the instrument inherent noise correction for channel power measurement.

Remote Command	[:SENSe] :POWer:NCORrection ON OFF 1 0 [:SENSe] :POWer:NCORrection?
Example	POW:NCOR ON POW:NCOR?
Remote Command Notes	FSP, FSU
Preset	OFF
State Saved	Saved in instrument state.

[SENSe<1|2>:]POWer:TRACe 1 to 3

Assigns the channel or adjacent channel power measurement to the indicated trace in the selected measurement window.

This command is accepted, but takes no action and reports no error.

Remote Command	[:SENSe] :POWer:TRACe <int> [:SENSe] :POWer:TRACe?
Example	SENS:POW:TRAC 1 SENS:POW:TRAC?
Remote Command Notes	FSP, FSU
State Saved	Saved in instrument state.

Min	1
Max	3

SENSe:ROSCillator Subsystem

The SENSe:ROSCillator subsystem controls the reference oscillator. The numeric suffix in SENSe is irrelevant for the commands in this subsystem.

[SENSe<1|2>:]ROSCillator:SOURce INTernal | EXTernal

Switches between an external or internal reference oscillator for frequency processing of external generators 1 and 2.

Remote Command	[:SENSe] :ROSCillator :SOURce INTernal EXTernal [:SENSe] :ROSCillator :SOURce?
Example	ROSC:SOUR EXT
Remote Command Notes	FSE, FSP, FSU
Preset	INT
State Saved	Yes.
Range	INTernal EXTernal

SENSe:SWEEP Subsystem

The SENSe:SWEEP subsystem controls the sweep parameters.

[SENSe<1|2>:]SWEEP:TIME 2.5ms to 16000s (frequency domain) | 1us to 16000s (time domain)

Specifies sweep time.

Remote Command	[:SENSe] :SWEEP :TIME <time> [:SENSe] :SWEEP :TIME?
Example	SWE:TIME 500ms
Remote Command Notes	FSE, FSP, FSU
Preset	1ms
State Saved	Saved in instrument state.
Min	1 ms
Max	4000 s

[SENSe<1|2>:]SWEep:TIME:AUTO ON | OFF

Automatically couples sweep time to the frequency span and bandwidth settings.

Remote Command	[:SENSe] :SWEep:TIME:AUTO OFF ON 0 1 [:SENSe] :SWEep:TIME:AUTO?
Example	SWE:TIME:AUTO OFF
Remote Command Notes	FSE, FSP, FSU
Preset	ON
State Saved	Saved in instrument state.

[SENSe<1|2>:]SWEep:COUNT 0 to 32767

Specifies the number of single sweeps that are used to calculate the average or maximum value.

SCPI mode supports the maximum value of 10000, not 32767 as in FSx.

Remote Command	[:SENSe] :SWEep:COUNT <int> [:SENSe] :SWEep:COUNT?
Example	SENS:SWE:COUN 1000 SENS:SWE:COUN?
Remote Command Notes	FSE, FSP, FSU
Preset	0
State Saved	Saved in instrument state.
Min	0
Max	10000

[SENSe<1|2>:]SWEep:EGATe ON | OFF

Switches on or off sweep control by an external gate signal.

Remote Command	[:SENSe] :SWEep:EGATe ON OFF 1 0 [:SENSe] :SWEep:EGATe?
Example	SWE:EGAT ON SWE:EGAT?
Remote Command Notes	FSE, FSP, FSU
Preset	OFF
State Saved	Saved in instrument state.

[SENSe<1|2>:]SWEep:EGATe:TYPE LEVel | EDGE

Specifies the external gate signal's triggering type (level or edge).

Remote Command	[:SENSe] :SWEep:EGATe:TYPE LEVel EDGE [:SENSe] :SWEep:EGATe:TYPE?
Example	SWE:EGAT:TYPE LEVel SWE:EGAT:TYPE?
Remote Command Notes	FSE, FSP, FSU
Preset	EDGE
State Saved	Saved in instrument state.
Range	LEVel EDGE

[SENSe<1|2>:]SWEep:EGATe:POLarity POSitive | NEGative

Specifies the external gate signal's polarity.

Remote Command	[:SENSe] :SWEep:EGATe:POLarity POSitive NEGative [:SENSe] :SWEep:EGATe:POLarityh°O
Example	SWE:EGAT:POL NEG SWE:EGAT:POL?
Remote Command Notes	FSE, FSP, FSU
Preset	POSitive
State Saved	Saved in instrument state.
Range	POSitive NEGative

[SENSe<1|2>:]SWEep:EGATe:HOLDoff 125 ns to 100 s

For edge triggering, specifies the instrument's holdoff time.

Remote Command	[:SENSe] :SWEep:EGATe:HOLDoff <time> [:SENSe] :SWEep:EGATe:HOLDoff ?
Example	SWE:EGAT:HOLD 0.0002 SWE:EGAT:HOLD?
Remote Command Notes	FSE, FSP, FSU
Couplings	0
State Saved	Saved in instrument state.
Min	125ns

Max	100s
-----	------

[SENSe<1|2>:]SWEep:EGATe:LENGth 0 to 100 s

For edge triggering, specifies the time interval during which the instrument sweeps.

Remote Command	[:SENSe] :SWEep:EGATe:LENGth <time> [:SENSe] :SWEep:EGATe:LENGth?70
Example	SWE:EGAT:LENG 1 SWE:EGAT:LENG?
Remote Command Notes	FSE, FSP, FSU
Preset	1.25e-7
State Saved	Saved in instrument state.
Min	0
Max	100s

[SENSe<1|2>:]SWEep:EGATe:SOURce EXTernal | IFPower | RFPower

Switches between an external gate signal and an IF power signal as the signal source for the gate mode.

Remote Command	[:SENSe] :SWEep:EGATe:SOURce EXTernal IFPower RFPower [:SENSe] :SWEep:EGATe:SOURce?
Example	SENS:SWE:EGAT:SOUR EXT SENS:SWE:EGAT:SOUR?
Remote Command Notes	FSE, FSP, FSU
Preset	EXTernal
Range	EXTernal IFPower RFPower

[SENSe<1|2>:]SWEep:POINts 125 to 8001

Specifies the number of measurement points for one sweep run.

Remote Command	[:SENSe] :SWEep:POINts <integer> [:SENSe] :SWEep:POINts?
Example	SWE:POIN 251 SWE:POIN?
Remote Command Notes	FSP, FSU
Preset	FSP: 501 FSU: 625

State Saved	Saved in instrument state.
Min	1
Max	8001

STATus Subsystem

The STATus subsystem provides the status reporting system commands. The status registers are not influenced by *RST.

STATus:OPERation[:EVENT]?

Queries the contents of the EVENT section in the STATus:OPERation register. After the readout, the contents of the EVENT section are deleted.

Remote Command	:STATus:OPERation[:EVENT]?
Example	STAT:OPER?
Remote Command Notes	FSE, FSP, FSU
Preset	0

STATus:OPERation:CONDition?

Queries the CONDition section in the STATus:OPERation register. The current hardware status is reflected in the value returned. The contents of the CONDition section are not deleted after the readout.

Remote Command	:STATus:OPERation:CONDition
Example	STAT:OPER:COND?
Remote Command Notes	FSE, FSP, FSU
Preset	0

STATus:OPERation:ENABLE

Specifies the bits of the ENABLE section in the STATus:OPERation register.

This mode supports the maximum value of 32767, not 65535 as in FSx.

Remote Command	:STATus:OPERation:ENABLE <integer> :STATus:OPERation:ENABLE?
Example	STAT:OPER:ENAB 1
Remote Command Notes	FSE, FSP, FSU
Preset	0
Min	0
Max	32767

STATus:OPERation:PTRansition

Specifies the edge detectors of all bits in the STATus:OPERation register from 0 to 1 for the transitions of the CONDition bit.

This mode supports the maximum value of 32767, not 65535 as FSx.

Remote Command	:STATus:OPERation:PTRansition <integer> :STATus:OPERation:PTRansition?
Example	STAT:OPER:PTR 1
Remote Command Notes	FSE, FSP, FSU
Preset	32767
Min	0
Max	32767
SCPI Status Bits/OPC Dependencies	Sequential command

STATus:OPERation:NTRansition

Specifies the edge detectors of all bits in the STATus:OPERation register from 1 to 0 for the transitions of the CONDition bit.

This mode supports the maximum value of 32767, not 65535 as in FSx.

Mode	All
Remote Command	:STATus:OPERation:NTRansition <integer> :STATus:OPERation:NTRansition?
Example	STAT:OPER:NTR 1
Remote Command Notes	FSE, FSP, FSU
Preset	0
Min	0
Max	32767
SCPI Status Bits/OPC Dependencies	Sequential command

STATus:PRESet

Resets the edge detectors and ENABLE parts of all registers to a defined value.

Remote Command	:STATus:PRESet
Example	STAT:PRES

Remote Command Notes	FSE, FSP, FSU
----------------------	---------------

STATus:QUEStionable[:EVENT]

Queries the contents of the EVENT section in the STATus:QUEStionable register. After the readout, the contents of the EVENT section are deleted.

Remote Command	:STATus:QUEStionable[:EVENT]?
Example	STAT:QUES?
Remote Command Notes	FSE, FSP, FSU
Preset	0

STATus:QUEStionable:CONDition

Queries the CONDition section in the STATus:QUEStionable register. The contents of the CONDition section are not deleted after a readout.

Remote Command	:STATus:QUEStionable:CONDition?
Example	STAT:QUES:COND?
Remote Command Notes	FSE, FSP, FSU
Preset	0

STATus:QUEStionable:ENABle

Specifies the bits of the ENABle section in the STATus-QUEStionable register. The

ENABle register selectively enables individual events in the associated EVENT section for the summary bit in the status byte.

This mode supports the maximum value of 32767, not 65535 as FSx.

Remote Command	:STATus:QUEStionable:ENABle <integer> :STATus:QUEStionable:ENABle?
Example	:STATus:QUEStionable:ENABle 16
Remote Command Notes	FSE, FSP, FSU
Preset	0
Min	0
Max	32767

STATus:QUEStionable:PTRansition

This mode supports the maximum value of 32767, not 65535 as in FSx.

Remote Command	:STATus:QUEStionable:PTRansition <integer>. :STATus:QUEStionable:PTRansition?
Example	STAT:QUES:PNTR 16
Remote Command Notes	FSE, FSP, FSU
Preset	0
Min	0
Max	32767

STATus:QUEStionable:NTRansition

Specifies the edge detectors of all bits in the STATus:OPERation register from 1 to 0 for the transitions of the CONDition bit.

This mode supports the maximum value of 32767, not 65535 as FSx.

Remote Command	:STATus:QUEStionable:NTRansition <integer>. :STATus:QUEStionable:NTRansition?
Example	STAT:QUES:NTR 16
Remote Command Notes	FSE, FSP, FSU
Preset	0
Min	0
Max	32767

STATus:QUEStionable:FREQuency[:EVENT]

Queries the contents of the EVENT section in the STATus:QUEStionable: FREQuency register.

Remote Command	:STATus:QUEStionable:FREQuency[:EVENT] ?
Example	STAT:QUES:FREQ?
Remote Command Notes	FSE, FSP, FSU
Preset	0

STATus:QUEStionable:FREQuency:CONDition

Queries the contents of the CONDition section in the

STATus:QUEStionable:FREQuency register. The contents of the CONDition section are not deleted

after a readout.

Remote Command	:STATus:QUESTionable:FREQuency:CONDition?
Example	STAT:QUES:FREQ:COND?
Remote Command Notes	FSE, FSP, FSU
Preset	0

STATus:QUESTionable:FREQuency:ENABle

Specifies the bits of the ENABle section in the STATus:QUESTionable:FREQuency register.

This mode supports the maximum value of 32767, not 65535 as FSx.

Remote Command	:STATus:QUESTionable:FREQuency:ENABle <integer> :STATus:QUESTionable:FREQuency:ENABle?
Example	STAT:QUES:FREQ:ENAB 2
Remote Command Notes	FSE, FSP, FSU
Preset	32767
Min	0
Max	32767

STATus:QUESTionable:FREQuency:PTRansition

Specifies the edge detectors of all bits in the STATus:QUESTionable:FREQuency register from 0 to 1 for the transitions of the CONDition bit.

This mode supports the maximum value of 32767, not 65535 as in FSx.

Remote Command	:STATus:QUESTionable:FREQuency:PTRansition <integer> :STATus:QUESTionable:FREQuency:PTRansition?
Example	STAT:QUES:FREQ:PTR 2
Remote Command Notes	FSE, FSP, FSU
Preset	32767
Min	0
Max	32767

STATus:QUESTionable:FREQuency:NTRansition

Specifies the edge detectors of all bits in the STATus:QUESTionable:FREQuency register from 1 to 0 for the transitions of the CONDition bit.

This mode supports the maximum value of 32767, not 65535 as in FSx.

Remote Command	:STATus:QUESTionable:FREQuency:NTRansition <integer> :STATus:QUESTionable:FREQuency:NTRansition?
Example	STAT:QUES:FREQ:NTR 2
Remote Command Notes	FSE, FSP, FSU
Preset	0
Min	0
Max	32767

STATus:QUEue[:NEXT]?

Returns the earliest entry to the error queue then deletes it.

Remote Command	:STATus:QUEue[:NEXT]?
Example	STAT:QUE?
Remote Command Notes	FSE, FSP, FSU

SYSTEM Subsystem

This subsystem includes commands for general functions.

SYSTEM:COMMunicate:GPIB[:SELF]:ADDRESS

Changes the analyzer's IEC/IEEE-bus address.

Changing this address may require further communication to use the new address.

Remote Command	:SYSTem:COMMunicate:GPIB[:SELF]:ADDRESS <integer> :SYSTem:COMMunicate:GPIB[:SELF]:ADDRESS?
Example	:SYST:COMM:GPIB:ADDR 17
Remote Command Notes	Note: Changing the Address on the GPIB port requires all further communication to use the new address.
Remote Command Notes	FSE, FSP, FSU
State Saved	No
Range	0 to 30

SYSTem:DATE

Sets or queries the date portion of the analyzer's internal clock.

Remote Command	:SYSTem:DATE <year>, <month>, <day> :SYSTem:DATE?
Example	:SYST:DATE 2006,05,26
Remote Command Notes	FSE, FSP, FSU

SYSTem:DISPlay:FPANel ON | OFF

Turns the on-screen display of front panel keys on or off.

Mode	All
Remote Command	SYSTem:DISPlay:FPANel ON OFF SYSTem:DISPlay:FPANel?
Example	:SYST:DISP:FPAN ON :SYST:DISP:FPAN?
Remote Command Notes	FSP, FSU
Preset	ON
State Saved	No

SYSTem:DISPlay:UPDate

This command is accepted, but takes no action and reports no error.

Remote Command	:SYSTem:DISPlay:UPDate ON OFF 1 0 :SYSTem:DISPlay:UPDate?
Example	SYST:DISP:UPD ON
Remote Command Notes	FSE, FSP, FSU
Preset	ON

SYSTem:ERRor?

Queries the first (oldest) entry in the error queue, then deletes that entry.

Remote Command	:SYSTem:ERRor [:NEXT] ?
Example	:SYST:ERR?
Remote Command Notes	FSE, FSP, FSU

SYSTem:ERRor:LIST?

This command is accepted, but takes no action and reports no error.

Remote Command	:SYSTem:ERRor:LIST?
Example	:SYST:ERR:LIST?
Remote Command Notes	FSP, FSU

SYSTem:ERRor:CLEar:ALL

This command is accepted, but takes no action and reports no error.

Remote Command	:SYSTem:ERRor:CLEar:ALL
Example	:MMEM:CLE:ALL
Remote Command Notes	FSP, FSU

SYSTem:PRESet

Initiates a reset of the analyzer.

Remote Command	:SYSTem:PRESet
Example	:SYST:PRES
Remote Command Notes	FSE, FSP, FSU

SYSTem:TIME

Sets or queries the time portion of the analyzer's internal clock. The sequence of entry is hour, minute, second.

Remote Command	:SYSTem:TIME <hour>, <minute>, <second> :SYSTem:TIME?
Example	:SYST:TIME 13,05,26
Remote Command Notes	FSE, FSP, FSU

SYSTem:VERSion

Queries the analyzer's SCPI version number.

This command is a query and therefore has no *RST value.

Remote Command	:SYSTem:VERSion?
Example	:SYST:VERS?
Remote Command Notes	FSE, FSP, FSU

TRACe Subsystem

TRACe<1|2>[:DATA] TRACE1 | TRACE2 | TRACE3 | SPURious | ABITstream | PWCDp | CTABLE, <block> | <numeric_value>

The command writes trace data from the control computer to the specified trace of the analyzer.

The TRACe[:DATA] command is of the form:

```
:TRACe:DATA <trace>, <data>
```

where <trace> may be one of the following parameters:

TRACE1, TRACE2, TRACE3, TRACE4, TRACE5, TRACE6

and where <data> may be either:

- ASCII data, consisting of a set of comma-separated values, or,
- REAL or INTeger, sent as a definite length block, with a header describing the data to follow.

The query reads the specified trace data from the analyzer to the control computer.

Remote Command	:TRACe [1] [:DATA] TRACE1 TRACE2 TRACE3 TRACE4 TRACE5 TRACE6, <data> :TRACe [1] [:DATA] ? TRACE1 TRACE2 TRACE3 TRACE4 TRACE5 TRACE6
Example	TRAC? TRACE1
Remote Command Notes	FSE, FSP, FSU

TRACe:IQ Subsystem

This subsystem includes commands for handling measured I/Q data.

TRACe<1|2>:IQ:DATA

This query causes a measurement to be performed, then returns a list of measurement results. The results are corrected for frequency response before being returned.

Before sending this query, specify the measurement settings using TRACe<1|2>:IQ:SET <filter type>,<rbw>,<sample rate>,<trigger source>,<trigger slope>,<pretrigger samples>,<# of samples>.

Remote Command	:TRACe [1] :IQ:DATA?
Example	TRAC:IQ:DATA?
Remote Command Notes	FSP, FSU

TRACe<1|2>:IQ:SET <filter type>,<rbw>,<sample rate>,<trigger source>,<trigger slope>,<pretrigger samples>,<# of samples>

Specifies the analyzer settings for measurement of I/Q data.

Remote Command	TRACe[1]:IQ:SET <filter type>,<rbw>,<sample rate>,<trigger source>,<trigger slope>,<pretrigger samples>,<# of samples>
Example	TRAC:IQ:SET NORM,10MHz,32MHz,EXT,POS,0,2048 TRAC:IQ:SET NORM,1MHz,4MHz,EXT,POS,1024,512
Remote Command Notes	FSP, FSU

TRACe:IQ:SRATe 15.625kHz to 32MHz

Specifies the sampling rate for I/Q data acquisition.

Remote Command	TRACe[1]:IQ:SRATe <freq> TRACe[1]:IQ:SRATe?
Example	TRAC:IQ:SRAT 4MHZ TRAC:IQ:SRAT?
Remote Command Notes	FSP, FSU
Preset	32MHz
State Saved	Saved in instrument state.
Min	15.625kHz
Max	32MHz

TRACe:IQ:STATe ON | OFF

Switches the current measurement between SWEEP mode and I/Q mode.

Remote Command	:TRACe:IQ:STATe ON OFF 1 0 :TRACe:IQ:STATe?
Example	TRAC:IQ:STAT ON
Remote Command Notes	FSP, FSU
Preset	OFF
State Saved	Saved in instrument state.

TRIGger Subsystem

This subsystem includes commands that control and synchronize the start of a sweep.

An external trigger signal can be supplied via the connector on the analyzer's rear panel.

TRIGger<1|2>[:SEQuence]:SOURce

Selects the trigger source.

This mode supports only parameter options: IMMEDIATE, EXTERNAL, VIDEO, RFPower.

It does **not** support parameter options: IFPower, TV, AF, FM, AM and PM.

Remote Command	:TRIGger [1] [:SEQuence] :SOURce IMMediate EXTErnal VIDeo IFPower RFPower TV AF FM AM PM :TRIGger [1] [:SEQuence] :SOURce?
Example	TRIG:SOUR EXT
Remote Command Notes	
Remote Command Notes	FSE, FSP, FSU
Preset	IMMediate
Range	IMMediate EXTErnal VIDeo IFPower RFPower TV AF FM AM PM

TRIGger<1|2>[:SEQuence]:LEVel:IFPower -30 to -10DBM

This command sets the level of the IF power trigger source.

This command is accepted, but takes no action and reports no error.

Remote Command	:TRIGger [1] [:SEQuence] :LEVel:IFPower <ampl> :TRIGger [1] [:SEQuence] :LEVel:IFPower?
Example	TRIG:LEV:IFP -20DBM
Remote Command Notes	FSP, FSU
Preset	-20dBm
State Saved	Saved in instrument state.
Min	30dBm
Max	-70dBm

TRIGger<1|2>[:SEQuence]:LEVel:RFPower

Sets the RF power trigger level.

This mode supports the level range:-200 to 100 dBm.

Remote Command	:TRIGger [1] [:SEQuence] :LEVel:RFPower <real> :TRIGger [1] [:SEQuence] :LEVel:RFPower?
Example	TRIG:LEV:RFP -20DBM
Remote Command Notes	FSP

Preset	-20dBm
State Saved	Saved in instrument state.
Min	-200dBm
Max	100dBm

TRIGger<1|2>[:SEQuence]:LEVel:VIDeo 0 to 100PCT

Sets the video trigger level.

Remote Command	TRIGger [1] [:SEQuence] :LEVel:VIDeo <real> TRIGger [1] [:SEQuence] :LEVel:VIDeo?
Example	TRIG:LEV:VID 40 TRIG:LEV:VID?
Remote Command Notes	FSE, FSP, FSU
Preset	50
State Saved	Saved in instrument state.
Min	0
Max	100

TRIGger<1|2>[:SEQuence]:HOLDoff

Specifies the trigger delay period.

Remote Command	:TRIGger [1] [:SEQuence] :HOLDoff <time> :TRIGger [1] [:SEQuence] :HOLDoff?
Example	TRIG:HOLD 100ms
Remote Command Notes	FSE, FSP, FSU
Preset	0s
State Saved	Saved in instrument state.
Min	0s
Max	0.5s

TRIGger<1|2>[:SEQuence]:SLOPe

Sets the trigger signal slope to positive or negative.

Compatible SCPI Commands
Supported SCPI Commands

The trigger slope setting applies to all trigger sources.

Remote Command	:TRIGger[1] [:SEquence]:SLOPe POSitive NEGative :TRIGger[1] [:SEquence]:SLOPe?
Example	TRIG:SLOP NEG
Remote Command Notes	FSE, FSP, FSU
Preset	POSitive
State Saved	Saved in instrument state.
Range	POSitive NEGative

Overview

This section is not intended to teach you everything about the SCPI (Standard Commands for Programmable Instruments) programming language. The SCPI Consortium or IEEE can provide that level of detailed information.

Programming with SCPI requires knowledge of:

- Computer programming languages, such as C, C++, and Microsoft® Visual Basic®
- The language of your instrument, such as MXA or FSP

Topics covered in this chapter include:

- “SCPI Language Basics” on page 150
- “Command Syntax” on page 151
- “Creating Valid Commands” on page 152
- “Special Characters in Commands” on page 153
- “Parameters in Commands” on page 153
- “Putting Multiple Commands on the Same Line” on page 156

For more information refer to:

IEEE Standard 488.1-2004, *IEEE Standard Digital Interface for Programmable Instrumentation*. New York, NY, 1998.

IEEE Standard 488.2-2004, *IEEE Standard Codes, Formats, Protocols and Comment Commands for Use with ANSI/IEEE Std 488.1-1987*. New York, NY, 1998.

SCPI Language Basics

SCPI is an ASCII-based instrument command language designed for test and measurement instruments, with the goal of reducing automatic test equipment (ATE) program development time.

SCPI accomplishes this goal by providing a consistent programming environment control and data usage. This consistent programming environment is achieved by the use of defined program messages, instrument responses, and data formats across all SCPI instruments.

By providing a consistent programming environment, replacing one SCPI instrument with another SCPI instrument in a system will usually requires less effort than with non-SCPI instrument.

SCPI is not a standard which completely provides for interchangeable instrumentation. SCPI helps move toward interchangeability by defining instrument commands and responses, but no functionality,

accuracy, resolution, and so on.

Table 4-1 Common Terms used in this Book

Term	Description
Command	An instruction. You combine commands to form messages that control instruments to complete a specified task. In general, a command consists of mnemonics (also known as keywords), parameters and punctuation.
Query	A special type of command. Queries instruct the instrument to make response data available to the controller. Query keywords always end with a question mark, ?.

Command Syntax

A command consists of keywords, parameters and punctuation. A typical command is made up of keywords set off by colons. The keywords are followed by parameters that can be followed by optional units.

Here is an example:

```
SENSe:FREQuency:STARt 1.5 MHZ
```

Table 4-2

Command Keywords	<p>Many commands have both a long and a short form: use either one. A combination of the two is not allowed. Consider the :FREQuency command for example:</p> <ul style="list-style-type: none"> • Short form is :FREQ • Long form is :FREQUENCY <p>SCPI is not case sensitive, so fREquEncy is just as valid as FREQUENCY, but FREQ and FREQUENCY are the only valid forms of the FREQuency command and FREQU is not a valid form.</p> <p>For example, Sens:Freq:Star 1.5 mhz is the same as SENSE:FREQUENCY:START 1.5 MHZ.</p> <p>In this documentation, upper case letters indicate the short form of the keyword. The lower case letters indicate the long form of the keyword.</p>
Punctuation	<ul style="list-style-type: none"> • A vertical bar " " dictates a choice of one element from a list. For example: <A> indicates that either A or B can be selected, but not both. • Square brackets "[]" indicates that the enclosed items are optional. • Angle brackets "< >" indicates a variable items to be entered to represent user choices. • Braces "{ }" can optionally be used in the command either not at all, once, or several times. • A question mark "?" after a subsystem command indicates that the command is a query. The returned information, <value> varies in format according to the type of the field.
Separator	<ul style="list-style-type: none"> • A colon ":" separates keywords of different levels. The colon before the root keyword is usually omitted. • A space separates a keyword and a parameter, as well as a parameter and a unit.

NOTE The command `SENS:FREQU:STAR` is not valid because `FREQU` is neither the short, nor the long form of the command.

Command Statement Rules Overview

Besides the standard notation of SCPI described above, please remember the following rules in programming:

- command statements read from left to right
- use either long form or short form of keywords, but do not use both
- no separating space between the keywords, only use a colon to separate keywords of different levels
- always separating a keyword from a variable with a space
- always separating a variable from its unit with a space (if variable has a unit)

Creating Valid Commands

Commands are not case sensitive and there are often many different ways of writing a particular command. These are examples of valid commands for a given command syntax:

Command Syntax	Sample Valid Commands
<code>[SENSe:]BANDwidth[:RESolution] <freq></code>	<p>The following sample commands are all identical. They all cause the same result.</p> <ul style="list-style-type: none">• <code>Sense:Band:Res 1700</code>• <code>BANDWIDTH:RESOLUTION 1.7e3</code>• <code>sens:band 1.7KHZ</code>• <code>SENS:band 1.7E3Hz</code>• <code>band 1.7kHz</code>• <code>bandwidth:RES 1.7e3Hz</code>
<code>MEASure:SPECTrum[n]?</code>	<ul style="list-style-type: none">• <code>MEAS:SPEC?</code>• <code>Meas:spec?</code>• <code>meas:spec3?</code> <p>The number 3 in the last meas example causes it to return different results than the commands above it. See the command description for more information.</p>
<code>[:SENSe]:DETEctor[:FUNCTION] NEGative POSitive SAMPLE</code>	<ul style="list-style-type: none">• <code>DET:FUNC neg</code>• <code>Detector:Func Pos</code>
<code>INITiate:CONTInuous ON OFF 1 0</code>	<p>The sample commands below are identical.</p> <ul style="list-style-type: none">• <code>INIT:CONT ON</code>• <code>init:continuous 1</code>

Special Characters in Commands

Special Character	Meaning	Example
	A vertical stroke between parameters indicates alternative choices. The effect of the command is different depending on which parameter is selected.	Command: TRIGger:SOURCE EXTernal INTernal LINE The choices are external, internal, and line. Ex: TRIG:SOURCE INT is one possible command choice.
	A vertical stroke between keywords indicates identical effects exist for both keywords. The command functions the same for either keyword. Only one of these keywords is used at a time.	Command: SENSe:BANDwidth BWIDTH: OFFSet Two identical commands are: Ex1: SENSE:BWIDTh:OFFSEt Ex2: SENSE:BAND:OFFSEt
[]	keywords in square brackets are optional when composing the command. These implied keywords are executed even if they are omitted.	Command: [SENSe:]BANDwidth[:RESolution]:AUTO The following commands are all valid and have identical effects: Ex1: bandwidth:auto Ex2: band:resolution:auto Ex3: sense:bandwidth:auto
<>	Angle brackets around a word, or words, indicates they are not to be used literally in the command. They represent the needed item.	Command: SENS:FREQ <freq> In this command example the word <freq> should be replaced by an actual frequency. Ex: SENS:FREQ 9.7MHz.
{ }	Parameters in braces can optionally be used in the command either not at all, once, or several times.	Command: MEASure:BW <freq>{,level} A valid command is: meas:BW 6 MHz, 3 dB, 60 dB

Parameters in Commands

There are four basic types of parameters: booleans, keywords, variables and arbitrary block program data.

OFF|ON|0|1

(Boolean)

This is a two state boolean-type parameter. The numeric value 0 is equivalent to OFF. Any numeric value other than 0 is equivalent to ON. The numeric values of 0 or 1 are commonly used in the command instead of OFF or ON. Queries of the parameter always return a numeric value of 0 or 1.

keyword

The keywords that are allowed for a particular command are defined in the command syntax description.

Overview

Units Numeric variables may include units. The valid units for a command depend on the variable type being used. See the following variable descriptions. The indicated default units are used if no units are sent. Units can follow the numerical value with, or without, a space.

Variable A variable can be entered in exponential format as well as standard numeric format. The appropriate range of the variable and its optional units are defined in the command description.

The following keywords may also be used in commands, but not all commands allow keyword variables.

- DEFault - resets the parameter to its default value.
- UP - increments the parameter.
- DOWN - decrements the parameter.
- MINimum - sets the parameter to the smallest possible value.
- MAXimum - sets the parameter to the largest possible value.

The numeric value for the function's MINimum, MAXimum, or DEFault can be queried by adding the keyword to the command in its query form. The keyword must be entered following the question mark.

Example query: `SENSE:FREQ:CENTER? MAX`

Variable Parameters

<ampl>	Is a rational number followed by optional units. The default units are dBm. Acceptable units include: dBm, dBmV, dB μ V.
<angle> <degrees>	Is a rational number followed by optional units. The default units are degrees. Acceptable units include: DEG, RAD.
<bit_pattern>	Specifies a series of bits rather than a numeric value. The bit series is the binary representation of a numeric value. There are no units. Bit patterns are most often specified as hexadecimal numbers, though octal, binary or decimal numbers may also be used. In the SCPI language these numbers are specified as: <ul style="list-style-type: none">• Hexadecimal, #Hdddd or #hdddd where 'd' represents a hexadecimal digit 0 to 9 and 'a' to 'f'. So #h14 can be used instead of the decimal number 20.• Octal, #Odddddd or #oddddd where 'd' represents an octal digit 0 to 7. So #o24 can be used instead of the decimal number 20.• Binary, #Bddddddddd or #bddddddddd where 'd' represents a 1 or 0. So #b10100 can be used instead of the decimal number 20.
<current>	Is a rational number followed by optional units. The default units are Amperes. Acceptable units include: A, mA, μ A, nA.
<freq> <bandwidth>	Is a positive rational number followed by optional units. The default unit is Hertz. Acceptable units include: Hz, kHz, MHz, GHz.
<integer>	is an integer value with no units.
<percent>	Is a rational number between 0 and 100. You can either use no units or use PCT.
<power>	Is a rational number followed by optional units. The default units are W. Acceptable units include: mAW, kW, W, mW, μ W, nW, pW.
<real>	Is a floating point number with no units.
<rel_power> <rel_ampl>	Is a positive rational number followed by optional units. The default units are dB. Acceptable units include: dB.
<string>	Is a series of alpha numeric characters.
<time> <seconds>	Is a rational number followed by optional units. The default units are seconds. Acceptable units include: ks, s, ms, us, ns.
<voltage>	Is a rational number followed by optional units. The default units are Volts. Acceptable units include: V, mV, μ V, nV

Block Program Data

Some parameters consist of a block of data. There are a few standard types of block data. Arbitrary blocks of program data can also be used.

Overview

<trace> Is an array of rational numbers corresponding to displayed trace data. See FORMat:DATA for information about available data formats.

A SCPI command often refers to a block of current trace data with a variable name such as: Trace1, TRACE2, or trace3, depending on which trace is being accessed.

<arbitrary block data> Consists of a block of data bytes. The first information sent in the block is an ASCII header beginning with #. The block is terminated with a semi-colon. The header can be used to determine how many bytes are in the data block. There are no units. (You do not get block data if your data type is ASCII, using FORMat :DATA ASCII command. Your data is comma separated ASCII values.

Block data example: suppose the header is #512320.

- The first digit in the header (5) tells you how many additional digits/bytes there are in the header.
- The 12320 means 12 thousand, 3 hundred, 20 data bytes follow the header.
- Divide this number of bytes by your current data format (bytes/data point), either 8 (for real,64), or 4 (for real,32). For this example, if you're using real64 then there are 1540 points in the block.

Putting Multiple Commands on the Same Line

Multiple commands can be written on the same line, reducing your code space requirement. To do this:

- Commands must be separated with a semicolon (;).
- If the commands are in different subsystems, the key word for the new subsystem must be preceded by a colon (:).
- If the commands are in the same subsystem, the full hierarchy of the command key words need not be included. The second command can start at the same key word level as the command that was just executed.

SCPI Termination and Separator Syntax

All binary trace and response data is terminated with <NL><END>, as defined in Section 8.5 of IEEE Standard 488.2-1992, *IEEE Standard Codes, Formats, Protocols and Common Commands for Use with ANSI/IEEE Std 488.1-1987*. New York, NY, 1992. (Although one intent of SCPI is to be interface independent, <END> is only defined for IEEE 488 operation.)

The following are some examples of good and bad commands. The examples are created from a theoretical instrument with the simple set of commands indicated below:

```
[ :SENSE]
    :POWER
        [:RF]
        :ATTenuation 40dB

:TRIGger
    [:SEQuence]
    :EXTErnal [1]
        :SLOPe
            POSitive

[:SENSE]
    :FREQuency
        :STARt
    :POWER
    [:RF]
        :MIXer
            :RANGe
            [:UPPer]
```

Bad Command	Good Command
PWR:ATT 40dB	POW:ATT 40dB
The short form of POWER is POW, not PWR.	
FREQ:STAR 30MHz;MIX:RANG -20dBm	FREQ:STAR 30MHz;POW:MIX:RANG -20dBm
The MIX:RANG command is in the same :SENSE subsystem as FREQ, but executing the FREQ command puts you back at the SENSE level. You must specify POW to get to the MIX:RANG command.	
FREQ:STAR 30MHz;POW:MIX RANG -20dBm	FREQ:STAR 30MHz;POW:MIX:RANG -20dBm
MIX and RANG require a colon to separate them.	
:POW:ATT 40dB;TRIG:FREQ:STAR 2.3GHz	:POW:ATT 40dB;:FREQ:STAR 2.3GHz
:FREQ:STAR is in the :SENSE subsystem, not the :TRIGGER subsystem.	
:POW:ATT?:FREQ:STAR?	:POW:ATT?;:FREQ:STAR?
:POW and FREQ are within the same :SENSE subsystem, but they are two separate commands, so they should be separated with a semicolon, not a colon.	
:POW:ATT -5dB;:FREQ:STAR 10MHz	:POW:ATT 5dB;:FREQ:STAR 10MHz
Attenuation cannot be a negative value.	

