



Agilent X-Series Signal Analyzer

This manual provides documentation
for the following analyzers:

PXA Signal Analyzer N9030A

MXA Signal Analyzer N9020A

EXA Signal Analyzer N9010A

CXA Signal Analyzer N9000A

N9081A & W9081A Bluetooth Measurement Application Measurement Guide

Notices

© Agilent Technologies, Inc. 2010, 2011

No part of this manual may be reproduced in any form or by any means (including electronic storage and retrieval or translation into a foreign language) without prior agreement and written consent from Agilent Technologies, Inc. as governed by United States and international copyright laws.

Trademark Acknowledgements

Microsoft® is a U.S. registered trademark of Microsoft Corporation.

Windows® and MS Windows® are U.S. registered trademarks of Microsoft Corporation.

Adobe Acrobat® and Reader® are U.S. registered trademarks of Adobe Systems Incorporated.

Java™ is a U.S. trademark of Sun Microsystems, Inc.

MATLAB® is a U.S. registered trademark of Math Works, Inc.

Norton Ghost™ is a U.S. trademark of Symantec Corporation.

Wikipedia® is a registered trademark of the Wikimedia Foundation.

Manual Part Number

N9081-90002

Print Date

August 2011

Supersedes: November 2010

Printed in USA

Agilent Technologies Inc.
1400 Fountaingrove Parkway
Santa Rosa, CA 95403

Warranty

The material contained in this document is provided "as is," and is subject to being changed, without notice, in future editions. Further, to the maximum extent permitted by applicable law, Agilent disclaims all warranties, either express or implied, with regard to this manual and any information contained herein, including but not limited to the implied warranties of merchantability and fitness for a particular purpose. Agilent shall not be liable for errors or for incidental or consequential damages in connection with the furnishing, use, or performance of this document or of any information contained herein. Should Agilent and the user have a separate written agreement with warranty terms covering the material in this document that conflict with these terms, the warranty terms in the separate agreement shall control.

Technology Licenses

The hardware and/or software described in this document are furnished under a license and may be used or copied only in accordance with the terms of such license.

Restricted Rights Legend

If software is for use in the performance of a U.S. Government prime contract or subcontract, Software is delivered and licensed

as "Commercial computer software" as defined in DFAR 252.227-7014 (June 1995), or as a "commercial item" as defined in FAR 2.101(a) or as "Restricted computer software" as defined in FAR 52.227-19 (June 1987) or any equivalent agency regulation or contract clause. Use, duplication or disclosure of Software is subject to Agilent Technologies' standard commercial license terms, and non-DOD Departments and Agencies of the U.S. Government will receive no greater than Restricted Rights as defined in FAR 52.227-19(c)(1-2) (June 1987). U.S. Government users will receive no greater than Limited Rights as defined in FAR 52.227-14 (June 1987) or DFAR 252.227-7015 (b)(2) (November 1995), as applicable in any technical data.

Safety Notices

CAUTION

A **CAUTION** notice denotes a hazard. It calls attention to an operating procedure, practice, or the like that, if not correctly performed or adhered to, could result in damage to the product or loss of important data. Do not proceed beyond a CAUTION notice until the indicated conditions are fully understood and met.

WARNING

A **WARNING** notice denotes a hazard. It calls attention to an operating procedure, practice, or the like that, if not correctly performed or adhered to, could result in personal injury or death. Do not proceed beyond a WARNING notice until the indicated conditions are fully understood and met.

Warranty

This Agilent technologies instrument product is warranted against defects in material and workmanship for a period of one year from the date of shipment. During the warranty period, Agilent Technologies will, at its option, either repair or replace products that prove to be defective.

For warranty service or repair, this product must be returned to a service facility designated by Agilent Technologies. Buyer shall prepay shipping charges to Agilent Technologies, and Agilent Technologies shall pay shipping charges to return the product to Buyer. For products returned to Agilent Technologies from another country, Buyer shall pay all shipping charges, duties, and taxes.

Where to Find the Latest Information

Documentation is updated periodically. For the latest information about these products, including instrument software upgrades, application information, and product information, browse to one of the following URLs, according to the name of your product:

<http://www.agilent.com/find/pxa>

<http://www.agilent.com/find/mxa>

<http://www.agilent.com/find/exa>

<http://www.agilent.com/find/cxa>

To receive the latest updates by email, subscribe to Agilent Email Updates at the following URL:

<http://www.agilent.com/find/emailupdates>

Information on preventing analyzer damage can be found at:

<http://www.agilent.com/find/tips>

Is your product software up-to-date?

Periodically, Agilent releases software updates to fix known defects and incorporate product enhancements. To search for software updates for your product, go to the Agilent Technical Support website at:

<http://www.agilent.com/find/techsupport>

Table of Contents

1 About Bluetooth Measurement Application

What Does the Agilent N9081A & W9081A Bluetooth Measurement Application Do?	10
Licenses for N9081A & W9081A Bluetooth Measurement Application	11

2 Making Bluetooth Measurements

Introduction	14
Making Transmit Analysis Measurements	16
Transmit Analysis Measurement Overview	16
Step 1 - Set Up the Test Equipment and DUT	16
Making the Initial Signal Connection	16
Step 2 - Select the Mode and Preset the Analyzer	17
Step 3 - Select the Mode Setup Parameters	18
Step 4 - Select Measurement	18
Step 5 - Select the Meas Setup Parameters	19
Step 6 - Configure the Display	19
View - Quad View	20
View - RF Envelope	21
View - Demod Waveform	22
View - RF Spectrum	23
Step 7 - Optimize your Results	24
Viewing Transmit Analysis Measurement Results	25
Example of One-Button-Measurement - Transmit Analysis	25
Example of Using SCPI Commands	26
Making Adjacent Channel Power Measurements	27
Adjacent Channel Power Measurement Overview	27
Step 1 - Set Up the Test Equipment and DUT	27
Making the Initial Signal Connection	27
Step 2 - Select the Mode and Preset the Analyzer	28
Step 3 - Select the Mode Setup Parameters	29

Contents

Step 4 - Select Measurement	29
Step 5 - Select the Meas Setup Parameters	30
Step 6 - Configure the Display	30
View - RF Envelope	31
View - RF Spectrum	32
Step 7 - Optimize your Results	32
Viewing Adjacent Channel Power Measurement Results	33
Example Measurement - Adjacent Channel Power	33
Example of Using SCPI Commands	34
Making Output Spectrum BW Measurements	35
Output Spectrum BW Measurement Overview	35
Step 1 - Set Up the Test Equipment and DUT	35
Making the Initial Signal Connection	35
Step 2 - Select the Mode and Preset the Analyzer	36
Step 3- Select the Mode Setup Parameters	37
Step 4- Select Measurement	37
Step 5 - Select the Meas Setup Parameters	38
Step 6 - Optimize your Results	38
Viewing Output Spectrum BW Measurement Results	39
Example of One-Button-Measurement - Output Spectrum BW	39
Example of Using SCPI Commands	40
Making EDR In-band Spurious Emissions Measurements	41
EDR In-band Spurious Emissions Measurement Overview	41
Step 1 - Set Up the Test Equipment and DUT	41
Making the Initial Signal Connection	41
Step 2 - Select the Mode and Preset the Analyzer	42
Step 3- Select the Mode Setup Parameters	43
Step 4- Select Measurement	43
Step 5 - Select the Meas Setup Parameters	44
Step 6 - Configure the Display	44
View - RF Envelope	45
View - RF Spectrum	46

Step 7 - Optimize your Results	46	
Viewing EDR In-band Spurious Emissions Measurement Results		47
Example Measurement - EDR In-band Spurious Emissions		47
Example of Using SCPI Commands	48	
Making LE In-band Emissions Measurements	49	
LE In-band Emissions Measurement Overview	49	
Step 1 - Set Up the Test Equipment and DUT	49	
Making the Initial Signal Connection	49	
Step 2 - Select the Mode and Preset the Analyzer	50	
Step 3 - Select the Mode Setup Parameters	51	
Step 4 - Select Measurement	51	
Step 5 - Select the Meas Setup Parameters	52	
Step 6 - Configure the Display	52	
View - RF Envelope	53	
View - RF Spectrum	54	
Step 7 - Optimize your Results	54	
Viewing LE In-band Emissions Measurement Results		55
Example Measurement - LE In-band Emissions		55
Example of Using SCPI Commands	56	
Making Monitor Spectrum Measurements	57	
Monitor Spectrum Measurement Overview	57	
Step 1 - Set Up the Test Equipment and DUT	57	
Making the Initial Signal Connection	57	
Step 2 - Select the Mode and Preset the Analyzer	58	
Step 3- Select the Mode Setup Parameters	59	
Step 4- Select Measurement	59	
Step 5 - Optimize your Results	59	
Viewing Monitor Spectrum Measurement Results		60
Example of One-Button-Measurement - Monitor Spectrum		60
Example of Using SCPI Commands	61	
Troubleshooting Bluetooth Measurements	62	

3 Concepts

Introduction	64
Understanding Bluetooth	65
What is Bluetooth	65
What is Bluetooth Enhanced Data Rate (EDR)	66
What is Bluetooth Low Energy (LE)	66
Bluetooth Radio Characteristics	66
Baseband Characteristics	68
The Bluetooth Radio Unit	69
The Bluetooth Link Control Unit and Link Management	70
Measurement Concepts and Results	73
Adjacent Channel Power Measurement	73
EDR In-band Spurious Emissions Measurement	73
LE In-band Emissions Measurement	74
Output Spectrum Bandwidth	74
Transmit Analysis Measurement	75
Output Power	76
Carrier Frequency Drift	77
Monitor Band/Channel Measurement	77
Initial Carrier Frequency Tolerance	78
Modulation Characteristics	78

1 About Bluetooth Measurement Application

This chapter provides overall information on the Agilent N9081A & W9081A Bluetooth Measurement Application and describes the measurements made by the analyzer.

What Does the Agilent N9081A & W9081A Bluetooth Measurement Application Do?

The Agilent N9081A & W9081A Bluetooth measurement application can be used to quickly ensure a product development conforms to regulatory requirements, as well as providing RF diagnostic and troubleshooting capability for a Bluetooth device.

The Bluetooth measurement application supports the following standards:

- Basic: Basic Bluetooth
- EDR: Enhanced Data Rate, which offers enhanced data rates by using Phase Modulation (DPSK) and Frequency Modulation (GFSK).
- Low Energy: LE, which is based on Basic Bluetooth with less power.

NOTE

You can select the standard by pressing **Mode Setup, Radio Standard** and select the geography by pressing **FREQ Channel, Geography**.

The Option N9081A & W9081A Bluetooth measurement application provides the following one-button measurements:

- Adjacent Channel Power (ACP) Measurement
- Enhanced Data Rate (EDR) In-band Spurious Emissions Measurement
- Low Energy (LE) In-band Emissions Measurement
- Monitor Spectrum Measurement
- Output Spectrum Bandwidth (OBW) Measurement
- Transmit Analysis Measurement

When you select the standard format for use, the instrument automatically makes measurements using the measurement methods and limits defined in the standards. Detailed measurement results are displayed allowing you to make further analysis.

NOTE

To reduce the number of variables and identify individual performance characteristics, frequency hopping is turned off for the above Bluetooth measurements.

Measurement parameters may be altered for specialized analysis via the **Meas Setup** menu.

The measurements conform to the following Bluetooth RF Test Specifications. Supported tests are briefed in the [Table 1-1 on page 11](#).

- Bluetooth Test Specification Ver. 1.2/2.0/2.0 + EDR/2.1/2.1 + EDR
- Bluetooth Low Energy RF Test Specifications: RF PHY 0.7 Draft

About Bluetooth Measurement Application
What Does the Agilent N9081A & W9081A Bluetooth Measurement Application Do?

Table 1-1 Supported Tests

Test Purposes		Measurement
TRM/CA/01/C	Output Power	Transmit Analysis (Radio Standard is Basic or Low Energy)
TRM/CA/05/C	Tx Output Spectrum - 20 dB Bandwidth	Output Spectrum BW
TRM/CA/06/C	TX Output Spectrum - Adjacent Channel Power	Adjacent Channel Power
TRM/CA/07/C	Modulation Characteristics	Transmit Analysis (Radio Standard is Basic or Low Energy)
TRM/CA/08/C	ICFT	Transmit Analysis (Radio Standard is Basic or Low Energy)
TRM/CA/09/C	Carrier Frequency Drift	Transmit Analysis (Radio Standard is Basic or Low Energy)
TRM/CA/10/C	EDR Relative Transmit Power	Transmit Analysis (Radio Standard is EDR)
TRM/CA/11/C	EDR Carrier Freq Stability and Mod Accuracy	Transmit Analysis (Radio Standard is EDR)
TRM/CA/12/C	EDR Differential Phase Encoding	Transmit Analysis (Radio Standard is EDR)
TRM/CA/13/C	EDR In-band Spurious Emissions	EDR In-band Spur Emissions
TRM-LE/CA/02/C	In-band Emissions	LE In-band Emissions

Licenses for N9081A & W9081A Bluetooth Measurement Application

The Option N9081A supports the following license types:

- N9081A-2FP: Fixed/Perpetual license which enables the Bluetooth application on a single X-Series instrument. It cannot be transported from the instrument.
- N9081A-2TP: Transportable/Perpetual license which enables the Bluetooth application on a single X-Series instrument.
- N9081A-2F3: Fixed/Time-limited license which is 3 month duration and cannot be transported from the instrument.

The Option W9081A supports the following license types:

- W081A-2FP: Fixed/Perpetual license which enables the Bluetooth application on a single X-Series instrument. It cannot be transported from the instrument.
- W9081A-2TP: Transportable/Perpetual license which enables the Bluetooth application on a single X-Series instrument.

About Bluetooth Measurement Application

What Does the Agilent N9081A & W9081A Bluetooth Measurement Application Do?

- W9081A-2F3: Fixed/Time-limited license which is 3 month duration and cannot be transported from the instrument.

2 Making Bluetooth Measurements

This chapter describes procedures to set up and perform measurements for analysis of Bluetooth signals. Details of the steps necessary for accurate signal analysis are provided as well.

Introduction

This chapter provides measurement procedures and also shows example results obtained using the Bluetooth measurement application.

There are six measurements available in this mode:

- **Transmit Analysis**
- **Adjacent Channel Power**
- **Output Spectrum BW**
- **EDR In-band Spurious Emissions**
- **LE In-band Emissions**
- **Monitor Spectrum**

NOTE

Before you can begin making measurements, make sure you have installed the required license.

1. Press **Mode**, and check to make sure that **Bluetooth** is available.
 2. Press **Meas**, and check to make sure that all the measurements are available.
-

The following main subjects are presented in this chapter:

- [“Making Transmit Analysis Measurements” on page 16](#)
This section describes steps to perform a Transmit Analysis measurement.
- [“Making Adjacent Channel Power Measurements” on page 27](#)
This section describes steps to perform an Adjacent Channel Power measurement.
- [“Making Output Spectrum BW Measurements” on page 35](#)
This section describes steps to perform an Output Spectrum BW measurement.
- [“Making EDR In-band Spurious Emissions Measurements” on page 41](#)
This section describes steps to perform an EDR In-band Spurious Emissions measurement.
- [“Making LE In-band Emissions Measurements” on page 49](#)
This section describes steps to perform a LE In-band Emissions measurement.
- [“Making Monitor Spectrum Measurements” on page 57](#)
This section describes steps to perform a Monitor Spectrum measurement.
- [“Troubleshooting Bluetooth Measurements” on page 62](#)
This section introduces error messages you may have when making a Bluetooth measurement.

NOTE

For more information on how to make an Output Spectrum BW measurement, please refer to *N9020A Signal Analyzer Measurement Guide*.

For more information on how to make a Monitor Spectrum measurement, please refer to *N9020A Signal Analyzer Measurement Guide*.

Making Transmit Analysis Measurements

Transmit Analysis Measurement Overview

Making successful measurements of Bluetooth signals is easy when you follow the main steps below. The procedure overview below provides links in blue to a detailed procedure and description for each step:

[“Step 1 - Set Up the Test Equipment and DUT” on page 16](#)

This step configures the analyzer connections for an RF measurement.

[“Step 2 - Select the Mode and Preset the Analyzer” on page 17](#)

This step assures the analyzer will be set to a known state, either the factory preset or a user-configured preset.

[“Step 3 - Select the Mode Setup Parameters” on page 18](#)

This step specifies the radio parameters and carriers configuration.

[“Step 4 - Select Measurement” on page 18](#)

This step allows you to make a transmit analysis measurement either by preset settings or desired settings.

[“Step 6 - Configure the Display” on page 19](#)

This step enables you to select different views of display.

[“Step 7 - Optimize your Results” on page 24](#)

Measurement results are available in graphical displays that can be presented in preset groups or adjusted to suit your specific application. Most results are also available in tabular format.

Step 1 - Set Up the Test Equipment and DUT


This step configures the analyzer connections for measuring Bluetooth signals.

Making the Initial Signal Connection

CAUTION

Before connecting a signal to the instrument, make sure the instrument can safely accept the signal level provided. The signal level limits are marked next to the connectors on the front panel.

Figure 2-1 Bluetooth Measurement Setup



Bluetooth_setup

The Bluetooth device under test (DUT) is connected to the RF input port by an appropriate RF cable as shown. The DUT may possibly be controlled externally by a computer. External attenuator, external trigger, and external reference are optional.

Click to Go Back to [“Making Transmit Analysis Measurements”](#) on page 16.

Step 2 - Select the Mode and Preset the Analyzer

This step assures the analyzer will be set to a known state, either the factory preset or a user-configured preset.

Step	Action	Notes
2a. Select the Mode.	Press MODE , Bluetooth .	
2b. Preset the analyzer.	<ul style="list-style-type: none"> a. Press the green Mode Preset button for the factory preset. b. To recall a user-defined Preset: Press User Preset, User Preset to preset the current mode. 	

Click to Go Back to [“Making Transmit Analysis Measurements”](#) on page 16.

Step 3 - Select the Mode Setup Parameters

This step is to select the radio standard and the frequency range for signals of your interest and select the default limit value for the Output Power measurements.

Step	Action	Notes
3a. Select Mode Setup menu.	Press Mode Setup , if needed.	
3b. Select radio standards menu and select the desired standard.	Press Radio Standard , and select the standard from the list.	
3c. Select the default output power limit.	a. Press Device , and select the desired power class from the list.	
3d. If necessary, select the frequency settings.	a. Press FREQ Channel , and input the desired center frequency or select the France bands or non-France bands by toggling Geography . The default setting is Others .	Channel number is coupled with Center Frequency.

Click to Go Back to [“Making Transmit Analysis Measurements” on page 16.](#)

Step 4 - Select Measurement

This step is to invoke the measurement you want to run.

Step	Action	Notes
4a. Select the transmit analysis measurement.	Press Meas, Transmit Analysis .	
4b. If desired, change the parameters for the measurement.	Press Meas Setup , then select the parameters you want to change and input your settings.	
4c. Select the transmit analysis measurement again or restart the measurement.	Press Meas, Transmit Analysis , or press Restart .	

Click to Go Back to [“Making Transmit Analysis Measurements” on page 16.](#)

Step 5 - Select the Meas Setup Parameters

This step is to adjust the measurement settings.

Step	Action	Notes
5a. Select the measurement setup menus.	Press Meas Setup .	
5b. If desired, change the synchronization method.	Press Meas Setup, Burst Sync , then select one from Preamble, RF Amptd or None .	Changing Burst Sync may only impact the Output Power measurement result.
5c. If desired, change the limit parameters.	Press Meas Setup, Limits . You can turn on or turn off the limit test by toggling Limit Test . You can also change the limit lines for your measurements.	The limit lines are preset according to the test spec.
5d. If desired, store the measurement results for $\Delta f1$ Avg or $\Delta f2$ Avg.	Press Meas Setup, Hold Result , then select $\Delta f1$ Avg or $\Delta f2$ Avg .	The selected result will not be updated until Hold Result is Off.
5e. If desired, change the start point or the end point when calculating the signal power.	Press Meas Setup, More , then select what you want to change and input your settings.	

Click to Go Back to [“Making Transmit Analysis Measurements”](#) on page 16.

Step 6 - Configure the Display

Measurement results are available in graphical displays that can be presented in preset groups or adjusted to suit your specific application. Most results are also available in tabular format.


Step	Action	Notes
6a. Select a view of results display.	a. Press View/Display , and select a view from a selection of useful combinations of related data trace displays.	

Making Bluetooth Measurements

Making Transmit Analysis Measurements


View - Quad View

The Quad View provides a combination view of RF Envelope, Demod Waveform, RF Spectrum graph and Results Metrics in four quadrants.



View - RF Envelope

The RF Envelope view provides a combination view of RF Envelope graph and Results Metrics. RF Envelop graph shows the Power vs. Time trace for a single channel.




View - Demod Waveform

NOTE If radio standard is Basic or Lower Energy, this view is Demod Waveform.
If radio standard is EDR, this view is Constellation view.

The Demod Waveform or Constellation view provides a combination view of Demod Waveform graph and Results Metrics.


For Basic and Low Energy, the modulation is GFSK and Demod Waveform shows the demodulated signal as a Frequency vs. Time trace.

For EDR, the modulation is DQPSK/D8PSK and Constellation view shows an I/Q measured polar vector trace.



View - RF Spectrum

The RF Spectrum view shows a Spectrum trace.



Step 7 - Optimize your Results

In this step you can finalize the presentation of your data by adjusting various parameters to make your measurement data more useful.

Step	Action	Notes
7a. Select a graph for optimization.	Use the "Next Window" key (boxes with arrows) below the display to move the green outline "focus" to activate a graph for customization.	
7b. Optimize the X Scale.	a. Press SPAN X Scale, Scale/Div , then use the front-panel keypad to input the scale/div, then press a units key, like μs , to complete the entry. OR b. Press SPAN X Scale, More, X Scale Auto to allow the instrument to complete the scaling.	
7c. Optimize the Y Scale.	a. Press AMPTD YScale, Scale/Div , then use the front-panel keypad to input the scale/div, then press a units key, like μs , to complete the entry. OR b. Press AMPTD YScale, Auto Scale to allow the instrument to complete the scaling.	
7d. Use a Marker to indicate and measure signal characteristics.	a. Press Marker, Select Marker , and select the number of the maker desired for your display. b. Press Marker Function, Select Marker to assign a function to the marker. c. Select either Band/Interval Power , or Band/Interval Density , to set the marker control function type. d. Press Marker, Delta to set a Delta marker on/off and adjust its position on the graph.	For more information on Markers see the Markers section under Measurement Setup in Help.
7e. Add titles, or make other changes to the measurement graphs.	a. Press View/Display, Display, System, Settings to adjust other parts of the display for all measurement windows. b. Press View/Display, Display, Annotation , to turn on/off various parts of the display annotation. c. Press View/Display, Display, Title, Change Title to create a title for the graph. d. Press View/Display, Display, Graticule to turn graticules on/off. e. Press View/Display, Display, Display Line to turn a reference line on/off and adjust its position on the graph.	For more information on Display settings see the Display section under Analyzer Setup in Help.

Click to Go Back to ["Making Transmit Analysis Measurements"](#) on page 16.

Viewing Transmit Analysis Measurement Results

Measurement results available include graphical displays of various trace data as well as tabular results available remotely. These graphic results may be viewed separately, or combined into views. For more information on Views see [“Step 6 - Configure the Display” on page 19](#).

Graphical Measurement Results

To help pin-point modulation problems, the Transmit Analysis measurement provides three different graphs and numeric results tables.

You can assign any available data to any of the traces displayed on the screen. The list below of data types may also be viewed in various formats, depending on the data, like constellation, spectrum, and so on.

Transmit Analysis Measurement Data includes three sets of test results combined into this measurement. Each set of results has limit testing to provide a Pass or Fail test result.

- Power vs. Time Results
- Frequency vs. Time Results
- Spectrum Results

For more information on what these data traces are, see the *N9081A Bluetooth Measurement Application User's and Programmer's Reference* section on Trace/Detector, Data.

Click to Go Back to [“Making Transmit Analysis Measurements” on page 16](#).

Example of One-Button-Measurement - Transmit Analysis

This example assumes the signal is correctly applied to the input.


Step 1. Press **Mode Preset**

Step 2. Press **Meas, Transmit Analysis**

Step 3. View the measurement default view. (See [Figure 2-2](#)).

Making Bluetooth Measurements Making Transmit Analysis Measurements

Figure 2-2 Transmit Analysis (QuadView)



If you have a problem, and get an error message, see *Instrument Messages*.

Example of Using SCPI Commands

Select Bluetooth mode

```
:INST BT
```

Select the Transmit Analysis measurement

```
CONF:TX
```

Get the measurement result array of 27 comma-separated results.

```
READ:TX1?
```

Making Adjacent Channel Power Measurements

Adjacent Channel Power Measurement Overview

Making successful measurements of Bluetooth signals is easy when you follow the main steps below. The procedure overview below provides links in blue to a detailed procedure and description for each step:

[“Step 1 - Set Up the Test Equipment and DUT” on page 16](#)

This step configures the analyzer connections for an RF measurement.

[“Step 2 - Select the Mode and Preset the Analyzer” on page 17](#)

This step assures the analyzer will be set to a known state, either the factory preset or a user-configured preset.

[“Step 3 - Select the Mode Setup Parameters” on page 18](#)

This step specifies the radio parameters and carriers configuration.

[“Step 4 - Select Measurement” on page 18](#)

This step allows you to make an adjacent channel power measurement either by preset settings or desired settings.

[“Step 5 - Select the Meas Setup Parameters” on page 19](#)

This step specifies the measurement setup parameters.

[“Step 6 - Configure the Display” on page 19](#)

This step enables you to select different views of display.

[“Step 7 - Optimize your Results” on page 24](#)

Measurement results are available in graphical displays that can be presented in preset groups or adjusted to suit your specific application. Most results are also available in tabular format.

Step 1 - Set Up the Test Equipment and DUT

This step configures the analyzer connections for measuring Bluetooth signals.


Making the Initial Signal Connection

CAUTION

Before connecting a signal to the instrument, make sure the instrument can safely accept the signal level provided. The signal level limits are marked next to the connectors on the front panel.

Making Bluetooth Measurements
Making Adjacent Channel Power Measurements

Figure 2-3 Bluetooth Measurement Setup



The device under test (DUT) is connected to the RF input port. The DUT may possibly be controlled externally by a computer. External attenuator, external trigger, and external reference are optional. Using the appropriate cables, connect the equipment as shown.

Click to Go Back to [“Making Adjacent Channel Power Measurements”](#) on page 27.

Step 2 - Select the Mode and Preset the Analyzer

This step assures the analyzer will be set to a known state, either the factory preset or a user-configured preset.

Step	Action	Notes
2a. Select the Mode.	Press MODE , Bluetooth .	
2b. Preset the analyzer.	<ul style="list-style-type: none"> a. Press the green Mode Preset button for the factory preset. b. To recall a user-defined Preset: Press User Preset, User Preset to preset the current mode. 	

Click to Go Back to [“Making Adjacent Channel Power Measurements”](#) on page 27.

Step 3 - Select the Mode Setup Parameters

This step is to select the radio standard for signals of your interest and select the default limit value for the Output Power measurements.

Step	Action	Notes
3a. Select Mode Setup menu.	Press Mode Setup , if needed.	
3b. Select radio standards menu and select the desired standard.	Press Radio Standard , and select the standard from the list.	
3c. Select the default output power limit.	a. Press Device , and select the desired power class from the list.	
3d. If necessary, select the frequency settings.	a. Press FREQ Channel , and input the desired center frequency or select the France bands or non-France bands by toggling Geography .	

Click to Go Back to ["Making Adjacent Channel Power Measurements"](#) on page 27.

Step 4 - Select Measurement

This step is to invoke the measurement you want to run.

Step	Action	Notes
4a. Select the transmit power measurement.	Press Meas, Adjacent Channel Power .	
OR		
4b. If desired, change the parameters for the measurement.	Press Meas Setup , then select the parameters you want to change and input your settings.	
4c. Select the adjacent channel power measurement again or restart the measurement.	Press Meas, Adjacent Channel Power , or press Restart .	

Click to Go Back to ["Making Adjacent Channel Power Measurements"](#) on page 27.

Step 5 - Select the Meas Setup Parameters

This step is to adjust the measurement settings.

Step	Action	Notes
5a. Select the measurement setup menus.	Press Meas Setup .	
5b. If desired, change the measurement method.	Press Meas Setup, Meas Method to toggle between Sweep and IQ Acquisition .	The test spec defines the measurement method as Sweep, the default method.
5c. If desired, change the limit parameters.	Press Meas Setup, Limits . You can turn on or turn off the limit test by toggling Limit Test . You can also change the limit lines for your measurements.	The limit lines are preset according to the test spec.
5d. If desired, preset the parameters to comply with the standard requirement.	Press Meas Setup, Preset Standard .	The couplings include Average settings, RBW, VBW, sweep time, detector settings and gate settings.

Click to Go Back to ["Making Adjacent Channel Power Measurements"](#) on page 27.

Step 6 - Configure the Display

Measurement results are available in graphical displays that can be presented in preset groups or adjusted to suit your specific application. Required results are also available in tabular format.


Step	Action	Notes
6a. Select a view of results display.	a. Press View/Display , and select a view from a selection of useful combinations of related data trace displays.	

View - RF Envelope

The RF Envelope view provides a combination view of RF Envelope graph and Results Metrics. RF Envelop graph shows the Power vs. Time trace for a single channel.

NOTE

This view is not available when Meas Method is set to Sweep.




Making Bluetooth Measurements

Making Adjacent Channel Power Measurements

View - RF Spectrum

The RF Spectrum view shows a Spectrum trace.



Step 7 - Optimize your Results

In this step you can finalize the presentation of your data by adjusting various parameters to make your measurement data more useful.

See [“Step 7 - Optimize your Results”](#) on page 24.

Click to Go Back to [“Making Adjacent Channel Power Measurements”](#) on page 27.

Viewing Adjacent Channel Power Measurement Results

Measurement results available include graphical displays of various trace data as well as tabular results available remotely. These graphic results may be viewed separately, or combined into views. For more information on Views see [“Step 6 - Configure the Display” on page 30](#).

Graphical Measurement Results

The adjacent channel power measurement provides two different graphical views with numeric results tables.

Adjacent Channel Power Measurement Data includes two sets of test results combined into this measurement. Each set of results has limit testing to provide a Pass or Fail test result.

- Power vs. Time Results
- Spectrum Results

For more information on what these data traces are, see the *N9081A Bluetooth Measurement Application User's and Programmer's Reference* section on Trace/Detector, Data.

Click to Go Back to [“Making Adjacent Channel Power Measurements” on page 27](#).

Example Measurement - Adjacent Channel Power

This example assumes the signal is correctly applied to the input.


Step 1. Press **Mode Preset**

Step 2. Press **Meas, Adjacent Channel Power**

Step 3. View the measurement default view. (See [Figure 2-4](#)).

Making Bluetooth Measurements Making Adjacent Channel Power Measurements

Figure 2-4 Adjacent Channel Power (RF Spectrum View)



If you have a problem, and get an error message, see *Instrument Messages*.

Example of Using SCPI Commands

Select Bluetooth mode

```
:INST BT
```

Select the ACP measurement

```
CONF:ACP
```

Get the measurement results array

```
READ:ACP?
```

Making Output Spectrum BW Measurements

Output Spectrum BW Measurement Overview

Making successful measurements of Bluetooth signals is easy when you follow the main steps below. The procedure overview below provides links in blue to a detailed procedure and description for each step:

[“Step 1 - Set Up the Test Equipment and DUT” on page 35](#)

This step configures the analyzer connections for an RF measurement.

[“Step 2 - Select the Mode and Preset the Analyzer” on page 36](#)

This step assures the analyzer will be set to a known state, either the factory preset or a user-configured preset.

[“Step 3- Select the Mode Setup Parameters” on page 37](#)

This step specifies the radio parameters and carriers configuration.

[“Step 4- Select Measurement” on page 37](#)

This step allows you to make an output spect measurement either by preset settings or desired settings.

[“Step 5 - Select the Meas Setup Parameters” on page 38](#)

This step specifies the measurement setup parameters.

[“Step 6 - Optimize your Results” on page 38](#)

This step enables you to view the result display.

Step 1 - Set Up the Test Equipment and DUT

This step configures the analyzer connections for measuring Bluetooth signals.


Making the Initial Signal Connection

CAUTION

Before connecting a signal to the instrument, make sure the instrument can safely accept the signal level provided. The signal level limits are marked next to the connectors on the front panel.

Making Bluetooth Measurements
 Making Output Spectrum BW Measurements

Figure 2-5 Bluetooth Measurement Setup



The Bluetooth device under test (DUT) is connected to the RF input port by an appropriate RF cable as shown. The DUT may possibly be controlled externally by a computer. External attenuator, external trigger, and external reference are optional.

Click to Go Back to [“Making Output Spectrum BW Measurements”](#) on page 35.

Step 2 - Select the Mode and Preset the Analyzer

This step assures the analyzer will be set to a known state, either the factory preset or a user-configured preset.

Step	Action	Notes
2a. Select the Mode.	Press MODE , Bluetooth .	
2b. Preset the analyzer.	a. Press the green Mode Preset button for the factory preset. b. To recall a user-defined Preset: Press User Preset , User Preset to preset the current mode.	

Click to Go Back to [“Making Output Spectrum BW Measurements”](#) on page 35.

Step 3- Select the Mode Setup Parameters

This step is to select the radio standard for signals of your interest and select the default limit value for the Output Power measurements.

Step	Action	Notes
3a. Select Mode Setup menu.	Press Mode Setup , if needed.	
3b. Select radio standards menu and select the desired standard.	Press Radio Standard , and select the standard from the list.	
3c. Select the default output power limit.	a. Press Device , and select the desired power class from the list.	
3d. If necessary, select the frequency settings.	a. Press FREQ Channel , and input the desired center frequency or select the France bands or non-France bands by toggling Geography . The default setting is Others .	Channel number is coupled with Center Frequency.

Click to Go Back to [“Making Output Spectrum BW Measurements”](#) on page 35.

Step 4- Select Measurement

This step is to invoke the measurement you want to run.

Step	Action	Notes
4a. Select the output spectrum BW measurement.	Press Meas, Output Spectrum BW .	
4b. If desired, change the parameters for the measurement.	Press Meas Setup , then select the parameters you want to change and input your settings.	
4c. Select the output spectrum BW measurement again or restart the measurement.	Press Meas, Output Spectrum BW , or press Restart .	

Click to Go Back to [“Making Output Spectrum BW Measurements”](#) on page 35.

Step 5 - Select the Meas Setup Parameters

This step is to adjust the measurement settings.

Step	Action	Notes
5a. Select the measurement setup menus.	Press Meas Setup .	
5b. If desired, change the measurement method.	Press Meas Setup, Meas Method to toggle between Sweep and IQ Acquisition .	The test spec defines the measurement method as Sweep, the default method.
5c. If desired, change the limit parameters.	Press Meas Setup, Limits . You can turn on or turn off the limit test by toggling Limit Test . You can also change the limit lines for your measurements.	The limit lines are preset according to the test spec.
5d. If desired, preset the parameters to comply with the standard requirement.	Press Meas Setup, Preset Standard .	The couplings include Average settings, RBW, VBW, sweep time, detector settings and gate settings.

Click to Go Back to [“Making Output Spectrum BW Measurements”](#) on page 35.

Step 6 - Optimize your Results

In this step you can finalize the presentation of your data by adjusting various parameters to make your measurement data more useful.

See [“Step 7 - Optimize your Results”](#) on page 24.

Click to Go Back to [“Making Output Spectrum BW Measurements”](#) on page 35.

Viewing Output Spectrum BW Measurement Results

Measurement results available include graphical displays of trace data as well as tabular results available.

Output Spectrum BW Measurement Data Includes:

- 20 dB Bandwidth
- Transmit Frequency Error
- Total Power
- Occupied Bandwidth
- OBW Power

For more information on what these data traces are, see the *N9081A Bluetooth Measurement Application User's and Programmer's Reference* section on Trace/Detector, Data.

Click to Go Back to ["Making Output Spectrum BW Measurements"](#) on page 35.

Example of One-Button-Measurement - Output Spectrum BW


This example assumes the signal is correctly applied to the input.

- Step 1.** Press **Mode Preset**
- Step 2.** Press **Meas, Output Spectrum BW**
- Step 3.** View the measurement result. (See [Figure 2-6](#)).

Making Bluetooth Measurements

Making Output Spectrum BW Measurements

Figure 2-6 Output Spectrum BW



If you have a problem, and get an error message, see *Instrument Messages*.

Example of Using SCPI Commands

Select Bluetooth mode

```
:INST BT
```

Select the Output Spectrum BW measurement

```
CONF:OBW
```

Get the measurement results array

```
READ:OBW?
```


Making EDR In-band Spurious Emissions Measurements

EDR In-band Spurious Emissions Measurement Overview

Making successful measurements of Bluetooth signals is easy when you follow the main steps below. The procedure overview below provides links in blue to a detailed procedure and description for each step:

[“Step 1 - Set Up the Test Equipment and DUT” on page 41](#)

This step configures the analyzer connections for making a RF measurement.

[“Step 2 - Select the Mode and Preset the Analyzer” on page 42](#)

This step assures the analyzer will be set to a known state, either the factory preset or a user-configured preset.

[“Step 3- Select the Mode Setup Parameters” on page 43](#)

This step specifies the radio parameters and carriers configuration.

[“Step 4- Select Measurement” on page 43](#)

This step allows you to make an EDR in-band spurious emissions measurement either by preset settings or desired settings.

[“Step 5 - Select the Meas Setup Parameters” on page 44](#)

This step specifies the measurement setup parameters.

[“Step 6 - Configure the Display” on page 44](#)

This step enables you to select different views of display.

[“Step 7 - Optimize your Results” on page 46](#)

Measurement results are available in graphical displays that can be presented in preset groups or adjusted to suit your specific application. Most results are also available in tabular format.

Step 1 - Set Up the Test Equipment and DUT


This step configures the analyzer connections for measuring Bluetooth signals.

Making the Initial Signal Connection

CAUTION

Before connecting a signal to the instrument, make sure the instrument can safely accept the signal level provided. The signal level limits are marked next to the connectors on the front panel.

Figure 2-7 EDR In-band Spurious Emissions Measurement Setup



Bluetooth_setup

The device under test (DUT) is connected to the RF input port of the analyzer. The DUT may possibly be controlled externally by a computer. External attenuator, external trigger, and external reference are optional. Using the appropriate cables, connect the equipment as shown.

Click to Go Back to [“Making EDR In-band Spurious Emissions Measurements”](#) on page 41.

Step 2 - Select the Mode and Preset the Analyzer

This step assures the analyzer will be set to a known state, either the factory preset or a user-configured preset.

Step	Action	Notes
2a. Select the Mode.	Press MODE , Bluetooth .	
2b. Preset the analyzer.	a. Press the green Mode Preset button for the factory preset. b. To recall a user-defined Preset: Press User Preset , User Preset to preset the current mode.	

Click to Go Back to [“Making EDR In-band Spurious Emissions Measurements”](#) on page 41.

Step 3- Select the Mode Setup Parameters

This step is to turn on the Tests you want to run.

Step	Action	Notes
3a. Select Mode Setup menu.	Press Mode Setup , if needed.	
3b. Select radio standards menu and select the desired standard.	Press Radio Standard , and select the standard from the list.	
3c. Select the default output power limit.	a. Press Device , and select the desired power class from the list.	
3d. If necessary, select the frequency settings.	a. Press FREQ Channel , and input the desired center frequency or select the France bands or non-France bands by toggling Geography .	

Click to Go Back to ["Making EDR In-band Spurious Emissions Measurements" on page 41](#)

Step 4- Select Measurement

This step is to invoke the measurement you want to run.

Step	Action	Notes
4a. Select the EDR in-band spurious emissions measurement.	Press Meas, EDR In-band Spur Emissions .	
4b. If desired, change the parameters for the measurement.	Press Meas Setup , then select the parameters you want to change and input your settings.	
4c. Select the EDR in-band spurious emissions measurement again or restart the measurement.	Press Meas, EDR In-band Spur Emissions , or press Restart .	

Click to Go Back to ["Making EDR In-band Spurious Emissions Measurements" on page 41](#).

Step 5 - Select the Meas Setup Parameters

This step is to adjust the measurement settings.

Step	Action	Notes
5a. Select the measurement setup menus.	Press Meas Setup .	
5b. If desired, change the measurement method.	Press Meas Setup, Meas Method to toggle between Sweep and IQ Acquisition .	The test spec defines the measurement method as Sweep, the default method.
5c. If desired, change the synchronization timing parameters.	Press Meas Setup, Guard Delay , and enter the guard interval for your test. Press Meas Setup, Sync Interval , and enter the interval for your test.	The Guard Delay defaults to a negative value because it's prior to S0.
5d. If desired, change the limit parameters.	Press Meas Setup, Limits . You can turn on or turn off the limit test by toggling Limit Test . You can also change the limit lines for your measurements.	The limit lines are preset according to the test spec.
5e. If desired, preset the parameters to comply with the standard requirement.	Press Meas Setup, Preset Standard .	The couplings include Average settings, RBW, VBW, sweep time, detector settings and gate settings.

Click to Go Back to ["Making EDR In-band Spurious Emissions Measurements"](#) on page 41.

Step 6 - Configure the Display

Measurement results are available in graphical displays that can be presented in preset groups or adjusted to suit your specific application. Required results are also available in tabular format.


Step	Action	Notes
6a. Select a view of results display.	Press View/Display , and select a view from a selection of useful combinations of related data trace displays.	

View - RF Envelope

The RF Envelope view provides a combination view of RF Envelope graph and Results Metrics. RF Envelop graph shows the Power vs. Time trace for a single channel.

NOTE

This view is not available when Meas Method is set to Sweep.




Making Bluetooth Measurements

Making EDR In-band Spurious Emissions Measurements

View - RF Spectrum

The RF Spectrum view shows a Spectrum trace.



Step 7 - Optimize your Results

In this step you can finalize the presentation of your data by adjusting various parameters to make your measurement data more useful.

See “Step 7 - Optimize your Results” on page 24.

Click to Go Back to “Making EDR In-band Spurious Emissions Measurements” on page 41.

Viewing EDR In-band Spurious Emissions Measurement Results

Measurement results available include graphical displays of various trace data as well as tabular results available remotely. These graphic results may be viewed separately, or combined into views. For more information on Views see [“Step 6 - Configure the Display” on page 30](#).

Graphical Measurement Results

The adjacent channel power measurement provides two different graphical views with numeric results tables.

EDR In-band Spurious Emissions Measurement Data includes two sets of test results combined into this measurement. Each set of results has limit testing to provide a Pass or Fail test result.

- Power vs. Time Results
- Spectrum Results

For more information on what these data traces are, see the *N9081A Bluetooth Measurement Application User's and Programmer's Reference* section on Trace/Detector, Data.

Click to Go Back to [“Making EDR In-band Spurious Emissions Measurements” on page 41](#).


Example Measurement - EDR In-band Spurious Emissions

This example assumes the signal is correctly applied to the input.

- Step 1.** Press **Mode Preset**
- Step 2.** Press **Meas, EDR In-band Spurious Emissions**
- Step 3.** View the measurement default view. (See [Figure 2-8](#)).

Making Bluetooth Measurements Making EDR In-band Spurious Emissions Measurements

Figure 2-8 EDR In-band Spurious Emissions RF Spectrum View (Default)



If you have a problem, and get an error message, see *Instrument Messages*.

Example of Using SCPI Commands

Select Bluetooth mode

```
:INST BT
```

Select the EDR In-band Spurious Emissions measurement

```
CONF:IBSP
```

Get the measurement results array

```
READ:IBSP2?
```


Making LE In-band Emissions Measurements

LE In-band Emissions Measurement Overview

“Step 1 - Set Up the Test Equipment and DUT” on page 49

This step configures the analyzer connections for making a EDR In-band Spurious Emissions Measurement.

“Step 2 - Select the Mode and Preset the Analyzer” on page 50

This step assures the analyzer will be set to a known state, either the factory preset or a user-configured preset.

“Step 3 - Select the Mode Setup Parameters” on page 51

This step specifies the radio parameters and carriers configuration.

“Step 4 - Select Measurement” on page 51

This step allows you to make a LE In-band Emissions Measurement either by preset settings or desired settings.

“Step 5 - Select the Meas Setup Parameters” on page 52

This step specifies the measurement setup parameters.

“Step 6 - Configure the Display” on page 52

This step enables you to select different views of display.

“Step 7 - Optimize your Results” on page 54

Measurement results are available in graphical displays that can be presented in preset groups or adjusted to suit your specific application. Most results are also available in tabular format.

Step 1 - Set Up the Test Equipment and DUT


This step configures the analyzer connections for measuring Bluetooth signals.

Making the Initial Signal Connection

CAUTION

Before connecting a signal to the instrument, make sure the instrument can safely accept the signal level provided. The signal level limits are marked next to the connectors on the front panel.

Figure 2-9 LE In-band Emissions Measurement Setup



The device under test (DUT) is connected to the RF input port of the analyzer. The DUT may possibly be controlled externally by a computer. External attenuator, external trigger, and external reference are optional. Using the appropriate cables, connect the equipment as shown.

Click to Go Back to [“Making LE In-band Emissions Measurements”](#) on page 49.

Step 2 - Select the Mode and Preset the Analyzer

This step assures the analyzer will be set to a known state, either the factory preset or a user-configured preset.

Step	Action	Notes
2a. Select the Mode.	Press MODE , Bluetooth .	
2b. Preset the analyzer.	a. Press the green Mode Preset button for the factory preset. b. To recall a user-defined Preset: Press User Preset , User Preset to preset the current mode.	

Click to Go Back to [“Making LE In-band Emissions Measurements”](#) on page 49.

Step 3 - Select the Mode Setup Parameters

This step is to turn on the Tests you want to run.

Step	Action	Notes
3a. Select Mode Setup menu.	Press Mode Setup , if needed.	
3b. Select radio standards menu and select the desired standard.	Press Radio Standard , and select the standard from the list.	
3c. Select the default output power limit.	Press Device , and select the desired power class from the list.	
3d. If necessary, select the frequency settings.	Press FREQ Channel , and input the desired center frequency or select the France bands or non-France bands by toggling Geography .	

Click to Go Back to [“Making LE In-band Emissions Measurements” on page 49](#)

Step 4 - Select Measurement

This step is to invoke the measurement you want to run.

Step	Action	Notes
4a. Select the LE in-band emissions measurement.	Press Meas, LE In-band Emissions .	
4b. If desired, change the parameters for the measurement.	Press Meas Setup , then select the parameters you want to change and input your settings.	
4c. Select the LE in-band emissions measurement again or restart the measurement.	Press Meas, LE In-band Emissions , or press Restart .	

Click to Go Back to [“Making LE In-band Emissions Measurements” on page 49](#).

Step 5 - Select the Meas Setup Parameters

This step is to adjust the measurement settings.

Step	Action	Notes
5a. Select the measurement setup menus.	Press Meas Setup .	
5b. If desired, change the measurement method.	Press Meas Setup, Meas Method to toggle between Sweep and IQ Acquisition .	The test spec defines the measurement method as Sweep, the default method.
5c. If desired, change the limit parameters.	Press Meas Setup, Limits . You can turn on or turn off the limit test by toggling Limit Test . You can also change the limit lines for your measurements.	The limit lines are preset according to the test spec.
5d. If desired, preset the parameters to comply with the standard requirement.	Press Meas Setup, Preset Standard .	The couplings include Average settings, RBW, VBW, sweep time, detector settings and gate settings.

Click to Go Back to ["Making LE In-band Emissions Measurements"](#) on page 49.

Step 6 - Configure the Display

Measurement results are available in graphical displays that can be presented in preset groups or adjusted to suit your specific application. Required results are also available in tabular format.


Step	Action	Notes
5a. Select a view of results display.	a. Press View/Display , and select a view from a selection of useful combinations of related data trace displays.	

View - RF Envelope

The RF Envelope view provides a combination view of RF Envelope graph and Results Metrics. RF Envelop graph shows the Power vs. Time trace for a single channel.

NOTE

This view is not available when Meas Method is set to Sweep.




Making Bluetooth Measurements

Making LE In-band Emissions Measurements

View - RF Spectrum

The RF Spectrum view shows a Spectrum trace.



Step 7 - Optimize your Results

In this step you can finalize the presentation of your data by adjusting various parameters to make your measurement data more useful.

See [“Step 7 - Optimize your Results”](#) on page 24.

Click to Go Back to [“Making LE In-band Emissions Measurements”](#) on page 49.

Viewing LE In-band Emissions Measurement Results

This section provides some examples of LE In-band Emissions Measurement results. Results available include graphical displays of various trace data as well as tabular results available remotely. These graphic results may be viewed separately, or combined into views. For more information on Views see [“Step 6 - Configure the Display” on page 52](#) and [“Step 7 - Optimize your Results” on page 54](#).

LE In-band Emissions Measurement Data includes two sets of test results combined into this measurement. Each set of results has limit testing to provide a Pass or Fail test result.

- Power vs. Time Results
- Spectrum Results

For more information on what these data traces are, see the *N9081A Bluetooth Measurement Application User's and Programmer's Reference* section on Trace/Detector, Data.

Click to Go Back to [“Making LE In-band Emissions Measurements” on page 49](#).


Example Measurement - LE In-band Emissions

This example assumes the signal is correctly applied to the input.

- Step 1.** Press **Mode Preset**
- Step 2.** Press **Meas, LE In-band Emissions**
- Step 3.** View the measurement default view. (See [Figure 2-10](#)).

Making Bluetooth Measurements Making LE In-band Emissions Measurements

Figure 2-10 LE In-band Emissions RF Spectrum View (Default)



If you have a problem, and get an error message, see *Instrument Messages*.

Example of Using SCPI Commands

Select Bluetooth mode

```
:INST BT
```

Select the LE In-band Emissions measurement

```
CONF:IBEM
```

Get the measurement results array

```
READ:IBEM2?
```


Making Monitor Spectrum Measurements

Monitor Spectrum Measurement Overview

Making successful measurements of Bluetooth signals is easy when you follow the main steps below. The procedure overview below provides links in blue to a detailed procedure and description for each step:

[“Step 1 - Set Up the Test Equipment and DUT” on page 57](#)

This step configures the analyzer connections for an RF measurement.

[“Step 2 - Select the Mode and Preset the Analyzer” on page 58](#)

This step assures the analyzer will be set to a known state, either the factory preset or a user-configured preset.

[“Step 3- Select the Mode Setup Parameters” on page 59](#)

This step specifies the radio parameters and carriers configuration.

[“Step 4- Select Measurement” on page 59](#)

This step allows you to make a transmit analysis measurement either by preset settings or desired settings.

[“Step 5 - Optimize your Results” on page 59](#)

Measurement results are available in graphical displays that can be presented in preset groups or adjusted to suit your specific application. Most results are also available in tabular format.

Step 1 - Set Up the Test Equipment and DUT

This step configures the analyzer connections for measuring Bluetooth signals.


Making the Initial Signal Connection

CAUTION

Before connecting a signal to the instrument, make sure the instrument can safely accept the signal level provided. The signal level limits are marked next to the connectors on the front panel.

Making Bluetooth Measurements
 Making Monitor Spectrum Measurements

Figure 2-11 Bluetooth Measurement Setup



Bluetooth_setup

The Bluetooth device under test (DUT) is connected to the RF input port by an appropriate RF cable as shown. The DUT may possibly be controlled externally by a computer. External attenuator, external trigger, and external reference are optional.

Click to Go Back to [“Making Monitor Spectrum Measurements”](#) on page 57.

Step 2 - Select the Mode and Preset the Analyzer

This step assures the analyzer will be set to a known state, either the factory preset or a user-configured preset.

Step	Action	Notes
2a. Select the Mode.	Press MODE, Bluetooth .	
2b. Preset the analyzer.	a. Press the green Mode Preset button for the factory preset. b. To recall a user-defined Preset: Press User Preset, User Preset to preset the current mode.	

Click to Go Back to [“Making Monitor Spectrum Measurements”](#) on page 57.

Step 3- Select the Mode Setup Parameters

This step is to select the radio standard for signals of your interest and select the default limit value for the Output Power measurements.

Step	Action	Notes
3a. Select Mode Setup menu.	Press Mode Setup , if needed.	
3b. Select radio standards menu and select the desired standard.	Press Radio Standard , and select the standard from the list.	
3c. Select the default output power limit.	a. Press Device , and select the desired power class from the list.	
3d. If necessary, select the frequency settings.	a. Press FREQ Channel , and input the desired center frequency or select the France bands or non-France bands by toggling Geography . The default setting is Others .	Channel number is coupled with Center Frequency.

Click to Go Back to [“Making Monitor Spectrum Measurements” on page 57.](#)

Step 4- Select Measurement

This step is to invoke the measurement you want to run.

Step	Action	Notes
4a. Select the monitor spectrum measurement.	Press Meas, Monitor Spectrum .	
4b. If desired, change the parameters for the measurement.	Press Meas Setup , then select the parameters you want to change and input your settings.	
4c. Select the monitor spectrum measurement again or restart the measurement.	Press Meas, Monitor Spectrum , or press Restart .	

Click to Go Back to [“Making Monitor Spectrum Measurements” on page 57.](#)

Step 5 - Optimize your Results

In this step you can finalize the presentation of your data by adjusting various parameters to make your measurement data more useful.

See [“Step 7 - Optimize your Results” on page 24.](#)

Click to Go Back to [“Making Monitor Spectrum Measurements” on page 57.](#)

Viewing Monitor Spectrum Measurement Results

Monitor Spectrum Measurement provides a graphic display only.

For more information on what these data traces are, see the *N9081A Bluetooth Measurement Application User's and Programmer's Reference* section on Trace/Detector, Data.


Click to Go Back to ["Making Monitor Spectrum Measurements"](#) on page 57.

Example of One-Button-Measurement - Monitor Spectrum

This example assumes the signal is correctly applied to the input.

- Step 1.** Press **Mode Preset**
- Step 2.** Press **Meas, Monitor Spectrum**
- Step 3.** View the measurement result. (See [Figure 2-12](#)).

Figure 2-12 *Monitor Spectrum*



If you have a problem, and get an error message, see *Instrument Messages*.

Example of Using SCPI Commands

Select Bluetooth mode

```
:INST BT
```

Select the Transmit Analysis measurement

```
CONF:MON
```

Get the measurement results array

```
READ:MON?
```

Troubleshooting Bluetooth Measurements

The following table provides some with descriptions you might encounter when making Bluetooth measurements.

Table 2-1

Error Numbers	Messages	Notes
207	Burst Not Found; Preamble not found	Packet preamble has not been detected.
207	Burst Not Found; Acq time too short	The data does not include the falling edge.
215	Sync error; EDR Sync not found	The capture data does not include EDR sync code.
215	Sync error; Preamble not found	The preamble code is not detected.
217	Demod error; Invalid packet type	Packet type detected is not recognized.
217	Demod error; Acq time too short	Not an entire Payload is captured, that is, the acquisition time is too short.
159	Settings Alert; Parm/data mismatch	The actual captured data length is not long enough, or, the offset from TxFrequency to center frequency is not an integer multiple of 1 MHz.

For more information on error messages see *Instrument Messages*.

3 Concepts

This chapter describes basics of Bluetooth technology and an explanation of how measurements are performed by the instrument.

Introduction

This chapter provides a variety of digital modulation concepts to help you to better understand the features and measurements performed by the signal analyzer.

[“Understanding Bluetooth” on page 65](#), explains the basic concepts of the Bluetooth standards.

[“Measurement Concepts and Results” on page 73](#), explains the methods by which a digital baseband signal is modulated onto an RF carrier.

Understanding Bluetooth

NOTE Bluetooth is a trademark of the Bluetooth SIG, Inc..

What is Bluetooth


Bluetooth is a limited range RF link technology. It targets at voice and data transmission between appliances for home or business environments. The need for interconnecting cables is eliminated. The system, typically now a single integrated component, consists of a radio, baseband link control, and link management.

The Bluetooth radio system may operate as either master or slave units. The connection topology is basically a star, with a single master and up to 7 slaves. Another 200+ slaves can be registered and in a non-communicating, power-saving mode.

This area of control is defined as a piconet. A master in one piconet may be a slave to a master from a different piconet. Similarly, multiple masters from different piconets may control a single slave. This network of piconets is referred to as a scatternet.

Figure 3-1 depicts two piconets comprising a scatternet. Units that are not part of either piconet remain in standby mode.

Figure 3-1 Bluetooth Network Topology



What is Bluetooth Enhanced Data Rate (EDR)

The key characteristic of Bluetooth enhanced data rate technology is that the transmitted data rate and modulation scheme are changed within the packet. The access code and packet header are transmitted with the GFSK modulation scheme (Basic Rate 1 Mbps) and the subsequent synchronization sequence, payload, and trailer sequence are transmitted using the $\pi/4$ DQPSK (2 Mbps) or 8DPSK (3 Mbps).

What is Bluetooth Low Energy (LE)


Bluetooth low energy technology is a new extension of Bluetooth technology. It is designed with two modes: stand-alone mode and dual mode. Small devices like watches and sports sensors are based on a stand-alone Bluetooth low energy implementation and consume minimal power. Dual mode implementations use parts of the existing Bluetooth hardware, sharing one physical radio and antenna and will basically keep the same power consumption as classic Bluetooth technology.

Bluetooth Radio Characteristics

Figure 3-2 is a block diagram for the Bluetooth radio system, showing the baseband controller and the RF transmitter and receiver sections.

Figure 3-2 Bluetooth Radio Block Diagram

Bluetooth radio block diagram



Bluetooth devices operate in the 2.4 to 2.485 GHz Industrial, Scientific and Medical (ISM) band. This band is unlicensed and available globally.

The design emphasis for Bluetooth device is on very low power, extremely low cost, and robust operation in the uncoordinated, interference dominated RF environment of the ISM band.

The operating range depends on the device class.

- Power Class 3: has a range of up to 1 meter
- Power Class 2: has a range of 10 meters
- Power Class 1: has a range of 100 meters with an additional 20 dB power amplifier

Table 3-1

Power Class	Operating Range	Maximum Power	Minimum Power	Power Control	Notes
3	up to 1 m	100 mW (20 dBm)	1 mW (0 dBm)	Yes	

Table 3-1

Power Class	Operating Range	Maximum Power	Minimum Power	Power Control	Notes
2	up to 10 m	2.5 mW (4 dBm)	0.25 mW (-6 dBm)	NA	nominal 1 mW (0 dBm)
1	100 m	1 mW (0 dBm)		NA	

Bluetooth low energy technology has a range of up to 200 meters.

Bluetooth devices are designed to have very low power consumption. The most commonly used radio is Class 2 and uses 2.5 mW of power.

Bluetooth low energy technology consumes between 1/2 and 1/100 the power of classic Bluetooth technology.

Bluetooth has two modulation modes: Basic Rate, a mandatory mode, which uses a shaped, binary FM modulation to minimize transceiver complexity, and Enhanced Data Rate (EDR), an optional mode, which uses PSK modulation and has two variants: $\pi/4$ DQPSK and 8DPSK. The symbol rate for all modulation schemes is 1 M s/sec.

For classic Bluetooth devices, the maximum data rates for a single link are 432.6/432.6 kbps symmetrical, or 721.0/57.6 kbps asymmetrical.

For Bluetooth low energy, the maximum data rates are 1 Mbps.

For EDR, the maximum data rates are 2 Mbps using $\pi/4$ DQPSK and 3 Mbps using 8DPSK.

The Bluetooth radio unit employs adaptive frequency hopping (AFH), usually at 1600 hops/sec. The signal is nominally at each hop frequency for one 625 μ s timeslot. Each pair of time consecutive slots constitute a time-division duplex (TDD) frame. In the TDD scheme used, the master transmits in even-numbered timeslots, and the slave transmits in odd-numbered timeslots.

A Bluetooth channel is divided into timeslots, each 625 μ s in length. There are 79 Bluetooth channels, each 1 MHz wide.

Baseband Characteristics

Voice or data communication is by packet based TDD. Data packets may extend over one, three, or five time slots, whereas voice packets are limited to a single time slot. Frequency hopping occurs at the end of the last time slot associated with the packet.

A Basic Rate packet, shown in [Figure 3-3](#), contains an access code, header and payload. The access code consists of a preamble, a sync word, and an optional trailer. The header contains piconet member address and packet information. The payload data, consisting of payload header, payload data and CRC, carries the user's voice or data information. The payload CRC (Cyclic Redundancy Check) is a 16-bit field at the

end of the payload that is used for a data integrity check. Depending on the packet type, a payload starts with a 1 (DH1) or 2 (DH3/5) byte header, and finishes with a 2 byte CRC.

Figure 3-3 *Bluetooth Basic Rate Packet Format*

Access Code			Header	Payload		
Preample	Sync Code	Trailer		Payload Header	Payload Data	CRC

A general format of Enhanced Data Rate packet is shown in Figure 3-4. The access code and header are identical in format and modulation to Basic Rate packets. It has a guard time and synchronization sequence following the header and two trailer symbols following the payload.

Figure 3-4 *Bluetooth Enhanced Data Rate Packet Format*

Access Code	Header	Guard	Sync	Enhanced Data Rate Payload	Trailer
-------------	--------	-------	------	----------------------------	---------

The Bluetooth Radio Unit

The Bluetooth radio unit is shown in Figure 3-2 as the transmitter and receiver sections of the block diagram. The transmitter upconverts the baseband information to the frequency-modulated carrier. Frequency hopping and bursting are performed at this level. Conversely, the receiver downconverts and demodulates the RF signal. Table 3-2 summarizes some of the key RF characteristics of Bluetooth.

Table 3-2 *Bluetooth RF Characteristics*

Characteristic	Specification	Notes
Carrier Frequency	2400 to 2483.5 MHz (ISM radio band)	$f=2402+k$ MHz, $k=0,\dots,78$
Modulation	0.5 BT Gaussian-filtered 2FSK at 1 Msymbol/s Modulation index: 0.28 to 0.35 (0.32 nominal)	Digital FM scheme The peak frequency deviation allowed is 175 kHz

Table 3-2 Bluetooth RF Characteristics

Characteristic	Specification	Notes
Hopping	<p>1600 hops/s (in normal operation)^a</p> <p>1 MHz channel spacing</p> <p>The system has 5 different hopping sequences:</p> <ol style="list-style-type: none"> 1) Page hopping sequence 2) Page response sequence 3) Inquiry sequence 4) Inquiry response sequence 5) Channel hopping sequence <p>The first four are restricted hopping sequences used during connection setup. The normal channel hopping sequence is pseudorandom based on the master clock value and device address.</p>	<p>The channel hopping sequence is designed to visit each frequency regularly and with roughly equal probability. It has a periodicity of 23 hours and 18 minutes.</p>
Transmit Power	<p>Power Class 1: 1 mW (0 dBm) to 100 mW (+20 dBm)</p> <p>Power Class 2: 0.25 mW (-6 dBm) to 2.5 mW (+4 dBm)</p> <p>Power Class 3: 1 mW (0 dBm)</p>	<p>Class 1 power control: +4 to +20 dBm (required) -30 to 0 dBm (optional)</p> <p>Class 2 power control: -30 to 0 dBm (optional)</p> <p>Class 3 power control: -30 to 0 dBm (optional)</p>
Operating Range	<p>10 cm to 10 m (100 m with Power Class 1)</p>	
Maximum Data Throughput	<p>The asynchronous channel can support an asymmetric link of maximally 721 kbps in either direction while permitting 57.6 kb/s in the return direction, or a 432.6 kbps symmetric link.</p>	<p>Data throughput is lower than the 1 Msymbol/s rate as a result of the overhead, which is inherent in the protocol.</p>

a. Hop speed may vary depending on packet length.

The Bluetooth Link Control Unit and Link Management

The Bluetooth link control unit, also known as the link controller, determines the state of the device. It is responsible for establishing the network connections as well as power efficiency, error correction, and encryption.

Bluetooth radios may operate as either master or slave units. The link manager sets up the connection between master and slave units and also determines the slave's power saving mode.

The link management software works with the link control unit. Devices communicate among each other through the link manager. [Table 3-3](#) provides a summary of the link control and management functions.

Table 3-3 Link Control and Management Functions

Characteristic	Specification	Notes
Network Connections	The master's link controller initiates the connection procedure and sets the power saving mode of the slave.	
Link Types	Two link types: <ul style="list-style-type: none"> • Synchronous Connection Oriented (SCO) type, primarily for voice • Asynchronous Connectionless (ACL) type, primarily for packet data 	Bluetooth can support an asynchronous data channel, up to three simultaneous synchronous voice channels, or a channel that simultaneously supports asynchronous data and synchronous voice. Time-Division Duplexing is used for full duplex operation.
Packet Types	NULL, POLL, FHS - System packets DM1, DM3, DM5 - Medium rate, error-protected data packets DH1, DH3, DH5 - High rate, non-protected data packets HV1, HV2, HV3 - Digitized audio, 3 levels of error protection DV - Mixed data and voice AUX1 - For other uses	The 1, 3 and 5 suffixes indicate the number of time slots occupied by the data burst. Nominal burst lengths: DH1–366 ms DH3–1622 ms DH5–2870 ms
Error Correction	Three error correction schemes: <ul style="list-style-type: none"> • 1/3 rate Forward Error Correction (FEC) code • 2/3 rate Forward Error Correction (FEC) code • Automatic repeat request (ARQ) scheme for data 	Error correction is provided by the Link Manager
Authentication	Challenge-response algorithm. Authentication may be unused, unidirectional, or bidirectional.	Authentication is provided by the Link Manager

Table 3-3 *Link Control and Management Functions*

Characteristic	Specification	Notes
Encryption	Stream cipher with secret key lengths of 0, 40, or 64 bits.	
Test Modes	Provides the ability to place the device into test loop-back mode and allows control of test parameters such as frequency settings, power control, and packet type.	

Measurement Concepts and Results

The N9081A & W9081A Bluetooth measurement application provides:

- Adjacent Channel Power (ACP) Measurement
- Enhanced Data Rate (EDR) In-band Spurious Emissions Measurement
- Low Energy (LE) In-band Emissions Measurement
- Monitor Spectrum Measurement
- Output Spectrum Bandwidth (OBW) Measurement
- Transmit Analysis Measurement

Adjacent Channel Power Measurement

The Adjacent Channel Power measurement is a low cost test for Bluetooth + LE devices to conform to the following Bluetooth specifications:

- Bluetooth 2.1 TRM/CA/06/C (TX Output Spectrum - Adjacent Channel Power) RF test specifications

The measurement is performed to verify the emissions levels within the operating frequency range conform to the limits. The power measurements covered by this test are total peak power for adjacent channels.

The measurement provides 79 scalar values of the transmit power per channel. The front panel results include:

- Tx Chan: transmit channel number
- Tx Chan Power: transmit channel power
- Adj Chan kHz Lower: adjacent 500 kHz lower channel power
- Adj Chan kHz Upper: adjacent 500 kHz upper channel power
- Number of Exceptions: pass/fail indicator for every channel

EDR In-band Spurious Emissions Measurement

The EDR In-band Spurious Emissions measurement is a low cost test for Bluetooth + EDR devices to conform to the following Bluetooth specifications:

- Bluetooth 2.1 TRM/CA/13/C (EDR - In-band Spurious Emissions) RF test specifications

The measurement verifies whether the level of unwanted signal within the used frequency band lies below the required level.

The analyzer is set to zero span with a sweep time of packet length, a sweep mode of Max Hold, a detector of Average, a RBW of 100 kHz and VBW of 300 kHz. By adjusting the Gate delay and Gate length, the analyzer records the signal only in those parts of the signal in which the device transmits DPSK-modulated data. The measurement is implemented by sweep.

The measurement provides 79 scalar values of the Tx power per channel. The front panel results include:

- Tx Chan: transmit channel number
- Tx Chan Power: transmit channel power
- Adj 500 kHz Lower
- Adj 500 kHz Upper
- Number of Exceptions: pass/fail indicator for every channel

LE In-band Emissions Measurement

The LE In-band Emissions measurement is a low cost test for Bluetooth + LE devices to conform to the following Bluetooth specifications:

- Bluetooth + LE RF PHY 0.7 (TRM-LE/CA-02-C) RF test specifications

The power measurements are total peak power for adjacent channels.

The measurement provides 80 scalar values of the Tx power per channel. The front panel results include:

- Tx Chan: transmit channel number
- Tx Chan Power: transmit channel power
- Adj Chan kHz Lower
- Adj Chan kHz Upper
- Number of Exceptions: pass/fail indicator for every channel

Output Spectrum Bandwidth

The output spectrum bandwidth measurement is used to verify if the emissions inside the operating frequency are within the limits.

The Bluetooth RF Specification defines a minimal bandwidth of 3 MHz, normally for a standard gaussian filter, which may cause around 4% frequency deviation when 0101 symbol sequence is used. Therefore it is recommended to use a higher bandwidth IF that has a flat amplitude characteristics and does not affect the frequency deviation of the signal under test.

To perform this measurement the analyzer is tuned to the channel to be measured and the span is set to 2 MHz. The peak of the current trace is identified. The measurement then places markers at the points highest and lowest in frequency in the current span where the signal drops -20dB from this peak value. The frequency between these two points is measured as the output spectrum bandwidth.

Transmit Analysis Measurement

The transmit analysis measurement is a low cost test, which combines multiple measurements in a single package, for Bluetooth + EDR devices to conform to the following Bluetooth specifications:

- Bluetooth + EDR Ver2.1 RF test specifications
- Bluetooth Low Energy RF test specifications

The measurement is performed on a single IQ data acquisition. If radio standard is Basic or Low Energy, it measures Output Power, Modulation Characteristics, Initial Carrier Frequency Tolerance (ICFT) and Carrier Frequency Drift. If radio standard is EDR, it measures Relative Transmit Power, Frequency Stability and Modulation Accuracy, and Differential Phase Decoding.

The results shown on the front panel depend upon the radio standard selected from **Mode Setup, Radio**. The following results are common to all standards:

- Packet Type
- Payload
- Payload Length

The following results are common to basic Bluetooth standard by pressing **Mode Setup, Radio, Basic**:

- Output Power
- GFSK Average Power
- GFSK Peak Power
- Modulation Characteristics:
 - $\Delta F1$ Avg for the 11110000 payload data pattern
 - $\Delta F2$ Avg for the 10101010 payload data pattern
 - Min $\Delta F1$ Max
 - Max $\Delta F1$ Max
 - Min $\Delta F2$ Max
 - Max ΔF Max
 - $\Delta F2 > 115$ kHz
 - $\Delta F2/\Delta F1$ Ratio
- Initial Carrier Frequency Tolerance (ICFT)
- Carrier Frequency Drift
 - Frequency Drift
 - Max Drift Rate

The following results are common to basic Bluetooth standard by pressing **Mode Setup, Radio, EDR**:

- Carrier Frequency Stability & Modulation Accuracy
 - Frequency Offset w_i
 - Frequency Offset w_0
 - Frequency Offset w_i+w_0
 - RMS DEVM
 - Peak DEVM
 - 99% DEVM for EDR modulation
- EDR Relative Transmit Power
 - GFSK Avg Power
 - DPSK Avg Power
 - Relative Power (GFSK Avg Power - DPSK Avg Power)
- EDR Differential Phase Encoding
 - BER
 - Bit Errors
 - Guard Interval

Output Power

Output power measurements are performed to ensure power levels are high enough to maintain links, yet low enough to minimize interference within the ISM band and to maximize battery life.

The power measurements covered by this test are average power and peak power for the specified channel or center frequency. The analyzer is set to zero span mode with a sweep time dependent on the packet type being measured. When the analyzer is triggered, it makes a sweep over the duration of the burst.


The peak power is calculated as the highest point in the burst.

The average power is calculated as the average power over 20% to 80% of the burst duration. You can choose a method to determine the burst duration:

1. Preamble - The position of p_0 is used to define the start of the burst.
2. RF Amptd - The burst duration is taken as the time between the leading and trailing 3 db points compared to the average power.
3. None - There is no synchronization process.

Figure 3-5 illustrates power and timing characteristics of a signal burst in the time domain.

Figure 3-5 Time Domain Power and Timing Analysis



Carrier Frequency Drift

The carrier frequency drift measurement checks the performance of the modulator circuitry and the stability of the Voltage Controlled Oscillator (VCO).

To make the measurement a demodulated signal is used with the payload data consisting of a repeating 4-bit 1010 sequence. The absolute frequencies of the 4 preamble bits are measured and integrated, providing the initial carrier frequency. The absolute frequencies of each successive 10-bit pattern in the payload are then measured and integrated.

The frequency drift is the maximum difference between the average frequency of the 4 preamble bits and the average frequency of any 10 bits in the payload field. The maximum drift rate applies to the difference between any two 10-bit groups separated by 50 μ s within the payload field.

Monitor Band/Channel Measurement

This measurement can be used to visually check either the Bluetooth band or individual channels. In Monitor Band, you can easily check the channel occupancy and flatness when frequency hopping is on. In Monitor Channel you can verify the correct spectral shape of the selected channel.

Any interfering signals may also be apparent when using this measurement.

This procedure scans the specified band or channels. By placing markers on the trace it is possible to check the band/channels for interference or other impairments. A Max Hold function enables monitoring over time. This is useful for dealing with hopping signals.

Initial Carrier Frequency Tolerance

The initial carrier frequency tolerance measurement is designed to verify the accuracy of the transmitter's initial carrier frequency.

This is measured by integrating over the frequency deviations of the packets first 4 bits (the preamble bits). The result is either a positive or negative number in Hz indicating the frequency difference from the specified nominal carrier frequency.

This measurement requires the signal to be demodulated to measure the frequency deviation of each symbol. After demodulation, the frequency offset of each of the preamble bits is measured and averaged.

Modulation Characteristics

Modulation characteristics is a frequency deviation measurement which is designed to verify both the modulator performance and the accuracy of the pre-modulation 0.5BT Gaussian Filter.

Two separate test signals are required for this measurement, each one containing an 8-bit repeating sequence in the payload. These repeating sequences are 11110000 and 10101010.

The measurement is performed in 2 stages, each stage requiring a different packet: one carrying the 11110000 payload, the other carrying the 10101010 payload.

Using the 11110000 payload

The average frequency over the first 8 bits in the payload is calculated and then the maximum deviation from this average over bits 2,3,6 & 7 is measured. The maximum for each repeating 8-bit sequence in the payload is measured in the same way, each time calculating a new average frequency over the respective 8 bits. Eventually, the average of all these maximums is calculated and shown in the results window as $\Delta f1$ Avg.

Using the 10101010 payload

The average frequency over the first 8 bits in the payload is calculated and then the maximum deviation from this average over all 8 bits is measured. The maximum for each repeating 8-bit sequence in the payload is measured in the same way, each time calculating a new average frequency over the respective 8 bits. Eventually the average of all these maximums is calculated and shown in the results window as $\Delta f2$ Avg.

Once the measurement has acquired values for $\Delta f1$ Avg and $\Delta f2$ Avg, the ratio of $\Delta f2$ Avg to $\Delta f1$ Avg is also displayed.

Since this measurement requires human interaction (to manually change test signals), it will display only either $\Delta f1$ Avg or $\Delta f2$ Avg the first time it is run, depending on the signal type. The first result must be "held" using the **Hold Result** parameter and the measurement restarted. At this point the other signal type should be supplied, and the measurement restarted.

If a remote query of all three results is requested after having obtained only one result, then 3 values will be returned, although only 1 result will be correct and the other 2 will contain NaN.