

Agilent U8202A Electronic Test Lead Kit

Operating Instructions

This test lead kit is designed for use with Agilent Digital Multimeter | DC Power Supply.

Assistance

For technical assistance, contact your nearest Agilent Sales Office or visit the Agilent website at www.agilent.com/find/assist for further information.

Symbols



Refer to instrument user's guide for additional safety information on test probe lead.

Ratings

Test Probe Lead	CAT II 1000 V, 15 A
Alligator Clips	CAT III 1000 V, 10A
Mini Grabber	CAT II 300 V, 3 A
Fine Tips Test Probe	CAT II 300 V, 3 A

Operation

Refer to instrument user's guide for further information.

The fine tips test probes, alligator clips, and mini grabber are plugged on to the end of the test probe lead.

WARNING

If test lead kit is used in a manner not specified by Agilent Technologies, the protection provided may be impaired. Instrument damage or personal injury may result.

Maintenance

If any portion of this test lead kit is worn or damaged, do not use.

