

4145B-07

S E R V I C E N O T E

SUPERSEDES: None

HP 4145B Semiconductor Parameter Analyzer

Serial Numbers: 2830J02614 / 9999J99999

New CRT Display for the HP 4145B

Situation:

The CRT display for the HP 4145B has been changed to HP part number 03562- 60151. The previous display was from the HP 1345A and can no longer be used due to the discontinuance of the HP 1345A. CRT Display part number 03562-60151 is from the HP 3562A Dynamic Signal Analyzer.

Solution/Action:

When you need to service the CRT display of an HP 4145B with serial number 2830J02614 and above, use the service information for the HP 3562A CRT Display. The Manual Supplement of the HP 3562A CRT Display is available from DMK as HP part number 03562-90013. The support strategy of the display is component level repair.

DATE: 08 October 1991

ADMINISTRATIVE INFORMATION

SERVICE NOTE CLASSIFICATION:

INFORMATION ONLY

AUTHOR:	ENTITY:	ADDITIONAL INFORMATION:
HS	3300	

© 1991 HEWLETT-PACKARD COMPANY
PRINTED IN U.S.A.

