

4156B-12

S E R V I C E N O T E

Supersedes:
NONE

4156B Precision Semiconductor Parameter Analyzer

Serial Numbers: JP10E00101/JP99Z99999

LCD Replaceable Unit Change

Parts Required:

P/N	Description	Qty.
04155-65010	LCD	1

Situation:

When the LCD becomes dark, we recommended replacing the LCD lamp (p/n 2090-0566). According to the part shortage and design problem, we changed the LCD replacement unit. When the LCD becomes dark, replace LCD (p/n 2090-0574) instead of LCD lamp (p/n 2090-0566). If you do not find the LCD (p/n 2090-0574) in your inventory, use another LCD (p/n 04155-65010) for the repair.

Solution/Action:

When the LCD becomes dark, replace LCD (p/n 2090-0574 or 04155-65010) instead of LCD lamp (p/n 2090-0566).

Note that p/n 04155-65010 should be used for 4155B/4156B only. Do not use this to repair the other instrument, which uses p/n 2090-0574.

ADMINISTRATIVE INFORMATION

SERVICE NOTE CLASSIFICATION:

INFORMATION ONLY

AUTHOR: SY PRODUCT LINE: 1H

ADDITIONAL INFORMATION: