

Agilent 4279A MHz C-V Meter

Manual Change

Agilent Part No. 04279-90001

February 2003

Printed in Japan

Miscellaneous Changes

The option system of the 4279A has changed since February 2003. Apply the following changes.

Old Option Number	New Option Number	Remark
Standard	800 (No Test Frequency Shift: Measurement frequency : 1MHz Standard)	1
003 (1% Frequency Shift)	same as the left number	
907 (Front Handle Kit)	same as the left number	
908 (Rack Mount Kit)	same as the left number	
909 (Rack Flange and Handle Kit)	same as the left number	
ABA (Add Operation Manual (English))	—	2
—	ABA (Add Operation Manual (English))	
ABJ (Add Operation Manual (Japanese))	—	
—	ABJ (Add Operation Manual (Japanese))	
0BW (Add Service Manual)	same as the left number	
0B0 (Delete Manual)	—	3
0B1 (Add Manual)	—	

1. In the previous system, an option for the 1% Frequency Shift was available only for 1% Frequency Shift. In the new option system, it is available for the 1% Frequency Shift and No Test Frequency Shift, requiring the customer to select either of them.
2. In the previous system, the option number is used to choose the language of the operation manual (standard accessory). In the new option system, it is used to add an operation manual (optional accessory) of the language the customer desires.
3. No selection of addition/deletion is required for the operation manual set because it is only available

as an optional accessory in the new option system. When 2 or more sets are required, specify option ABA,ABJ * (number of required sets).