

S E R V I C E N O T E

SUPERSEDES: None

Agilent 4286A RF LCR Meter

Serial Numbers: 0000A0000 / 99999Z99999

Part Information about 3 dB attenuator

To Be Performed By: Agilent-Qualified Personnel

Parts Required:

P/N	Description	Quantity
0955-1309	Attenuator,3dB	1

Situation:

3 dB attenuator in 4286A has been changed from 0955-0701 to 0955-1309.

Solution / Action:

When replacing 3 dB attenuator for 4286A, use the new part 0955-1309.

Note:

The new attenuator shouldn't be used in 4291A/B, because of maximum frequency limitation.

DATE: February 2001

ADMINISTRATIVE INFORMATION

SERVICE NOTE CLASSIFICATION:

INFORMATION ONLYAUTHOR:
AYENTITY:
3355

ADDITIONAL INFORMATION:

