

Agilent 4291B RF Impedance/Material Analyzer Calibration Kit Contents List

Calibration Kit Serial No. _____

Calibration Kit Contents:

50 ohm Termination ¹	04291-65006	Serial No. _____
0 ohm Termination	04191-85300	
0 S Termination	04191-85302	
Low-Loss Capacitor	04291-60042	

1: Note: If this 50 ohm termination is replaced, the measurement accuracy test described in the *Service Manual* must be done.



Agilent Technologies

Agilent Part Number. 04291-90120
Printed in JAPAN February 2000