

Agilent 4338B Milliohm Meter

Manual Change

Agilent Part No. 04338-90004

May 2003

Printed in Japan

Miscellaneous Changes

1. The option system of the 4338B has changed since May 2003. Apply the following changes.

Old Option Number	New Option Number	remarks
Standard	800 (Standard Milliohm Meter)	
1CM (Rack Mount Kit)	same as the left number	
1CN (Handle Kit)	same as the left number	
AB0 (Taiwan – Chinese localization)	–	1,2
–	AB0 (Taiwan – Chinese localization)	
AB1 (Korea – Korean localization)	–	
–	AB1 (Korea – Korean localization)	
ABA (U.S. – English localization)	–	2
–	ABA (U.S. – English localization)	
ABD (Germany – German localization)	–	1,2
–	ABD (Germany – German localization)	
ABF (France – French localization)	–	
–	ABF (France – French localization)	
ABJ (Japan – Japanese localization)	–	2
–	ABJ (Japan – Japanese localization)	
0BW (Add Service Manual)	same as the left number	
0B0 (Delete Operation Manual)	–	3
0B1 (Add Operation Manual)	–	

1. The Chinese/Korean /German/French operation manual has been edited by translating part of the whole English operation manual. For detailed information on the 4338B, also refer to the English operation manual of the 4338B.
2. In the previous system, the option number is used to choose the language of the operation manual (standard accessory). In the new option system, it is used to add an operation manual (optional accessory) of the language the customer desires.
3. No selection of addition/deletion is required for the operation manual because it is only available as an optional accessory in the new option system.

2. CD-ROM that contains the contents of the 4338B Operation Manual is furnished as a standard content.

© Copyright Agilent Technologies Japan, Ltd. 2003