

3589A-04

HP 3589A Spectrum/Network Analyzer

Serial Numbers:

0000A00000/9999Z99999

Y2K Update Information

To Be Performed By:

Customer

SITUATION:

Product requires operator intervention.

The hp-3589A has an internal clock which provides time and date 'stamping' functions for the display and stored files.

The hp-3589A was designed to provide a day/month/hour time-stamp. It does not have a 'YEAR' tracking feature. You can enter the year/month/day information into the instrument and it is quite happy to keep track of everything up to 23:59 hours on 31 Dec, YR after which it rolls-over to 00:00 hours 01 Jan, YR where YR has not incremented from the last entry.

The non-rolling date stamp does not affect the normal operation of the 3589A instrument. It is simply a date-stamping feature which appears on the display and on stored files and can be edited (changed) by the operator.

SOLUTION/ACTION:

To change the YR date at year end:

1. Press [Special Fctn] [NON-VOL SETUP]
2. Press [TIME HHMM] <time> [ENTER]
3. Press [DATE MMDDYY] <date> [ENTER]

Enter <time> and <date> using the numeric keypad. Use a 24 hour format for <time> and a month-day-year format for <date>. The [ENTER] softkey is displayed when you begin entering either parameter.

EXAMPLES:

8:05 am-

Press [TIME HHMM] 0805 [ENTER]

3:42 pm-

Press [TIME HHMM] 1542 [ENTER]

August 3, 1991-

Press [DATE MMDDYY] 080391 [ENTER]

DATE: July 1998

```

*****
*
*              INFORMATION ONLY
*              -----
*
*  AUTHOR/ENTITY:  ___DWH/A100_____
*

```

* ADDITIONAL INFO: _____ *
*
