

S E R V I C E N O T E

SUPERSEDES: None

5348A Microwave Counter/Power Meter

Serial Numbers: 2924A00101 / 3009A00503

Front Membrane and/or Frame Replacement Kit**Duplicate Service Notes:**

5347A-09

5348A-09

To Be Performed By: Agilent-Qualified Personnel or Qualified Customers**Parts Required:** P/N 05348-67004 Membrane Panel/Front Panel Replacement Kit**Situation:**

The original membrane assembly has not been as reliable as intended. It has been redesigned, but will not fit inside the original frame casting, which had holes at the bottom of the handles which were later plugged with plastic.

Solution/Action:

A replacement kit includes the new membrane panel, and the new casting. Since the only difference in this assembly when used on the 5347A and 5348A is the Model Number label, both labels are included in the kit. If future replacement of the membrane is required, part number 05348-60203 can be ordered.

DATE: 04 May 1994

ADMINISTRATIVE INFORMATION

SERVICE NOTE CLASSIFICATION:					
MODIFICATION RECOMMENDED					
ACTION CATEGORY:	<input type="checkbox"/> IMMEDIATELY <input checked="" type="checkbox"/> ON SPECIFIED FAILURE <input type="checkbox"/> AGREEABLE TIME	STANDARDS:	Labor 3.0 Hour		
LOCATION CATEGORY:	<input checked="" type="checkbox"/> CUSTOMER INSTALLABLE <input type="checkbox"/> ON-SITE <input checked="" type="checkbox"/> SERVICE CENTER	SERVICE INVENTORY:	<input type="checkbox"/> RETURN <input type="checkbox"/> SCRAP <input checked="" type="checkbox"/> SEE TEXT	USED PARTS:	<input type="checkbox"/> RETURN <input checked="" type="checkbox"/> SCRAP <input type="checkbox"/> SEE TEXT
AVAILABILITY:	PRODUCT'S SUPPORT LIFE	AGILENT RESPONSIBLE UNTIL: 04 May 1996			
AUTHOR: CJF	ENTITY: 0200	ADDITIONAL INFORMATION: Installation Note is 05348-90017			