

S E R V I C E N O T E

SUPERSEDES: NONE

54600A Oscilloscope

Duplicate Service Note: 54600A-05, 54601A-05

User and Service Guide, part number 54600-90901 has an error on page 82 regarding output of dc calibrator (Limits)**Situation:**

Manual presently states, Output of dc calibrator Limits: 4.9990 V to 5.0010 V

Solution/Action:

Manual will be changed to read. Output of dc calibrator Limits: 4.990 V to 5.010 V

Correct voltage is specified on page 66. 5.000 +/- 10mv.

DATE: 11 NOVEMBER 1991

ADMINISTRATIVE INFORMATION

SERVICE NOTE CLASSIFICATION:

INFORMATION ONLY

AUTHOR:

NCW

ENTITY:

0800

ADDITIONAL INFORMATION:

Refer to PCO 09-28720