

S E R V I C E N O T E

SUPERSEDES: NONE

54601A Oscilloscope

User and Service Guide, part number 54600-90901 has an error on page 82 regarding output of dc calibrator (Limits)

Duplicate Service Notes:

54600A-05

54601A-05

Situation:

Manual presently states. Output of dc calibrator Limits: 4.9990 V to 5.0010 V

Solution/Action:

Manual will be changed to read. Output of dc calibrator Limits: 4.990 V to 5.010 V

Correct voltage is specified on page 66. 5.000 +/- 10mv

DATE: 11 NOVEMBER 1991

ADMINISTRATIVE INFORMATION

SERVICE NOTE CLASSIFICATION:

INFORMATION ONLY

AUTHOR:

NCW

ENTITY:

0800

ADDITIONAL INFORMATION: