

## S E R V I C E N O T E

SUPERSEDES: NONE

**54610B Oscilloscope****Serial Numbers:** 0000A00000 / US99999999**Internal trigger specification change****Situation:**

The 54610B internal trigger sensitivity specification are as follows:

DC to 100 MHz 0.50 div or 5.0 mV

DC to 500 MHz 1 div or 10 mV

The DC to 100 MHz specification modification is due to customer feedback. Customers wanted to be certain that ECL signals could be used to trigger the 54610B, and the new specification (DC to 100 MHz) will reflect this.

**Solution/Action:**

Performance verification of the 54610B should be changed to reflect the new specifications.

DATE: January 1996

## ADMINISTRATIVE INFORMATION

SERVICE NOTE CLASSIFICATION:

**INFORMATION ONLY**

<b>AUTHOR:</b> CLD	<b>ENTITY:</b> 0840	<b>ADDITIONAL INFORMATION:</b>
-----------------------	------------------------	--------------------------------