# DSO81204B-03 S E R V I C E N O T E

Supersedes: NONE

## 12 GHZ 4 CH Infiniium Oscilloscope

Serial Numbers: My00000000 - MY46001400 Sg00000000 - SG46001400

The purpose of this Service Note is to notify service personnel that the original CPU PCA assembly has been replaced in these affected units. This new CPU PCA assembly is compatible with the old assembly.

Part number D81004-68809 has been replaced by part number 0960-2652.

| Parts Required: |                                     |      |
|-----------------|-------------------------------------|------|
| P/N             | Description                         | Qty. |
| 0960-2562       | CPU PCA assembly - Micro ATX 915GHZ | 1    |

### ADMINISTRATIVE INFORMATION

SERVICE NOTE CLASSIFICATION:

## **INFORMATION ONLY**

AUTHOR: DPM PRODUCT LINE: PL1A

ADDITIONAL INFORMATION:

© AGILENT TECHNOLOGIES, INC. 2007 PRINTED IN U.S.A.





June 1, 2007

#### Situation:

The original CPU PCA assembly D81004-68809 is no longer available from the vendor. The compatible replacement part number is 0960-2652.

#### Solution/Action:

Whenever a replacement CPU PCA assembly for the D81004-68809 is required order part number 0960-2652. This is a direct replacement PCA CPU assembly. This change will be made in future revisions of the Service Guide. This change only applies to the Infiniium Oscilloscopes.