

S E R V I C E N O T E

SUPERSEDES: 8153A11

Lightwave Multimeter Mainframe

Serial Numbers: 0000A00000 / 9999Z99999

Powersupply Board 08153-66503 is now adjusted and tested at factory level.

To Be Performed By: Agilent-Qualified Personnel

Solution / Action:

The replacement powersupply board is adjusted and tested at factory level. In order to avoid ambiguity, the partnumber 08153-68703 has been obseleted. Order 08153-66503 for power-supply boards now.

DATE: April 2000

ADMINISTRATIVE INFORMATION

SERVICE NOTE CLASSIFICATION:		
INFORMATION ONLY		
AUTHOR: JH	ENTITY: OCMD	ADDITIONAL INFORMATION:

