

16505A-01A

Serial Numbers:

0000A00000 / 9999Z99999

Supersedes:

16505A-01 Dated 02 Nov 1998

Y2K update information. Product requires Software update revision 1.40

To Be Performed By: HP-Qualified Personnel, and customer

Parts Required:

HP P/N	Description	Quantity
16505-68714	SW Update CD-ROM v1.40	1

December 1996 Serial Number:

US36420311

Situation:

The HP 16505A Prototype Analyzer User Registration password will yield an "invalid" status after 31 Dec 1999 even though a valid password is entered. Consequently the User Registration window will always appear after the instruments boots.

Otherwise the HP 16505A will operate as expected after 31 Dec 1999, including all visible date-related functions.

Solution / Action:

Operating system software revision 1.40 will be sent free of charge to users who have registered by contacting the HP Password Center. Users who have not registered may request a free copy of the software revision at internet URL

<http://software.col.hp.com/16505A>

Install operating system software revision 1.40 using the procedure provided with the update CD-ROM. A SCSI CD-ROM drive is required to install the update.

Operating system software version 1.40 disables the User Registration utility because HP is not planning any more software revisions for the HP 16505A product.

Date:

27 Aug 1999

```
*****
*
*           MODIFICATION RECOMMENDED           *
*           -----                           *
*
*           STANDARDS: LABOR HOURS _0.5_____ *
* ACTION CATEGORY: ON SPECIFIED FAILURE ___ IMMEDIATELY ___ *
*           AGREEABLE TIME XX                 *
* LOCATION CATEGORY: CUSTOMER INSTALLABLE XX   ON-SITE XX *
*           HP LOCATION XX                   *
*           AVAILABILITY: PRODUCT'S SUPPORT LIFE *
* SERVICE INVENTORY: RETURN ___ SCRAP ___ SEE TEXT ___ *
*           USED PARTS: RETURN ___ SCRAP ___ SEE TEXT ___ *
* HP RESPONSIBLE UNTIL: _02 Nov 2000_         *
*           AUTHOR/ENTITY: _RB/0800_____ *
*           ADDITIONAL INFO: _____ *
*****
```

HP has made every effort to ensure the accuracy of our product testing. However, because each customer's environment is

different from HP's laboratory test environment, it is the customer's responsibility to validate the Year 2000 readiness of these products in their own environment. Therefore, information about the Year 2000 status of HP products is provided "as is" without warranties of any kind and is subject to change without notice. HP makes no representation or warranty respecting the accuracy or reliability of information about non-HP products. Such information, if any, was provided by the manufacturers of those products and customers are urged to contact the manufacturer directly to verify Year 2000 readiness. The information provided here constitutes a Year 2000 Readiness Disclosure for purposes of the Year 2000 Information and Readiness Disclosure Act.