

## IMPORTANT - DO NOT DISCARD

### Agilent 16700A-Series Frame Compatibility with Agilent logic analyzer measurement modules

**Your Frame MAY require an upgrade to operate with  
some Agilent logic analyzer measurement modules**

#### Hardware Upgrade

Frames with the following serial numbers need to be upgraded to operate with 16715A, 16716A, 16717A, 16718A, 16719A, 16740A, 16741A, 16742A, 16750A/B, 16751A/B, 16752A/B, 16753A, 16754A, 16755A, 16756A, or 16760A measurement modules. Frames with the serial numbers above this range have shipped from the factory with the upgrade installed.

Agilent Model Number	Serial Numbers That Require a Frame Upgrade
Agilent 16700A Mainframe	US37030051 through US38490155
Agilent 16701A Expander	US37200051 through US39020113
Agilent 16702A Mainframe	US37380051 through US38490154

Contact your local Agilent Service Center to determine shipment location and to schedule your frame upgrade. Also contact them if you have any questions regarding this program. Reference Interoffice Service Memo (IOSM) 16700-1299-01 when you call.

Service center contact information is found on the World Wide Web at <http://www.agilent.com>.

- At the top of the page select **Contact Us**.
- Select **Test & Measurement**.
- Select the appropriate country.

Note: Your frame's serial number will not be changed during the upgrade. An additional sticker will be added to the frame's back panel designating that the upgrade has been installed. For the 16700A and 16702A frames, the sticker has part number 16700-68711. For the 16701A expander frame, the sticker has part number 16701-68703.

#### Software Upgrade

You must upgrade your Agilent 16700A-series system software to Version A.02.50.00 or higher in order for the Agilent 16715/16/17/18/19/40/41/42/50/51/52/60 measurement modules to be recognized and operate. Version A.02.70.00 or higher is required in order for the Agilent 16753/54/55/56 measurement modules to be recognized and operate. The Version A.02.00.00 CD-ROM (or higher) is included with this module shipment. Refer to the instructions provided with the CD-ROM for the system software update process. The system software is not upgraded during the hardware upgrade.

**Please fill in the following information and include this Installation Note with your frame when you send it in for the upgrade.**

<b>Model Number</b> (Circle one)	Agilent 16700A	Agilent 16700A/ 16701A system	Agilent 16702A	Agilent 16702A/ 16701A system	Agilent 16701A
<b>Serial Number(s)</b>	_____				

