Caution

Initial Registration of E5070B/E5071B

When you start up the E5070B/E5071B for the first time or after executing system recovery, you need to perform the initial registration of the Windows 2000 operating system of the E5070B/E5071B.

You cannot use the front panel keys during the initial registration of the E5070B/E5071B therefore connect the mouse and the keyboard before turning on the power.

For more information on the initial registration, see "Initial Registration of E5070B/E5071B" in Chapter 2 of the E5070B/E5071B Installation and Quick Start Guide.

Agilent Part No. E5070-90008