

## S E R V I C E N O T E

SUPERSEDES: None

**437B Power Meter****Serial Numbers:** 0000A00000 / 3110A99999**Preferred Replacement for A3BT1 and A3MP12.****To Be Performed By:** Agilent-Qualified Personnel**Parts Required:** Battery 1420-0338**Situation:**

The old battery and holder (p/n 1420-0341 and 1400-1465) may develop contact problems after prolonged use.

**Solution/Action:**

P/N 1420-0338 is the direct and preferred replacement for A3BT1 and A3MP12. If A3BT1 or A3MP12 fails, replace both the battery and the holder with a new solder-in-battery Part Number 1420-0338.

DATE: 15 October 1991

## ADMINISTRATIVE INFORMATION

SERVICE NOTE CLASSIFICATION:		
<b>MODIFICATION AVAILABLE</b>		
ACTION CATEGORY:	<b>AGREEABLE TIME</b>	PERFORMANCE ENHANCEMENT <input checked="" type="checkbox"/> SERVICE/RELIABILITY ENHANCEMENT
LOCATION CATEGORY:	<input checked="" type="checkbox"/> CUSTOMER INSTALLABLE <input type="checkbox"/> ON-SITE <input checked="" type="checkbox"/> SERVICE CENTER	AVAILABLE UNTIL: Product's Support Life
AUTHOR: H.J.J.	ENTITY: 0400	ADDITIONAL INFORMATION: