

S E R V I C E N O T E

SUPERSEDES: None

8481A Power Sensor

Serial Numbers: 0000A00000 / 9999Z99999

Restored replacement cartridge will no longer be available.

To Be Performed By: Agilent-Qualified Personnel

Situation:

Due to the high cost of refurbishing this part, the 08481-69009 Restored Cartridge is being obsoleted.

Solution / Action:

Customers should order the 08481-60009 Replacement Cartridge.

DATE: June 2001

ADMINISTRATIVE INFORMATION

SERVICE NOTE CLASSIFICATION:		
INFORMATION ONLY		
AUTHOR: TC	ENTITY: BEST PGU	ADDITIONAL INFORMATION: