

S E R V I C E N O T E

SUPERSEDES: NONE

8990A Peak Power Analyzer

Serial Numbers: 0000A00000 / 9999Z99999

Save A1 control board assembly top and bottom shields**Duplicate Service Notes:**

8990A-01

8991A-01

8992A-01

Situation:

The two top and bottom shields for the channels 1 and 4 inputs are not structured to the A1 control board. They will not be on the replacement board.

Solution/Action:

After removing the A1 Control board assembly: Remove both the top and bottom shields from channels 1 and 4. Save all hardware. Install these shields onto the replacement board prior to installing the board into the unit.

DATE: 23 November 1993

ADMINISTRATIVE INFORMATION

SERVICE NOTE CLASSIFICATION:

INFORMATION ONLY

AUTHOR:

EM

ENTITY:

5300

ADDITIONAL INFORMATION:

