

**E4418B-19**

**S E R V I C E N O T E**

Supersedes:  
None

E4418B with Option 001 only

Serial Numbers: MY45A00000/MY45Z99999

**Battery Assembly**

To Be Performed By: Agilent-Qualified Personnel or Customer

**Parts Required:**

P/N	Description	Qty.
E4418-61026	Battery Assy	1

ADMINISTRATIVE INFORMATION

SERVICE NOTE CLASSIFICATION:		
<b>MODIFICATION AVAILABLE</b>		
ACTION CATEGORY:	AGREEABLE TIME	<input checked="" type="checkbox"/> PERFORMANCE ENHANCEMENT <input type="checkbox"/> SERVICE / RELIABILITY ENHANCEMENT
LOCATION CATEGORY:	<input checked="" type="checkbox"/> CUSTOMER INSTALLABLE <input type="checkbox"/> ON-SITE <input checked="" type="checkbox"/> SERVICE CENTER	AVAILABILITY: NOVEMBER 30, 2008
AUTHOR: CFL    PRODUCT LINE: WC		
ADDITIONAL INFORMATION:		

© AGILENT TECHNOLOGIES, INC. 2007  
PRINTED IN U.S.A.



December 6, 2007

**Situation:**

There will be a tight fit when performing the battery assembly in EPM series power meters option 001. Excessive force might be needed for the battery pack installation. However, it will not damage the power meter and the battery pack.

**Solution/Action:**

Please insert the battery pack at the correct direction as below. For the ease of pulling out the battery pack from the meter, users can use some tools such as flat head screw drivers. This must be done slowly and with care.

