

S E R V I C E N O T E

SUPERSEDES: 8664A-01

8664A Synthesized signal generator

Serial Numbers: 0000A00000 / 3035A00185

I/O assembly replacement requires modification to reference assembly

Duplicate Service Notes:

8665A-09 Serial Numbers: 0000A00000 / 3026A00585

8665B-07 Serial Numbers: 0000A00000 / 3020A00192

To Be Performed By: Agilent-Qualified Personnel

Situation:

When an I/O assembly (08665-69104, 08665-60204 or 08665-60304) is replaced in instruments (serial numbers noted above) a jumper on the Reference assembly (P/N 1258-0209) must be removed to ensure proper operation of the reference oscillator.

NOTE:

Billings related to replacement of the I/O or Reference assemblies are not covered by this service note.

Solution / Action:

Remove J17 on the I/O assembly. This jumper is located at the front edge of the I/O assembly near the J11 Coax Discriminator connector.

DATE: August 1998

ADMINISTRATIVE INFORMATION

| | | | |
|---------------------------------|---|---|---|
| SERVICE NOTE CLASSIFICATION: | | | |
| MODIFICATION RECOMMENDED | | | |
| ACTION CATEGORY: | <input type="checkbox"/> IMMEDIATELY <input checked="" type="checkbox"/> ON SPECIFIED FAILURE <input type="checkbox"/> AGREEABLE TIME | STANDARDS: | LABOR 0.1 Hours |
| LOCATION CATEGORY: | <input type="checkbox"/> CUSTOMER INSTALLABLE <input type="checkbox"/> ON-SITE <input checked="" type="checkbox"/> SERVICE CENTER | SERVICE INVENTORY: | <input type="checkbox"/> RETURN <input checked="" type="checkbox"/> SCRAP <input type="checkbox"/> SEE TEXT |
| AVAILABILITY: | PRODUCT'S SUPPORT LIFE | USED PARTS: | <input type="checkbox"/> RETURN <input checked="" type="checkbox"/> SCRAP <input type="checkbox"/> SEE TEXT |
| AUTHOR: MW | ENTITY: 5320 | AGILENT RESPONSIBLE UNTIL: August 1, 2003 | |
| | | ADDITIONAL INFORMATION: None | |