8664A-09

S	Е	R	V	Ι	С	Е	Ν	0	Т	Е
							SUPERSE	DES: Non	e	
8	8664A Synthesized Signal Generator									
Serial Numbers: 0000A00000 / 9999Z99999										
New power switch part number										
Γ	Duplicate Service Notes:									
8 8 8 8 8	643A-05 644A-06 644B-06 645A-07 664A-09 665A-10 665B-08									
Т	To Be Performed By: Agilent-Qualified Personnel									
S	Situation:									
	The old part number (3101-2720) for the power switch on these instruments has been changed in production for one with increased reliability.									
S	Solution / Action:									
C	Order part number 83750-60040 instead.									
	DATE: March 1999									

ADMINISTRATIVE INFORMATION

SERVICE NOTE CLASSIFICA	TION:							
INFORMATION ONLY								
AUTHOR:	ENTITY:	ADDITIONAL INFORMATION:						
MW	5320							
© 1999 AGILENT TEC PRINTED IN U.S.A.	HNOLOGIES	Agilent Technologies						

1 of 1