

INFORMATION ONLY – DOES NOT COMMUNICATE
A MODIFICATION OR SAFETY CONDITION

E8257D-05

S E R V I C E N O T E

Supersedes:
None

E8257D PSG Analog Signal Generator

Serial Numbers: [0000A00000 / 9999Z99999]

A7 Reference Part Number (Non-UNR)

Parts Required:

P/N	Description	Qty.
None		

ADMINISTRATIVE INFORMATION

SERVICE NOTE CLASSIFICATION: INFORMATION ONLY
AUTHOR: DMc PRODUCT LINE: PL15
ADDITIONAL INFORMATION:

© AGILENT TECHNOLOGIES, INC. 2005
PRINTED IN U.S.A.



July 22, 2005

Situation:

The Service Guide replacement parts section lists the A7 Reference for non-UNR instruments as part number E8251-60073. When shipped from the factory the E82x7D instruments may have an E4423-60018 A7 Reference. The goal was to use the E4423-60018 reference assembly in both the E82x7D and E44xxB models. Because of slight differences in the way header data loads when the board is installed in each instrument, a replacement assembly with data loaded in one model will not work in the other model.

Solution/Action:

When replacing A7 Reference in non-UNR instruments, order the part number in the manual. At the time this service note was written the standard reference replacement part number was E8251-60073. Check the service guide for the correct replacement part number.