

INFORMATION ONLY – DOES NOT COMMUNICATE
A MODIFICATION OR SAFETY CONDITION

E8257D-13

S E R V I C E N O T E

Supersedes:
None

E8257D Analog Signal Generator

Serial Numbers: [0000A00000 / 9999Z99999]

The correct part number when replacing E4423-60018 boards (standard reference boards) in the field.

Parts Required:

P/N	Description	Qty.
none		

ADMINISTRATIVE INFORMATION

SERVICE NOTE CLASSIFICATION:

INFORMATION ONLY

AUTHOR: DMc PRODUCT LINE: PL15

ADDITIONAL INFORMATION:

© AGILENT TECHNOLOGIES, INC. 2006
PRINTED IN U.S.A.



Agilent Technologies

February 15, 2006

Situation:

The E4423-60018 reference board is used by several products families including the E82x7D models as the standard board. This is possible because the specific board data is downloaded during production into the board once the board is installed in the instrument. But data stored for one instrument family does not work in other instruments. Therefore, the standard reference board (E4423-60018) replacement part number is not the same part number as the installed part number (E8251-60073).

Solution/Action:

When replacing any assembly, always check the online service guide for the correct replacement part number. Currently, the replacement part number for E4423-60018 in E82x7D products is E8251-60073.