

S E R V I C E N O T E

SUPERSEDES: None

8562A Portable Spectrum Analyzer

Serial Numbers: 0000A00000 / 3043A00535

YTF Replacement for Option T01**Situation:**

The standard A10 Yig Tuned Filter (YTF) was shipped in approximately the first 50 option T01 instruments. The part number for the option T01 YTF has been changed from 0955-0277 TO 0955-0521. The standard YTF will work as the replacement in most instruments, but not all.

The YTF and the mixer both provide some LO isolation. There are some combinations of these two assemblies that will not provide enough isolation to satisfy the instrument's conducted LO requirements. Using the new YTF will keep this from happening.

Solution/Action:

When replacing the YTF in any option T01 instruments, the new YTF, part number 0955-0521, should be used.

DATE: 23 October 1992

ADMINISTRATIVE INFORMATION

SERVICE NOTE CLASSIFICATION:

INFORMATION ONLY

AUTHOR: MLG	ENTITY: 5300	ADDITIONAL INFORMATION:
----------------	-----------------	-------------------------