


## Agilent RF Preselector, 20 Hz to 2 GHz For the 8566B or 8568B Spectrum Analyzer Data Sheet

### Your High-Performance Spectrum Analyzer can also be a Test Receiver

Add the 85685A RF preselector to the 8566B or 8568B general-purpose spectrum analyzer, and you have a test receiver for specialized applications. The RF preselector enhances these high-performance spectrum analyzers with tracking filters and preamplifiers that cover the 20 Hz to 2 GHz range. The spectrum analyzer/test receiver is sensitive to low-level signals while providing overload protection from out-of-band signals. The result, for tests in the presence of high-level interference, is a measurement range 30 dB greater than that of the spectrum analyzer alone. (The improvement is even greater in the presence of broadband pulse-interference.)

The test receiver is easy to operate; simply use the spectrum analyzer controls without concern for preselector settings. The preselector automatically adjusts input-filter tracking, and the spectrum analyzer reports preselector operating conditions on the CRT display. For remote operation with a computer, the receiver system is fully GPIB programmable (GPIB is hardware, software, documentation, and support for IEEE-488 and IEC-625, worldwide standards for interfacing instruments.)



The 85685A preselector adds:

- Automatic filter tracking
- Input overload protection
- Low system noise

to the 8566B and 8568B spectrum analyzers.




**Agilent Technologies**

## EMI Testing According to CISPR Publication 16 Recommendations

Combining the 85685A RF preselector with the 8566B or 8568B spectrum analyzer and the 85650A quasi-peak adapter produces an EMI receiver with the characteristics recommended in CISPR publication 16<sup>1</sup>. The RF preselector provides the measurement sensitivity and overload protection needed for FCC and VDE radiated emission testing at open sites. For commercial and MIL-STD conducted EMI tests, the low-frequency input tolerates large impulses and line impedance stabilization network (LISN) transients. A built-in calibrator ensures  $\pm 2.0$  dB absolute-amplitude accuracy as required by the FCC and VDE. For measurement confidence, a linearity check tests for system overload and distortion.

The sample display shows results for VDE 0871 emission testing at an open site. To prevent overload from the high-level ambient signals, a spectrum analyzer needs input attenuation. This however, increases displayed noise, which can mask the low-level emissions. Using the RF preselector in this test eliminates the need for adding attenuation. The noise of the spectrum analyzer, as shown, exceeds the VDE limit, but the spectrum analyzer/test receiver noise is more than 10 dB below the lowest limit.

1. Comite International Special Des Perturbations Radio-Electriques (CISPR) publication 16 is the "CISPR specification for radio interference measuring apparatus and measurement methods."



### Signal Monitoring in the Presence of High-Level Interference

The RF preselector gives the 8566B or the 8568B spectrum analyzer the sensitivity and selectivity of a receiver. The spectrum analyzer/test receiver measures signals within the preselector filter passband and rejects out-of-band interference by 40 dB. Even in the presence of interference with a total input power of +10 dBm, your test receiver can monitor signals down to -150 dBm.

The RF preselector improves signal reception for the broadband signal environment of a mobile test station. In the 150 – 170 MHz business band, the RF preselector reduces interference from radio and TV broadcast stations and from amateur radios. For the 800 – 900 MHz cellular radio band, the spectrum analyzer/test receiver makes field strength measurements on cellular base-station transmitters while rejecting interference from cellular mobile transmitters.

### Transmitter Spurious Testing with Fast Wideband Measurements

Reduce transmitter-spurious test time by adding the RF preselector to your 8566B or 8568B spectrum analyzer. The preselector decreases input overload from the out-of-band carrier and thereby increases the range for measuring low-level signals. Compared to the spectrum analyzer alone, the spectrum analyzer/test receiver uses a wider resolution bandwidth for the same measurement range, resulting in a faster sweep time. For example, to measure spurious signals over a 1 GHz frequency range at -100 dB relative to the transmitter carrier, a spectrum analyzer by itself takes 3000 seconds, but when configured as a test receiver it requires only 3 seconds.

## Specifications

Specifications describe the instrument's warranted performance over the temperature range of 0°C to +55°C unless otherwise noted. Typical values for the specified parameters represent probable but non-warranted performance. Nominal values provide useful but non-warranted information about functional performance.

### Test Receiver

The following system specifications apply when the 85685A RF preselector (RFP) operates with the 8566B or 8568B spectrum analyzer (SA) and with or without the 85650A quasi-peak adapter (QPA). The test receiver system is fully GPIB programmable. Refer to the specific instrument data sheets for more detailed information.


Test receiver frequency	
Parameter	Specification
<b>Frequency range</b>	20 Hz – 1.5 GHz with 8568B 20 Hz – 2.0 GHz with 8566B
RFP bypass <sup>1</sup>	20 Hz – 18 GHz (nominal) with 8568B
SA frequency resolution	1 Hz
SA frequency accuracy	$\pm(\text{tuned frequency} \times \text{freq. reference error})$ Hz for SA zero span
SA frequency reference error	Stability <sup>2</sup> $< 10^{-8}$ , Drift $< 2 \times 10^{-7}$ /year
<b>SA Frequency selectivity<sup>3</sup></b> 3 dB resolution BW	10 Hz – 3 MHz in 1, 3, 10 sequence

Test receiver amplitude	
Parameter	Specification
Measurement range	-150 to +30 dBm (-43 to +137 dB $\mu$ V) (nominal)
<b>Gain compression (0 dB atten. for RFP and SA)</b> Inside RFP passband Outside RFP passband	$< 1$ dB (nominal) for input $\leq -30$ dBm $< 1$ dB (nominal) for input $\leq +10$ dBm
<b>Displayed average noise level</b> (0 dB atten. for RFP and SA 10 Hz resolution BW)	$< -132$ dBm (-25 dB $\mu$ V) for input 9 kHz – 1 MHz <sup>4</sup> $< -150$ dBm (-43 dB $\mu$ V) for input 1 – 1500 MHz $< -147$ dBm (-40 dB $\mu$ V) for input $\leq 2.0$ GHz
<b>Broadband signal sensitivity<sup>5,6</sup></b> (0 dB atten. for RFP and SA, 10 Hz resolution BW)	$< 31$ dB $\mu$ V/MHz (nominal) for input $< 1$ MHz <sup>4</sup> $< 13$ dB $\mu$ V/MHz (nominal) for input = 1 – 1500 MHz $< 16$ dB $\mu$ V/MHz (nominal) for input $\leq 2.0$ GHz
<b>Second harmonic distortion</b> (0 dB atten. for RFP and SA) For -35 dBm input signal: Within RFP passband Outside RFP passband	$< -55$ dBc (nominal), Equiv. SOI = +20 dBm $< -75$ dBc (nominal), Equiv. SOI = +40 dBm
<b>Third order intermodulation</b> (0 dB atten. for RFP and SA) For two -35 dBm input signals: Within RFP passband Outside RFP passband	$< -40$ dBc (nominal), Equiv. SOI = -15 dBm $< -110$ dBc (nominal), Equiv. SOI = +20 dBm
<b>Residual responses</b> (RFP input terminated, 0 dB atten. for RFP and SA)	$< -90$ dB for input 2 kHz – 1 MHz $< -112$ dB for input $\geq 1$ MHz $< -120$ dB (typical) for input $\geq 1$ MHz
<b>Amplitude accuracy<sup>6</sup></b> (+10° to + 30°C, 10 dB SA atten. 0 – 20 dB RFP atten.)  RFP bypass	$\pm 2.0$ dB for input $< 1.0$ GHz $\pm 3.0$ dB for input $\geq 1.0$ GHz  <b>SA specification plus:</b> $\pm 0.5$ dB (nominal) for input $\leq 2.0$ GHz $\pm 1.0$ dB (nominal) for input $> 2.0$ GHz

## CISPR EMI Receiver

The following system specifications apply when the 85685A RF preselector (RFP) operates with the 8566B or 8568B spectrum analyzer (SA) and with the 85650A quasi-peak adapter (QPA). The EMI receiver system has the characteristics recommended in CISPR publication 16 and is fully GPIB programmable. Refer to the specific instrument data sheets for more information.


CISPR EMI receiver frequency	
Parameter	Specification
Frequency range	10 kHz – 1 GHz
SA frequency resolution	1 Hz
SA frequency accuracy	$\pm(\text{tuned frequency} \times \text{freq. reference error}) \text{ Hz}$
SA frequency reference error	Stability <sup>2</sup> < $10^{-8}$ , Drift < $2 \times 10^{-7}$ /year
<b>SA Frequency selectivity<sup>9</sup></b> CISPR 6 dB bandwidth (QPA)	200 Hz for input = 10 – 150 kHz (band A) 9 kHz for input = 150 kHz – 30 MHz (band B) 120 kHz for input = 300 – 1000 MHz (band C, D)



**Make dedicated receiver measurements with a general-purpose spectrum analyzer**

CISPR EMI Amplitude	
Parameter	Specification
Measurement range	+137 dB $\mu$ V to noise floor (refer to amplitude sensitivity table)
<b>Amplitude accuracy<sup>8</sup></b> (+10° to +30°C, 0 – 20 dB RFP atten.)	$\pm 2.0$ dB for 10 dB SA atten. $\pm 2.5$ dB for 0 dB SA atten.
Calibration <sup>10</sup>	Automatic calibration for absolute amplitude with comb generator (RFP)
Linearity check	3 dB input attenuator test for system overload and distortion (RFP)
Impulsive signal tolerance	Input #1 withstands LISN transients (RFP) Input #2 withstands CISPR calib. pulses (RFP)
Detection <sup>9</sup>	Peak and average (SA), quasi-peak (QPA)
Audio output	Speaker with volume control (QPA)

CISPR EMI Amplitude sensitivity			
EMI receiver		Specification	
Frequency band	Resolution bandwidth (6 dB)	Average noise <sup>6</sup> (RFP/SA atten. = 0 dB)	Quasi-peak noise (RFP/SA atten. = 0 dB)
10 – 150 kHz (band A)	200 Hz	-12 dB $\mu$ V <sup>4</sup> -17 dB $\mu$ V typical	-15 dB $\mu$ V (nominal)
150 kHz – 1 MHz (band B)	9 kHz	+4 dB $\mu$ V -1 dB $\mu$ V typical	0 dB $\mu$ V (nominal)
1 – 30 MHz (band B)	9 kHz	-13 dB $\mu$ V -16 dB $\mu$ V typical	-11 dB $\mu$ V (nominal)
30 – 1000 MHz (band C, D)	120 kHz	-1 dB $\mu$ V -6 dB $\mu$ V typical	0 dB $\mu$ V (nominal)



## 85685A RF Preselector

The following instrument specifications apply to the 85685A RF preselector (RFP).

85685A RF preselector input		
Parameter	Specification	
	Input #1	Input #2
<b>Frequency range</b> RFP bypass <sup>1</sup>	20 Hz – 50 MHz	20 MHz – 2 GHz DC – 18 GHz (nom.)
<b>Connector type</b> Fuse blow time	BNC, 50 (nominal) < 0.1 sec for > + 35 dBm	Type N, 50 (nom.) NA
<b>Maximum safe input power</b> Average power Impulsive signals ≥ 10 dB RFP atten. ≥ 20 dB RFP atten.	+30 dBm (1 Watt)  100 Watts peak for a 10 μs pulse 2,000 Watts peak for a 10 μs pulse <sup>11</sup> (nom.)	+30 dBm (1 Watt)  NA
DC voltage	0 Volts	
<b>Standing wave ratio (SWR)</b> <sup>12</sup> 10 dB RFP atten. 0 dB RFP atten. RFP bypass	< 1.5:1 < 1.5:1 (nominal) NA	< 1.5:1 < 2.0:1 (nominal) < 1.5:1 (nominal)
<b>Attenuator range</b> Linearity check	0 – 50 dB in 10 dB steps (nominal) Add 3 dB (nominal)	
<b>Filter characteristics</b> Fixed filter BW (3 dB) Tracking filter BW (3 dB)  Stopband attenuation	BW range = 10 kHz• 1 MHz (nom.) for input < 2 MHz 10% (nom.) of tuned freq. for input = 2 – 55 MHz 5% (nom.) of tuned freq. for input = 55 – 500 MHz 20 MHz (nom.) for input = 500 – 2000 MHz > 40 dB (nominal)	
Preamplifier gain	20 dB (nominal) for 0 dB RFP atten.	

85685A RF preselector output	
Parameter	Specification
<b>Comb generator</b> Line spacing Line amplitude SWR	100 kHz, 500 kHz, 1 MHz, 5 MHz (nominal) -40 to -60 dBm (nominal) < 1.5:1 (nominal)
<b>GPIB interface*</b> Private BUS Public BUS interface functions (public BUS)	Digital BUS and sweep control between RFP and SA IEEE 488-1978 BUS SH1, AH1, T6, L4, SR1, RL1, PPO, DC1, DT1, C0, E1



<b>General</b>	
<b>Parameter</b>	<b>Specification</b>
Temperature	0° to + 55°C operating -40° to +75°C storage
Warm-up time	30 minutes after cold start
Temp. equilibrium	2 hours after warm-up
EMI	Conducted and radiated interference is in compliance with MIL-STD 461A methods in CE03 and RE02, CISPR publication 11 (1975), and Messempefaenger-Postverfuegung 526/527/79 (Kennzeichnung Mit F-Nummer/Funkschutzzeichen).
Power requirements	100, 120, 220, or 240 V RMS +5%, -10%; 50 –400 Hz; 35 VA (nominal)
Weight	19 kg (42 lbs) net (nominal)
Size	133 mm (5.25") H x 425 mm (16.75") W x 508 mm (20.0") D (nominal)
<b>Firmware compatibility</b> for RFP operation with:	
8566B, firmware data $\geq$ 5.5.85	Compatible
8566B, firmware data < 5.5.85	85685A Option 066 required
8566A, any firmware date	8566A + 01K retrofit kit required
8568B, firmware data $\geq$ 14.1.85	Compatible
8568B, firmware data < 14.1.85	85685A Option 068 required
8568A, any firmware date	8568A + 01K retrofit kit required

## Product Information

### 85685A RF preselector

- Option 010, Rack mount slide kit
- Option 908, Rack flange kit without handles
- Option 910, Extra manual
- Option 913, Rack flange kit with handles

### Firmware compatibility options (see compatibility specifications above)

- 85685A Option 066, Firmware upgrade kit\* for 8566B
- 85685A Option 068, Firmware upgrade kit\* for 8568B
- 8566A + 01K Retrofit kit
- 8568A + 01K Retrofit kit
- 8566B spectrum analyzer
- 8568B spectrum analyzer
- 85650A Quasi-peak adapter

### Accessories

- 85864B EMI measurement software
- Transit case, part number 1540-0663
- EMI accessories<sup>13</sup>: Antennas, LISNs, current probes, and amplifiers

## Footnotes:

1. The RFP bypass switch connects input #2 directly to the RF output.
2. Refer to spectrum analyzer test and adjustment manual for the correct frequency reference adjustment.
3. Refer to spectrum analyzer data sheet for filter BW accuracy and filter shape.
4. 17 dB higher with 8566B for frequency < 50 kHz.
5. Sensitivity is approximately equal to system peak noise level.
6. Peak noise is nominally 8 dB higher than average noise.
7. Measurement sensitivity is normalized to 1 MHz resolution BW.
8. Specifications apply after system calibration with the comb generator in the 85685A. Refer to 85685A operating manual for correct amplitude measurement technique.
9. The system meets CISPR publication 16 recommendations for resolution BW, detector time constant, and detector pulse response.
10. System calibration can be verified with an external CISPR standard pulse calibration generator.
11. Typical LISN transients.
12. SWR is measured at the tuned frequency of the spectrum analyzer.
13. Refer to the EMI measurement solutions guide (5954-2702) for more information.



## Agilent Email Updates

[www.agilent.com/find/emailupdates](http://www.agilent.com/find/emailupdates)

Get the latest information on the products and applications you select.



## Agilent Open

[www.agilent.com/find/open](http://www.agilent.com/find/open)

Agilent Open simplifies the process of connecting and programming test systems to help engineers design, validate and manufacture electronic products. Agilent offers open connectivity for a broad range of system-ready instruments, open industry software, PC-standard I/O and global support, which are combined to more easily integrate test system development.



## Agilent Direct

[www.agilent.com/find/agilentdirect](http://www.agilent.com/find/agilentdirect)

Quickly choose and use your test equipment solutions with confidence.

## www.agilent.com

### Agilent Technologies' Test and Measurement Support, Services, and Assistance

Agilent Technologies aims to maximize the value you receive, while minimizing your risk and problems. We strive to ensure that you get the test and measurement capabilities you paid for and obtain the support you need. Our extensive support resources and services can help you choose the right Agilent products for your applications and apply them successfully. Every instrument and system we sell has a global warranty. Two concepts underlie Agilent's overall support policy: "Our Promise" and "Your Advantage."

### Our Promise

Our Promise means your Agilent test and measurement equipment will meet its advertised performance and functionality. When you are choosing new equipment, we will help you with product information, including realistic performance specifications and practical recommendations from experienced test engineers. When you receive your new Agilent equipment, we can help verify that it works properly and help with initial product operation.

### Your Advantage

Your Advantage means that Agilent offers a wide range of additional expert test and measurement services, which you can purchase according to your unique technical and business needs. Solve problems efficiently and gain a competitive edge by contracting with us for calibration, extra-cost upgrades, out-of-warranty repairs, and onsite education and training, as well as design, system integration, project management, and other professional engineering services. Experienced Agilent engineers and technicians worldwide can help you maximize your productivity, optimize the return on investment of your Agilent instruments and systems, and obtain dependable measurement accuracy for the life of those products.

### Agilent T&M Software and Connectivity

Agilent's Test and Measurement software and connectivity products, solutions and developer network allows you to take time out of connecting your instruments to your computer with tools based on PC standards, so you can focus on your tasks, not on your connections. Visit [www.agilent.com/find/connectivity](http://www.agilent.com/find/connectivity) for more information.

For more information on Agilent Technologies' products, applications or services, please contact your local Agilent office.

### Phone or Fax

#### United States:

(tel) 800 829 4444  
(fax) 800 829 4433

#### Canada:

(tel) 877 894 4414  
(fax) 800 746 4866

#### China:

(tel) 800 810 0189  
(fax) 800 820 2816

#### Europe:

(tel) 31 20 547 2111

#### Japan:

(tel) (81) 426 56 7832  
(fax) (81) 426 56 7840

#### Korea:

(tel) (080) 769 0800  
(fax) (080)769 0900

#### Latin America:

(tel) (305) 269 7500

#### Taiwan:

(tel) 0800 047 866  
(fax) 0800 286 331

#### Other Asia Pacific

#### Countries:

(tel) (65) 6375 8100  
(fax) (65) 6755 0042

Email: [tm\\_ap@agilent.com](mailto:tm_ap@agilent.com)

Contacts revised: 05/05/05

The complete list is available at:

[www.agilent.com/find/contactus](http://www.agilent.com/find/contactus)

Product specifications and descriptions in this document subject to change without notice.

© Agilent Technologies, Inc. 2005  
Printed in USA, October 5, 2005  
5954-2708



Agilent Technologies