

Installation Note

Agilent E4406A VSA Transmitter Tester Handle Removal Kit



Agilent Part Number E4406-90163

Printed in USA November 2000

Notice.

The information contained in this document is subject to change without notice.

Agilent Technologies makes no warranty of any kind with regard to this material, including but not limited to, the implied warranties of merchantability and fitness for a particular purpose. Agilent Technologies shall not be liable for errors contained herein or for incidental or consequential damages in connection with the furnishing, performance, or use of this material.

Description

NOTE This installation note applies to instruments with serial numbers US3845010 through US40061423, with the following exceptions:

US40061107

US40061362

US40061376

US40061383

US40061388

US40061389

As part of its quality monitoring program, Agilent Technologies has discovered that handles located on the left side of the instrument (when viewed from the front panel) may separate from the chassis when exposed to irregular handling conditions. Irregular handling conditions are situations where forces exerted to the handle exceed 3.5 times the instrument's weight. This discovery was made during a laboratory test only. No customer has reported any problem with the handles under any condition.

The handle removal kit prevents the installation of a handle on the left side of the instrument.

Kit Contents

Item	Part Number	Description
1	E4406-90163	Installation Note
2	E4406-80040	Adhesive Hole Covers (2)
3	0515-4301	SMM 5.0 Screw (1)

Equipment Required

Item	Part Number	Description
1	8710-1615	T-20 Torx Driver

Installation Procedure

1. Remove the left-side strap handle from instruments that fall within the serial prefixes listed on page 3.
2. Use the T-20 Torx driver to insert the SMM 5.0 screw (0515-4301) into the rear handle mounting hole. This is the hole located closest to the rear frame. This safety precaution ensures that another handle is not installed on the left-side of the instrument
3. Cover the handle mounting holes with two adhesive covers (E4406-80040).

NOTE For more information about this service issue please visit our web site by following these steps:

1. Access our web site at http://www.agilent.com/find/tm_servicenotes.
 2. Enter "E4406A"
 3. Click "search"
 4. Click "E4406A-06-S"
-

Agilent Technologies Sales and Services Offices

UNITED STATES		
<p>Instrument Support Center Agilent Technologies Company (800) 403-0801</p>		
EUROPEAN FIELD OPERATIONS		
<p>Headquarters Agilent Technologies S.A. 150, Route du Nant-d'Avril 1217 Meyrin 2/ Geneva Switzerland (41 22) 780.8111</p>	<p>France Agilent Technologies France 1 Avenue Du Canada Zone D'Activite De Courtaboeuf F-91947 Les Ulis Cedex France (33 1) 69 82 60 60</p>	<p>Germany Agilent Technologies Technologies GmbH Agilent Technologies Strasse 61352 Bad Homburg v.d.H Germany (49 6172) 16-0</p>
<p>Great Britain Agilent Technologies Ltd. Eskdale Road, Winnersh Triangle Wokingham, Berkshire RG41 5DZ England (44 118) 9696622</p>		
INTERCON FIELD OPERATIONS		
<p>Headquarters Agilent Technologies Company 3495 Deer Creek Rd. Palo Alto, CA 94304-1316 USA (415) 857-5027</p>	<p>Australia Agilent Technologies Australia Ltd. 31-41 Joseph Street Blackburn, Victoria 3130 (61 3) 895-2895</p>	<p>Canada Agilent Technologies (Canada) Ltd. 17500 South Service Road Trans-Canada Highway Kirkland, Quebec H9J 2X8 Canada (514) 697-4232</p>
<p>Japan Agilent Technologies Japan, Ltd. Measurement Assistance Center 9-1, Takakura-Cho, Hachioji-Shi, Tokyo 192-8510, Japan TEL (81) -426-56-7832 FAX (81) -426-56-7840</p>	<p>Singapore Agilent Technologies Singapore (Pte.) Ltd. 150 Beach Road #29-00 Gateway West Singapore 0718 (65) 291-9088</p>	<p>Taiwan Agilent Technologies Taiwan 8th Floor, H-P Building 337 Fu Hsing North Road Taipei, Taiwan (886 2) 712-0404</p>
<p>China China Agilent Technologies Co. 38 Bei San Huan X1 Road Shuang Yu Shu Hai Dian District Beijing, China (86 1) 256-6888</p>		