

# Programming Code Compatibility Suite Guide - Option 266

## Agilent Technologies PSA Spectrum Analyzers

This manual provides documentation for the following instruments  
with Option 266 Installed:

### PSA Series

E4440A (3 Hz - 26.5 GHz)

E4443A (3 Hz - 6.7 GHz)

E4445A (3 Hz - 13.2 GHz)

E4446A (3 Hz - 44.0 GHz)

E4447A (3 Hz - 42.98 GHz)

E4448A (3 Hz - 50.0 GHz)



**Agilent Technologies**

**Manufacturing Part Number: E4440-90352**

**Supersedes: E4440-90316**

**Printed in USA**

**May 2007**

© Copyright 2001 – 2007 Agilent Technologies, Inc.

---

## **Notice**

The information contained in this document is subject to change without notice.

Agilent Technologies makes no warranty of any kind with regard to this material, including but not limited to, the implied warranties of merchantability and fitness for a particular purpose. Agilent Technologies shall not be liable for errors contained herein or for incidental or consequential damages in connection with the furnishing, performance, or use of this material.

---

## **Where to Find the Latest Information**

Documentation is updated periodically. For the latest information about Agilent PSA Spectrum Analyzers, including firmware upgrades, software upgrades, application information, and product information, please visit the Internet URL listed below:

<http://www.agilent.com/find/psa>

<b>1. Getting Started</b>	
Option 266 Description	26
Option 266 Limitations	26
Hardware and Firmware Requirements for Option 266	28
Installing Option 266 - PSA Series Analyzers Only	31
Installing Optional Measurement Personalities	31
Configuring Option 266 on PSA Analyzers	38
The Configure Remote Lang Screen Menu - PSA Analyzers	41
Running Software that Requires SCPI Commands	46
Service and Calibration	48
Documentation for Option 266	49
Spectrum Analyzers with Option 266	49
Spectrum Analyzer Updates	49
<b>2. Legacy Analyzer Command List</b>	
Table of All Legacy Analyzer Commands	52
<b>3. Hints and Tips</b>	
A Few Helpful Hints and Tips	84
<b>4. Programming Commands</b>	
Command Syntax	88
Programming Command Descriptions	90
A1 [one]	
Clear Write for Trace A	91
Syntax	91
Description	91
A2 [two]	
Maximum Hold for Trace A	92
Syntax	92
Description	92
A3 [three]	
View Mode for Trace A	93
Syntax	93
Description	93
A4 [four]	
Blank Trace A	94
Syntax	94
Description	94
ACP	
Adjacent Channel Power	95
Syntax	95
Description	95
ACPALPHA	
Adjacent Channel Power Alpha Weighting	96
Syntax	96
Description	96
ACPALTCH	

---

# Contents

Adjacent Channel Power Alternate Channels	.97
Syntax	.97
Description	.97
<b>ACPBRPER</b>	
Adjacent Channel Power Burst Period	.98
Syntax	.98
Description	.98
<b>ACPBRWID</b>	
Adjacent Channel Power Burst Width	.99
Syntax	.99
Description	.99
<b>ACPBW</b>	
Adjacent Channel Power Bandwidth	.100
Syntax	.100
Description	.100
<b>ACPCOMPUTE</b>	
Adjacent Channel Power Compute	.101
Syntax	.101
Description	.101
<b>ACPFRQWT</b>	
Adjacent Channel Power Frequency Weighting	.102
Syntax	.102
Description	.102
<b>ACPLOWER</b>	
Lower Adjacent Channel Power	.103
Syntax	.103
Description	.103
<b>ACPMAX</b>	
Maximum Adjacent Channel Power	.104
Syntax	.104
<b>ACPMEAS</b>	
Measure Adjacent Channel Power	.105
Syntax	.105
Description	.105
<b>ACPMSTATE</b>	
Adjacent Channel Power Measurement State	.106
Syntax	.106
Description	.106
<b>ACPPAR</b>	
Adjacent Channel Power Manual or Auto	.107
Syntax	.107
Description	.107
<b>ACPPWRTX</b>	
Adjacent Channel Power Total Power Transmitted	.108
Syntax	.108
Description	.108
<b>ACPRSLTS</b>	
Adjacent Channel Power Measurement Results	.109
Syntax	.109

Description .....	109
ACPSP	
Adjacent Channel Power Channel Spacing .....	110
Syntax .....	110
Description .....	110
ACPT	
Adjacent Channel Power T Weighting .....	111
Syntax .....	111
Description .....	111
ACPUPPER	
Upper Adjacent Channel Power .....	112
Syntax .....	112
Description .....	112
ADJALL	
LO and IF Adjustments .....	113
Syntax .....	113
Description .....	113
AMB	
A minus B into A .....	114
Syntax .....	114
Description .....	114
AMBPL	
(A minus B) plus Display Line into A .....	115
Syntax .....	115
Description .....	115
ANNOT	
Annotation .....	117
Syntax .....	117
Description .....	117
APB	
Trace A Plus Trace B to A .....	118
Syntax .....	118
Description .....	118
AT	
Input Attenuation .....	119
Syntax .....	119
Description .....	120
AUNITS	
Absolute Amplitude Units .....	121
Syntax .....	121
Description .....	121
AUTO	
Auto Couple .....	123
Syntax .....	123
Description .....	123
AUTOCP	
Auto Coupled .....	125
Syntax .....	125
Description .....	125

---

# Contents

AXB	
Exchange Trace A and Trace B	126
Syntax	126
Description	126
B1 [one]	
Clear Write for Trace B	127
Syntax	127
Description	127
B2 [two]	
Maximum Hold for Trace B	128
Syntax	128
Description	128
B3 [three]	
View Mode for Trace B	129
Syntax	129
Description	129
B4 [four]	
Blank Trace B	130
Syntax	130
Description	130
BL	
Trace B minus Display Line to Trace B	131
Syntax	131
Description	131
BLANK	
Blank Trace	132
Syntax	132
Description	132
BML	
Trace B Minus Display Line	133
Syntax	133
Description	133
BTC	
Transfer Trace B to Trace C	134
Syntax	134
Description	134
BXC	
Exchange Trace B and Trace C	135
Syntax	135
Description	135
C1 [one]	
Set A Minus B Mode Off	136
Syntax	136
Description	136
C2 [two]	
A Minus B Into A	137
Syntax	137
Description	137
CA	

Couple Attenuation .....	138
Syntax .....	138
Description .....	138
CAL	
Calibration .....	139
Syntax .....	139
Description .....	139
CARRON	
Carrier On Power .....	140
Syntax .....	140
Description .....	140
CF	
Center Frequency .....	141
Syntax .....	141
Description .....	141
CHANNEL	
Channel Selection .....	142
Syntax .....	142
Description .....	142
CHANPWR	
Channel Power .....	143
Syntax .....	143
Description .....	143
CHP	
Channel Power .....	144
Syntax .....	144
Description .....	144
CHPWRBW	
Channel Power Bandwidth .....	145
Syntax .....	145
Description .....	145
CLRAVG	
Clear Average .....	146
Syntax .....	146
Description .....	146
CLRW	
Clear Write .....	147
Syntax .....	147
Description .....	147
CLS	
Clear Status Byte .....	148
Syntax .....	148
Description .....	148
CONTS	
Continuous Sweep .....	149
Syntax .....	149
Description .....	149
CORREK	
Correction Factors On .....	150

---

# Contents

Syntax .....	150
Description .....	150
<b>COUPLE</b>	
Input Coupling .....	151
Syntax .....	151
Description .....	151
<b>CR</b>	
Couple Resolution Bandwidth .....	152
Syntax .....	152
Description .....	152
<b>CS</b>	
Couple Frequency Step Size .....	153
Syntax .....	153
Description .....	153
<b>CT</b>	
Couple Sweep Time .....	154
Syntax .....	154
Description .....	154
<b>CV</b>	
Couple Video Bandwidth .....	155
Syntax .....	155
Description .....	155
<b>DA</b>	
Display Address .....	156
Syntax .....	156
Description .....	156
<b>DELMKBW</b>	
Occupied Power Bandwidth Within Delta Marker .....	157
Syntax .....	157
Description .....	157
<b>DET</b>	
Detection Mode .....	158
Syntax .....	158
Description .....	158
<b>DL</b>	
Display Line .....	160
Syntax .....	160
Description .....	161
<b>DLE</b>	
Display Line Enable .....	162
Syntax .....	162
Description .....	162
<b>DLYSWP</b>	
Delay Sweep .....	163
Syntax .....	163
Description .....	163
<b>DONE</b>	
Done .....	164
Syntax .....	164



Description .....	164
E1[one]	
Peak Marker .....	165
Syntax .....	165
Description .....	165
E2 [two]	
Marker to Center Frequency .....	166
Syntax .....	166
Description .....	166
E3 [three]	
Delta Marker Step Size .....	167
Syntax .....	167
Description .....	167
E4 [four]	
Marker to Reference Level .....	168
Syntax .....	168
Description .....	168
EDITDONE	
End of Limit Line Edits .....	169
Syntax .....	169
Description .....	169
EDITLIML	
Edit Limit Line .....	170
Syntax .....	170
Description .....	170
ERR	
Error .....	171
Syntax .....	171
Description .....	171
ET	
Elapsed Time .....	172
Syntax .....	172
Description .....	172
EX	
Exchange Trace A and Trace B .....	173
Syntax .....	173
Description .....	173
FA	
Start Frequency .....	174
Syntax .....	174
Description .....	174
FB	
Stop Frequency .....	175
Syntax .....	175
Description .....	175
FDSP	
Frequency Display Off .....	176
Syntax .....	176
Description .....	176

---

# Contents

<b>FOFFSET</b>	
Frequency Offset .....	177
Syntax .....	177
Description .....	177
<b>FPKA</b>	
Fast Preselector Peak .....	178
Syntax .....	178
Description .....	178
<b>FREF</b>	
Frequency Reference .....	179
Syntax .....	179
Description .....	179
<b>FS</b>	
Full Span .....	180
Syntax .....	180
Description .....	180
<b>GRAT</b>	
Graticule .....	184
Syntax .....	184
Description .....	184
<b>HD</b>	
Hold Data Entry .....	185
Syntax .....	185
Description .....	185
<b>I1 [one]</b>	
Set RF Coupling to DC .....	186
Syntax .....	186
Description .....	186
<b>I2 [two]</b>	
Set RF Coupling to AC .....	188
Syntax .....	188
Description .....	188
<b>ID</b>	
Identify .....	190
Syntax .....	190
Description .....	190
<b>IP</b>	
Instrument Preset .....	191
Syntax .....	191
Description .....	191
<b>KS,</b>	
Mixer Level .....	192
Syntax .....	192
Description .....	192
<b>KS=</b>	
Marker Counter Resolution .....	193
Syntax .....	193
Description .....	193
<b>KSA</b>	

Amplitude in dBm	194
Syntax	194
Description	194
KSa	
Normal Detection	195
Syntax	195
Description	195
KSB	
Amplitude in dBmV	196
Syntax	196
Description	196
KSb	
Positive Peak Detection	197
Syntax	197
Description	197
KSC	
Amplitude in dBuV	198
Syntax	198
Description	198
KSc	
A Plus B to A	199
Syntax	199
Description	199
KSD	
Amplitude in Volts	200
Syntax	200
Description	200
KSd	
Negative Peak Detection	201
Syntax	201
Description	201
KSE	
Title Mode	202
Syntax	202
Description	202
KSe	
Sample Detection	203
Syntax	203
Description	203
KSG	
Video Averaging On	204
Syntax	204
Description	204
KSg	
Display Off	205
Syntax	205
Description	205
KSH	
Video Averaging Off	206

---

# Contents

Syntax .....	206
Description .....	206
<b>KSh</b>	
Display On .....	207
Syntax .....	207
Description .....	207
<b>KSI</b>	
Extend Analyzer Reference Level .....	208
Syntax .....	208
Description .....	208
<b>KSi</b>	
Exchange Trace B and Trace C .....	209
Syntax .....	209
Description .....	209
<b>KSj</b>	
View Trace C .....	210
Syntax .....	210
Description .....	210
<b>KSK</b>	
Marker to Next Peak .....	211
Syntax .....	211
Description .....	211
<b>KSk</b>	
Blank Trace C .....	212
Syntax .....	212
Description .....	212
<b>KSL</b>	
Marker Noise Off .....	213
Syntax .....	213
Description .....	213
<b>KSI</b>	
Transfer Trace B to Trace C .....	214
Syntax .....	214
Description .....	214
<b>KSM</b>	
Marker Noise On .....	215
Syntax .....	215
Description .....	215
<b>KSm</b>	
Graticule Off .....	217
Syntax .....	217
Description .....	217
<b>KSN</b>	
Marker Minimum .....	218
Syntax .....	218
Description .....	218
<b>KSn</b>	
Graticule On .....	219
Syntax .....	219

Description .....	219
<b>KSO</b>	
Marker Span .....	220
Syntax .....	220
Description .....	220
<b>KSo</b>	
Annotation Off .....	221
Syntax .....	221
Description .....	221
<b>KSp</b>	
Annotation On .....	222
Syntax .....	222
Description .....	222
<b>KST</b>	
Fast Preset .....	223
Syntax .....	223
Description .....	223
<b>KSV</b>	
Frequency Offset .....	224
Syntax .....	224
Description .....	224
<b>KSx</b>	
External Trigger .....	225
Syntax .....	225
Description .....	225
<b>KSy</b>	
Video Trigger .....	226
Syntax .....	226
Description .....	226
<b>KSZ</b>	
Reference Level Offset .....	227
Syntax .....	227
Description .....	227
<b>L0 [zero]</b>	
Display Line Off .....	228
Syntax .....	228
Description .....	228
<b>LF</b>	
Low Frequency Preset .....	229
Syntax .....	229
Description .....	229
<b>LG</b>	
Logarithmic Scale .....	230
Syntax .....	230
Description .....	230
<b>LIMD</b>	
Limit Line Delta Value .....	232
Syntax .....	232
Description .....	232

---

# Contents

<b>LIMF</b>	
Limit Line Frequency Value .....	233
Syntax .....	233
Description .....	233
<b>LIMIDEL</b>	
Delete Limit Line Table .....	234
Syntax .....	234
Description .....	234
<b>LIMIDISP</b>	
Limit Line Display .....	235
Syntax .....	235
Description .....	235
<b>LIMIFAIL</b>	
Limits Failed .....	236
Syntax .....	236
Description .....	236
<b>LIMIFT</b>	
Select Frequency or Time Limit Line .....	237
Syntax .....	237
Description .....	237
<b>LIMIPURGE</b>	
Delete Current Limit Line .....	238
Syntax .....	238
Description .....	238
<b>LIMIREL</b>	
Relative Limit Lines .....	239
Syntax .....	239
Description .....	239
<b>LIMITEST</b>	
Enable Limit Line Testing .....	240
Syntax .....	240
Description .....	240
<b>LIML</b>	
Lower-Limit Amplitude .....	241
Syntax .....	241
Description .....	241
<b>LIMM</b>	
Limit Middle-Amplitude .....	242
Syntax .....	242
Description .....	242
<b>LIMU</b>	
Upper-Limit Amplitude .....	243
Syntax .....	243
Description .....	243
<b>LN</b>	
Linear Scale .....	244
Syntax .....	244
Description .....	244
<b>LSPAN</b>	

Previous Span	245
Syntax	245
Description	245
M1 [one]	
Marker Off	246
Syntax	246
Description	246
M2 [two]	
Marker Normal	247
Syntax	247
Description	247
M3 [three]	
Delta Marker	248
Syntax	248
Description	248
M4 [four]	
Marker Zoom	250
Syntax	250
Description	250
MA	
Marker Amplitude Output	251
Syntax	251
Description	251
MC0 [zero]	
Marker Frequency Counter Off	252
Syntax	252
Description	252
MC1 [one]	
Marker Frequency Counter On	253
Syntax	253
Description	253
MDS	
Measurement Data Size	254
Syntax	254
Description	254
MEAN	
Trace Mean	255
Syntax	255
Description	255
MEANPWR	
Mean Power measurement	256
Syntax	256
Description	256
MEASOFF	
Measurement Off	257
Syntax	257
Description	257
MF	
Marker Frequency Output	258

---

# Contents

Syntax .....	258
Description .....	258
<b>MINH</b>	
Minimum Hold .....	259
Syntax .....	259
Description .....	259
<b>MINPOS</b>	
Minimum X Position .....	260
Syntax .....	260
Description .....	260
<b>MKA</b>	
Marker Amplitude .....	261
Syntax .....	261
Description .....	261
<b>MKACT</b>	
Activate Marker .....	262
Syntax .....	262
Description .....	262
<b>MKBW</b>	
Marker Bandwidth .....	263
Syntax .....	263
Description .....	263
<b>MKCF</b>	
Marker to Center Frequency .....	264
Syntax .....	264
Description .....	264
<b>MKD</b>	
Marker Delta .....	265
Syntax .....	265
Description .....	265
<b>MKF</b>	
Marker Frequency .....	267
Syntax .....	267
Description .....	267
<b>MKFC</b>	
Marker Counter .....	268
Syntax .....	268
Description .....	268
<b>MKFCR</b>	
Marker Counter Resolution .....	269
Syntax .....	269
Description .....	269
<b>MKMIN</b>	
Marker Minimum .....	270
Syntax .....	270
Description .....	270
<b>MKN</b>	
Marker Normal .....	271
Syntax .....	271



Description .....	271
<b>MKNOISE</b>	
Marker Noise .....	272
Syntax .....	272
Description .....	272
<b>MKOFF</b>	
Marker Off .....	274
Syntax .....	274
Description .....	274
<b>MKP</b>	
Marker Position .....	275
Syntax .....	275
Description .....	275
<b>MKPK</b>	
Marker Peak .....	276
Syntax .....	276
Description .....	276
<b>MKPT</b>	
Marker Threshold .....	277
Syntax .....	277
Description .....	277
<b>MKPX</b>	
Marker Peak Excursion .....	278
Syntax .....	278
Description .....	278
<b>MKREAD</b>	
Marker Readout .....	279
Syntax .....	279
Description .....	279
<b>MKRL</b>	
Marker to Reference Level .....	281
Syntax .....	281
Description .....	281
<b>MKSP</b>	
Marker to Span .....	282
Syntax .....	282
Description .....	282
<b>MKSS</b>	
Marker to Step Size .....	283
Syntax .....	283
Description .....	283
<b>MKT</b>	
Marker Time .....	284
Syntax .....	284
Description .....	284
<b>MKTBL</b>	
Marker Table .....	285
Syntax .....	285
Description .....	285

---

# Contents

<b>MKTRACE</b>	
Marker Trace .....	286
Syntax .....	286
Description .....	286
<b>MKTRACK</b>	
Marker Track .....	287
Syntax .....	287
Description .....	287
<b>MKTYPE</b>	
Marker Type .....	288
Syntax .....	288
Description .....	288
<b>ML</b>	
Mixer Level .....	289
Syntax .....	289
Description .....	289
<b>MT0 [zero]</b>	
Marker Track Off .....	291
Syntax .....	291
Description .....	291
<b>MT1 [one]</b>	
Marker Track On .....	292
Syntax .....	292
Description .....	292
<b>MXMH</b>	
Maximum Hold .....	293
Syntax .....	293
Description .....	293
<b>O1 [one]</b>	
Format - Display Units .....	294
Syntax .....	294
Description .....	294
<b>O2 [two]</b>	
Format - Two 8-Bit Bytes .....	295
Syntax .....	295
Description .....	295
<b>O3 [three]</b>	
Format - Real Amplitude Units .....	296
Syntax .....	296
Description .....	296
<b>O4 [four]</b>	
Format - One 8-Bit Byte .....	297
Syntax .....	297
Description .....	297
<b>OCCUP</b>	
Percent Occupied Power Bandwidth .....	298
Syntax .....	298
Description .....	298
<b>OL</b>	

Output Learn String	299
Syntax	299
Description	299
OT	
Output Trace Annotations	300
Syntax	300
Description	300
PEAKS	
Peaks	301
Syntax	301
Description	301
PKPOS	
Peak Position	302
Syntax	302
Description	302
PLOT	
Plot	303
Syntax	303
Description	303
PP	
Preselector Peak	304
Syntax	304
Description	304
PREAMPG	
External Preamplifier Gain	305
Syntax	305
Description	305
PRINT	
Print	306
Syntax	306
Description	306
PWRBW	
Power Bandwidth	307
Syntax	307
Description	307
PWRUPTIME	
Power Up Time	308
Syntax	308
Description	308
Q0 [zero]	
EMI Peak Detection	309
Syntax	309
Description	309
Q1 [one]	
Quasi-Peak Detection	310
Syntax	310
Description	310
R1 [one]	
Illegal Command SRQ	311

---

# Contents

Syntax .....	311
Description .....	311
R2 [two]	
End-of-Sweep SRQ .....	312
Syntax .....	312
Description .....	312
R3 [three]	
Hardware Broken SRQ .....	313
Syntax .....	313
Description .....	313
R4 [four]	
Units-Key-Pressed SRQ .....	314
Syntax .....	314
Description .....	314
RB	
Resolution Bandwidth .....	315
Syntax .....	315
Description .....	316
RBR	
Resolution Bandwidth Ratio .....	317
Syntax .....	317
Description .....	317
RC	
Recall State .....	318
Syntax .....	318
Description .....	318
RCLS	
Recall State .....	319
Syntax .....	319
Description .....	319
REV	
Revision .....	320
Syntax .....	320
Description .....	320
RL	
Reference Level .....	321
Syntax .....	321
Description .....	321
RMS	
Root Mean Square Value .....	323
Syntax .....	323
Description .....	323
ROFFSET	
Reference Level Offset .....	324
Syntax .....	324
Description .....	324
RQS	
Service Request Mask .....	325
Syntax .....	325

Description .....	325
S1[one]	
Continuous Sweep .....	326
Syntax .....	326
Description .....	326
S2 [two]	
Single Sweep .....	327
Syntax .....	327
Description .....	327
SADD	
Add Limit Line Segment .....	328
Syntax .....	328
Description .....	328
SAVES	
Save State .....	329
Syntax .....	329
Description .....	329
SDEL	
Delete Limit Line Segment .....	330
Syntax .....	330
Description .....	330
SDON	
Terminate Limit Line Segment .....	331
Syntax .....	331
Description .....	331
SEDI	
Activate Limit Line Segment .....	332
Syntax .....	332
Description .....	332
SENDER	
Segment Entry for Frequency Limit Lines .....	333
Syntax .....	333
Description .....	333
SER	
Serial Number .....	334
Syntax .....	334
Description .....	334
SETDATE	
Set Date .....	335
Syntax .....	335
Description .....	335
SETTIME	
Set Time .....	336
Syntax .....	336
Description .....	336
SMOOTH	
Smooth Trace .....	337
Syntax .....	337
Description .....	337

---

# Contents

SNGLS	
Single Sweep .....	338
Syntax .....	338
Description .....	338
SP	
Frequency Span .....	339
Syntax .....	339
Description .....	340
SRQ	
User-Defined SRQ .....	341
Syntax .....	341
Description .....	341
SS	
Center Frequency Step Size .....	342
Syntax .....	342
Description .....	343
ST	
Sweep Time .....	344
Syntax .....	344
Description .....	345
STB	
Status Byte Query .....	346
Syntax .....	346
Description .....	346
STDEV	
Standard Deviation of Trace Amplitudes .....	347
Syntax .....	347
Description .....	347
SV	
Save State .....	348
Syntax .....	348
Description .....	348
T1 [one]	
Free Run Trigger .....	349
Syntax .....	349
Description .....	349
T2 [two]	
Line Trigger .....	350
Syntax .....	350
Description .....	350
T3 [three]	
External Trigger .....	351
Syntax .....	351
Description .....	351
T4 [four]	
Video Trigger .....	352
Syntax .....	352
Description .....	352
TA	

Trace A	353
Syntax	353
Description	353
TB	
Trace B	354
Syntax	354
Description	354
TDF	
Trace Data Format	355
Syntax	355
Description	355
TH	
Threshold	356
Syntax	356
Description	356
TIMEDATE	
Time Date	357
Syntax	357
Description	357
TITLE	
Title	358
Syntax	358
Description	358
TM	
Trigger Mode	359
Syntax	359
Description	359
TRA	
Trace Data Input and Output	360
Syntax	360
Description	360
TRB	
Trace Data Input and Output	361
Syntax	361
Description	361
TRC	
Trace Data Input and Output	362
Syntax	362
Description	362
TRDSP	
Trace Display	363
Syntax	363
Description	363
TRIGPOL	
Trigger Polarity	364
Syntax	364
Description	364
TRPRST	
Trace Preset	365

---

# Contents

Syntax .....	365
Description .....	365
TRSTAT	
Trace State .....	366
Syntax .....	366
Description .....	366
TS	
Take Sweep .....	367
Syntax .....	367
Description .....	367
VAVG	
Video Average .....	368
Syntax .....	368
Description .....	368
VB	
Video Bandwidth .....	370
Syntax .....	370
Description .....	371
VBO	
Video Bandwidth Coupling Offset .....	372
Syntax .....	372
Description .....	372
VBR	
Video Bandwidth to Resolution Bandwidth Ratio .....	373
Syntax .....	373
Description .....	373
VIEW	
View Trace .....	374
Syntax .....	374
Description .....	374
VTL	
Video Trigger Level .....	375
Syntax .....	375
Description .....	375
XCH	
Exchange .....	376
Syntax .....	376
Description .....	376
<b>5. A Brief Introduction to the SCPI Language</b>	
SCPI Language Basics .....	378
Command Keywords and Syntax .....	378
Creating Valid Commands .....	378
Special Characters in Commands .....	379
Parameters in Commands .....	381
Putting Multiple Commands on the Same Line .....	383



---

**1 Getting Started**

---

## Option 266 Description

Option 266 (ESA/PSA Series Programming Code Compatibility Suite) for Agilent Technologies' ESA-E series, ESA-L series and PSA series of spectrum analyzers allows the analyzer to be controlled using many of the remote programming commands from the following analyzers:

- 8560 E/EC Series Portable Spectrum Analyzers, comprising:
  - 8560E
  - 8560EC
  - 8561E
  - 8561EC
  - 8563E
  - 8563EC
  - 8564E
  - 8564EC
  - 8565E
  - 8565EC
- 8566A/B
- 8568A/B
- 8590 E/L Series Portable Spectrum Analyzers, comprising:
  - 8590L
  - 8591E
  - 8592L
  - 8593E
  - 8594E
  - 8594L
  - 8595E
  - 8596E

---

### NOTE

The 8566A/B and the 8568A/B are not considered part of the 8560 series of analyzers.

Option 266 is designed to replace these analyzers in many automated systems with minimal or no modification to the currently used measurement software.

## Option 266 Limitations

The Agilent Option 266 (ESA/PSA Series Programming Code Compatibility Suite) has been designed to emulate as closely as possible the operation of the specified spectrum analyzers. It is not, however, intended as an absolute direct replacement for these analyzers.

### **Supported commands**

Only a subset of the 8566/8568/8560 Series/8590 Series commands is supported in this option (through a GPIB interface). These supported commands were determined by feedback from our customers combined with technical considerations and constraints.

### **Predefined Functions**

In the 8566/8568/8560 Series/8590 Series analyzers, a “predefined function” is an analyzer command that returns a number that can be operated on by other analyzer commands. “Predefined variables” follow the same idea, except the value to be passed as a parameter to the next command is stored in a variable.

Option 266 does not support this type of behavior, so any commands that originally acted as predefined functions or variables, or that allowed predefined functions or variables as arguments in the 8566/8568/8560 Series/8590 Series analyzers, will no longer do so.

### **User-defined Functions**

No user-defined functions, traces, or variables (FUNCDEF, TRDEF or VARDEF) can be used as arguments or commands in programs controlling any analyzer running Option 266. In addition, the behavior of certain commands that rely on the “active functions” (UP, DN, etc.) may be slightly different.

## Hardware and Firmware Requirements for Option 266

One of the following Agilent spectrum analyzers and associated hardware options is required to run Option 266.

**Table 1-1 Compatible Agilent ESA-E Series Spectrum Analyzers**

Analyzer Model Number	Upper Frequency Limit	Firmware	ESA/PSA Series Programming Code Compatibility Suite Personality	B72 (Extended Memory Option)	A4H (GPIB and Parallel Interface Option)
E4401B	1.5 GHz	Rev A.14.00 or later	Option 266 Rev B.00.00 or later	Required	Required
E4402B	3.0 GHz	Rev A.14.00 or later	Option 266 Rev B.00.00 or later	Required	Required
E4404B	6.7 GHz	Rev A.14.00 or later	Option 266 Rev B.00.00 or later	Required	Required
E4405B	13.2 GHz	Rev A.14.00 or later	Option 266 Rev B.00.00 or later	Required	Required
E4407B	26.5 GHz	Rev A.14.00 or later	Option 266 Rev B.00.00 or later	Required	Required

**Table 1-2 Compatible Agilent ESA-L Series Spectrum Analyzers**

Analyzer Model Number	Upper Frequency Limit	Firmware	ESA/PSA Series Programming Code Compatibility Suite <sup>a</sup>	B72 (Extended Memory Option)	A4H (GPIB and Parallel Interface Option)
E4411B	1.5 GHz	Rev A.14.00 or later	Option 266 Rev B.00.00 or later <sup>a</sup>	Required	Required
E4403B	3.0 GHz	Rev A.14.00 or later	Option 266 Rev B.00.00 or later <sup>a</sup>	Required	Required
E4408B	26.5 GHz	Rev A.14.00 or later	Option 266 Rev B.00.00 or later <sup>a</sup>	Required	Required

a. Only 8590 Series code compatibility is supported on the ESA-L series of analyzers. There is no support for 8566, 8568, or 8560 Series code compatibility.

**Table 1-3**      **Compatible Agilent PSA Series Spectrum Analyzers**

<b>Analyzer Model Number</b>	<b>Upper Frequency Limit</b>	<b>Firmware</b>	<b>ESA/PSA Series Programming Code Compatibility Suite Personality</b>
E4440A	26.5 GHz	A.10.00 or later	Option 266
E4443A	6.7 GHz	A.10.00 or later	Option 266
E4445A	13.2 GHz	A.10.00 or later	Option 266
E4446A	44.0 GHz	A.10.00 or later	Option 266
E4447A	42.98 GHz	A.10.00 or later	Option 266
E4448A	50.0 GHz	A.10.00 or later	Option 266

For maximum compatibility, you should select a PSA Series analyzer that best matches the frequency range of your chosen remote language. The frequency limits of the remote languages are listed below.

**Table 1-4 Frequency Ranges of the Remote Languages**

Remote Language	Start Frequency	Stop Frequency
8560E/EC	0 Hz	2.9 GHz
8561E/EC	0 Hz	6.5 GHz
8562E/EC	0 Hz	13.2 GHz
8563E/EC	0 Hz	26.5 GHz
8564E/EC	0 Hz	40.0 GHz
8565E/EC	0 Hz	50.0 GHz
8566A	0 Hz	22.0 GHz
8566B	0 Hz	22.0 GHz
8568A	0 Hz	1.5 GHz
8568B	0 Hz	1.5 GHz
8590L	0 Hz	1.8 GHz
8591E	0 Hz	1.8 GHz
8592L <sup>a</sup>	0 Hz	22.0 GHz
8593E <sup>a</sup>	0 Hz	22.0 GHz
8594E	0 Hz	2.9 GHz
8595E	0 Hz	6.5 GHz
8596E <sup>a</sup>	0 Hz	12.8 GHz

- a. The command HNLOCK (Harmonic Lock) is not supported in Option 266. Thus, the different frequency spans associated with the various harmonic bands are also not supported.

---

## Installing Option 266 - PSA Series Analyzers Only

You must load the desired personality option into the instrument memory. Loading can be done from a firmware CD-ROM or from the internet location. An automatic loading program comes with the files and runs from your PC.

### Installing Optional Measurement Personalities

When you install a measurement personality, you need to follow a three step process:

1. Determine whether your memory capacity is sufficient to contain all the options you want to load. If not, decide which options you want to install now, and consider upgrading your memory. Details follow in [“Do You Have Enough Memory to Load All Your Personality Options?”](#) on page 1-31.
2. Install the measurement personality firmware into the instrument memory. Details follow in [“Loading an Optional Measurement Personality”](#) on page 35.
3. Enter a license key that activates the measurement personality. Details follow in [“Obtaining and Installing a License Key”](#) on page 35.

Adding measurement personalities requires the purchase of an upgrade kit for the desired option. The upgrade kit contains the measurement personality firmware and an entitlement certificate that is used to generate a license key from the internet website. A separate license key is required for each option on a specific instrument serial number and host ID.

For the latest information on Agilent Spectrum Analyzer options and upgrade kits, visit the following web location:

[http://www.agilent.com/find/sa\\_upgrades](http://www.agilent.com/find/sa_upgrades)

### Do You Have Enough Memory to Load All Your Personality Options?

If you do not have memory limitations then you can skip ahead to the next section [“Loading an Optional Measurement Personality”](#) on page 1-35. If after installing your options you get error messages relating to memory issues, you can return to this section to learn more about how to optimize your configuration.

If you have 64 MBytes of memory installed in your instrument, you should have enough memory to install at least four optional personalities, with plenty of memory for data and states.

The optional measurement personalities require different amounts of memory. So the number of personalities that you can load varies. This is also impacted by how much data you need to save. If you are having memory errors you must swap the applications in or out of memory as needed. If you only have 48 MBytes of memory, you can upgrade your hardware to 64 MBytes.

Additional memory can be added to any PSA Series analyzer by installing Option 115. With this option installed, you can install all currently available measurement personalities in your analyzer and still have memory space to store more state and trace files than would otherwise be possible.

To see the size of your installed memory for PSA Series Spectrum Analyzers:

1. Ensure that the spectrum analyzer is in spectrum analyzer mode because this can affect the screen size.
2. Press **System, More, Show Hdw.**
3. Read Flash Memory size in the table. If Option 115 is installed, the table will also show Compact Flash Type and Compact Flash Size.

PSA Flash Memory Size	Available Memory Without Option B7J and Option 122 or 140	Available Memory With Option B7J and Option 122 or 140
64 Mbytes	32.5 MBytes	30.0 MBytes
48 Mbytes	16.9 MBytes	14.3 MBytes

PSA Compact Flash Memory Size	Available Additional Memory for Measurement Personalities
512 Mbytes (Opt. 115)	512 MBytes

If you have 48 MBytes of memory, and you want to install more than 3 optional personalities, you may need to manage your memory resources. The following section, [“How to Predict Your Memory Requirements” on page 1-33](#), will help you decide how to configure your installed options to provide optimal operation.



## How to Predict Your Memory Requirements

If you plan to install many optional personalities, you should review your memory requirements, so you can determine whether you have enough memory (unless you have a PSA Series with Option 115). There is an Agilent “Memory Calculator” available online that can help you do this, or you can make a calculated approximation using the information that follows. You will need to know your instrument’s installed memory size as determined in the previous section and then select your desired applications.

---

**NOTE**

If you have a PSA Series analyzer with Option 115, there is adequate memory to install all of the available optional personalities in your instrument.

To calculate the available memory on your PSA, see:  
<http://sa.tm.agilent.com/PSA/memory/>

Select the “Memory Calculator” link. You can try any combination of available personalities to see if your desired configuration is compatible with your installed memory.

---

**NOTE**

After loading all your optional measurement personalities, you should have a reserve of ~2 MBytes memory to facilitate mode switching. Less available memory will increase mode switching time. For example, if you employ excessive free memory by saving files of states and/or data, your mode switching time can increase to more than a minute.

You can manually estimate your total memory requirements by adding up the memory allocations described in the following steps. Compare the desired total with the available memory that you identified in the previous section.

1. Program memory - Select option requirements from the table “[Measurement Personality Options and Memory Required](#)” on page 1-34.
2. Shared libraries require 7.72 MBytes.
3. Recommended mode swap space is 2 MBytes.
4. Screens - .gif files need 20-25 kBytes each.
5. State memory - State file sizes range from 21 kB for SA mode to 40 kB for W-CDMA. The state of every mode accessed since power-on will be saved in the state file. File sizes can exceed 150 kB each when several modes are accessed, for each state file saved.

---

**TIP**

State memory retains settings for all states accessed before the **Save State** command. To reduce this usage to a minimum, reduce the modes accessed before the **Save State** is executed. You can set the PSA to boot into a selected mode by accessing the desired mode, then pressing the **System, Power On/Preset, Power On** keys and toggle the setting to **Last**.

### Measurement Personality Options and Memory Required

Personality Options for PSA Series Spectrum Analyzers <sup>a</sup>	Option	File Size (PSA Rev: A.10)
cdmaOne measurement personality	<b>BAC</b>	1.91 Mbytes
NADC and PDC measurement personalities (not available separately)	<b>BAE</b>	2.43 Mbytes
W-CDMA or W-CDMA, HSDPA, HSUPA measurement personality	<b>BAF, 210</b>	5.38 Mbytes <sup>b</sup>
cdma2000 or cdma2000 w/ 1xEV-DV measurement personality	<b>B78, 214</b>	4.00 Mbytes <sup>b</sup>
1xEV-DO measurement personality	<b>204</b>	5.61 Mbytes <sup>b</sup>
GSM (with EDGE) measurement personality	<b>202</b>	3.56 Mbytes <sup>b</sup>
Shared measurement library <sup>b</sup>	n/a	7.72 Mbytes
Phase Noise measurement personality	<b>226</b>	2.82 Mbytes <sup>c</sup>
Noise Figure measurement personality	<b>219</b>	4.68 Mbytes <sup>c</sup>
Basic measurement personality with digital demod hardware	<b>B7J</b>	Cannot be deleted (2.64 Mbytes)
Programming Code Compatibility Suite <sup>d</sup> (8560 Series, 8590 Series, and 8566/8568)	<b>266</b>	1.18 Mbytes <sup>c</sup>
TD-SCDMA Power measurement personality	<b>211</b>	5.47 Mbytes <sup>c</sup>
TD-SCDMA Modulation Analysis or TD-SCDMA Modulation Analysis w/ HSDPA/8PSK measurement personality	<b>212, 213</b>	1.82 Mbytes
Flexible Digital Modulation Analysis	<b>241</b>	2.11 Mbytes <sup>b</sup>
WLAN measurement personality	<b>217</b>	3.24 Mbytes <sup>b</sup>
External Source Control	<b>215</b>	0.72 Mbytes <sup>c</sup>
Measuring Receiver Personality (available with Option 23A - Trigger support for AM/FM/PM and Option 23B - CCITT filter)	<b>233</b>	2.91 Mbytes <sup>b</sup>
EMC Analyzer	<b>239</b>	4.06 Mbytes <sup>b</sup>

- a. Available as of the print date of this guide.
- b. Many PSA Series personality options use a 7.72 Mbyte shared measurement library. If you are loading multiple personalities that use this library, you only need to add this memory allocation once.
- c. Shared measurement library allocation not required.
- d. This is a no charge option that does not require a license key.

**Memory Upgrade Kits** The PSA 64 MByte Memory Upgrade kit part number is E4440AU-ANE. The PSA Compact Flash Upgrade kit part number is E4440AU-115.

For more information about memory upgrade kits contact your local sales office, service office, or see:

[http://www.agilent.com/find/sa\\_upgrades](http://www.agilent.com/find/sa_upgrades)

### Loading an Optional Measurement Personality

You must use a PC to load the desired personality option into the instrument memory. Loading can be done from a firmware CD-ROM or by downloading the update program from the internet. An automatic loading program comes with the files and runs from your PC.

You can check the Agilent internet website for the latest PSA firmware versions available for downloading:

[http://www.agilent.com/find/psa\\_firmware](http://www.agilent.com/find/psa_firmware)

---

#### NOTE

When you add a new option, or update an existing option, you will get the updated versions of all your current options as they are all reloaded simultaneously. This process may also require you to update the instrument core firmware so that it is compatible with the new option.

Depending on your installed hardware memory, you may not be able to fit all of the available measurement personalities in instrument memory at the same time. You may need to delete an existing option file from memory and load the one you want. Use the automatic update program that is provided with the files. Refer to the table showing “[Measurement Personality Options and Memory Required](#)” on [page 1-34](#). The approximate memory requirements for the options are listed in this table. These numbers are worst case examples. Some options share components and libraries, therefore the total memory usage of multiple options may not be exactly equal to the combined total.

### Obtaining and Installing a License Key

If you purchase an optional personality that requires installation, you will receive an “Entitlement Certificate” which may be redeemed for a license key specific to one instrument. Follow the instructions that accompany the certificate to obtain your license key.

To install a license key for the selected personality option, use the following procedure:

---

#### NOTE

You can also use this procedure to reinstall a license key that has been deleted during an uninstall process, or lost due to a memory failure.

For PSA:

1. Press **System, More, More, Licensing, Option** to access the alpha editor. Use this alpha editor to enter letters (upper-case), and the front-panel numeric keys to enter numbers for the option

designation. You will validate your option entry in the active function area of the display. Then, press the **Enter** key.

2. Press **License Key** to enter the letters and digits of your license key. You will validate your license key entry in the active function area of the display. Then, press the **Enter** key.
3. Press the **Activate License** key.

### Viewing a License Key

Measurement personalities purchased with your instrument have been installed and activated at the factory before shipment. The instrument requires a **License Key** unique to every measurement personality purchased. The license key is a hexadecimal number specific to your measurement personality, instrument serial number and host ID. It enables you to install, or reactivate that particular personality.

Use the following procedure to display the license key unique to your personality option that is already installed in your PSA:

Press **System, More, More, Licensing, Show License**. The **System, Personality** key displays the personalities loaded, version information, and whether the personality is licensed.

---

#### NOTE

*You will want to keep a copy of your license key in a secure location. Press **System, More**, then **Licensing, Show License**, and print out a copy of the display that shows the license numbers. If you should lose your license key, call your nearest Agilent Technologies service or sales office for assistance.*

---

### Using the Delete License Key on PSA

This key will make the option unavailable for use, but will not delete it from memory. Write down the 12-digit license key for the option before you delete it. If you want to use that measurement personality later, you will need the license key to reactivate the personality firmware.

---

#### NOTE

Using the **Delete License** key does not remove the personality from the instrument memory, and does not free memory to be available to install another option. If you need to free memory to install another option, refer to the instructions for loading firmware updates located at the URL : <http://www.agilent.com/find/psa/>

---

1. Press **System, More, More, Licensing, Option**. Pressing the **Option** key will activate the alpha editor menu. Use the alpha editor to enter the letters (upper-case) and the front-panel numeric keyboard to enter the digits (if required) for the option, then press the **Enter** key. As you enter the option, you will see your entry in the active function area of the display.
2. Press **Delete License** to remove the license key from memory.

## Ordering Optional Measurement Personalities

When you order a personality option, you will receive an entitlement certificate. Then you will need to go to the Web site to redeem your entitlement certificate for a license key. You will need to provide your instrument serial number and host ID, and the entitlement certificate number.


<b>Required Information:</b>	<b>Front Panel Key Path:</b>
Model #: (Ex. E4440A)	
Host ID: _____	<b>System, Show System</b>
Instrument Serial Number: _____	<b>System, Show System</b>

---


## Configuring Option 266 on PSA Analyzers

[Figure 1-1](#) shows the menu map to configure Option 266 (Programming Code Compatibility Suite) on your spectrum analyzer. Pressing **System, Config I/O, Configure Remote Lang, Language** on a PSA Series analyzer allows you to select the remote programming language for the instrument you are replacing.

**Figure 1-1 System Menu Map for Option 266 on PSA Series Analyzers**




**Figure 1-2 Language Link Menu Map for All Analyzers**



**NOTE** The language softkeys are not in alphanumeric order.

**Figure 1-3** Error Handling Link Menu Map



**Figure 1-4** Preferences Link Menu Map



Getting Started



## The Configure Remote Lang Screen Menu - PSA Analyzers

### Configure Remote Lang

This key displays a menu allowing you to select the remote language you wish to use, and to determine how command error messages are stored and displayed.

### Language

This key allows you to select which remote programming language you wish to use.

- SCPI** Selects the **SCPI** remote programming language. This is the default setting after installation.
- HP8560E/EC** Selects the **HP8560E/EC** remote programming language and sets the response to the remote programming command 'ID?' to **HP8560E**. It also performs an instrument preset and sets Span, Trace Points, couplings, VBW/RBW ratio, and Span/RBW ratio appropriately as shown in [Table 1-5 on page 44](#).
- HP8561E/EC** Selects the **HP8561E/EC** remote programming language and sets the response to the remote programming command 'ID?' to **HP8561E**. It also performs an instrument preset and sets Span, Trace Points, couplings, VBW/RBW ratio, and Span/RBW ratio appropriately as shown in [Table 1-5 on page 44](#).
- HP8562E/EC** Selects the **HP8562E/EC** remote programming language and sets the response to the remote programming command 'ID?' to **HP8562E**. It also performs an instrument preset and sets Span, Trace Points, couplings, VBW/RBW ratio, and Span/RBW ratio appropriately as shown in [Table 1-5 on page 44](#).
- HP8563E/EC** Selects the **HP8563E/EC** remote programming language and sets the response to the remote programming command 'ID?' to **HP8563E**. It also performs an instrument preset and sets Span, Trace Points, couplings, VBW/RBW ratio, and Span/RBW ratio appropriately as shown in [Table 1-5 on page 44](#).
- HP8564E/EC** Selects the **HP8564E/EC** remote programming language and sets the response to the remote programming command 'ID?' to **HP8564E**. It also performs an instrument preset and sets Span, Trace Points, couplings, VBW/RBW ratio, and Span/RBW ratio appropriately as shown in [Table 1-5 on page 44](#).
- HP8565E/EC** Selects the **HP8565E/EC** remote programming language and sets the response to the remote programming command 'ID?' to **HP8565E**. It also performs an

	instrument preset and sets Span, Trace Points, couplings, VBW/RBW ratio, and Span/RBW ratio appropriately as shown in <a href="#">Table 1-5 on page 44</a> .
HP8566A	Selects the <b>HP8566A</b> remote programming language and sets the response to the remote programming command 'ID?' to <b>HP8566A</b> . It also performs an instrument preset and sets Span, Trace Points, couplings, VBW/RBW ratio, and Span/RBW ratio appropriately as shown in <a href="#">Table 1-5 on page 44</a> .
HP8566B	Selects the <b>HP8566B</b> remote programming language and sets the response to the remote programming command 'ID?' to <b>HP8566B</b> . It also performs an instrument preset and sets Span, Trace Points, couplings, VBW/RBW ratio, and Span/RBW ratio appropriately as shown in <a href="#">Table 1-5 on page 44</a> .
HP8568A	Selects the <b>HP8568A</b> remote programming language and sets the response to the remote programming command 'ID?' to <b>HP8568A</b> . It also performs an instrument preset and sets Span, Trace Points, couplings, VBW/RBW ratio, and Span/RBW ratio appropriately as shown in <a href="#">Table 1-5 on page 44</a> .
HP8568B	Selects the <b>HP8568B</b> remote programming language and sets the response to the remote programming command 'ID?' to <b>HP8568B</b> . It also performs an instrument preset and sets Span, Trace Points, couplings, VBW/RBW ratio, and Span/RBW ratio appropriately as shown in <a href="#">Table 1-5 on page 44</a> .
HP8590L	Selects the <b>HP8590L</b> remote programming language and sets the response to the remote programming command 'ID?' to <b>HP8590L</b> . It also performs an instrument preset and sets Span, Trace Points, couplings, VBW/RBW ratio, and Span/RBW ratio appropriately as shown in <a href="#">Table 1-5 on page 44</a> .
HP8591E	Selects the <b>HP8591E</b> remote programming language and sets the response to the remote programming command 'ID?' to <b>HP8591E</b> . It also performs an instrument preset and sets Span, Trace Points, couplings, VBW/RBW ratio, and Span/RBW ratio appropriately as shown in <a href="#">Table 1-5 on page 44</a> .
HP8592L	Selects the <b>HP8592L</b> remote programming language and sets the response to the remote programming command 'ID?' to <b>HP8592L</b> . It also performs an instrument preset and sets Span, Trace Points, couplings, VBW/RBW ratio, and Span/RBW ratio appropriately as shown in <a href="#">Table 1-5 on page 44</a> .
HP8593E	Selects the <b>HP8593E</b> remote programming language and

sets the response to the remote programming command 'ID?' to **HP8593E**. It also performs an instrument preset and sets Span, Trace Points, couplings, VBW/RBW ratio, and Span/RBW ratio appropriately as shown in [Table 1-5 on page 44](#).

**HP8594E** Selects the **HP8594E** remote programming language and sets the response to the remote programming command 'ID?' to **HP8594E**. It also performs an instrument preset and sets Span, Trace Points, couplings, VBW/RBW ratio, and Span/RBW ratio appropriately as shown in [Table 1-5 on page 44](#).

**HP8594L** Selects the **HP8594L** remote programming language and sets the response to the remote programming command 'ID?' to **HP8594L**. It also performs an instrument preset and sets Span, Trace Points, couplings, VBW/RBW ratio, and Span/RBW ratio appropriately as shown in [Table 1-5 on page 44](#).

**HP8595E** Selects the **HP8595E** remote programming language and sets the response to the remote programming command 'ID?' to **HP8595E**. It also performs an instrument preset and sets Span, Trace Points, couplings, VBW/RBW ratio, and Span/RBW ratio appropriately as shown in [Table 1-5 on page 44](#).

**HP8596E** Selects the **HP8596E** remote programming language and sets the response to the remote programming command 'ID?' to **HP8596E**. It also performs an instrument preset and sets Span, Trace Points, couplings, VBW/RBW ratio, and Span/RBW ratio appropriately as shown in [Table 1-5 on page 44](#).

---

**NOTE** Setting the remote language to anything other than 'SCPI' does not affect the response to the SCPI command '\*IDN?'. This command will still return the model number and firmware version number of the PSA Series spectrum analyzer.

---

**Table 1-5 Span, Trace Points, Couplings, VBW/RBW Ratio, and Span/RBW Ratio Settings**

Remote Language	Start Freq.	Stop Freq.	Number of Trace Points	RF Coupling	VBW/RBW Ratio	Span/RBW Ratio
HP8560E/EC	0 Hz	2.9 GHz	601	AC	1	91
HP8561E/EC	0 Hz	6.5 GHz	601	AC	1	91
HP8562E/EC	0 Hz	13.2 GHz	601	AC	1	91
HP8563E/EC	0 Hz	26.5 GHz	601	AC	1	91
HP8564E/EC	0 Hz	40 GHz	601	AC	1	91
HP8565E/EC	0 Hz	50 GHz	601	AC	1	91
HP8566A	2 GHz	22 GHz	1001	DC	3 (VBW one step wider than RBW)	106
HP8566B	2 GHz	22 GHz	1001	DC	3 (VBW one step wider than RBW)	106
HP8568A	0 Hz	1.5 GHz	1001	DC	3 (VBW one step wider than RBW)	106
HP8568B	0 Hz	1.5 GHz	1001	DC	3 (VBW one step wider than RBW)	106
HP8590L	0 Hz	1.8 GHz	401	DC	0.3	106
HP8591E	0 Hz	1.8 GHz	401	DC	0.3	106
HP8592L	2.75 GHz	22 GHz	401	DC	0.3	106
HP8593E	2.75 GHz	22 GHz	401	DC	0.3	106
HP8594E	0 Hz	2.9 GHz	401	AC	0.3	106
HP8594L	0 Hz	2.9 GHz	401	DC	0.3	106
HP8595E	0 Hz	6.5 GHz	401	AC	0.3	106
HP8596E	0 Hz	12.8 GHz	401	AC	0.3	106

<b>Command Err</b>	This key determines whether or not command errors are displayed on the screen. When set to <b>On</b> , error messages generated by unrecognized commands or command arguments are displayed on the screen. When set to <b>Off</b> , error messages generated by unrecognized commands are not displayed. The current setting is underlined on the key label, and this setting is toggled each time the key is pressed.
<b>Cmd Error Log</b>	<p>This key sets command error logging <b>On</b> or <b>Off</b>. When set to <b>On</b>, all error messages are stored in a log file, regardless of whether they have been displayed on the screen. When set to <b>Off</b>, no further command error messages are written to the log file.</p> <p>The log file is an ASCII text file called LOGFILE.TXT on the C: drive. It has a maximum size of 32 KB. Once it has reached its maximum size, no further error messages will be recorded, but a message will be displayed prompting you to clear the log file. Switching <b>Cmd Error Log</b> to <b>Off</b> does not clear the log file. Only the <b>Clear Command Error Log</b> softkey will clear the error log.</p>
<b>Clear Command Error Log</b>	This key allows you to clear all messages from the command error log file LOGFILE.TXT. You will be asked to press the key a second time to confirm your decision to clear the log file.
<b>Limit RBW/VBW</b>	This key toggles the <b>Limit RBW/VBW</b> between On and Off. Setting it to On causes the range of values for resolution and video bandwidths to be limited, dependent on the remote language selected.
<b>NOTE</b>	<p>This restriction on RBW and VBW range changes to use the base PSA range of bandwidths if the detector type is set to Quasi Peak, MIL Peak, EMI Average, or Average.</p> <p>Setting this key to Off causes the resolution bandwidth and video bandwidth filters to use the base PSA range of values for all remote languages.</p>
<b>Atten Offset</b>	This key toggles a supplemental attenuation On and Off. The default state is Off. Setting it to On sets a fixed 12 dB supplemental attenuation in the hardware to prevent too great a signal at the input to the mixer (0 dBm maximum) for reference level settings of up to +12 dBm. The selected condition is preserved when <b>Preset</b> is pressed or power is cycled.
<b>Save Reg 7</b>	This key enables or disables the saving of the Preset State in Register 7 (RC 7). Setting this to On causes the Preset State to be saved in Register 7, but it can slow your analyzer down. Setting this to Off will make your analyzer run faster, but the Preset State is not saved in Register 7.

---

## Running Software that Requires SCPI Commands

When a Remote Language other than SCPI has been selected, you will only have access to a very small subset of SCPI commands. If you are not familiar with the SCPI remote programming language, [Chapter 5](#), “A Brief Introduction to the SCPI Language,” on page 377 contains some useful information.

The SCPI commands available while using other remote languages are:

- **\*IDN?**  
Queries and returns the instrument identification string.
- **\*RST**  
Performs an instrument preset.
- **:SYSTem:LANGUage SCPI | HP8560E | HP8561E | HP8562E | HP8563E | HP8564E | HP8565E | HP8566A | HP8566B | HP8568A | HP8568B | HP8590L | HP8591E | HP8592L | HP8593E | HP8594E | HP8594L | HP8595E | HP8596E**  
Sets the current remote language. This command is only available if you have Option 266 installed on your analyzer.

---

### NOTE

Agilent Technologies recommends that you do not repeatedly swap to and from the SCPI language within your programs.

- **:SYSTem:LANGUage?**  
Queries and returns the current remote language. This command is only available if you have Option 266 installed on your analyzer.
- **[:SENSe]:SWEep:TYPE AUTO|SWEep|FFT**  
Sets the Sweep Type. This command is only valid on the PSA Series of analyzers.
- **[:SENSe]:SWEep:TYPE?**  
Queries and returns the Sweep Type. This command is only valid on the PSA Series of analyzers.
- **[:SENSe]:SWEep:TYPE:AUTO:RULEs SPEed|DRANge**  
Sets the auto rule setting for FFT and Sweep Type. This command is only valid on the PSA Series of analyzers.
- **[:SENSe]:SWEep:TYPE:AUTO:RULEs?**  
Queries and returns the auto rule setting for FFT and Sweep Type. This command is only valid on the PSA Series of analyzers.
- **:SYSTem:OPTion?**

Returns a list of installed options.

If Option 266 is installed on your analyzer, the string “266” will appear in the returned string. In the following example, options B7J, 266, and 110 are all installed.

Example: “B7J,266,110”

To return the analyzer to its full PSA SCPI capability, you must specify the remote language as SCPI.

## Service and Calibration

Since the Performance Verification and Adjustment Software uses the SCPI command language, you will need to set Remote Language to SCPI prior to calibration or service of your Agilent spectrum analyzer.

When your analyzer is returned from an Agilent Technologies service center, you may have to reinstall Option 266. Refer to [“Installing Option 266 - PSA Series Analyzers Only”](#) on page 31.



---

## Documentation for Option 266

### Spectrum Analyzers with Option 266

When you purchase your PSA Series spectrum analyzer with the Programming Code Compatibility Suite (Option 266), you will receive this manual - the *Programming Code Compatibility Suite Guide*. For information on ESA series analyzers and other related documentation, refer to the ESA web site at <http://www.agilent.com/find/esa/>. For information on PSA series analyzers and other related documentation, refer to the PSA web site at <http://www.agilent.com/find/psa/>.

This *Programming Code Compatibility Suite Guide* is not designed to be a comprehensive guide to all legacy commands. It gives brief descriptions of the supported commands, and highlights important functional or behavioral differences that you should be aware of when transferring your existing code to your PSA Series analyzer. For a fuller description of these commands, refer to the manuals supplied with your original analyzer.

### Spectrum Analyzer Updates

For the latest information about this instrument, including software upgrades, application information, and product information, please visit the URL below.

#### Updating the Firmware and Software

Updated versions of the Agilent Spectrum Analyzers' firmware and software will be available via several sources. Information on the latest firmware and software revision can be accessed through the following URL:

<http://www.agilent.com/find/psa/>

---

**NOTE**

If you have received Option 266 as an upgrade, the latest version of the analyzer's firmware has been included and should be loaded into your instrument before installing the Option 266.

---



---

## **2 Legacy Analyzer Command List**

## Table of All Legacy Analyzer Commands

The following table lists all of the original programming commands from the legacy analyzers (that is, the 8566A/B, the 8568A/B, the 8560 Series and the 8590 Series), and indicates which are supported in Option 266. Refer to the alphabetical listing of commands in [Chapter 4](#), “Programming Commands,” on [page 87](#) for more detailed information about each supported command.

**Table 2-1 Alphanumeric List of all Legacy Commands Showing their Option 266 Support**

Command	Description	8566	8568	8590 Series	8560 Series	Page for Further Details
A1	Clear-writes trace A	Yes	Yes	Yes	Yes	<a href="#">Page 91</a>
A2	Max Holds trace A	Yes	Yes	Yes	Yes	<a href="#">Page 92</a>
A3	View trace A	Yes	Yes	Yes	Yes	<a href="#">Page 93</a>
A4	Blanks trace A	Yes	Yes	Yes	Yes	<a href="#">Page 94</a>
ABORT	Interrupt operation of all user-defined functions	N/A	N/A	No	No	
ABS	Absolute	No	No	No	No	
ACP	Performs the adjacent channel power measurement	N/A	N/A	Yes	N/A	<a href="#">Page 95</a>
ACPACCL	Accelerate adjacent channel power measurement	N/A	N/A	N/A	No	
ACPALPHA	Adjacent channel power alpha weighting	N/A	N/A	N/A	Yes	<a href="#">Page 96</a>
ACPALTCH	Adjacent channel power alternate channels	N/A	N/A	N/A	Yes	<a href="#">Page 97</a>
ACPBRPER	Adjacent channel power burst period	N/A	N/A	N/A	Yes	<a href="#">Page 98</a>
ACPBRWID	Adjacent channel power burst width	N/A	N/A	N/A	Yes	<a href="#">Page 99</a>
ACPBW	Specifies channel bandwidth for ACP measurement	N/A	N/A	Yes	Yes	<a href="#">Page 100</a>
ACPCOMPUTE	Compute adjacent channel power	N/A	N/A	N/A	Yes	<a href="#">Page 101</a>

**Table 2-1                    Alphanumeric List of all Legacy Commands Showing their  
Option 266 Support**

Command	Description	8566	8568	8590 Series	8560 Series	Page for Further Details
ACPCONTM	Performs ACP measurement in continuous sweep	N/A	N/A	No	N/A	
ACPE	Adjacent channel power extended	N/A	N/A	No	N/A	
ACPERR	ACP measurement error query	N/A	N/A	No	N/A	
ACPFREQWT	Adjacent channel power frequency weighting	N/A	N/A	N/A	Yes	<a href="#">Page 102</a>
ACPGR	Adjacent channel power graph on or off	N/A	N/A	No	N/A	
ACPGGRAPH	Compute adjacent channel power graph	N/A	N/A	No	No	
ACPLOWER	Lower adjacent channel power	N/A	N/A	Yes	Yes	<a href="#">Page 103</a>
ACPMAX	Maximum adjacent channel power	N/A	N/A	Yes	Yes	<a href="#">Page 104</a>
ACPMEAS	Measure adjacent channel power	N/A	N/A	Yes	Yes	<a href="#">Page 105</a>
ACPMETHOD	Adjacent channel power measurement method	N/A	N/A	N/A	No	
ACPMK	Adjacent channel power marker on or off	N/A	N/A	No	N/A	
ACPMSTATE	Adjacent channel power measurement state	N/A	N/A	Yes	Yes	<a href="#">Page 106</a>
ACPPAR	ACP manual or auto	N/A	N/A	Yes	N/A	<a href="#">Page 107</a>
ACPPWR TX	Total power transmitted	N/A	N/A	Yes	Yes	<a href="#">Page 108</a>
ACPRSLTS	Adjacent channel power measurement results	N/A	N/A	Yes	Yes	<a href="#">Page 109</a>
ACPSNGLM	Performs ACP measurement in single sweep	N/A	N/A	No	N/A	
ACPSP	Channel spacing	N/A	N/A	Yes	Yes	<a href="#">Page 110</a>
ACPT	Adjacent channel power T weighting	N/A	N/A	N/A	Yes	<a href="#">Page 111</a>

**Table 2-1 Alphanumeric List of all Legacy Commands Showing their Option 266 Support**

Command	Description	8566	8568	8590 Series	8560 Series	Page for Further Details
ACPUPPER	Upper adjacent channel power	N/A	N/A	Yes	Yes	<a href="#">Page 112</a>
ACTDEF	Give user-defined function active status	N/A	N/A	No	N/A	
ACTVF	Active function	N/A	N/A	No	N/A	
ACTVFUNC	Creates a user defined active function	N/A	N/A	N/A	No	
ADD	Add	No	No	No	No	
ADJALL	LO & IF adjustment	N/A	N/A	N/A	Yes	<a href="#">Page 113</a>
ADJCRT	Adjust CRT alignment	N/A	N/A	N/A	No	
ADJIF	Adjust IF	N/A	N/A	N/A	No	
AMB	Trace A – trace B -> trace A	Yes	Yes	Yes	Yes	<a href="#">Page 114</a>
AMBPL	Trace A – trace B + Display Line -> trace A	Yes	Yes	Yes	Yes	<a href="#">Page 115</a>
AMPCOR	Applies amplitude correction at specified frequencies	N/A	N/A	No	No	
AMPCORDATA	Amplitude correction data	N/A	N/A	N/A	No	
AMPCORSAVE	Save amplitude correction data	N/A	N/A	N/A	No	
AMPCORSIZE	Amplitude correction data array size	N/A	N/A	N/A	No	
AMPCORRCL	Amplitude correction recall	N/A	N/A	N/A	No	
AMPLEN	Amplitude correction length	N/A	N/A	No	N/A	
ANLGPLUS	Turns on or off the Analog+ display mode	N/A	N/A	No	N/A	
ANNOT	Display Annotation	Yes	Yes	Yes	Yes	<a href="#">Page 117</a>
APB	Trace A + trace B -> trace A	Yes	Yes	Yes	Yes	<a href="#">Page 118</a>
ARRAYDEF	Defines an array	N/A	N/A	N/A	No	
AT	Input Attenuation	Yes	Yes	Yes	Yes	<a href="#">Page 119</a>
AUNITS	Amplitude Units	Yes	Yes	Yes	Yes	<a href="#">Page 121</a>

**Table 2-1                    Alphanumeric List of all Legacy Commands Showing their  
Option 266 Support**

Command	Description	8566	8568	8590 Series	8560 Series	Page for Further Details
AUTO	Auto couple	N/A	N/A	Yes	N/A	<a href="#">Page 123</a>
AUTOCP	Auto couple	N/A	N/A	N/A	Yes	<a href="#">Page 125</a>
AUTOEXEC	Turns on or off the function defined with AUTOFUNC	N/A	N/A	N/A	No	
AUTOFUNC	Defines a function for automatic execution	N/A	N/A	N/A	No	
AUTOSAVE	Automatically saves trace	N/A	N/A	N/A	No	
AVG	Average	No	No	No	No	
AXB	Exchange Traces A & B	Yes	Yes	Yes	Yes	<a href="#">Page 126</a>
B1	Clear-writes trace B	Yes	Yes	Yes	Yes	<a href="#">Page 127</a>
B2	Max Holds trace B	Yes	Yes	Yes	Yes	<a href="#">Page 128</a>
B3	View trace B	Yes	Yes	Yes	Yes	<a href="#">Page 129</a>
B4	Blanks trace B	Yes	Yes	Yes	Yes	<a href="#">Page 130</a>
BAUDRATE	Baud rate of spectrum analyzer	N/A	N/A	No	N/A	
BIT	Return or receive state of bit	N/A	N/A	No	N/A	
BITF	Bit flag	N/A	N/A	No	N/A	
BL	Trace B – Display line -> trace B	Yes	Yes	Yes	N/A	<a href="#">Page 131</a>
BLANK	Blanks specified trace	Yes	Yes	Yes	Yes	<a href="#">Page 132</a>
BML	Trace B – Display line -> trace B	Yes	Yes	Yes	Yes	<a href="#">Page 133</a>
BRD	Bus Read	No	No	N/A	N/A	
BTC	Transfer trace B to C	Yes	Yes	Yes	N/A	<a href="#">Page 134</a>
BWR	Bus Write	No	No	N/A	N/A	
BXC	Exchange Traces B & C	Yes	Yes	Yes	N/A	<a href="#">Page 135</a>
C1	Turns off A - B	Yes	Yes	Yes	Yes	<a href="#">Page 136</a>
C2	A – B -> A	Yes	Yes	Yes	Yes	<a href="#">Page 137</a>
CA	Couples Attenuation	Yes	Yes	Yes	Yes	<a href="#">Page 138</a>
CAL	Calibrate	N/A	N/A	Yes	N/A	<a href="#">Page 139</a>

**Table 2-1 Alphanumeric List of all Legacy Commands Showing their Option 266 Support**

Command	Description	8566	8568	8590 Series	8560 Series	Page for Further Details
CARDLOAD	Copies data from memory card to module memory	N/A	N/A	N/A	No	
CARDSTORE	Copies data to memory card	N/A	N/A	N/A	No	
CARROFF	Carrier off power	N/A	N/A	N/A	No	
CARRON	Carrier on power	N/A	N/A	N/A	Yes	<a href="#">Page 140</a>
CAT	Catalog	N/A	N/A	No	N/A	
CATALOG	Catalog	N/A	N/A	N/A	No	
CF	Center Frequency	Yes	Yes	Yes	Yes	<a href="#">Page 141</a>
CHANNEL	Channel selection	N/A	N/A	N/A	Yes	<a href="#">Page 142</a>
CHANPWR	Channel power	N/A	N/A	N/A	Yes	<a href="#">Page 143</a>
CHP	Performs the channel power measurement	N/A	N/A	Yes	N/A	<a href="#">Page 144</a>
CHPGR	Channel power graph on or off	N/A	N/A	No	N/A	
CHPWRBW	Channel power bandwidth	N/A	N/A	N/A	Yes	<a href="#">Page 145</a>
CLRAVG	Reset avg. counter to 1	Yes	Yes	Yes	N/A	<a href="#">Page 146</a>
CLRBOX	Clears a rectangular area on the analyzer display	N/A	N/A	No	N/A	
CLRDSP	Clear display	N/A	N/A	No	No	
CLRSCHED	Clears autosave & autoexec schedule buffer	N/A	N/A	N/A	No	
CLRW	Clear-writes specified trace	Yes	Yes	Yes	Yes	<a href="#">Page 147</a>
CLS	Clear status byte	N/A	N/A	Yes	N/A	<a href="#">Page 148</a>
CMDERRQ	Command error query	N/A	N/A	No	N/A	
CNF	Confidence test	N/A	N/A	No	N/A	
CNTLA	Auxiliary interface control line A	N/A	N/A	No	No	
CNTLB	Auxiliary interface control line B	N/A	N/A	No	No	
CNTLC	Auxiliary interface control line C	N/A	N/A	No	No	



**Table 2-1            Alphanumeric List of all Legacy Commands Showing their  
Option 266 Support**

Command	Description	8566	8568	8590 Series	8560 Series	Page for Further Details
CNTLD	Auxiliary interface control line D	N/A	N/A	No	No	
CNTLI	Auxiliary interface control line input	N/A	N/A	No	No	
CNVLOSS	Selects ref level offset to calibrate amplitude display	No	N/A	N/A	No	
COMB	Turns the comb generator on or off	N/A	N/A	No	N/A	
COMPRESS	Compress	No	No	No	N/A	
CONCAT	Concat	No	No	No	N/A	
CONTS	Continuous sweep mode	Yes	Yes	Yes	Yes	<a href="#">Page 149</a>
CORREK	Correction factors on	N/A	N/A	Yes	N/A	<a href="#">Page 150</a>
COUPLE	Selects AC or DC coupling	N/A	N/A	Yes	Yes	<a href="#">Page 151</a>
CR	Couples Resolution BW	Yes	Yes	Yes	Yes	<a href="#">Page 152</a>
CRTHPOS	Horizontal position of CRT display	N/A	N/A	No	N/A	
CRTVPOS	Vertical position of CRT display	N/A	N/A	No	N/A	
CS	Couples Step Size	Yes	Yes	Yes	Yes	<a href="#">Page 153</a>
CT	Couples Sweep Time	Yes	Yes	Yes	Yes	<a href="#">Page 154</a>
CTA	Converts display units to dBm	No	No	No	N/A	
CTM	Converts dBm to display units	No	No	No	N/A	
CTRLHPIB	Allows SA to control HP-IB	N/A	N/A	N/A	No	
CV	Couples Video Bandwidth	Yes	Yes	Yes	Yes	<a href="#">Page 155</a>
D1	Sets display to normal size	No	No	N/A	N/A	
D2	Sets display to full CRT size	No	No	N/A	N/A	
D3	Sets display to expanded size	No	No	N/A	N/A	
DA	Display Memory Address	Yes	Yes	Yes	N/A	<a href="#">Page 156</a>
DATEMODE	Set the date display format	N/A	N/A	No	No	

**Table 2-1 Alphanumeric List of all Legacy Commands Showing their Option 266 Support**

Command	Description	8566	8568	8590 Series	8560 Series	Page for Further Details
DD	Display write binary	No	No	N/A	N/A	
DELMKBW	Occupied power bandwidth within delta marker	N/A	N/A	N/A	Yes	<a href="#">Page 157</a>
DEMOD	Turns the demodulator on or off	N/A	N/A	No	No	
DEMODAGC	Demodulation automatic gain control	N/A	N/A	N/A	No	
DEMODT	Demodulation time	N/A	N/A	N/A	No	
DET	Detection Mode	Yes	Yes	Yes	Yes	<a href="#">Page 158</a>
DISPOSE	Frees Memory	No	No	No	No	
DIV	Divide	No	No	No	No	
DL	Display Line Level	Yes	Yes	Yes	Yes	<a href="#">Page 160</a>
DLE	Turns the display line on/off	Yes	Yes	N/A	N/A	<a href="#">Page 162</a>
DLYSWP	Delay sweep	N/A	N/A	N/A	Yes	<a href="#">Page 163</a>
DN	Reduces the active function by applicable step size	N/A	N/A	No	N/A	
DONE	Synchronizing function	Yes	Yes	Yes	Yes	<a href="#">Page 164</a>
DOTDENS	Sets the dot density value in Analog+ display mode	N/A	N/A	No	N/A	
DR	Display Memory Address Read	No	No	N/A	N/A	
DRAWBOX	Draws a rectangular box on analyzer display	N/A	N/A	No	N/A	
DSPLY	Display	No	No	No	No	
DT	Define Terminator	No	No	No	N/A	
DW	Display Memory Address Write	No	No	N/A	N/A	
E1	Active marker to maximum signal	Yes	Yes	Yes	Yes	<a href="#">Page 165</a>
E2	Active marker to center frequency	Yes	Yes	Yes	Yes	<a href="#">Page 166</a>
E3	Active marker frequency to CF step size	Yes	Yes	Yes	Yes	<a href="#">Page 167</a>

**Table 2-1            Alphanumeric List of all Legacy Commands Showing their  
Option 266 Support**

Command	Description	8566	8568	8590 Series	8560 Series	Page for Further Details
E4	Active marker to reference level	Yes	Yes	Yes	Yes	<a href="#">Page 168</a>
EDITDONE	Indicates limit line editing is complete	N/A	N/A	N/A	Yes	<a href="#">Page 169</a>
EDITLIML	Allows current limit line to be edited	N/A	N/A	N/A	Yes	<a href="#">Page 170</a>
EE	Enable entry	No	No	No	N/A	
EK	Enable knob	No	No	No	N/A	
ELSE	Conditional Programming (If...then...else...endif)	No	No	N/A	No	
EM	Erase trace C memory	No	No	No	No	
ENDIF	Conditional Programming (If...then...else...endif)	No	No	N/A	N/A	
ENTER	Enter from HP-IB	No	No	No	No	
EP	Enter parameter function	N/A	N/A	No	N/A	
ERASE	User memory & registers erased	No	No	No	N/A	
ERR	Queries the error queue	Yes	Yes	N/A	Yes	<a href="#">Page 171</a>
ET	Elapsed time	N/A	N/A	N/A	Yes	<a href="#">Page 172</a>
EX	Exchanges trace A & B	Yes	Yes	Yes	Yes	<a href="#">Page 173</a>
EXP	Exponential	No	No	No	No	
EXTMXR	Presets external mixing mode	No	N/A	N/A	No	
FA	Start frequency	Yes	Yes	Yes	Yes	<a href="#">Page 174</a>
FB	Stop frequency	Yes	Yes	Yes	Yes	<a href="#">Page 175</a>
FDIAG	Frequency diagnostics	N/A	N/A	N/A	No	
FDSP	Frequency display off	N/A	N/A	N/A	Yes	<a href="#">Page 176</a>
FFT	Fast fourier transform	No	No	No	No	
FFTAUTO	Marker to Auto FFT	N/A	N/A	No	N/A	
FFTCLIP	FFT signal clipped	N/A	N/A	No	N/A	
FFTCONTS	FFT continuous sweep	N/A	N/A	No	N/A	

**Table 2-1 Alphanumeric List of all Legacy Commands Showing their Option 266 Support**

Command	Description	8566	8568	8590 Series	8560 Series	Page for Further Details
FFTKNL	Fast fourier transform kernel	No	No	N/A	N/A	
FFTMKR	FFT markers	N/A	N/A	No	N/A	
FFTMM	FFT marker to midscreen	N/A	N/A	No	N/A	
FFTMS	FFT marker to FFT stop frequency	N/A	N/A	No	N/A	
FFTOFF	FFT off	N/A	N/A	No	N/A	
FFTPCTAM	FFT percent amplitude modulation	N/A	N/A	No	N/A	
FFTPCTAMR	FFT percent amplitude modulation readout	N/A	N/A	No	N/A	
FFTSNGLS	FFT single sweep	N/A	N/A	No	N/A	
FFTSTAT	FFT status	N/A	N/A	No	N/A	
FFTSTOP	FFT stop frequency	N/A	N/A	No	N/A	
FMGAIN	FM gain	N/A	N/A	No	N/A	
FOFFSET	Frequency offset	Yes	Yes	Yes	Yes	<a href="#">Page 177</a>
FORMAT	Erase & format the selected memory device	N/A	N/A	No	No	
FPKA	Fast preselector peak	Yes	N/A	N/A	N/A	<a href="#">Page 178</a>
FREF	Frequency reference	N/A	N/A	N/A	Yes	<a href="#">Page 179</a>
FS	Full frequency span	Yes	Yes	Yes	Yes	<a href="#">Page 180</a>
FULBAND	Set start/stop freq for ext mixing bands	No	N/A	N/A	No	
FUNCDEF	Function definition	No	No	No	No	
GATE	Turn time-gating on or off	N/A	N/A	No	No	
GATECTL	Gate control	N/A	N/A	No	No	
GC	Gate preset	N/A	N/A	No	N/A	
GD	Gate delay	N/A	N/A	No	No	
GDRVCLPAR	Clear pulse parameters	N/A	N/A	No	N/A	
GDRVGDEL	Gate Delay for the frequency window	N/A	N/A	No	N/A	

**Table 2-1                    Alphanumeric List of all Legacy Commands Showing their  
Option 266 Support**

Command	Description	8566	8568	8590 Series	8560 Series	Page for Further Details
GDRVLEN	Gate length for frequency & time windows	N/A	N/A	No	N/A	
GDRVGT	Turns gate in frequency window on or off	N/A	N/A	No	N/A	
GDRVGTIM	Gate trigger to marker position for time window	N/A	N/A	No	N/A	
GDRVPRI	Pulse repetition interval	N/A	N/A	No	N/A	
GDRVPWID	Pulse width	N/A	N/A	No	N/A	
GDRVRBW	Couple resolution bandwidth to pulse width	N/A	N/A	No	N/A	
GDRVREFE	Enter reference edge	N/A	N/A	No	N/A	
GDRVST	Couple sweep time to pulse repetition interval	N/A	N/A	No	N/A	
GDRVSWAP	Update the time or frequency window	N/A	N/A	No	N/A	
GDRVSWDE	Delay sweep for time window	N/A	N/A	No	N/A	
GDRVSWP	Sweep time for the time window	N/A	N/A	No	N/A	
GDRVUTIL	Turns the gate utility on or off	N/A	N/A	No	N/A	
GDRVVBW	Couple video bandwidth to the gate length	N/A	N/A	No	N/A	
GETPLOT	Get plot	N/A	N/A	No	N/A	
GETPRNT	Get print	N/A	N/A	No	N/A	
GL	Gate length	N/A	N/A	No	No	
GP	Sets the polarity (positive/negative) of the gate trigger	N/A	N/A	No	No	
GR	Plot GPIB input as Graphs	No	No	No	N/A	
GRAT	Graticule on/off	Yes	Yes	Yes	Yes	<a href="#">Page 184</a>
HAVE	Checks for options installed	N/A	N/A	No	N/A	
HD	Holds data entry	No	No	Yes	No	<a href="#">Page 185</a>

**Table 2-1 Alphanumeric List of all Legacy Commands Showing their Option 266 Support**

Command	Description	8566	8568	8590 Series	8560 Series	Page for Further Details
HN	Harmonic number	N/A	N/A	No	N/A	
HNLOCK	Harmonic lock	No	N/A	No	No	
HNUNLK	Harmonic band unlock	No	N/A	No	No	
IB	Input to trace B memory	No	No	No	N/A	
I1	Sets the RF coupling to AC	N/A	Yes	N/A	N/A	<a href="#">Page 186</a>
I2	Sets the RF coupling to DC	N/A	Yes	N/A	N/A	<a href="#">Page 188</a>
ID	Instrument identification	Yes	Yes	Yes	Yes	<a href="#">Page 190</a>
IDCF	Identified signal to center frequency	N/A	N/A	N/A	No	
IDFREQ	Identified signal frequency	N/A	N/A	N/A	No	
IDSTAT	Signal identifier status	No	N/A	N/A	N/A	
IF	Conditional Programming (If...then...else...endif)	No	No	No	No	
IFTKNL	16 bit discrete fourier transform	No	No	N/A	N/A	
INT	Integer	No	No	No	No	
INZ	Input impedance	N/A	N/A	No	N/A	
IP	Instrument preset	Yes	Yes	Yes	Yes	<a href="#">Page 191</a>
KEYCLR	Clear user defined keys	N/A	N/A	No	No	
KEYCMD	Define function & label of softkey	N/A	N/A	No	N/A	
KEYDEF	Assign function to soft key	No	No	No	No	
KEYENH	Key enhance	N/A	N/A	No	N/A	
KEYEXC	Executes specified soft key	No	No	No	N/A	
KEYLBL	Relabels softkey without changing its function	N/A	N/A	No	N/A	
KS,	Mixer level	Yes	Yes	N/A	N/A	<a href="#">Page 192</a>
KS=	HP8566: Selects factory preselector setting HP8568: Marker counter frequency resolution	No	Yes	No	No	<a href="#">Page 193</a>
KS(	Locks the save registers	No	No	N/A	N/A	

**Table 2-1                    Alphanumeric List of all Legacy Commands Showing their  
Option 266 Support**

Command	Description	8566	8568	8590 Series	8560 Series	Page for Further Details
KS)	Unlocks the save registers	No	No	N/A	N/A	
KS>	Specifies preamp gain for signal input 2	N/A	No	N/A	N/A	
KS<	Specifies preamp gain for signal input 1	N/A	No	N/A	N/A	
KS	Display memory address write	No	No	N/A	N/A	
KS#	Turns off YTX self-heating correction	No	N/A	N/A	N/A	
KS/	Allows preselector to be peaked manually	No	N/A	N/A	N/A	
KS123	Returns up to 1001 words display memory	No	No	N/A	N/A	
KS125	Writes up to 1001 display memory words	No	No	N/A	N/A	
KS126	Returns every Nth value of a trace	No	No	N/A	N/A	
KS127	Sets analyzer to accept binary display write	No	No	N/A	N/A	
KS39	Writes display memory address in fast binary	No	No	N/A	N/A	
KS43	Sets SRQ 102 when frequency limit exceeded	No	No	N/A	N/A	
KS91	Returns the amplitude error	No	No	N/A	N/A	
KS92	Specifies value DL, TH, active mkr in display units	No	No	N/A	N/A	
KS94	Returns code for harmonic number in binary	No	No	N/A	N/A	
KSA	Sets amplitude units to dBm	Yes	Yes	Yes	N/A	<a href="#">Page 194</a>
KSa	Selects normal detection	Yes	Yes	N/A	N/A	<a href="#">Page 195</a>
KSB	Sets amplitude units to dBmV	Yes	Yes	Yes	N/A	<a href="#">Page 196</a>

**Table 2-1 Alphanumeric List of all Legacy Commands Showing their Option 266 Support**

Command	Description	8566	8568	8590 Series	8560 Series	Page for Further Details
KSb	Selects positive peak detection	Yes	Yes	N/A	N/A	<a href="#">Page 197</a>
KSC	Sets amplitude units to dBuV	Yes	Yes	Yes	N/A	<a href="#">Page 198</a>
KSc	Trace A + trace B -> trace A	Yes	Yes	Yes	N/A	<a href="#">Page 199</a>
KSD	Sets amplitude units to V	Yes	Yes	Yes	N/A	<a href="#">Page 200</a>
KSd	Selects negative peak detection	Yes	Yes	N/A	N/A	<a href="#">Page 201</a>
KSE	Sets the analyzer title mode	Yes	Yes	Yes	N/A	<a href="#">Page 202</a>
KSe	Selects sample detection	Yes	Yes	N/A	N/A	<a href="#">Page 203</a>
KSF	HP8566: Shifts the YTO HP8568: Measures the Sweep Time	No	No	N/A	N/A	
KSf	Recover last instrument state at power on	No	No	N/A	N/A	
KSG	Turns on video averaging	Yes	Yes	Yes	N/A	<a href="#">Page 204</a>
KSG	Turns off the display	Yes	Yes	N/A	N/A	<a href="#">Page 205</a>
KSH	Turns off video averaging	Yes	Yes	Yes	N/A	<a href="#">Page 206</a>
KSh	Turns on the display	Yes	Yes	N/A	N/A	<a href="#">Page 207</a>
KSI	Allows the reference level to be extended	No	No	N/A	N/A	<a href="#">Page 208</a>
KSi	Exchanges traces B & C	Yes	Yes	Yes	N/A	<a href="#">Page 209</a>
KSJ	Manual control of DACs	No	No	N/A	N/A	
KSj	Views trace C	Yes	Yes	N/A	N/A	<a href="#">Page 210</a>
KSK	HP8566: Active Mkr to next highest peak HP8568: Counts pilot IF at marker	Yes	No	N/A	N/A	<a href="#">Page 211</a>
KSk	Blanks trace C	Yes	Yes	N/A	N/A	<a href="#">Page 212</a>
KSL	Turns off marker noise function	Yes	Yes	N/A	N/A	<a href="#">Page 213</a>
KSl	Moves trace B into trace C	Yes	Yes	Yes	N/A	<a href="#">Page 214</a>



**Table 2-1      Alphanumeric List of all Legacy Commands Showing their  
Option 266 Support**

Command	Description	8566	8568	8590 Series	8560 Series	Page for Further Details
KSM	Turns on marker noise function	Yes	Yes	Yes	N/A	<a href="#">Page 215</a>
KSm	Turns off the graticule	Yes	Yes	Yes	N/A	<a href="#">Page 217</a>
KSN	Marker minimum value detected	Yes	No	N/A	N/A	<a href="#">Page 218</a>
KSn	Turns on the graticule	Yes	Yes	Yes	N/A	<a href="#">Page 219</a>
KSO	Marker span	Yes	Yes	Yes	N/A	<a href="#">Page 220</a>
KSo	Turns off the annotation	Yes	Yes	Yes	N/A	<a href="#">Page 221</a>
KSP	GPIB address	No	No	N/A	N/A	
KSp	Turns on the annotation	Yes	Yes	Yes	N/A	<a href="#">Page 222</a>
KSQ	Unlocks frequency band	No	No	N/A	N/A	
KSq	Decouples IF gain and input attenuation	No	No	N/A	N/A	
KSR	Turns on service diagnostics	No	No	N/A	N/A	
KSr	Sets service request 102	No	No	N/A	N/A	
KSS	HP8566: Fast GPIB operation HP8568: Determine second LO frequency	No	No	N/A	N/A	
KST	HP8566: Fast preset HP8568: Shifts second LO down	Yes	No	N/A	N/A	<a href="#">Page 223</a>
KSt	HP8566: Locks frequency band HP8568: Continues sweep from marker	No	No	N/A	N/A	
KSU	HP8566: External mixer preset HP8568: Shift second LO up	No	No	N/A	N/A	
KSu	Stops the sweep at the active marker	No	No	N/A	N/A	
KSV	Frequency offset	Yes	Yes	N/A	N/A	<a href="#">Page 224</a>

Legacy Analyzer Command List

**Table 2-1 Alphanumeric List of all Legacy Commands Showing their Option 266 Support**

Command	Description	8566	8568	8590 Series	8560 Series	Page for Further Details
KSv	HP8566: External mixer frequency identifier HP8568: Inhibits phase lock	No	No	N/A	N/A	
KSW	Amplitude error correction routine	No	No	N/A	N/A	
KSw	Displays amplitude error correction routine	No	No	N/A	N/A	
KSX	Amplitude correction factors on	No	No	N/A	N/A	
KSx	Sets trigger mode to external	Yes	Yes	N/A	N/A	<a href="#">Page 225</a>
KSY	Amplitude correction factors off	No	No	N/A	N/A	
KSy	Sets trigger mode to video	Yes	Yes	N/A	N/A	<a href="#">Page 226</a>
KSZ	Reference level offset	Yes	Yes	Yes	N/A	<a href="#">Page 227</a>
KSz	Sets the display storage address	No	No	N/A	N/A	
L0	Turns off the display line	Yes	Yes	Yes	Yes	<a href="#">Page 228</a>
LB	Writes text label	No	No	No	No	
LCLVAR	Defines a local variable for use	N/A	N/A	N/A	No	
LF	Preset 0–2.5GHz	Yes	N/A	Yes	N/A	<a href="#">Page 229</a>
LG	Selects log scale	Yes	Yes	Yes	Yes	<a href="#">Page 230</a>
LIMD	Delta amplitude value for limit line segment	N/A	N/A	N/A	Yes	<a href="#">Page 232</a>
LIMF	Frequency value for limit-line segment	N/A	N/A	N/A	Yes	<a href="#">Page 233</a>
LIMIDEL	Erase contents of limit line table	N/A	N/A	Yes	N/A	<a href="#">Page 234</a>
LIMIDISP	Controls when the limit line(s) are displayed	N/A	N/A	Yes	N/A	<a href="#">Page 235</a>
LIMIFAIL	Limit line fail	N/A	N/A	Yes	Yes	<a href="#">Page 236</a>

**Table 2-1                    Alphanumeric List of all Legacy Commands Showing their  
Option 266 Support**

Command	Description	8566	8568	8590 Series	8560 Series	Page for Further Details
LIMIFT	Select frequency or time limit line	N/A	N/A	Yes	N/A	<a href="#">Page 237</a>
LIMIHI	Upper limit	N/A	N/A	No	N/A	
LIMILINE	Limit line	N/A	N/A	No	N/A	
LIMILO	Lower limit	N/A	N/A	No	N/A	
LIMIMIRROR	Mirror limit line	N/A	N/A	No	N/A	
LIMIMODE	Limit line entry mode	N/A	N/A	No	N/A	
LIMIPURGE	Disposes of current limit line, not limit line table	N/A	N/A	N/A	Yes	<a href="#">Page 238</a>
LIMIRCL	Load stored limit line into limit line table	N/A	N/A	N/A	No	
LIMIREL	Determine whether limit line values absolute/relative	N/A	N/A	Yes	Yes	<a href="#">Page 239</a>
LIMISAV	Save contents of limit line table for recall	N/A	N/A	N/A	No	
LIMISEG	Define slope & offset of limit line segments	N/A	N/A	No	N/A	
LIMISEGT	Enter limit line segment for sweep time	N/A	N/A	No	N/A	
LIMITEST	Compare active trace data to limit line parameters	N/A	N/A	Yes	Yes	<a href="#">Page 240</a>
LIML	Amplitude value for limit line segment in lower limit line	N/A	N/A	N/A	Yes	<a href="#">Page 241</a>
LIMM	Middle amplitude value for limit-line segment	N/A	N/A	N/A	Yes	<a href="#">Page 242</a>
LIMTFL	Specifies a flat limit-line segment	N/A	N/A	N/A	No	
LIMTSL	Specifies a sloped limit-line segment	N/A	N/A	N/A	No	
LIMU	Amplitude value for limit line segment in upper limit line	N/A	N/A	N/A	Yes	<a href="#">Page 243</a>
LINFILL	Line fill	N/A	N/A	No	N/A	

**Table 2-1 Alphanumeric List of all Legacy Commands Showing their Option 266 Support**

Command	Description	8566	8568	8590 Series	8560 Series	Page for Further Details
LL	Provides lower left recorder output voltage at rear	No	No	N/A	N/A	
LN	Selects linear scale	Yes	Yes	Yes	Yes	<a href="#">Page 244</a>
LOAD	Load article/file into internal memory	N/A	N/A	No	N/A	
LOG	Log	No	No	No	No	
LOLIMOFF	LO Limit Off	No	No	N/A	N/A	
LSPAN	Last span	N/A	N/A	Yes	N/A	<a href="#">Page 245</a>
M1	Turns off all markers	Yes	Yes	Yes	Yes	<a href="#">Page 246</a>
M2	Marker Normal	Yes	Yes	Yes	Yes	<a href="#">Page 247</a>
M3	Marker Delta	Yes	Yes	Yes	Yes	<a href="#">Page 248</a>
M4	Marker zoom	Yes	Yes	Yes	N/A	<a href="#">Page 250</a>
MA	Returns the amplitude of active marker	Yes	Yes	Yes	Yes	<a href="#">Page 251</a>
MBIAS	Mixer bias	N/A	N/A	N/A	No	
MBRD	Processor memory block read	No	No	N/A	N/A	
MBWR	Processor memory block write	No	No	N/A	N/A	
MC0	Turns off the marker frequency counter	N/A	Yes	Yes	N/A	<a href="#">Page 252</a>
MC1	Turns on the marker frequency counter	N/A	Yes	Yes	N/A	<a href="#">Page 253</a>
MDS	Measurement data size	Yes	Yes	Yes	N/A	<a href="#">Page 254</a>
MDU	Measurement data units	No	No	No	N/A	
MEAN	Returns mean value of trace in display units	Yes	Yes	Yes	Yes	<a href="#">Page 255</a>
MEANPWR	Mean power measurement	N/A	N/A	N/A	Yes	<a href="#">Page 256</a>
MEANTH	Trace mean above threshold	N/A	N/A	No	N/A	
MEAS	Measurement status	N/A	N/A	N/A	No	
MEASOFF	Measurement off	N/A	N/A	Yes	N/A	<a href="#">Page 257</a>
MEASURE	Measure mode	N/A	N/A	No	N/A	

**Table 2-1            Alphanumeric List of all Legacy Commands Showing their  
Option 266 Support**

Command	Description	8566	8568	8590 Series	8560 Series	Page for Further Details
MEM	Returns amount of memory available	No	No	No	No	
MENU	Menu	N/A	N/A	No	No	
MERGE	Merge two traces	No	No	No	N/A	
MF	Returns frequency of the active marker	Yes	Yes	Yes	Yes	<a href="#">Page 258</a>
MIN	Minimum	No	No	No	No	
MINH	Min Hold	N/A	N/A	Yes	Yes	<a href="#">Page 259</a>
MINPOS	Returns the minimum position in the trace	Yes	Yes	Yes	N/A	<a href="#">Page 260</a>
MIRROR	Mirror image of the trace	No	No	No	N/A	
MKA	Amplitude of the active marker	Yes	Yes	Yes	Yes	<a href="#">Page 261</a>
MKACT	Specifies the active marker	Yes	Yes	Yes	N/A	<a href="#">Page 262</a>
MKACTV	Marker as the active function	N/A	N/A	No	N/A	
MKBW	Marker bandwidth	N/A	N/A	Yes	Yes	<a href="#">Page 263</a>
MKCF	Moves the active marker to center frequency	Yes	Yes	Yes	Yes	<a href="#">Page 264</a>
MKCHEDGE	Marker to channel edge	N/A	N/A	N/A	No	
MKCONT	Continues sweeping from the marker after stop	No	No	No	N/A	
MKD	Delta marker	Yes	Yes	Yes	Yes	<a href="#">Page 265</a>
MKDELCHBW	Delta markers to channel power bandwidth	N/A	N/A	N/A	No	
MKDLMODE	Marker delta display line mode	N/A	N/A	No	N/A	
MKDR	Reciprocal of marker delta	N/A	N/A	N/A	No	
MKF	Specifies the frequency of the active marker	Yes	Yes	Yes	Yes	<a href="#">Page 267</a>
MKFC	Turns the marker frequency counter on or off	N/A	Yes	Yes	Yes	<a href="#">Page 268</a>

**Table 2-1 Alphanumeric List of all Legacy Commands Showing their Option 266 Support**

Command	Description	8566	8568	8590 Series	8560 Series	Page for Further Details
MKFCR	Specifies the marker frequency counter resolution	N/A	Yes	Yes	Yes	<a href="#">Page 269</a>
MKMCF	Marker mean to center frequency	N/A	N/A	N/A	No	
MKMIN	Moves active marker to minimum signal detected	Yes	Yes	Yes	Yes	<a href="#">Page 270</a>
MKN	Normal marker	Yes	Yes	Yes	Yes	<a href="#">Page 271</a>
MKNOISE	Marker noise function	Yes	Yes	Yes	Yes	<a href="#">Page 272</a>
MKOFF	Turns all markers or the active marker off	Yes	Yes	Yes	Yes	<a href="#">Page 274</a>
MKP	Specifies the horizontal position of the marker	Yes	Yes	Yes	N/A	<a href="#">Page 275</a>
MKPAUSE	Pauses the sweep at the active marker	No	No	No	N/A	
MKPK	Marker peak	Yes	Yes	Yes	Yes	<a href="#">Page 276</a>
MKPT	Marker peak threshold	N/A	N/A	N/A	Yes	<a href="#">Page 277</a>
MKPX	Marker peak excursion	Yes	Yes	Yes	Yes	<a href="#">Page 278</a>
MKREAD	Specifies marker readout mode	Yes	Yes	Yes	N/A	<a href="#">Page 279</a>
MKRL	Moves the active marker to reference level	Yes	Yes	Yes	Yes	<a href="#">Page 281</a>
MKSP	Marker span	Yes	Yes	Yes	Yes	<a href="#">Page 282</a>
MKSS	Marker step size	Yes	Yes	Yes	Yes	<a href="#">Page 283</a>
MKSTOP	Stops the sweep at the active marker	No	No	No	N/A	
MKT	Position marker in units of time	N/A	N/A	N/A	Yes	<a href="#">Page 284</a>
MKTBL	Marker table	N/A	N/A	No	N/A	<a href="#">Page 285</a>
MKTRACE	Marker trace	Yes	Yes	Yes	N/A	<a href="#">Page 286</a>
MKTRACK	Turns the marker signal track on or off	Yes	Yes	Yes	Yes	<a href="#">Page 287</a>
MKTYPE	Specifies the type of active marker to be used	Yes	Yes	No	N/A	<a href="#">Page 288</a>

**Table 2-1                    Alphanumeric List of all Legacy Commands Showing their  
Option 266 Support**

Command	Description	8566	8568	8590 Series	8560 Series	Page for Further Details
ML	Mixer Level	Yes	Yes	Yes	Yes	<a href="#">Page 289</a>
MOD	Modulo	No	No	No	No	
MODE	Mode of operation	N/A	N/A	No	N/A	
MODRCLT	Recalls trace from module memory	N/A	N/A	N/A	No	
MODSAVT	Saves trace in module memory	N/A	N/A	N/A	No	
MOV	Move	No	No	No	No	
MPY	Multiply	No	No	No	No	
MRD	Memory Read	No	No	N/A	N/A	
MRDB	Memory read byte	No	No	N/A	N/A	
MSDEV	Specifies mass storage device	N/A	N/A	N/A	No	
MSI	Mass storage interface	N/A	N/A	No	N/A	
MT0	Turns off marker signal track	Yes	Yes	Yes	Yes	<a href="#">Page 291</a>
MT1	Turns on marker signal track	Yes	Yes	Yes	Yes	<a href="#">Page 292</a>
MWR	Memory Write	No	No	N/A	N/A	
MWRB	Memory write byte	No	No	N/A	N/A	
MXM	Maximum	No	No	No	No	
MXMH	Max Hold	Yes	Yes	Yes	Yes	<a href="#">Page 293</a>
MXRMODE	Mixer mode	N/A	N/A	N/A	No	
NDB	Number of dB	N/A	N/A	No	N/A	
NDBPNT	Turns the N dB points function on or off	N/A	N/A	No	N/A	
NDBPNTR	N dB points bandwidth	N/A	N/A	No	N/A	
NORMLIZE	Normalize trace data	N/A	N/A	N/A	No	
NRL	Normalized reference level	N/A	N/A	No	No	
NRPOS	Normalized reference position	N/A	N/A	N/A	No	

**Table 2-1 Alphanumeric List of all Legacy Commands Showing their Option 266 Support**

Command	Description	8566	8568	8590 Series	8560 Series	Page for Further Details
NSTART	Start harmonic	No	N/A	N/A	N/A	
NSTOP	Stop harmonic	No	N/A	N/A	N/A	
O1	Output format	Yes	Yes	Yes	N/A	<a href="#">Page 294</a>
O2	Output format	Yes	Yes	Yes	N/A	<a href="#">Page 295</a>
O3	Output format	Yes	Yes	Yes	N/A	<a href="#">Page 296</a>
O4	Output format	Yes	Yes	Yes	N/A	<a href="#">Page 297</a>
OA	Returns the active function value	N/A	N/A	No	N/A	
OBW	Occupied bandwidth	N/A	N/A	No	N/A	
OBWBW	Bandwidth measured by occupied bandwidth	N/A	N/A	No	N/A	
OBWFERR	Occupied bandwidth transmit frequency error	N/A	N/A	No	N/A	
OBWLOWER	Relative lower frequency limit of occupied bandwidth	N/A	N/A	No	N/A	
OBWPCT	Occupied bandwidth percent	N/A	N/A	No	N/A	
OBWPWR	Total power in the occupied bandwidth	N/A	N/A	No	N/A	
OBWUPPER	Relative upper frequency limit of occupied bandwidth	N/A	N/A	No	N/A	
OCCUP	Percent occupied power bandwidth	N/A	N/A	N/A	Yes	<a href="#">Page 298</a>
OL	Output learn string	Yes	Yes	No	N/A	<a href="#">Page 299</a>
ONCYCLE	On cycle	N/A	N/A	No	N/A	
ONDELAY	On delay	N/A	N/A	No	N/A	
ONEOS	On end of sweep	No	No	No	No	
ONMKR	On marker pause	N/A	N/A	No	N/A	
ONMKRU	On marker update	N/A	N/A	No	N/A	
ONPWRUP	On power up	N/A	N/A	No	N/A	
ONSRQ	On service request	N/A	N/A	No	N/A	
ONSWP	On sweep	No	No	No	N/A	



**Table 2-1            Alphanumeric List of all Legacy Commands Showing their  
Option 266 Support**

Command	Description	8566	8568	8590 Series	8560 Series	Page for Further Details
ONTIME	On time	N/A	N/A	No	N/A	
OP	Output parameters	No	No	No	No	
OR	Set position of origin	N/A	N/A	N/A	No	
OT	Output trace annotations	Yes	Yes	N/A	N/A	<a href="#">Page 300</a>
OUTPUT	Output - sending data to the GPIB from function	No	No	No	No	
PA	Plot absolute	No	No	No	No	
PARSTAT	Parallel status	N/A	N/A	No	N/A	
PCTAM	Turns the percent AM measurement on or off	N/A	N/A	No	N/A	
PCTAMR	Percent AM response	N/A	N/A	No	N/A	
PD	Pen down	No	No	No	No	
PDA	Probability distribution amplitude	No	No	No	No	
PDF	Probability distribution frequency	No	No	No	No	
PEAKS	Sorts the signal peaks by amplitude/frequency	Yes	Yes	Yes	Yes	<a href="#">Page 301</a>
PKDLMODE	Peak table delta display line mode	N/A	N/A	No	N/A	
PKPOS	Peak position	Yes	Yes	Yes	N/A	<a href="#">Page 302</a>
PKRES	Peak result	N/A	N/A	No	N/A	
PKSORT	Selects how to sort signal peaks listed in peak table	N/A	N/A	No	N/A	
PKTBL	Turns the peak table on or off	N/A	N/A	No	N/A	
PKZMOK	Peak zoom okay	N/A	N/A	No	N/A	
PKZOOM	Peak zoom	N/A	N/A	No	N/A	
PLOT	Prints the screen	Yes	Yes	Yes	Yes	<a href="#">Page 303</a>
PLOTORG	Display origins	N/A	N/A	N/A	No	
PLOTSRC	Plot source	N/A	N/A	N/A	No	
PLTPRT	Plot port	N/A	N/A	No	N/A	

**Table 2-1 Alphanumeric List of all Legacy Commands Showing their Option 266 Support**

Command	Description	8566	8568	8590 Series	8560 Series	Page for Further Details
POWERON	Power on state	N/A	N/A	No	N/A	
PP	Peaks the preselector	Yes	N/A	Yes	Yes	<a href="#">Page 304</a>
PR	Plot relative	No	No	No	No	
PREAMPG	External preamplifier gain	N/A	N/A	Yes	N/A	<a href="#">Page 305</a>
PREFX	Change user memory entries file prefix	N/A	N/A	No	N/A	
PRINT	Print	N/A	N/A	Yes	Yes	<a href="#">Page 306</a>
PRNPRT	Print port	N/A	N/A	No	N/A	
PRNTADRS	Print address	N/A	N/A	No	N/A	
PS	Skip page	No	No	N/A	N/A	
PSDAC	Preselector DAC number	N/A	N/A	N/A	No	
PSTATE	Protect state	N/A	N/A	No	No	
PU	Pen up	No	No	No	No	
PURGE	Purge file	N/A	N/A	No	N/A	
PWRBW	Power bandwidth	Yes	Yes	Yes	Yes	<a href="#">Page 307</a>
PWRUPTIME	Power up time	N/A	N/A	Yes	N/A	<a href="#">Page 308</a>
Q0	Sets detector to EMI Peak detection	Yes	Yes	N/A	N/A	<a href="#">Page 309</a>
Q1	Sets detector to Quasi Peak detection	Yes	Yes	N/A	N/A	<a href="#">Page 310</a>
R1	Resets service request 140	Yes	Yes	Yes	N/A	<a href="#">Page 311</a>
R2	Allows service request 140 & 104	Yes	Yes	Yes	N/A	<a href="#">Page 312</a>
R3	Allows service request 140 & 110	Yes	Yes	Yes	N/A	<a href="#">Page 313</a>
R4	Allows service request 140 & 102	Yes	Yes	Yes	N/A	<a href="#">Page 314</a>
RB	Resolution bandwidth	Yes	Yes	Yes	Yes	<a href="#">Page 315</a>
RBR	Resolution bandwidth/Span ratio	N/A	N/A	N/A	Yes	<a href="#">Page 317</a>
RC	Recalls state register	Yes	Yes	Yes	Yes	<a href="#">Page 318</a>

**Table 2-1            Alphanumeric List of all Legacy Commands Showing their  
Option 266 Support**

Command	Description	8566	8568	8590 Series	8560 Series	Page for Further Details
RCLOSCAL	Recall open/short average	N/A	N/A	N/A	No	
RCLS	Recall state	Yes	Yes	Yes	Yes	<a href="#">Page 319</a>
RCLT	Recall trace	N/A	N/A	No	No	
RCLTHRU	Recall internal thru-reference trace into trace B	N/A	N/A	N/A	No	
RELHPIB	Release control of GPIB	N/A	N/A	No	No	
REPEAT	Conditional Programming (Repeat .. Until ...)	No	No	No	No	
RESETRL	Reset reference level	N/A	N/A	No	N/A	
RETURN	Return to user defined function origination point	N/A	N/A	No	No	
REV	Returns the revision string to the controller	Yes	Yes	Yes	Yes	<a href="#">Page 320</a>
RL	Reference level	Yes	Yes	Yes	Yes	<a href="#">Page 321</a>
RLCAL	Reference level calibration	N/A	N/A	N/A	No	
RLPOS	Reference level position	N/A	N/A	No	N/A	
RMS	Root mean square	Yes	Yes	Yes	Yes	<a href="#">Page 323</a>
ROFFSET	Reference level offset	Yes	Yes	Yes	Yes	<a href="#">Page 324</a>
RQS	SRQ mask	Yes	Yes	Yes	Yes	<a href="#">Page 325</a>
S1	Continuous sweep mode	Yes	Yes	Yes	Yes	<a href="#">Page 326</a>
S2	Single sweep mode	Yes	Yes	Yes	Yes	<a href="#">Page 327</a>
SADD	Adds a limit line segment	N/A	N/A	N/A	Yes	<a href="#">Page 328</a>
SAVEMENU	Save menu	N/A	N/A	No	N/A	
SAVES	Saves analyzer state to specified register	Yes	Yes	Yes	Yes	<a href="#">Page 329</a>
SAVET	Save trace	N/A	N/A	No	No	
SAVRCLF	Save or recall flag	N/A	N/A	No	N/A	
SAVRCLN	Save or recall number	N/A	N/A	No	N/A	
SAVRCLW	Save or recall data	N/A	N/A	No	N/A	
SDEL	Deletes a limit line segment	N/A	N/A	N/A	Yes	<a href="#">Page 330</a>

Legacy Analyzer Command List

**Table 2-1 Alphanumeric List of all Legacy Commands Showing their Option 266 Support**

Command	Description	8566	8568	8590 Series	8560 Series	Page for Further Details
SDON	Indicates limit line segment is done	N/A	N/A	N/A	Yes	<a href="#">Page 331</a>
SEDI	Edits limit line segment	N/A	N/A	N/A	Yes	<a href="#">Page 332</a>
SEGDEL	Delete specified segment from limit line tables	N/A	N/A	No	N/A	
SENER	Segment entry for frequency limit lines	N/A	N/A	Yes	No	<a href="#">Page 333</a>
SENER_T	Segment entry for sweep time limit lines	N/A	N/A	No	N/A	
SER	Serial number	N/A	N/A	Yes	Yes	<a href="#">Page 334</a>
SETDATE	Set the date of spectrum analyzer	N/A	N/A	Yes	Yes	<a href="#">Page 335</a>
SETTIME	Set the time of spectrum analyzer	N/A	N/A	Yes	Yes	<a href="#">Page 336</a>
SHOWMENU	Shows menu	N/A	N/A	N/A	No	
SIGDEL	Signal amplitude delta	No	N/A	N/A	N/A	
SIGID	External mixing frequency bands signal identifier	No	N/A	N/A	No	
SKYCLR	Clears user softkey	N/A	N/A	N/A	No	
SKYDEF	Defines user softkey	N/A	N/A	N/A	No	
SMOOTH	Smooths given trace over specified number points	Yes	Yes	Yes	Yes	<a href="#">Page 337</a>
SNGLS	Single sweep mode	Yes	Yes	Yes	Yes	<a href="#">Page 338</a>
SP	Frequency Span	Yes	Yes	Yes	Yes	<a href="#">Page 339</a>
SPEAKER	Turns the internal speaker on or off	N/A	N/A	No	N/A	
SPZOOM	Span Zoom	N/A	N/A	No	N/A	
SQLCH	Sets the squelch threshold	N/A	N/A	No	N/A	
SQR	Square root	No	No	No	No	
SQUELCH	Adjusts squelch level	N/A	N/A	N/A	No	
SRCALC	Selects internal or external level control	N/A	N/A	No	No	

**Table 2-1                    Alphanumeric List of all Legacy Commands Showing their  
Option 266 Support**

<b>Command</b>	<b>Description</b>	<b>8566</b>	<b>8568</b>	<b>8590 Series</b>	<b>8560 Series</b>	<b>Page for Further Details</b>
SRCAT	Attenuate source output level	N/A	N/A	No	N/A	
SRCCRSTK	Coarse tracking adjust	N/A	N/A	N/A	No	
SRCFINTK	Fine tracking adjust	N/A	N/A	N/A	No	
SRCNORM	Source normalization	N/A	N/A	No	N/A	
SRCPOFS	Offset source power level	N/A	N/A	No	No	
SRCPSTP	Select source power step size	N/A	N/A	No	No	
SRCPSWP	Select sweep range of source output	N/A	N/A	No	No	
SRCPWR	Select source power level	N/A	N/A	No	No	
SRCTK	Adjust tracking of source output with SA sweep	N/A	N/A	No	N/A	
SRCTKPK	Auto adjust tracking of source output with SA sweep	N/A	N/A	No	No	
SRQ	Service request	Yes	Yes	Yes	Yes	<a href="#">Page 341</a>
SS	Frequency Step Size	Yes	Yes	Yes	Yes	<a href="#">Page 342</a>
ST	Sweep Time	Yes	Yes	Yes	Yes	<a href="#">Page 344</a>
STB	Status byte query	N/A	N/A	Yes	Yes	<a href="#">Page 346</a>
STDEV	Standard deviation of trace amplitude	Yes	Yes	Yes	Yes	<a href="#">Page 347</a>
STOR	Store file	N/A	N/A	No	N/A	
STOREOPEN	Save current instrument state	N/A	N/A	N/A	No	
STORESHORT	Store short	N/A	N/A	N/A	No	
STORETHRU	Store thru-calibration trace in trace B	N/A	N/A	N/A	No	
SUB	Subtract	No	No	No	No	
SUM	Sum of trace element amplitudes in display units	No	No	No	No	
SUMSQR	Squares trace element amplitudes & returns sum	No	No	No	No	

**Table 2-1 Alphanumeric List of all Legacy Commands Showing their Option 266 Support**

Command	Description	8566	8568	8590 Series	8560 Series	Page for Further Details
SV	Saves state	Yes	Yes	Yes	Yes	<a href="#">Page 348</a>
SW	Skip to next control instruction	No	No	N/A	N/A	
SWPCPL	Sweep couple	N/A	N/A	No	No	
SWPOUT	Sweep output	N/A	N/A	N/A	No	
SYNCMODE	Synchronize mode	N/A	N/A	No	N/A	
T0	Turns the threshold level off	No	No	No	N/A	
T1	Sets the trigger mode to free run	Yes	Yes	Yes	Yes	<a href="#">Page 349</a>
T2	Sets the trigger mode to line	Yes	Yes	Yes	Yes	<a href="#">Page 350</a>
T3	Sets the trigger mode to external	Yes	Yes	Yes	Yes	<a href="#">Page 351</a>
T4	Sets the trigger mode to video	Yes	Yes	Yes	Yes	<a href="#">Page 352</a>
T7	Sets the trigger mode to level	N/A	N/A	No	N/A	
T8	Sets the trigger mode to edge	N/A	N/A	No	N/A	
TA	Returns trace A amplitude values to controller	Yes	Yes	Yes	Yes	<a href="#">Page 353</a>
TB	Returns trace B amplitude values to controller	Yes	Yes	Yes	Yes	<a href="#">Page 354</a>
TDF	Trace data format	Yes	Yes	Yes	Yes	<a href="#">Page 355</a>
TEXT	Writes text on the analyzer screen	No	No	No	No	
TH	Threshold	Yes	Yes	Yes	Yes	<a href="#">Page 356</a>
THE	Turns the threshold on or off	No	No	N/A	N/A	
THEN	Conditional Programming (If...then...else...endif)	No	No	N/A	No	
TIMEDATE	Allows setting of time & date for analyzer	N/A	N/A	Yes	Yes	<a href="#">Page 357</a>

**Table 2-1            Alphanumeric List of all Legacy Commands Showing their  
Option 266 Support**

Command	Description	8566	8568	8590 Series	8560 Series	Page for Further Details
TIMEDSP	Enables display of time & data on analyzer display	N/A	N/A	No	N/A	
TITLE	Title entry	N/A	N/A	Yes	Yes	<a href="#">Page 358</a>
TM	Trigger Mode	Yes	Yes	Yes	Yes	<a href="#">Page 359</a>
TOI	Third order intermodulation measurement	N/A	N/A	No	N/A	
TOIR	Third order intermodulation response	N/A	N/A	No	N/A	
TRA	Returns trace A amplitude values to controller	Yes	Yes	Yes	Yes	<a href="#">Page 360</a>
TRB	Returns trace B amplitude values to controller	Yes	Yes	Yes	Yes	<a href="#">Page 361</a>
TRC	Returns trace C amplitude values to controller	N/A	N/A	Yes	N/A	<a href="#">Page 362</a>
TRCMEM	Trace C memory	N/A	N/A	No	N/A	
TRDEF	Trace define	No	No	No	No	
TRDSP	Trace display	Yes	Yes	Yes	N/A	<a href="#">Page 363</a>
TRGRPH	Trace graph display	No	No	No	N/A	
TRIGPOL	Trigger polarity	N/A	N/A	N/A	Yes	<a href="#">Page 364</a>
TRMATH	Executes specified trace math at end of sweep	No	No	No	N/A	
TRPRST	Sets trace operations to their preset values	No	No	Yes	N/A	<a href="#">Page 365</a>
TRSTAT	Returns current trace states to controller	Yes	Yes	Yes	N/A	<a href="#">Page 366</a>
TS	Takes a sweep	Yes	Yes	Yes	Yes	<a href="#">Page 367</a>
TVLINE	Selects which horizontal line of video to trigger on	N/A	N/A	No	N/A	
TVLSFRM	Selects the type of video frame to trigger on	N/A	N/A	No	N/A	
TVSTND	TV standard	N/A	N/A	No	N/A	
TVSYNC	Selects polarity of video modulation to trigger on	N/A	N/A	No	N/A	

**Table 2-1 Alphanumeric List of all Legacy Commands Showing their Option 266 Support**

Command	Description	8566	8568	8590 Series	8560 Series	Page for Further Details
TWNDOW	Formats trace information for FFT.	N/A	N/A	No	No	
UNTIL	Conditional Programming (Repeat...Until...)	No	No	N/A	No	
UP	Increases active function value by applicable step	N/A	N/A	No	N/A	
UR	Upper right x-y recorder output voltage at rear	No	No	N/A	N/A	
USTATE	Configures user defined states	No	No	No	N/A	
VARDEF	Variable definition	No	No	No	No	
VARIANCE	Returns the amplitude variance of specified trace	No	No	No	No	
VAVG	Turns video averaging on or off	Yes	Yes	Yes	Yes	<a href="#">Page 368</a>
VB	Video Bandwidth	Yes	Yes	Yes	Yes	<a href="#">Page 370</a>
VBO	Video Bandwidth Coupling Offset	Yes	Yes	N/A	N/A	<a href="#">Page 372</a>
VBR	Video Bandwidth Ratio	N/A	N/A	Yes	Yes	<a href="#">Page 373</a>
VIEW	Stores and views the specified trace	Yes	Yes	Yes	Yes	<a href="#">Page 374</a>
VTL	Video trigger level	N/A	N/A	N/A	Yes	<a href="#">Page 375</a>
WAIT	Suspend program operation for specified time	N/A	N/A	No	N/A	
WINNEXT	Next window	N/A	N/A	No	N/A	
WINOFF	Turns off the window display mode	N/A	N/A	No	N/A	
WINON	Turns on the window display mode	N/A	N/A	No	N/A	
WINZOOM	Window zoom	N/A	N/A	No	N/A	
XCH	Exchanges the two specified traces.	Yes	Yes	Yes	N/A	<a href="#">Page 376</a>
ZMKCNTR	Zone marker at center frequency	N/A	N/A	No	N/A	



**Table 2-1            Alphanumeric List of all Legacy Commands Showing their  
Option 266 Support**

<b>Command</b>	<b>Description</b>	<b>8566</b>	<b>8568</b>	<b>8590 Series</b>	<b>8560 Series</b>	<b>Page for Further Details</b>
ZMKPKNL	Zone marker for next peak left	N/A	N/A	No	N/A	
ZMKPKNR	Zone marker for next peak right	N/A	N/A	No	N/A	
ZMKSPAN	Zone marker span	N/A	N/A	No	N/A	

Legacy Analyzer Command List  
**Table of All Legacy Analyzer Commands**



## A Few Helpful Hints and Tips

These pages lists a few hints and tips that will help you get the most from your analyzer and Option 266.

- **Speed** - Option 266 will run faster on PSA Series analyzers than it will on ESA Series analyzers.
- **Compatibility - speed and consistency** - for best compatibility with your legacy analyzer, Option 266 should be used on the analyzer whose frequency range most closely matches the frequency range of your legacy analyzer. For example, the best match for the 8565E with its 50 GHz upper frequency limit is the PSA E4448A analyzer which also has an upper frequency limit of 50 GHz.
- **Compatibility and Sweep Times** - for best compatibility between PSA Series analyzers and the legacy analyzers, use the Manual Swept mode for 8566A/B, 8568A/B, or 8590 Series analyzers. Manual Swept mode is the default setting on PSA Series analyzers with Option 266 installed.

When analyzing stationary signals, you can change to the Best Speed setting, which is accessed from the Auto Couple hardkey and the FFT & Sweep menu. This results in faster sweep times on a PSA analyzer than on the legacy analyzers because of the PSA's better performance. In the majority of applications, this faster speed would be desirable, but that is not always the case.

When you are using the Best Speed setting, you cannot change the sweep time manually as the sweep time is always coupled to give the fastest sweep times based on the current settings.

- **Time-out (1)** - Agilent suggests that you increase the timeout when performing MA and MF commands to allow previous marker functions to complete correctly.
- **Time-out (2)** - Agilent recommends increasing the timeout on a serial poll (**SPOLL**) due to differences in Sweep Times on some settings. Note, however, that this may not be necessary when using the Best Speed setting on the FFT & Sweep menu (accessed from the Auto Couple hardkey).
- **Synchronization (1)** - to synchronize after an IP command, Agilent recommends that you use the DONE command. We also suggest that the DONE command is used in conjunction with a timeout of about 5 seconds in case the analyzer starts to auto align. Alternatively, you could set auto alignment to Off. To set auto alignment to Off, press **System, Alignments, Auto Align** on the front panel.
- **Synchronization (2)** - Agilent recommends that synchronization (using the DONE command) is used with marker functions when signal tracking is turned on.

- **AC and DC Coupling** - The 8568A/B has two RF input ports:
  - DC Coupled (with a BNC connector) to handle a frequency range of 100 Hz to 1.5 GHz
  - AC Coupled (with an N Type connector) to handle a frequency range of 100 kHz to 1.5 GHz

The 8560 Series and the 8590 Series of legacy analyzers only have one RF input port, and support AC and DC coupling through the COUPLE command (page 151).

The PSA series analyzers only have a single RF input port.

When using PSA models E4440A, E4443A, and E4445A with the UKB option, you must use DC coupling to see calibrated frequencies of less than 10 MHz. Signals of less than 10 MHz are not calibrated when using AC coupling on these analyzers.

---

**NOTE**

PSA models E4446A, E4447A, and E4448A do not allow AC coupling.

- **Overloading** - if you are overloading the analyzer, connect a 10 dB attenuator to the RF input and set the **Ext Amp Gain** (external amplifier gain) to -10 dB. This softkey is accessed via the Amplitude hardkey and the **More 1 of 3** softkey.
- **Instrument Presets** - consecutive instrument presets (using the IP command, that is, issuing a command such as IP;IP;IP;) are not required.
- **SCPI Language** - Agilent recommends that you do not repeatedly swap to and from the SCPI language within your programs.

Hints and Tips  
**A Few Helpful Hints and Tips**

---


# 4

## Programming Commands

This chapter lists all the supported 8566A/B, 8568A/B, 8560 Series, and 8590 Series compatible commands in alphanumeric order, and gives brief details on their syntax and operation. For more detailed information on these commands, see your 8566A/B, 8568A/B, 8560 Series or 8590 Series Operating and Programming Manual.

## Command Syntax

Command syntax is represented pictorially.



cu114e

- Ovals enclose command mnemonics. The command mnemonic must be entered exactly as shown.
- Circles and ovals surround secondary keywords or special numbers and characters. The characters in circles and ovals are considered reserved words and must be entered exactly as shown.
- Rectangles contain the description of a syntax element defined in [Table 4-1, “Syntax Elements.”](#)
- A loop above a syntax element indicates that the syntax element can be repeated.
- Solid lines represent the recommended path.
- Dotted lines indicate an optional path for bypassing secondary keywords or using alternate units.
- Arrows and curved intersections indicate command path direction.
- Semicolons are the recommended command terminators. Using semicolons makes programs easier to read, prevents command misinterpretation, and is recommended by IEEE Standard 728.

### NOTE

Uppercase is recommended for entering all commands unless otherwise noted.

Syntax Elements are shown in the syntax diagrams as elements within rectangles. In the syntax diagrams, characters and secondary keywords are shown within circles or ovals. Characters and secondary keywords must be entered exactly as shown.



**Table 4-1 Syntax Elements**

Syntax Component	Definition/Range
Analyzer command	Any spectrum analyzer command in this chapter, with required parameters and terminators.
Character	Sp a b c d e f g h i j k l m n o p q r s t u v w x y z databyte.
Character & EOI	8-bit byte containing only character data and followed by end-or-identify (EOI) condition, where the EOI control line on GPIB is asserted to indicate the end of the transmission. END signifies the EOI condition.
Character string	A list of characters.
Data byte	8-bit byte containing numeric or character data.
Data byte & EOI	8-bit byte containing numeric or character data followed by end-or-identify (EOI) condition, where the EOI control line on GPIB is asserted to indicate the end of the transmission. END signifies the EOI condition.
Delimiter	\ @ ^ \$ % ; ! Matching characters that mark the beginning and end of a character string, or a list of spectrum analyzer commands. Choose delimiting characters that are not used within the string they delimit.
Digit	0 1 2 3 4 5 6 7 8 9
lsb length	Represents the least significant byte of a two-byte word that describes the number of bytes returned or transmitted. See msb length.
msb length	Represents the most significant byte of a two-byte word that describes the number of bytes returned or transmitted. See lsb length.
Number	Expressed as integer, decimal, or in exponential (E) form.  Real Number Range: $\pm 1.797693134862315 \times 10^{308}$ , including 0. Up to 15 significant figures allowed.  Numbers may be as small as $\pm 2.225073858507202 \times 10^{-308}$  Integer Number Range: -32,768 through +32,767
Output termination	Carriage return (C <sub>R</sub> ) and line feed (L <sub>F</sub> ), with end-or-identify (EOI) condition. ASCII codes 13 (carriage return) and 10 (line feed) is sent via GPIB, then the end-or-identify control line on GPIB sets to indicate the end of the transmission.
Units	Represent standard scientific units.  Frequency Units: GZ, GHZ, MZ, MHZ, KZ, KHZ, HZ Amplitude Units: DB, DBMV, DM, DBM, DBUV, V, MV, UV Time Units: SC, S, MS, US

---

## Programming Command Descriptions

All supported commands are listed here, along with their descriptions and cross-references to similar commands.

This chapter is not designed to be a comprehensive guide to all 8566A/B, 8568A/B, 8560 Series, or 8590 Series commands. It gives brief descriptions of the supported commands, and highlights important functional or behavioral differences that you should be aware of when transferring existing 8566A/B, 8568A/B, 8560 Series or 8590 Series code to your PSA Series analyzer. For a fuller description of the commands, refer to the 8566A/B, 8568A/B, 8560 Series or 8590 Series Operating and Programming Manual.

In the descriptions of the commands, TRA corresponds to Trace 1, TRB corresponds to Trace 2, and TRC corresponds to Trace 3.

To avoid confusion between numbers and letters, all commands that incorporate numbers have had the number spelled out and placed in square brackets after the command. For example, the command 'I1' is shown as 'I1 [*one*]' - that is, the capital letter 'I' followed by the number '1', and then the word 'one' italicized in square brackets. The italicized word in brackets does not form part of the command.

---

## A1 [one] Clear Write for Trace A

### Syntax



### Description

The A1 command sets Trace A to clear write. That is, it continuously displays any signal present at the spectrum analyzer input. The A1 command initially clears Trace A, setting all elements to zero. The sweep trigger then signals the start of the sweep, and Trace A is continually updated as the sweep progresses. Subsequent sweeps send new amplitude information to the display addresses.

---

**NOTE**

The functions of the command A1 are identical to the CLRW TRA command ([page 147](#)).

---

---

## A2 [*two*] Maximum Hold for Trace A

### Syntax



### Description

The A2 command updates each trace element with the maximum level detected during the period that the trace has been active.

---

**NOTE**

The functions of the command A2 are identical to the MXMH TRA command ([page 293](#)).

---

---

## A3 [*three*] View Mode for Trace A

### Syntax



### Description

The A3 command displays Trace A and then stops the sweep if no other traces are active. Trace A does not get updated with new data.

---

#### NOTE


The functions of the command A3 are identical to the VIEW TRA command ([page 374](#)).

---

---

## A4 [*four*] Blank Trace A

### Syntax



### Description

The A4 command blanks Trace A and stops the sweep if no other traces are active. Trace A is not updated.

---

**NOTE**

The functions of the command A4 are identical to the BLANK TRA command ([page 132](#)).

---

---

## ACP Adjacent Channel Power

### Syntax



### Description

ACP measures the power of the carrier and the power of the channels that are adjacent to the carrier, and then computes a power ratio for each of the adjacent channels, using the carrier power as a reference. ACP performs the adjacent channel power measurement using the values for channel spacing (ACPSP) and channel bandwidth (ACPBW).

---

#### NOTE

Option 266 supports this measurement using the ANALOG method only.


---

---

## ACPALPHA

### Adjacent Channel Power Alpha Weighting

#### Syntax



#### Description

This command sets the alpha weighting for ACP measurements.


Range: Any real number between 0 and 1



---

## ACPALTCH Adjacent Channel Power Alternate Channels

### Syntax



### Description


The ACPALTCH command sets the number of alternate channels to be measured by an adjacent channel power measurement to either 0, 1, or 2. The number of alternate channels is used with the ACPRSLTS command ([page 109](#)).

Range: 0, 1, or 2.

Default value: 0.

## ACPBRPER Adjacent Channel Power Burst Period

### Syntax



### Description


The ACPBRPER command sets the cycle time (period) of the burst RF signal. The cycle time is needed to set the sweep times when using the peak, two bandwidth, burst power, and gated methods for adjacent channel power measurements.

#### NOTE

Option 266 supports the ACP measurement using the ANALOG method only.

## ACPBRWID Adjacent Channel Power Burst Width

### Syntax



### Description

The ACPBRWID command sets the on-time (pulse width) of the burst RF signal. The pulse width is needed to set the gating times when using the gated method for adjacent channel power measurements.

Range            5  $\mu$ s to 9.5 seconds

---


**NOTE**

Option 266 supports the ACP measurement using the ANALOG method only.

---

## ACPBW Adjacent Channel Power Bandwidth

### Syntax



### Description

The ACPBW command sets the bandwidth of the channels as an active function for the ACPMEAS ([page 105](#)) and ACPCOMPUTE ([page 101](#)) commands. The channel bandwidth cannot be greater than the channel spacing. If the channel bandwidth is greater than the channel spacing, the measurement is not performed.

---

## ACPCOMPUTE Adjacent Channel Power Compute

### Syntax

ACPCOMPUTE → ;

### Description

Calculates the ACP of a transmitter based on data on the display. This function does not make a new measurement before computing. The measurement must have been made with ANALOG or PEAK method selected so the appropriate data is available for the calculation.

---


#### NOTE

Option 266 supports the ACP measurement using the ANALOG method only.

---

## ACPFQRWT Adjacent Channel Power Frequency Weighting

### Syntax



### Description

The ACPFRQWT command is used to control the frequency weighting when making an Adjacent Channel Power measurement. Weighting is not used in the measurement if OFF has been selected.

Root-raised-cosine weighting is selected with the RRCOS parameter.

Default value: OFF


#### NOTE

Option 266 supports the ACP measurement using the ANALOG method only.

---

## ACPLOWER Lower Adjacent Channel Power

### Syntax



### Description

The ACPLOWER query command returns the power ratio result of the Adjacent Channel Power measurement for the lower frequency channel.

---

**NOTE**


Option 266 supports the ACP measurement using the ANALOG method only.

---

---

## ACPMAX Maximum Adjacent Channel Power

### Syntax



The ACPMAX query command returns the maximum adjacent channel power of the adjacent channel power measurement.

---

**NOTE** Option 266 supports the ACP measurement using the ANALOG method only.

---



---

## ACPMEAS Measure Adjacent Channel Power

### Syntax

ACPMEAS → ;

### Description

The ACPMEAS command makes a measurement and calculates the adjacent channel power (ACP) of a transmitter. The measurement determines the leakage power that is in the channels adjacent to the carrier. The result is the ratio of the leakage power in the channel adjacent to the total power transmitted by the transmitter.

---

#### NOTE


Option 266 supports the ACP measurement using the ANALOG method only.

---

---

## ACPMSTATE Adjacent Channel Power Measurement State

### Syntax



### Description

Sets the parameters of the measurement state to either the default state (determined by the setup) or the current state. The state parameters that could change between the default state and a current state include:

- Resolution bandwidth
- Video bandwidth
- Span
- Sweep time
- Detector mode
- Gating parameters
- Trigger parameters
- Video averaging

Default value: DFLT

---


**NOTE**

Option 266 supports the ACP measurement using the ANALOG method only.

---

## ACPPAR Adjacent Channel Power Manual or Auto

### Syntax



### Description

Determines whether the spectrum analyzer settings for the ACP (page 95), CHP (page 144) or OBW (currently not supported) measurements are set automatically or manually.

If ACPPAR is set to 1 (automatic), the analyzer does the following before making the measurement:

- Performs the Trace Preset (TRPRST (page 365)) command.
- Changes Trigger Mode to Free Run.
- Changes Detector Mode to Sample.
- Changes the amplitude scale to 10 dB per division.
- Sets the frequency span, resolution bandwidth, video bandwidth, center frequency step size and sweep time based on the channel spacing (ACPSP (page 110)) and channel bandwidth (ACPBW (page 100)).
- Takes a sweep

Default value: 1 (auto)

#### NOTE


Option 266 supports the ACP measurement using the ANALOG method only.

---

## ACPPWRTX

### Adjacent Channel Power Total Power Transmitted

#### Syntax



#### Description

The ACPPWRTX query command returns the result of the total power transmitted calculation of the adjacent channel power measurement. The measurement must be made with the analog or burst power method selected.

---

**NOTE**


Option 266 supports the ACP measurement using the ANALOG method only.

---

---

## ACPRSLTS Adjacent Channel Power Measurement Results

### Syntax



### Description

Returns an array of power data resulting from an ACP measurement of an RF signal. The number of alternate channel pairs selected by the ACPALTCCH ([page 97](#)) command determines the size of the array.

---


#### NOTE

Option 266 supports the ACP measurement using the ANALOG method only.

---

## ACPSP Adjacent Channel Power Channel Spacing

### Syntax




### Description

Sets the channel spacing for the ACPMEAS ([page 105](#)) and ACP COMPUTE ([page 101](#)) commands.

**NOTE** Option 266 supports the ACP measurement using the ANALOG method only.

## ACPT Adjacent Channel Power T Weighting

### Syntax



### Description

The ACPT command is used to set the T used in weighting for an adjacent channel power measurement.

#### NOTE


Option 266 supports the ACP measurement using the ANALOG method only.

---

## ACPUPPER

### Upper Adjacent Channel Power

#### Syntax



#### Description

The ACPUPPER query command returns the power ratio result of the adjacent channel power measurement for the upper frequency channel.

---

**NOTE**

Option 266 supports the ACP measurement using the ANALOG method only.

---



---

## ADJALL LO and IF Adjustments

### Syntax

ADJALL → ;


### Description

The ADJALL command activates the RF local oscillator (LO) and intermediate frequency (IF) alignment routines. These are the same routines that occur when the spectrum analyzer is switched on. They are also the same routines that are performed when you press **System**, **Alignments**, **Align Now**, **All**.

Commands following ADJALL are not executed until after the analyzer has finished the alignment routines.

## AMB A minus B into A

### Syntax



### Description

The AMB command subtracts the points in Trace B from the corresponding points in Trace A, and sends the results to Trace A. Thus, AMB can restore the original trace after an APB (page 118) or a KSc (page 199) command has been executed.

The query command AMB? returns different responses depending on the language being used. The 8560 Series languages return either a **1** or a **0** to indicate the On or Off status. The 8566, 8568, and the 8590 Series languages all return either **ON** or **OFF**.

---

**NOTE** On the legacy analyzers, the AMB command operates continuously. That is, it continuously updates Trace A to display the results of Trace A minus Trace B until AMB is switched off.

This is not the case with Option 266 - Code Compatibility. The AMB command is performed once only, using the trace data available at the time of execution.

---

**NOTE** The functions of the command AMB are identical to the C2 [two] command (page 137).

---


## AMBPL (A minus B) plus Display Line into A

### Syntax

#### 8566 and 8568 Remote Language



#### 8560 Series and 8590 Series Remote Language



### Description

The AMBPL command does a point-by-point subtraction of Trace B from Trace A, and then adds the display line point values to the difference. The results are sent to Trace A.

---

#### NOTE

On the legacy analyzers, the AMBPL command operates continuously. That is, it continuously updates Trace A to display the results of Trace A minus Trace B until AMBPL is switched off.

This is not the case with Option 266 - Code Compatibility. The AMBPL command is performed once only, using the trace data available at the time of execution.

---

---

**NOTE**

The query command AMBPL? returns different responses depending on the language being used.

8560 Series      Query response is either 1 or 0, indicating ON or OFF state.

8566A            Query response is either ON or OFF.

8566B            Query response is either ON or OFF.

8568A            Query response is either ON or OFF.


8568B            Query response is either ON or OFF.

8590 Series      Query response is either ON or OFF.

---

# ANNOT Annotation

## Syntax



Preset State: ANNOT ON

## Description

The ANNOT command turns on or off all annotation on the spectrum analyzer display. Softkey labels are not affected by this command and remain displayed.

### NOTE

The functions of the command ANNOT are identical to the KSo command ([page 221](#)) and KSp command ([page 222](#)). Note also that these two alternative commands, KSo and KSp, are only valid when the remote language is either HP8566A, HP8566B, HP8568A, or HP8568B.

---

## APB

### Trace A Plus Trace B to A

#### Syntax



#### Description

The APB command does a point-by-point addition of Trace A and Trace B, and sends the results to Trace A. Thus, APB can restore the original trace after an AMB (page 114) or a C2 (page 137) command has been executed.

---

**NOTE**

The functions of the command APB are identical to the KSc command (page 199). Note also that the alternative command, KSc, is only valid when the remote language is either HP8566A, HP8566B, HP8568A, or HP8568B.

---

**NOTE**

On the legacy analyzers, the AMB command operates continuously. That is, it continuously updates Trace A to display the results of Trace A minus Trace B until AMB is switched off.


This is not the case with Option 266 - Code Compatibility. The AMB command is performed once only, using the trace data available at the time of execution.

---


# AT Input Attenuation

## Syntax

### 8560 Series Remote Language



### 8566 and 8568 Remote Language



### 8590 Series Remote Language




Table 4-2

Item	Description/Default	Range
Number	Any real number or integer. If the value you enter is not a valid value for the analyzer you are using, it will switch automatically to the closest valid setting. Default units are dB.	0 to 70 dB specified absolutely and 10 to 70 dB in 10 dB steps

Preset State: 10 dB

Step Increment: 10 dB

### Description

Specifies the RF input attenuation.

Although any attenuation level from 0 dB to 70 dB in PSA series analyzers can be specified using absolute values, you can never set attenuation below 10 dB using the DOWN steps. This is a safety feature to prevent inadvertent setting of attenuation to a level that could damage the analyzer.

---

**CAUTION** Signal levels above +30 dBm will damage the spectrum analyzer.

---



---

**NOTE** You cannot step down below 10 dB. To set levels below 10 dB, you must specify the attenuation absolutely. For example, to set attenuation to 0 dB, you must use the command **AT 0DB**.


---




# AUNITS Absolute Amplitude Units

## Syntax


### 8560 Series Remote Language



### 8566 and 8568 Remote Language



### 8590 Series Remote Language



## Description

Specifies the amplitude readout units for the reference level, the marker, and the display line.

**AUNITS Absolute Amplitude Units**

---

**NOTE**

If your selected remote language is any of the 8560 Series and you use either the AUTO or the MAN parameter, a warning will be displayed informing you that the command is not supported with either of these parameters.

---

---


**NOTE**

The functions of the command AUNITS are identical to the commands KSA ([page 194](#)), KSB ([page 196](#)), KSC ([page 198](#)), and KSD ([page 200](#)). Note also that these four alternative commands, KSA, KSB, KSC, and KSD are only valid when the remote language is either HP8566A, HP8566B, HP8568A, or HP8568B.

---

# AUTO Auto Couple

## Syntax



## Description

The AUTO command couples the active functions automatically. Sending the command HD; AUTO will cause all functions to be auto coupled.

### NOTE

On the legacy analyzers, if the currently active function *is not* represented by one of the keywords listed below when the command AUTO is executed, all functions are auto coupled. If the active function *is* represented by one of the keywords below, only that function is auto coupled.

This does not happen with Option 266 Programming Code Compatibility Suite on the PSA Series analyzers because they do not recognize active functions.

## Keywords Used in the Command

AT	Couples attenuation to the reference level.
DL	Turns the display line off but does not change the value of the line.
MKA	Turns the marker off.
MKD	Turns the delta marker off. It does not turn the current marker off.
MKFCR	Deactivates the use of user-supplied counter resolution value, but the value remains unchanged.
MKN	Turns the marker off.
RB	Couples the resolution bandwidth to the frequency span.
SRCPSTP	Sets the source power step to 0 (zero).
SRCPSWP	Turns power sweep off.
SCRPOWER	Turns source power off.
SS	Couples the step size to the frequency span.

## Programming Commands

**AUTO Auto Couple**

ST	Couples the sweep time to the frequency span.
TH	Turns the display of the threshold off, but does not change its value or prevent its usage in peak searching.
VAVG	Stops averaging.
VB	Couples the video bandwidth to the resolution bandwidth.
VBR	Sets the video to bandwidth ratio to 0.3.

---

## AUTOPL Auto Coupled

### Syntax

AUTOPL → ;

### Description

Sets video bandwidth, resolution bandwidth, input attenuation, sweep time and center frequency step-size to coupled mode.

---

## AXB Exchange Trace A and Trace B

### Syntax



### Description

This command exchanges Trace A and Trace B, point by point.

---

**NOTE**

The functions of the command AXB are identical to the EX command ([page 173](#)) and to the XCH TRA,TRB command ([page 376](#)).

---

---

## **B1 [one] Clear Write for Trace B**

### **Syntax**



### **Description**

The B1 command sets Trace B to clear write. That is, it continuously displays any signal present at the spectrum analyzer input. The B1 command initially clears Trace B, setting all elements to zero. The sweep trigger then signals the start of the sweep, and Trace B is continually updated as the sweep progresses. Subsequent sweeps send new amplitude information to the display addresses.

---

#### **NOTE**

The functions of the command B1 are identical to the CLRW TRB command ([page 147](#)).

---

---

## B2 [*two*] Maximum Hold for Trace B

### Syntax



### Description

The B2 command updates each trace element with the maximum level detected while the trace is active.

---

**NOTE**

The functions of the command B2 are identical to the MXMH TRB command ([page 293](#)).

---



---

## B3 [*three*] View Mode for Trace B

### Syntax



### Description

The B3 command displays Trace B and then stops the sweep if no other traces are active. Trace B does not get updated.

---

#### NOTE

The functions of the command B3 are identical to the VIEW TRB command ([page 374](#)).

---

---

## B4 [*four*] Blank Trace B

### Syntax



### Description

The B4 command blanks Trace B and stops the sweep if no other traces are active. Trace B is not updated.

---

**NOTE**

The functions of the command B4 are identical to the BLANK TRB command ([page 132](#)).

---

---

## BL Trace B minus Display Line to Trace B

### Syntax



### Description

The BL command subtracts the display line from Trace B and sends the results to Trace B.

---

#### NOTE

The command BL is calculated differently depending on the language being used.

8560 Series	The calculation is performed in units of dBm.
8566A	The calculation is performed in display units.
8566B	The calculation is performed in display units.
8568A	The calculation is performed in display units.
8568B	The calculation is performed in display units.
8590 Series	The calculation is performed in display units.

---

#### NOTE


The functions of the command BL are identical to the BML command ([page 133](#)).

---


## BLANK Blank Trace

### Syntax

#### 8560 Series Remote Language



#### 8566, 8568, and 8590 Series Remote Language



Preset State: BLANK TRB, BLANK TRC

### Description

Blanks Trace 1, trace 2, or trace 3, and stops taking new data into the specified trace. TRA corresponds to Trace 1, TRB corresponds to Trace 2, and TRC corresponds to Trace 3.

#### NOTE

The functions of the command BLANK are identical to the A4 command ([page 94](#)), the B4 command ([page 130](#)), and KSk command ([page 212](#)).

---

## BML Trace B Minus Display Line

### Syntax



### Description

The BML command subtracts the display line from trace B (point by point), and sends the difference to trace B. Trace B corresponds to Trace 2.

---

**NOTE** Remote language 8560 - the BML command is performed in dBm units.  
Remote languages 8566A, HP8566B, HP8568A, 8568B, and the 8590 Series - the BML command is performed in display units.

---

**NOTE** The functions of the command BML are identical to the BL command ([page 131](#)).

---

---

## BTC

### Transfer Trace B to Trace C

#### Syntax



#### Description

The BTC command transfers Trace B data to Trace C

---

#### NOTE

Trace C cannot be an active trace. This means that the data in Trace C cannot be updated as the analyzer sweeps. To ensure that the current settings of the analyzer are reflected in the data transferred from Trace B to Trace C, you must follow the four step process below.

- Select single sweep mode (S2 or SNGLS command)
- Select the desired analyzer settings
- Take one complete sweep
- Transfer the data

---

#### NOTE

The functions of the command BTC are identical to the KSI command ([page 214](#)).

---

## BXC Exchange Trace B and Trace C

### Syntax

BXC → ;

### Description

The BXC command exchanges Trace B data with Trace C data.

---

#### NOTE

Trace C must not be an active trace. This means that the data in Trace C cannot be updated as the analyzer sweeps. To ensure that the current settings of the analyzer are reflected in the data exchanged between Trace B and Trace C, you must follow the four step process below.

- Select single sweep mode (S2 or SNGLS command)
- Select the desired analyzer settings
- Take one complete sweep
- Exchange the data

---


#### NOTE

The functions of the command BXC are identical to the KSi command ([page 209](#)) and to the XCH TRB,TRC command ([page 376](#)).

---

## C1 [*one*] Set A Minus B Mode Off

### Syntax



### Description

The C1 command turns the A Minus B mode off. That is, it switches off the functionality that was switched on by the C2 command ([page 137](#)) or by the AMB ON command ([page 114](#)).

---

#### NOTE

The functions of the command C1 are identical to the AMB OFF command ([page 114](#)).

---



---

## **C2 [two] A Minus B Into A**

### **Syntax**



### **Description**

The C2 command subtracts the points in Trace B from the corresponding points in Trace A, and sends the results to Trace A. Thus, if your input signal remains unchanged, C2 can restore the original trace after an APB command ([page 118](#)) or a KSc ([page 199](#)) command has been executed.

---

**NOTE** The functions of the command C2 are identical to the AMB ON command ([page 114](#)).

---

**NOTE** On the legacy analyzers, the C2 command operates continuously. That is, it continuously updates Trace A to display the results of Trace A minus Trace B until C2 is switched off by issuing the C1 command.


This is not the case with Option 266 - Programming Code Compatibility. The C2 command is performed once only, using the trace data available at the time of execution.

---

---

## CA Couple Attenuation

### Syntax



### Description

During normal operation, the spectrum analyzer's input attenuation is coupled to the reference level. This coupling keeps the mixer input at a level such that a continuous wave signal displayed at the reference level is at or below -10 dBm (or the value specified in the ML command.)

The CA command sets the threshold to -10 dBm (or to the value specified by the ML command [\(page 289\)](#) or the KS, command [\(page 192\)](#)). The counterpart to the CA command is the AT command [\(page 119\)](#), which allows levels less than the threshold value at the mixer input.

---

## CAL Calibration

### Syntax




### Description

Calibrates the logarithmic and step gain amplifiers, the attenuator and the amplitude and frequency of the resolution bandwidth filters.

---

## CARRON Carrier On Power

### Syntax




### Description

Measures the average power of the carrier during the portion of time when it is on and within 20dB of its peak level.

# CF Center Frequency

## Syntax



**Table 4-3**

Item	Description/Default	Range
REAL	Any real or integer number. Default unit is HZ.	Frequency range of the spectrum analyzer

## Description

The CF command specifies the value of the center frequency.

The step size depends on whether the frequency has been coupled to the span width using the CS command (page 153). When coupled, the step size is 10% of the span, or one major graticule division; when uncoupled, the step size is determined by the SS command (page 342).


**NOTE**

Although the spectrum analyzer allows entry of frequencies not in the specified frequency range, using frequencies outside the frequency span of the spectrum analyzer is not recommended and is not warranted to meet specifications.

---

## CHANNEL Channel Selection

### Syntax



### Description

Increments or decrements the spectrum analyzer center frequency by one channel spacing.


---

**NOTE** The channel spacing value is set using the ACPSP command ([page 110](#)).

---

## CHANPWR Channel Power

### Syntax



### Description

Measures the power within the specified channel bandwidth.

#### NOTE

If no channel bandwidth is specified in the command, the channel bandwidth is set using the CHPWRBR command ([page 145](#)).

---

## CHP Channel Power

### Syntax



### Description

The CHP command performs the channel power measurement.

---

**NOTE**


The channel bandwidth is set with the ACPBW command ([page 100](#)).  
Channel spacing is set with the ACPSP command ([page 110](#)).

---



## CHPWRBW Channel Power Bandwidth

### Syntax




### Description

Queries or sets the current value of the channel power bandwidth. Channel power can be measured with the CHANPWR command [\(page 143\)](#).

---

## CLRAVG Clear Average

### Syntax



```
CLRAVG → ;
```

### Description


The command restarts the VAVG command by resetting the number of averaged sweeps to one. The video averaging routine resets the number of sweeps, but does not stop video averaging. Use “VAVG OFF;” to stop video averaging.

## CLRW Clear Write


Clears the specified trace and enables trace data acquisition.

### Syntax

#### 8560 Series Remote Language



#### 8566, 8568, and 8590 Series Remote Language



Preset State: CLRW TRA

### Description

The CLRW command places the indicated trace in clear-write mode. Data acquisition begins at the next sweep. (See the TS command [\(page 367\)](#) for more information about data acquisition.)

TRA corresponds to Trace 1, TRB corresponds to Trace 2, and TRC corresponds to Trace 3.

#### NOTE

The functions of the command CLRW are identical to the A1 command [\(page 91\)](#) and B1 command [\(page 127\)](#).

---

## CLS Clear Status Byte

### Syntax

A diagram showing the command 'CLS' in a rounded rectangle, followed by an arrow pointing to a semicolon ';' inside a circle.

```
CLS → ;
```

### Description

Clears all the status bits from the status byte.

---

## CONTS Continuous Sweep

### Syntax

CONTS → ;

Preset State: CONTS

### Description

The CONTS command sets the spectrum analyzer to continuous sweep mode. In the continuous sweep mode, the spectrum analyzer takes its next sweep as soon as possible after the current sweep (as long as the trigger conditions are met). A sweep may temporarily be interrupted by data entries made over the remote interface or from the front panel.

---

#### NOTE


The functions of the command CONTS are identical to the S1 command ([page 326](#)).

---

---

## CORREK Correction Factors On

### Syntax




### Description

The CORREK command queries the state of the analyzer frequency corrections. It returns a “1” if the correction factors are on, a “0” if they are off.

---

## COUPLE Input Coupling

### Syntax



### Description

The COUPLE command selects AC or DC coupling.

---

**NOTE**

PSA models E4446A, E4447A, and E4448A do not allow AC coupling.

---

## CR Couple Resolution Bandwidth

### Syntax



### Description

The CR command couples the resolution bandwidth to the video bandwidth and to the sweep time.

The counterpart to the CR command is the RB command ([page 315](#)) which breaks the coupling. Use the CR command to re-establish coupling after executing an RB command.

---

**NOTE**

CR uses the coupling settings from the PSA analyzer. These may differ from the settings that you would have seen on the legacy analyzer being emulated here.

---



---

## CS Couple Frequency Step Size

### Syntax



### Description

The CS command couples the center frequency step size to the span width so that the step size equals 10% of the span width, or one major graticule division.

The counterpart to the CS command is the SS command ([page 342](#)) which breaks the coupling. Use the CS command to re-establish coupling after an SS command has been executed.

---

## CT Couple Sweep Time

### Syntax



### Description

The CT command couples the sweep time to the span, resolution bandwidth and video bandwidth.

The counterpart to the CT command is the ST command ([page 344](#)) which breaks the coupling. Use the CT command to re-establish coupling after an ST command has been executed.

---

## CV Couple Video Bandwidth

### Syntax



### Description

The CV command couples the video bandwidth to the resolution bandwidth.

The counterpart to the CV command is the VB command ([page 370](#)) which breaks the coupling. Use the CV command to re-establish coupling after executing a VB command.

---

#### NOTE


CV uses the coupling settings from the PSA analyzer. These may differ from the settings that you would have seen on the legacy analyzer being emulated here.

---

---

## DA Display Address

### Syntax



### Description

The DA command returns the contents of the given display address to the controller

---


**NOTE**

This command only supports the use of the DA 1, DA 1025, and DA 3073; these display addresses contain the trace data and are equivalent to using the commands TRA? ([page 360](#)), TRB? ([page 361](#)), TRC? ([page 362](#)), TA ([page 353](#)) and TB ([page 354](#)).

---

## DELMKBW Occupied Power Bandwidth Within Delta Marker

### Syntax



### Description

Calculates the OBW with respect to the power between the displayed delta markers. The power between the displayed markers is then used as the reference, rather than using the total power in the frequency span as is done in the PWRBW ([page 307](#)) command.


#### NOTE

If the DELMKBW command is used when no marker is active, a delta marker is activated at the center frequency, and the returned bandwidth is 0. If the active marker is a normal marker when the DELMKBW command is used, the marker type is changed to delta, and the returned bandwidth is 0.


# DET Detection Mode

## Syntax


### 8566 and 8568 Remote Language



### 8560 Series Remote Language



### 8590 Series Remote Language



Preset State: DET NRM

## Description

The DET command selects the type of spectrum analyzer detection (positive-peak, negative peak, sample, normal, and so on).

<b>POS</b>	enables positive-peak detection, which displays the maximum video signal detected over a number of instantaneous samples for a particular frequency.
<b>SMP</b>	enables sample detection, which uses the instantaneous video signal value. Video averaging and noise-level markers, when activated, activate sample detection automatically.
<b>NEG</b>	enables negative peak detection in sweep times of less than or equal to 200 ms.
<b>NRM</b>	enables the ' <i>rosenfell</i> ' detection algorithm that selectively chooses between positive and negative values.
<b>QPD</b>	enables quasi-peak detection for EMC measurements.
<b>EPK</b>	enables EMI peak detection for EMC measurements.
<b>MPK</b>	enables MIL peak detection to meet military specifications when making EMC measurements.
<b>AVG</b>	enables EMI average detection for EMC measurements.
<b>AVE</b>	enables average peak detection (non-EMC measurements).
<b>AUTO</b>	sets the detector function selection to auto.
<b>MAN</b>	sets the detector function selection to manual.

---


**NOTE** The functions of the DET command are identical to the KSa command ([page 195](#)), the KSb command ([page 197](#)), the KSd command ([page 201](#)), the KSe command ([page 203](#)), the Q0 command ([page 309](#)), and the Q1 command ([page 310](#)).

---


# DL Display Line

## Syntax

### 8566 and 8568 Remote Language



### 8590 Series and 8590 Series Remote Language



**Table 4-4**

Item	Description/Default	Range
NUMBER	Any real or integer number. Default units are dBm.	Dependent on the reference level



Preset State: DL OFF

Step Increment: 1 major graticule division


### **Description**

Defines the level of the display line and displays it on the spectrum analyzer screen.

---

## DLE Display Line Enable

### Syntax




### Description

The DLE command enables or disables the display line.

## DLYSWP Delay Sweep

### Syntax



### Description

Delays the start of the sweep until the specified time after the trigger event has elapsed.

#### NOTE


A query response of 0 indicates that DLYSWP is switched off.

## DONE


### Done

### Syntax

#### 8566 and 8568 Remote Language



#### 8590 Series and 8590 Series



### Description

Allows you to determine when the spectrum analyzer has parsed a list of spectrum analyzer commands and has executed all commands prior to and including DONE. The DONE command returns a value of “1” when all commands in a command string or command list have been completed.

If a take sweep (TS command ([page 367](#))) precedes the command list, the TS command acts as a synchronizing function since the command list execution begins after the sweep has been completed.

---

## E1[one] Peak Marker

### Syntax



### Description

The E1 command positions the marker at the signal peak.

---

#### NOTE

The functions of the E1 command are identical to MKPK (no secondary keyword) and MKPK HI ([page 276](#)).

---

## E2 [*two*] Marker to Center Frequency

### Syntax



### Description

The E2 command positions the marker on the screen at the center frequency position.

---

**NOTE**

The functions of the E2 command are identical to the MKCF command ([page 264](#)).

---

---

## E3 [*three*] Delta Marker Step Size

### Syntax



### Description

The E3 command establishes the center frequency step size as being the frequency difference between the delta marker and the active marker.

---

#### NOTE

The functions of the E3 command are identical to the MKSS command ([page 283](#)).

---

---

## E4 [*four*] Marker to Reference Level

### Syntax



### Description

The E4 command moves the active marker to the reference level.

---

**NOTE**

The functions of the E4 command are identical to the MKRL command ([page 281](#)).

---



---

## EDITDONE End of Limit Line Edits

### Syntax

EDITDONE → ;

### Description

The EDITDONE command is used at the completion of limit-line editing within the EDITLIML command ([page 170](#)).

---

## EDITLIML

### Edit Limit Line

#### Syntax



```
EDITLIML ;
```

#### Description

The EDITLIML command turns the currently active limit line off, and puts the analyzer into limit-line edit mode.

---

## ERR Error

### Syntax



### Description

The ERR command returns an integer list of error codes to the controller.

---

## ET Elapsed Time

### Syntax



### Description

The ET command returns to the controller the elapsed time (in hours) of analyzer operation.

---

## EX Exchange Trace A and Trace B

### Syntax



### Description

This command exchanges Trace A and Trace B, point by point.


---

#### NOTE

The functions of the EX command are identical to the AXB command ([page 126](#)) and to the XCH TRA,TRB command ([page 376](#)).

## FA Start Frequency

### Syntax



**Table 4-5**

Item	Description/Default	Range
REAL	Any real or integer number. Default unit is Hz.	Frequency range of the spectrum analyzer

Step Increment: Frequency span divided by 10

### Description


The FA command specifies the start frequency value. The start frequency is equal to the center frequency minus (the span divided by two) ( $FA = CF - SP/2$ ). Changing the start frequency changes the center frequency and span.

**NOTE**

The OA parameter only returns the current value to the controller. It does not set the active function to the start frequency.

# FB Stop Frequency

## Syntax



**Table 4-6**

Item	Description/Default	Range
REAL	Any real or integer number. Default unit is Hz.	Frequency range of the spectrum analyzer

Step Increment: Frequency span divided by 10

## Description

The FB command specifies the stop frequency value. The stop frequency is equal to the center frequency plus the span divided by two ( $FB = CF + SP/2$ ). Changing the stop frequency changes the center frequency and span.


**NOTE**

The OA parameter only returns the current value to the controller. It does not set the active function to the stop frequency.

---

## FDSP Frequency Display Off

### Syntax



### Description

The FDSP command turns the frequency annotation OFF.

---

#### NOTE

It is not possible to enable or disable the frequency annotation alone, leaving other annotations unaffected. Thus, the FDSP command behaves in the same way as ANNOT (page 117). If the FDSP command has been used to disable the frequency annotation, sending the ANNOT ON command will not re-enable the display annotation. The display annotation will only be displayed by sending the IP (page 191) command.

---



# FOFFSET Frequency Offset

## Syntax




Table 4-7

Item	Description/Default	Range
REAL	Any real or integer number. Default unit is Hz.	

Preset State: 0 Hz

## Description

The FOFFSET command selects a value that offsets the frequency scale for all absolute frequency readouts (for example, center frequency). Relative values such as span and marker delta are not offset.

When an offset is in effect, it is displayed beneath the bottom graticule line on the spectrum analyzer screen.

Execute “FOFFSET 0;” or “IP;” to turn off the offset.

### NOTE

The functions of the FOFFSET command are identical to the KSV command ([page 224](#)).

---

## FPKA Fast Preselector Peak

### Syntax



### Description

The FPKA command automatically adjusts the preselector frequency to yield the greatest signal level at the active marker. It returns the amplitude of the active marker.

---

**NOTE**


The FPKA command is only available when the analyzer's upper frequency limit is greater than 3 GHz. The command is not supported in analyzers with an upper frequency limit of 3 GHz or less, and will return an error message when used.

---

---

## FREF Frequency Reference

### Syntax



### Description

The FREF command specifies whether an external source or an internal source is being used.

---

**NOTE**

The FREF command is only supported on the PSA Series of analyzers. It is not supported on any of the ESA Series.

---

---

## FS Full Span

### Syntax



### Description

The FS command sets the frequency span of the spectrum analyzer to full span. Resolution bandwidth, video bandwidth, and sweep time are all set to auto-coupled.

---

**NOTE** The functions of the FS command are identical to the LF command ([page 229](#)).

---

**NOTE** Whenever the frequency range of the analyzer you are using does not match the remote language's own range, the span will be limited by the capabilities of the analyzer. The tables on the following pages list the frequency ranges for all the supported remote languages when running on any of Agilent's ESA-E Series, ESA-L Series and PSA Series of analyzers.

---

**Table 4-8 ESA-E Series - Frequency Ranges Set by the FS Command**

	<b>E4401B</b>	<b>E4402B</b>	<b>E4404B</b>	<b>E4405B</b>	<b>E4407B</b>
<b>Remote Language</b>	<b>Frequency Range</b>	<b>Frequency Range</b>	<b>Frequency Range</b>	<b>Frequency Range</b>	<b>Frequency Range</b>
8560E	0 Hz - 1.5 GHz	0 Hz - 2.9 GHz	0 Hz - 2.9 GHz	0 Hz - 2.9 GHz	0 Hz - 2.9 GHz
8561E	0 Hz - 1.5 GHz	0 Hz - 3.0 GHz	0 Hz - 6.5 GHz	0 Hz - 6.5 GHz	0 Hz - 6.5 GHz
8562E	0 Hz - 1.5 GHz	0 Hz - 3.0 GHz	0 Hz - 6.7 GHz	0 Hz - 13.2 GHz	0 Hz - 13.2 GHz
8563E	0 Hz - 1.5 GHz	0 Hz - 3.0 GHz	0 Hz - 6.7 GHz	0 Hz - 13.2 GHz	0 Hz - 26.5 GHz
8564E	0 Hz - 1.5 GHz	0 Hz - 3.0 GHz	0 Hz - 6.7 GHz	0 Hz - 13.2 GHz	0 Hz - 26.5 GHz
8565E	0 Hz - 1.5 GHz	0 Hz - 3.0 GHz	0 Hz - 6.7 GHz	0 Hz - 13.2 GHz	0 Hz - 26.5 GHz
8566A	0 Hz - 1.5 GHz	0 Hz - 2.5 GHz	0 Hz - 2.5 GHz	0 Hz - 2.5 GHz	0 Hz - 2.5 GHz
8566B	0 Hz - 1.5 GHz	0 Hz - 2.5 GHz	0 Hz - 2.5 GHz	0 Hz - 2.5 GHz	0 Hz - 2.5 GHz
8568A	0 Hz - 1.5 GHz	0 Hz - 1.5 GHz	0 Hz - 1.5 GHz	0 Hz - 1.5 GHz	0 Hz - 1.5 GHz
8568B	0 Hz - 1.5 GHz	0 Hz - 1.5 GHz	0 Hz - 1.5 GHz	0 Hz - 1.5 GHz	0 Hz - 1.5 GHz
8590L	0 Hz - 1.5 GHz	0 Hz - 1.8 GHz	0 Hz - 1.8 GHz	0 Hz - 1.8 GHz	0 Hz - 1.8 GHz
8591E	0 Hz - 1.5 GHz	0 Hz - 1.8 GHz	0 Hz - 1.8 GHz	0 Hz - 1.8 GHz	0 Hz - 1.8 GHz
8592L <sup>a</sup>	Not a suitable combination	2.75 GHz - 3.0 GHz	2.75 GHz - 6.7 GHz	2.75 GHz - 13.2 GHz	2.75 GHz - 22.0 GHz
8593E <sup>a</sup>	Not a suitable combination	2.75 GHz - 3.0 GHz	2.75 GHz - 6.7 GHz	2.75 GHz - 13.2 GHz	2.75 GHz - 22.0 GHz
8594E	0 Hz - 1.5 GHz	0 Hz - 2.9 GHz	0 Hz - 2.9 GHz	0 Hz - 2.9 GHz	0 Hz - 2.9 GHz
8595E	0 Hz - 1.5 GHz	0 Hz - 3.0 GHz	0 Hz - 6.5 GHz	0 Hz - 6.5 GHz	0 Hz - 6.5 GHz
8596E <sup>a</sup>	0 Hz - 1.5 GHz	0 Hz - 3.0 GHz	0 Hz - 6.7 GHz	0 Hz - 12.8 GHz	0 Hz - 12.8 GHz

a. The command HNLOCK (Harmonic Lock) is not supported in Option 266. Thus, the different frequency spans associated with the various harmonic bands are also not supported.

**Table 4-9**      **ESA-L Series - Frequency Ranges Set by the FS Command**

	<b>E4411B</b>	<b>E4403B</b>	<b>E4408B</b>
<b>Remote Language</b>	<b>Frequency Range</b>	<b>Frequency Range</b>	<b>Frequency Range</b>
8590L	0 Hz - 1.5 GHz	0 Hz - 1.8 GHz	0 Hz - 1.8 GHz
8591E	0 Hz - 1.5 GHz	0 Hz - 1.8 GHz	0 Hz - 1.8 GHz
8592L <sup>a</sup>	Not a suitable combination	2.75 GHz - 3.0 GHz	2.75 GHz - 22.0 GHz
8593E <sup>a</sup>	Not a suitable combination	2.75 GHz - 3.0 GHz	2.75 GHz - 22.0 GHz
8594E	0 Hz - 1.5 GHz	0 Hz - 2.9 GHz	0 Hz - 2.9 GHz
8595E	0 Hz - 1.5 GHz	0 Hz - 3.0 GHz	0 Hz - 6.5 GHz
8596E <sup>a</sup>	0 Hz - 1.5 GHz	0 Hz - 3.0 GHz	0 Hz - 12.8 GHz

a. The command HNLOCK (Harmonic Lock) is not supported in Option 266. Thus, the different frequency spans associated with the various harmonic bands are also not supported.


**Table 4-10 PSA Series - Frequency Ranges Set by the FS Command**

	<b>E4440A</b>	<b>E4443A</b>	<b>E4445A</b>	<b>E4446A</b>	<b>E4447A</b>	<b>E4448A</b>
<b>Remote Language</b>	<b>Frequency Range</b>	<b>Frequency Range</b>	<b>Frequency Range</b>	<b>Frequency Range</b>	<b>Frequency Range</b>	<b>Frequency Range</b>
8560E	0 Hz - 2.9 GHz	0 Hz - 2.9 GHz	0 Hz - 2.9 GHz	0 Hz - 2.9 GHz	0 Hz - 2.9 GHz	0 Hz - 2.9 GHz
8561E	0 Hz - 6.5 GHz	0 Hz - 6.5 GHz	0 Hz - 6.5 GHz	0 Hz - 6.5 GHz	0 Hz - 6.5 GHz	0 Hz - 6.5 GHz
8562E	0 Hz - 13.2 GHz	0 Hz - 6.7 GHz	0 Hz - 13.2 GHz	0 Hz - 13.2 GHz	0 Hz - 13.2 GHz	0 Hz - 13.2 GHz
8563E	0 Hz - 26.5 GHz	0 Hz - 6.7 GHz	0 Hz - 13.2 GHz	0 Hz - 26.5 GHz	0 Hz - 26.5 GHz	0 Hz - 26.5 GHz
8564E	0 Hz - 26.5 GHz	0 Hz - 6.7 GHz	0 Hz - 13.2 GHz	0 Hz - 40.0 GHz	0 Hz - 40.0 GHz	0 Hz - 40.0 GHz
8565E	0 Hz - 26.5 GHz	0 Hz - 6.7 GHz	0 Hz - 13.2 GHz	0 Hz - 44.0 GHz	0 Hz - 42.98 GHz	0 Hz - 50.0 GHz
8566A	0 Hz - 2.5 GHz	0 Hz - 2.5 GHz	0 Hz - 2.5 GHz	0 Hz - 2.5 GHz	0 Hz - 2.5 GHz	0 Hz - 2.5 GHz
8566B	0 Hz - 2.5 GHz	0 Hz - 2.5 GHz	0 Hz - 2.5 GHz	0 Hz - 2.5 GHz	0 Hz - 2.5 GHz	0 Hz - 2.5 GHz
8568A	0 Hz - 1.5 GHz	0 Hz - 1.5 GHz	0 Hz - 1.5 GHz	0 Hz - 1.5 GHz	0 Hz - 1.5 GHz	0 Hz - 1.5 GHz
8568B	0 Hz - 1.5 GHz	0 Hz - 1.5 GHz	0 Hz - 1.5 GHz	0 Hz - 1.5 GHz	0 Hz - 1.5 GHz	0 Hz - 1.5 GHz
8590L	0 Hz - 1.8 GHz	0 Hz - 1.8 GHz	0 Hz - 1.8 GHz	0 Hz - 1.8 GHz	0 Hz - 1.8 GHz	0 Hz - 1.8 GHz
8591E	0 Hz - 1.8 GHz	0 Hz - 1.8 GHz	0 Hz - 1.8 GHz	0 Hz - 1.8 GHz	0 Hz - 1.8 GHz	0 Hz - 1.8 GHz
8592L <sup>a</sup>	2.75 GHz - 22.0 GHz	2.75 GHz - 6.7 GHz	2.75 GHz - 13.2 GHz	2.75 GHz - 22.0 GHz	2.75 GHz - 22.0 GHz	2.75 GHz - 22.0 GHz
8593E <sup>a</sup>	2.75 GHz - 22.0 GHz	2.75 GHz - 6.7 GHz	2.75 GHz - 13.2 GHz	2.75 GHz - 22.0 GHz	2.75 GHz - 22.0 GHz	2.75 GHz - 22.0 GHz
8594E	0 Hz - 1.5 GHz	0 Hz - 2.9 GHz	0 Hz - 2.9 GHz	0 Hz - 2.9 GHz	0 Hz - 2.9 GHz	0 Hz - 2.9 GHz
8595E	0 Hz - 1.5 GHz	0 Hz - 6.5 GHz	0 Hz - 6.5 GHz	0 Hz - 6.5 GHz	0 Hz - 6.5 GHz	0 Hz - 6.5 GHz
8596E <sup>a</sup>	0 Hz - 1.5 GHz	0 Hz - 6.7 GHz	0 Hz - 12.8 GHz	0 Hz - 12.8 GHz	0 Hz - 12.8 GHz	0 Hz - 12.8 GHz

a. The command HNLOCK (Harmonic Lock) is not supported in Option 266. Thus, the different frequency spans associated with the various harmonic bands are also not supported.

# GRAT Graticule

## Syntax



Preset State: GRAT ON

## Description

Turns the graticule on or off.

---

**NOTE** The functions of the GRAT command are identical to the KSm command ([page 217](#)) and the KSn command ([page 219](#)).

---

**NOTE** The query command GRAT? returns different responses depending on the language being used.


- |             |  |
|-------------|--|
| 8560 Series | Query response is either 1 or 0, indicating ON or OFF state. |
| 8566A       | Query response is either ON or OFF.                          |
| 8566B       | Query response is either ON or OFF.                          |
| 8568A       | Query response is either ON or OFF.                          |
| 8568B       | Query response is either ON or OFF.                          |
| 8590 Series | Query response is either ON or OFF.                          |
-



---

## HD Hold Data Entry

### Syntax



### Description

Disables data entry via the spectrum analyzer numeric keypad, knob, or step keys. The active function readout is blanked, and any active function is deactivated.

---

#### NOTE

This command will only be supported when followed by the parameter AUTO ([page 123](#)). All functions will then be auto-coupled.

---

## I1 [one] Set RF Coupling to DC

### Syntax

### Description

The I1 [one] command sets the RF coupling to DC if your analyzer is capable of being switched. If your analyzer cannot be switched to DC coupling, the command will have no effect but an error message will not be generated.

Whether or not your analyzer can be AC coupled, DC coupled, or both depends on both the analyzer's model number, and on whether or not Option UKB (Low Frequency Extension) has been installed. The tables below list the frequency specifications for all ESA-E series and PSA analyzers for both DC and AC coupling.

**Table 4-11 8568A/B Analyzer Frequency Coupling Specifications**

Analyzer Model	DC Coupled Range		AC Coupled Range	
	Min. Freq.	Max. Freq.	Min. Freq.	Max. Freq.
8568A/B	100 Hz	1.5 GHz	100 kHz	1.5 GHz

**Table 4-12 PSA Series Analyzer Frequency Coupling Specifications**

Analyzer Model (PSA series)	DC Coupled Range		AC Coupled Range	
	Min. Freq.	Max. Freq.	Min. Freq.	Max. Freq.
E4440A	3 Hz	26.5 GHz	10 MHz	26.5 GHz
E4443A	3 Hz	6.7 GHz	10 MHz	6.7 GHz
E4445A	3 Hz	13.2 GHz	10 MHz	13.2 GHz
E4446A	3 Hz	44 GHz	N/A	N/A
E4447A	3 Hz	42.98 GHz	N/A	N/A
E4448A	3 Hz	50 GHz	N/A	N/A

**Table 4-13**      **ESA Series Analyzer Frequency Coupling Specifications**

Analyzer Model (ESA-E series)	Installed Options	DC Coupled Range		AC Coupled Range	
		Min. Freq.	Max. Freq.	Min. Freq.	Max. Freq.
E4401B		N/A	N/A	9 kHz	1.5 GHz
E4402B		N/A	N/A	9 kHz	3.0 GHz
E4402B	UKB	100 Hz	3.0 GHz	100 kHz	3.0 GHz
E4404B		9 kHz	6.7 GHz	100 kHz	6.7 GHz
E4404B	UKB	100 Hz	6.7 GHz	100 kHz	6.7 GHz
E4405B		9 kHz	13.2 GHz	100 kHz	13.2 GHz
E4405B	UKB	100 Hz	13.2 GHz	100 kHz	13.2 GHz
E4407B		9 kHz	26.5 GHz	N/A	N/A
E4407B	UKB	100 Hz	26.5 GHz	10 MHz	26.5 GHz

**NOTE**      The ESA Series and PSA Series analyzers only have a single RF input port.

## I2 [two] Set RF Coupling to AC

### Syntax

```
I2 → ;
```

### Description

The I2 [two] command sets the RF coupling to AC if your analyzer is capable of being switched. If your analyzer cannot be switched to AC coupling, the command will have no effect but an error message will not be generated.

Whether or not your analyzer can be AC coupled, DC coupled, or both depends on both the analyzer's model number, and on whether or not Option UKB (Low Frequency Extension) has been installed. The tables below list the frequency specifications for all ESA-E series and PSA analyzers for both DC and AC coupling.

**Table 4-14 8568A/B Analyzer Frequency Coupling Specifications**

Analyzer Model	DC Coupled Range		AC Coupled Range	
	Min. Freq.	Max. Freq.	Min. Freq.	Max. Freq.
8568A/B	100 Hz	1.5 GHz	100 kHz	1.5 GHz

**Table 4-15 PSA Series Analyzer Frequency Coupling Specifications**

Analyzer Model (PSA series)	DC Coupled Range		AC Coupled Range	
	Min. Freq.	Max. Freq.	Min. Freq.	Max. Freq.
E4440A	3 Hz	26.5 GHz	10 MHz	26.5 GHz
E4443A	3 Hz	6.7 GHz	10 MHz	6.7 GHz
E4445A	3 Hz	13.2 GHz	10 MHz	13.2 GHz
E4446A	3 Hz	44 GHz	N/A	N/A
E4447A	3 Hz	42.98 GHz	N/A	N/A
E4448A	3 Hz	50 GHz	N/A	N/A

**Table 4-16**      **ESA Series Analyzer Frequency Coupling Specifications**


		DC Coupled Range		AC Coupled Range	
Analyzer Model (ESA-E series)	Installed Options	Min. Freq.	Max. Freq.	Min. Freq.	Max. Freq.
E4401B		N/A	N/A	9 kHz	1.5 GHz
E4402B		N/A	N/A	9 kHz	3.0 GHz
E4402B	UKB	100 Hz	3.0 GHz	100 kHz	3.0 GHz
E4404B		9 kHz	6.7 GHz	100 kHz	6.7 GHz
E4404B	UKB	100 Hz	6.7 GHz	100 kHz	6.7 GHz
E4405B		9 kHz	13.2 GHz	100 kHz	13.2 GHz
E4405B	UKB	100 Hz	13.2 GHz	100 kHz	13.2 GHz
E4407B		9 kHz	26.5 GHz	N/A	N/A
E4407B	UKB	100 Hz	26.5 GHz	10 MHz	26.5 GHz

**NOTE**      The ESA Series and PSA Series analyzers only have a single RF input port.

---

## ID Identify

### Syntax



### Description

The ID command returns the current remote language to the controller (for example, “HP8566B”).

---

**NOTE**

The response value is determined by your remote language selection. This is configured via the front-panel menu selection for remote ID. See [“Configuring Option 266 on PSA Analyzers” on page 38](#) for details.

---

---

## IP Instrument Preset

### Syntax



### Description

Performs an instrument preset, setting the analyzer back to its factory settings.

Instrument preset automatically occurs when you turn on the spectrum analyzer. IP is a good starting point for many measurement processes. When IP is executed remotely, the spectrum analyzer does not necessarily execute a complete sweep, however. You should execute a take sweep (TS) to ensure that the trace data is valid after an IP.

---

**NOTE** The functions of the IP command are identical to the KST command ([page 223](#)).

---


---

**NOTE** If the external amplifier gain has been set, executing an IP command will not reset this value. This is to protect the analyzer.

---

## KS, Mixer Level

### Syntax



### Description

The KS, command specifies the maximum signal level that is applied to the input mixer for a signal that is equal to or below the reference level.

The effective mixer level is equal to the reference level minus the input attenuator setting. When KS, is activated, the effective mixer level can be set from -10 dBm to -70 dBm in 10 dB steps.

As the reference level is changed, the coupled input attenuator automatically changes to limit the maximum signal at the mixer input to your specified setting for signals less than or equal to the reference level.

---

**NOTE** If the external amplifier gain has been set, executing an IP command will not reset this value. This is to protect the analyzer.

---


**NOTE** The functions of the KS, command are identical to the ML command ([page 289](#)).

---



## KS= Marker Counter Resolution

### Syntax



**NOTE** The marker counter can be specified in time units when operating in the time domain.

### Description

Specifies the resolution of the marker frequency counter.

**NOTE** The functions of the KS= command are identical to the MKFCR command ([page 269](#)).

**NOTE** On PSA analyzers, setting the marker frequency resolution will cause the Gate Time to change. The Gate Time is calculated using the following formula:

$$\text{Gate Time} = \frac{1}{\text{Resolution Value}}$$

For restrictions on the Gate Time, refer to the *PSA User's and Programmer's Reference, Volume 1*.

**NOTE** This command is only supported in the 8568 remote language. The same KS= command has a different function (selects the factory preselector setting) in the 8566 language, and is not supported.

---

## KSA Amplitude in dBm

### Syntax



### Description

The KSA command sets the amplitude readout (reference level, marker, display line and threshold) to dBm units.

---

**NOTE**

The functions of the KSA command are identical to the AUNITS DBM command ([page 121](#)).

---

---

## KSa Normal Detection

### Syntax



### Description

The KSa command selects normal input detection. That is, it enables the *rosenfell* detection algorithm that selectively chooses between positive and negative values.

---

**NOTE** Normal detection is only possible in PSA series analyzers. The KSa command will default to positive peak detection (KSb) ([page 197](#)) when executed on an ESA analyzer.

---

**NOTE** The functions of the KSa command are identical to the DET NRM command ([page 158](#)).

---

---

## KSB Amplitude in dBmV

### Syntax



### Description

The KSB command sets the amplitude readout (reference level, marker, display line and threshold) to dBmV units.

---

**NOTE**

The functions of the KSB command are identical to the AUNITS DBMV command ([page 121](#)).

---

---

## KSb Positive Peak Detection

### Syntax



### Description

The KSb command enables positive peak input detection for displaying trace information. Trace elements are only updated when the detected signal level is greater than the previous signal level.

---

#### NOTE

The functions of the KSb command are identical to the DET POS command ([page 158](#)).

---

---

## KSC Amplitude in dBuV

### Syntax



```
KSC ;
```

### Description

The KSC command sets the amplitude readout (reference level, marker, display line and threshold) to dBuV units.

---

**NOTE**

The functions of the KSC command are identical to the AUNITS DBUV command ([page 121](#)).

---

---

## KSc A Plus B to A

### Syntax



### Description

The KSc command does a point-by-point addition of Trace A and Trace B, and sends the results to Trace A. Thus, if your input signal remains unchanged, KSc can restore the original trace after an AMB or a C2 command has been executed.

---

#### NOTE

The functions of the command KSc are identical to the APB command ([page 118](#)).

---

---

## KSD Amplitude in Volts

### Syntax



### Description

The KSD command sets the amplitude readout (reference level, marker, display line and threshold) to voltage units.

---

**NOTE**

The functions of the KSD command are identical to the AUNITS V command ([page 121](#)).

---



---

## KSd Negative Peak Detection

### Syntax



### Description

The KSd command selects negative-peak input detection for displaying trace information. Each trace element is updated with the minimum value detected during the sweep.

---

#### NOTE


The functions of the command KSd are identical to the DET NEG command ([page 158](#)).

---

---

## KSE Title Mode

### Syntax



### Description

The KSE command activates the title mode, writing a message to the top line of the display.

---

**NOTE**


The only characters that Option 266 will accept as a terminator are the '@' character and the carriage return.

---

---

## KSe Sample Detection

### Syntax



### Description

The KSe command selects sample input detection for displaying trace information.

---


#### NOTE

The functions of the command KSe are identical to the DET SMP command ([page 158](#)).

---

## KSG Video Averaging On

### Syntax



### Description

The KSG command enables video averaging. The averaged trace is displayed in Trace A.

If video averaging is off when either KSG? or KSG OA is sent to the instrument, video averaging is turned on and the current average count is returned to the controller.

#### NOTE

The functions of the KSG command are identical to the VAVG ON command ([page 368](#)).

---

## K<sub>Sg</sub> Display Off

### Syntax



### Description

The K<sub>Sg</sub> command turns the analyzer's display Off.

---

#### NOTE


On the legacy spectrum analyzers, this command turned the CRT beam power off to avoid unnecessary wear on the CRT. Although this command is supported, displays used on the PSA Series analyzers have a much longer life than the CRTs used in the legacy spectrum analyzers.

---

---

## KSH Video Averaging Off

### Syntax



### Description

The KSH command switches video averaging Off.

---

**NOTE**

The functions of the KSH command are identical to the VAVG OFF command ([page 368](#)).

---

---

## KSh Display On

### Syntax



### Description

The KSh command turns the analyzer's display On.

---

#### NOTE

On the early models of spectrum analyzers, CRT beam power was often switched Off to prevent wear of the CRT. This command (KSh) was used to turn the CRT beam power on again. Although this command is supported, displays used on the PSA Series analyzers have a much longer life than the CRTs used in the legacy spectrum analyzers.

---

---

## KSI Extend Analyzer Reference Level

### Syntax



```
KSI ;
```

### Description

The KSI command was used in the legacy analyzers to extend the analyzer reference level range. As the PSA analyzer already has a minimum reference level of  $-170$  dBm and a maximum reference level of  $+30$  dBm, this command has no effect in Option 266 Programming Code Compatibility Suite. Issuing this command will not generate an error in Option 266 Programming Code Compatibility Suite.



---

## KSi Exchange Trace B and Trace C

### Syntax



### Description

The KSi command exchanges Trace B data with Trace C data.

---

#### NOTE

Trace C cannot be an active trace. This means that the data in Trace C cannot be updated as the analyzer sweeps. To ensure that the current settings of the analyzer are reflected in the data exchanged between Trace B and Trace C, you must follow the four step process below.

- Select single sweep mode (S2 ([page 327](#)) or SNGLS command ([page 338](#)))
- Select the desired analyzer settings
- Take one complete sweep using the TS command ([page 367](#))
- Exchange the data

---


#### NOTE

The functions of the command KSi are identical to the BXC command ([page 135](#)) and the XCH TRB,TRC command ([page 376](#)).

---

## KSj View Trace C

### Syntax



### Description

The KSj command displays Trace C.

---

**NOTE**

The functions of the command KSj are identical to the VIEW TRC command ([page 374](#)).

---

---

## KSK Marker to Next Peak

### Syntax



### Description

If there is a marker on the screen, the KSK command moves this marker to the next signal peak of lower amplitude.

---

#### NOTE

The functions of the KSK command are similar to the MKPK NH command ([page 276](#)), except that KSK does not take into account the marker peak threshold value or the marker peak excursion value. For more details on marker peak threshold, see the MKPT command ([page 277](#)) and the TH command ([page 356](#)). For more details on marker peak excursion, see the MKPX command ([page 278](#)).

---

---

## KSk Blank Trace C

### Syntax



### Description

The KSk command blanks Trace C.

---

**NOTE**

The functions of the command KSk are identical to the BLANK TRC command ([page 132](#))

---

---

## KSL Marker Noise Off

### Syntax



### Description

The KSL command disables the noise density function which displays the RMS noise density at the marker. KSL does not blank the marker.

---

#### NOTE

The functions of the KSL command are identical to the MKNOISE OFF command ([page 272](#)).

---

---

## KSI Transfer Trace B to Trace C

### Syntax



### Description

The KSI command transfers Trace B data to Trace C

---

**NOTE**

Trace C cannot be an active trace. This means that the data in Trace C cannot be updated as the analyzer sweeps. To ensure that the current settings of the analyzer are reflected in the data transferred from Trace B to Trace C, you must follow the four step process below.

- Select single sweep mode (S2 ([page 327](#)) or SNGLS command ([page 338](#)))
- Select the desired analyzer settings
- Take one complete sweep using the TS command ([page 367](#))
- Transfer the data


---

**NOTE**

The functions of the command KSI are identical to the BTC command ([page 134](#)).

## KSM Marker Noise On

### Syntax



### Description

The KSM command displays the noise density at the marker. The noise density is normalized to a 1Hz bandwidth.

**NOTE** The functions of the KSM command are identical to the MKNOISE ON command ([page 272](#)).

**NOTE** Some differences in marker noise may be seen between the legacy analyzers and the PSA Series analyzers due to the greater dynamic range of the PSA Series.

**NOTE** *All legacy analyzer languages:* If either the M3 command or the MKD command is executed with the marker noise function active (MKNOISE ON or KSM), the marker amplitude displayed and returned by the MKA? command ([page 261](#)) or the MA command ([page 251](#)) is the difference between the noise densities at the reference marker and at the delta marker position.

*8566, 8568 and 8590 Series only:* If either the M3 command or the MKD command is executed before marker noise has been activated (using the MKNOISE ON or KSM commands), the marker noise amplitude that is displayed on the screen is the difference between the carrier wave power and the noise density at the delta marker position. Regardless of the order in which Marker Noise and Delta Marker are activated, the marker amplitude displayed and returned by the MKA? command ([page 261](#)) or the MA command ([page 251](#)) is the difference between the noise densities at the reference marker and at the delta marker. That is, the value returned by MKA? and MA does not always agree with that displayed on the screen of the PSA Series analyzer. These returned values will only be correct as long as there has been no change in either the delta marker state or the marker noise state from the front panel.

*8560 Series only:* If either the M3 command or the MKD command is executed before marker noise has been activated (using the

## Programming Commands

**KSM Marker Noise On**

MKNOISE ON or KSM commands), the marker noise amplitude that is displayed on the screen is the difference between the carrier wave power and the noise density at the delta marker position. The value returned by the MKA? or MA command is the difference between the carrier wave power and the noise density at the delta marker position. That is, the value returned by MKA? and MA will agree with that displayed on the screen of the PSA Series analyzer.

---

**NOTE**

The nominal noise bandwidth to RBW ratio for the RBW filters used in the ESA series analyzers is 1.128 for RBWs of 1 kHz and higher, and 1.055 for RBWs of less than 1 kHz.

The nominal noise bandwidth to RBW ratio for PSA analyzers is 1.055 for all RBWs. The nominal ratio for the 8566 family of analyzers is 1.128 for RBWs 100 kHz and higher, and 1.114 for RBWs of 30kHz and lower.

---



---

## KSm Graticule Off

### Syntax



### Description

The KSm command blanks the graticule on the analyzer display.

---

#### NOTE

The functions of the command KSm are identical to the GRAT OFF command ([page 184](#)).

---

## KSN Marker Minimum

### Syntax



### Description

The KSN command moves the marker to the minimum value detected.

---

**NOTE**

The functions of the KSN command are identical to the MKMIN command ([page 270](#)).

---

---

## KSn Graticule On

### Syntax



### Description

The KSn command turns on the graticule on the analyzer display.

---

#### NOTE

The functions of the command KSn are identical to the GRAT ON command ([page 184](#)).

---

## KSO Marker Span

### Syntax



### Description

The KSO command operates only when the delta marker is On (see MKD [\(page 265\)](#) or M3 [\(page 248\)](#)). When the delta marker is on and the KSO command is executed, the left marker specifies the start frequency, and the right marker specifies the stop frequency. If the delta marker is off, the command does nothing.

---

**NOTE** If the active marker is not a delta marker, there is no change in its position.

---

---


**NOTE** The functions of the KSO command are identical to the MKSP command [\(page 282\)](#).

---

---

## KSo Annotation Off

### Syntax



### Description

The KSo command blanks the annotation on the analyzer display.

---

**NOTE**

The functions of the command KSo are identical to the ANNOT OFF command ([page 117](#)).

---

## KSp Annotation On

### Syntax



### Description

The KSp command activates the annotation on the analyzer display.

---

**NOTE**


The functions of the command KSp are identical to the ANNOT ON command ([page 117](#)).

---

---

## KST Fast Preset

### Syntax



### Description

Performs an instrument preset, setting the analyzer back to its factory settings.

---


#### NOTE

There is no fast preset for ESA and PSA series analyzers. Instead, the Code Compatibility software performs an instrument preset (IP) when the KST command is issued. The functions of the command KST are therefore identical to the IP command ([page 191](#)).

---

## KSV Frequency Offset

### Syntax



### Description

The KSV command allows you to specify a value that offsets the frequency scale for all absolute frequency readouts, for example, center frequency. Relative values, for example, span and delta marker, are not offset.

#### NOTE

The functions of the KSV command are identical to the FOFFSET command ([page 177](#)).



## KSx External Trigger

### Syntax



### Description

The KSx command activates the normal external trigger mode. When the KSx command is executed, the RF input signal is only displayed when the external trigger level exceeds the trigger threshold level.

#### NOTE

If an 8566A/B or an 8568A/B analyzer is in zero span and the sweep time is less than 20 msec, the display only gets refreshed when a fresh trace has been taken. This can cause the displayed trace to flicker.

Equally, if any of the 8560-E Series of analyzers, that is the 8560E, 8561E, 8563E, 8564E and the 8565E, are in zero span and with a sweep time of less than 50 msec, and they do not have Option 007 - Fast Digitized Time Domain installed, the display only gets refreshed when a fresh trace has been taken. This can cause the displayed trace to flicker.

In Option 266, all traces are displayed continuously, so all traces are therefore free of flicker.

#### NOTE

The functions of the command KSx are identical to the TM EXT command ([page 359](#)).

---

## KSy Video Trigger

### Syntax



### Description

The KSy command activates the normal video trigger mode. When the KSy command is executed, the RF input signal is only displayed when the video trigger signal, which is internally triggered off the input signal, exceeds the trigger threshold level.

---

#### NOTE

If an 8566A/B or an 8568A/B analyzer is in zero span and the sweep time is less than 20 msec, the display only gets refreshed when a fresh trace has been taken. This can cause the displayed trace to flicker.

Equally, if any of the 8560-E Series of analyzers, that is the 8560E, 8561E, 8563E, 8564E and the 8565E, are in zero span and with a sweep time of less than 50 msec, and they do not have Option 007 - Fast Digitized Time Domain installed, the display only gets refreshed when a fresh trace has been taken. This can cause the displayed trace to flicker.

In Option 266, all traces are displayed continuously, so all traces are therefore free of flicker.

---


#### NOTE

The functions of the command KSy are identical to the TM VID command ([page 359](#)) and to the T4 command ([page 352](#)).

---

# KSZ Reference Level Offset

## Syntax



## Description

The KSZ command offsets all amplitude readouts on the display but without affecting the trace.

Once activated, the KSZ command displays the amplitude offset on the left side of the screen.

Entering KSZ 0 or presetting the spectrum analyzer eliminates an amplitude offset.

**NOTE**

The functions of the KSZ command are identical to the ROFFSET command ([page 324](#)).

---

## L0 [zero] Display Line Off

### Syntax



### Description

The L0 [zero] command disables the display line.

---

**NOTE**

The functions of the L0 [zero] command are identical to the DLE OFF command ([page 162](#)).

---

---

## LF Low Frequency Preset

### Syntax



### Description

Performs a low frequency preset. That is, it selects a Start Frequency of 0 Hz and a Stop Frequency of 2.5 GHz, a Reference Level of 0dBm, and sets all coupled functions to automatic.

---


#### NOTE

If you are using an E4401B or E4411B analyzer, the Stop Frequency will be clipped to the limits of the analyzer, that is 1.5 GHz.

---

# LG Logarithmic Scale

## Syntax



## Description

Specifies the amplitude (vertical graticule divisions) as logarithmic units, without changing the reference level. The integer ranges vary between the different remote languages. The following table lists the ranges for each remote language.

**Table 4-17**

Remote Language	Integer Range using the LG Command
8560E/EC	1, 2, 5, and 10
8561E/EC	1, 2, 5, and 10
8562E/EC	1, 2, 5, and 10
8563E/EC	1, 2, 5, and 10
8564E/EC	1, 2, 5, and 10
8565E/EC	1, 2, 5, and 10
8566A	1, 2, 5, and 10
8566B	1, 2, 5, and 10
8568A	1, 2, 5, and 10
8568B	1, 2, 5, and 10
8590L	0.1 dB to 20.0 dB
8591E	0.1 dB to 20.0 dB
8592L	0.1 dB to 20.0 dB


**Table 4-17**

<b>Remote Language</b>	<b>Integer Range using the LG Command</b>
8593E	0.1 dB to 20.0 dB
8594E	0.1 dB to 20.0 dB
8594L	0.1 dB to 20.0 dB
8595E	0.1 dB to 20.0 dB
8596E	0.1 dB to 20.0 dB

---

## LIMD Limit Line Delta Value

### Syntax



### Description

The LIMD command is used to enter the delta value for the amplitude of a limit line segment.

---

**NOTE**


The response to the query command LIMD? is the delta value for the segment currently selected with the SEDI command ([page 332](#)).

---



## LIMF Limit Line Frequency Value

### Syntax



### Description

The LIMF command is used to enter a frequency value for a limit-line segment.

#### NOTE

The response to the query command LIMF? is the delta value for the segment currently selected with the SEDI command ([page 332](#)).

---

## LIMIDEL Delete Limit Line Table

### Syntax




```
LIMIDEL ;
```

### Description

The LIMIDEL command deletes all upper and lower segments in the current limit-line table.

## LIMIDISP Limit Line Display

### Syntax




### Description

Controls the display of the limit line (or limit lines).

---

## LIMIFAIL Limits Failed

### Syntax



### Description

The LIMIFAIL command returns a number between 0 and 3 which specifies whether the active trace passed or failed the upper and lower limit line tests. The meanings of the returned numbers are shown below.

**Table 4-18**


**Results of the LIMIFAIL Command**

Result	Meaning
0	The active trace passed both the upper and the lower limit tests.
1	The active trace failed the lower limit test.
2	The active trace failed the upper limit test.
3	The active trace failed both the upper and the lower limit tests.

---

## LIMIFT Select Frequency or Time Limit Line

### Syntax



### Description

The LIMIFT command determines how the limit-line segments are defined. They can be defined according to frequency, or according to the sweep time setting of the spectrum analyzer.

---

## LIMIPURGE Delete Current Limit Line

### Syntax



```
LIMIPURGE ;
```


### Description

The LIMIPURGE command deletes the current limit line.

---

## LIMIREL Relative Limit Lines

### Syntax




### Description

Specifies whether the current limit lines are fixed or relative.

## LIMITEST Enable Limit Line Testing

### Syntax



### Description


The LIMITEST command compares trace A with the current limit line data.



---

## LIML Lower-Limit Amplitude

### Syntax




### Description

The LIML command is used within the SEDI command ([page 332](#)) to assign the lower-limit amplitude value to a limit-line segment.

## LIMM Limit Middle-Amplitude

### Syntax




### Description

The LIMM command is used within the SEDI command ([page 332](#)) to assign the middle amplitude value to a limit-line segment.

## LIMU Upper-Limit Amplitude

### Syntax



### Description

The LIMU command is used within the SEDI command ([page 332](#)) to assign the upper-limit amplitude value to a limit-line segment.

---

## LN Linear Scale

### Syntax



### Description

Scales the amplitude (vertical graticule divisions) proportional to the input voltage, without changing the reference level. The bottom line of the graticule represents 0 volts.

---

## LSPAN Previous Span

### Syntax



### Description

The LSPAN command changes the spectrum analyzer's span to the previous span setting.

---

## M1 [*one*] Marker Off

### Syntax



### Description

The M1 [*one*] command blanks any markers showing on the display.

---


**NOTE**

The functions of the M1 [*one*] command are identical to the MKOFF ALL command ([page 274](#)).

---

## M2 [two] Marker Normal

### Syntax



### Description

The M2 [two] command moves the active marker to the marker frequency. If the active marker type is not currently normal (for example, it is delta or peak), the M2 command will change it to a normal marker.

---

**NOTE** The functions of the M2 command are identical to the MKN command (page 271).


---

**NOTE** Model numbers 8566, 8588, and the 8590 Series only: If the active marker has not been declared with MKACT, a normal marker is turned on and this active marker is assumed to be marker number 1 [one].

---

## M3 [three] Delta Marker

### Syntax



### Description

The M3 [three] command computes the frequency and amplitude difference between the active marker and the delta (or difference) marker.

If a delta marker is not displayed on the screen, the M3 command places one at the specified frequency or on the right hand edge of the display. If an active marker is not displayed on the screen, the M3 command places an active marker at the center of the screen.

#### NOTE

The active marker is the number 1 marker unless otherwise specified by the MKACT command ([page 262](#)).



---

**NOTE**

*All legacy analyzer languages:* If the M3 command is executed with the marker noise function active (MKNOISE ON or KSM), the marker amplitude displayed and returned by the MKA? command ([page 261](#)) or the MA command ([page 251](#)) is the difference between the noise densities at the reference marker and at the delta marker position.

*8566, 8568 and 8590 Series only:* If the M3 command is executed before marker noise has been activated (using the MKNOISE ON or KSM commands), the marker noise amplitude that is displayed on the screen is the difference between the carrier wave power and the noise density at the delta marker position. Regardless of the order in which Marker Noise and Delta Marker are activated, the marker amplitude displayed and returned by the MKA? command ([page 261](#)) or the MA command ([page 251](#)) is the difference between the noise densities at the reference marker and at the delta marker. That is, the value returned by MKA? and MA does not always agree with that displayed on the screen of the PSA Series analyzer. These returned values will only be correct as long as there has been no change in either the delta marker state or the marker noise state from the front panel.

*8560 Series only:* If the M3 command is executed before marker noise has been activated (using the MKNOISE ON or KSM commands), the marker noise amplitude that is displayed on the screen is the difference between the carrier wave power and the noise density at the delta marker position. The value returned by the MKA? or MA command is the difference between the carrier wave power and the noise density at the delta marker position. That is, the value returned by MKA? and MA will agree with that displayed on the screen of the PSA Series analyzer.

---

**NOTE**

The nominal noise bandwidth to RBW ratio for the RBW filters used in the ESA series analyzers is 1.128 for RBWs of 1 kHz and higher, and 1.055 for RBWs of less than 1 kHz.

The nominal ratio for PSA analyzers is 1.055 for all RBWs. The nominal ratio for the 8566 family of analyzers is 1.128 for RBWs 100 kHz and higher, and 1.114 for RBWs of 30kHz and lower.

---


**NOTE**

The functions of the M3 command are identical to the MKD command ([page 265](#)).

---

## M4 [four] Marker Zoom

### Syntax



### Description

The M4 [four] command, when specifying either the UP or DN parameter, increases or decreases the frequency span by one step. When specifying a numeric value, the M4 command moves the markers horizontal (X) position to the specified position in frequency or time.

**NOTE**

The OA option only returns the current value to the controller; it does not set the active function to the active marker.

---

## MA Marker Amplitude Output

### Syntax



### Description

The MA command returns the amplitude level of the active marker if the marker is on the screen. If both the active marker and the delta marker are displayed, the MA command returns the amplitude difference between the two markers.

---

**NOTE** The format of the returned data when using the MKA command is dependent on the currently set trace data format (refer to TDF, MDS, O1, O2, O3, or O4) when the selected remote language is either 8566 or 8590. When the selected remote language is 8560, the marker amplitude is always returned as an ASCII value (TDF P).

---

**NOTE** The functions of the MA command are identical to the MKA command ([page 261](#)).

---

---

## MC0 [zero] Marker Frequency Counter Off

### Syntax



### Description

The MC0 [zero] command turns the marker frequency counter off.

---

**NOTE**

The functions of the MC0 [zero] command are identical to the MKFC OFF command ([page 268](#)).

---

---

## MC1 [one] Marker Frequency Counter On

### Syntax



### Description

The MC1 [one] command turns the marker frequency counter on.


---

#### NOTE

The functions of the MC1 [one] command are identical to the MKFC ON command ([page 268](#)).

## MDS Measurement Data Size

### Syntax



### Description

The MDS command formats binary data in one of the following formats:


- B** selects a data size of one 8-bit byte.
- W** selects a data size of one word, which is two 8-bit bytes.

If no keyword is specified in the command, the default value of W is assumed.

---

## MEAN Trace Mean

### Syntax



### Description

Returns the mean value of the specified trace in display units.

---

#### NOTE


TRA corresponds to Trace 1, TRB corresponds to Trace 2, and TRC corresponds to Trace 3.

---

## MEANPWR

### Mean Power measurement

#### Syntax



#### Description

The MEANPWR command measures the average power of the carrier during that portion of the time when it is on. The on state is defined as the time when the signal is within a selected number of dB of its peak level. The range of amplitudes that is defined as the on state can be set with the command. The amplitude range is set relative to the peak value of the signal.

#### NOTE

The MEANPWR command is similar to the CARRON command ([page 140](#)), except that the CARRON command defines 'on' as that time when the signal is within 20dB of its peak level.



---

## MEASOFF Measurement Off

### Syntax

MEASOFF → ;

### Description

Turns the current measurement off if the current measurement is ACP (page 95) or CHP (page 144).

---

#### NOTE

If ACPPAR (page 107) is set to automatic, the MEASOFF command returns the following settings back to their pre-measurement states:

- Frequency span
- Resolution bandwidth
- Video bandwidth
- Center frequency step size
- Sweep time
- Detector mode
- Amplitude scale

It does not do any of the following:


- Change the values of ACPSP or ACPBW
- Restore trace elements
- Restore trigger mode
- Restore Amplitude units
- Restore any trace math functions

---


## MF Marker Frequency Output

### Syntax

#### 8566 and 8568 Remote Language



#### 8560 Series and 8590 Series Remote Language



### Description

Returns the frequency (or time) of the on-screen active marker. If both an active marker and the delta marker are on the screen, the frequency difference is returned.

---

**NOTE**



8566 and 8568 only: If the active marker has marker frequency count set to *On* when using the MF? command, the marker frequency count value is returned to the controller.

---

---

## MINH Minimum Hold

### Syntax




### Description

The MINH command updates the chosen trace with the minimum signal level detected at each trace-data point from subsequent sweeps.

---

## MINPOS Minimum X Position

### Syntax





### Description

The MINPOS command returns the *X* co-ordinate value that corresponds to the minimum amplitude of the specified trace.

## MKA Marker Amplitude

### Syntax



### Description

**8566 and 8568:** Specifies the amplitude of the active marker in dBm when the active marker is the fixed or amplitude type (refer to the MKTYPE command).

**8560 and 8590:** Returns the amplitude of the active marker. For further details, refer to the MA command.

---

**NOTE** The format of the returned data when using the MKA command is dependent on the currently set trace data format (refer to TDF, MDS, O1, O2, O3, or O4) when the selected remote language is either 8566 or 8590. When the selected remote language is 8560, the marker amplitude is always returned as an ASCII value (TDF P).

---

**NOTE** The functions of the MKA command are identical to the MA command ([page 251](#)).


---

**NOTE** In the 8566, 8568, and in the 8590 Series analyzers, the MKA command can be used with a numeric argument that places the marker at the specified amplitude on the screen. For the 8566 and 8568, the MKA command will only accept a value in PSA Series analyzers with firmware Rev. A.08.08 or later. In Option 266 Programming Code Compatibility, MKA is a query only command. That is, the command will not accept arguments of any type, and can only be used in the form MKA?.

---

## MKACT Activate Marker

### Syntax



**Table 4-19**

Item	Description/Default	Range
Marker Number	Any valid integer. Default value is 1.	1 to 4


### Description

MKACT specifies the active marker. There can be four different markers, but only one marker can be active at any time.

---

## MKBW Marker Bandwidth

### Syntax




### Description

Returns the bandwidth at the specified power level relative to an on-screen marker (if present) or the signal peak (if no on-screen marker is present).

---

## MKCF Marker to Center Frequency

### Syntax



### Description

Sets the center frequency equal to the marker frequency and moves the marker to the center of the screen.

---

**NOTE**


The functions of the MKCF command are identical to the E2 command ([page 166](#)).

---



## MKD Marker Delta

### Syntax



Step Increment: by 1/10 of the frequency span

### Description

The MKD command computes the frequency and amplitude difference of the active marker and the delta marker. These values are displayed in the screen.

If a delta marker is not displayed on the screen, the MKD command places one at the specified frequency or on the right hand edge of the display. If an active marker is not displayed on the screen, the MKD command places an active marker at the center of the screen.

#### NOTE

The active marker is the number 1 marker unless otherwise specified by the MKACT command ([page 262](#)).

---

**NOTE**

*All legacy analyzer languages:* If the MKD command is executed with the marker noise function active (MKNOISE ON or KSM), the marker amplitude displayed and returned by the MKA? command ([page 261](#)) or the MA command ([page 251](#)) is the difference between the noise densities at the reference marker and at the delta marker position.

*8566, 8568 and 8590 Series only:* If the MKD command is executed before marker noise has been activated (using the MKNOISE ON or KSM commands), the marker noise amplitude that is displayed on the screen is the difference between the carrier wave power and the noise density at the delta marker position. Regardless of the order in which Marker Noise and Delta Marker are activated, the marker amplitude displayed and returned by the MKA? command ([page 261](#)) or the MA command ([page 251](#)) is the difference between the noise densities at the reference marker and at the delta marker. That is, the value returned by MKA? and MA does not always agree with that displayed on the screen of the PSA Series analyzer. These returned values will only be correct as long as there has been no change in either the delta marker state or the marker noise state from the front panel.

*8560 Series only:* If the MKD command is executed before marker noise has been activated (using the MKNOISE ON or KSM commands), the marker noise amplitude that is displayed on the screen is the difference between the carrier wave power and the noise density at the delta marker position. The value returned by the MKA? or MA command is the difference between the carrier wave power and the noise density at the delta marker position. That is, the value returned by MKA? and MA will agree with that displayed on the screen of the PSA Series analyzer.

---

**NOTE**

The nominal noise bandwidth to RBW ratio for the RBW filters used in the ESA series analyzers is 1.128 for RBWs of 1 kHz and higher, and 1.055 for RBWs of less than 1 kHz.

The nominal ratio for PSA analyzers is 1.055 for all RBWs. The nominal ratio for the 8566 family of analyzers is 1.128 for RBWs 100 kHz and higher, and 1.114 for RBWs of 30kHz and lower.

---


**NOTE**

The functions of the MKD command are identical to the M3 command ([page 248](#)).

---

## MKF Marker Frequency

### Syntax



### Description

Specifies the frequency value of the active marker.

#### NOTE


With the 8560 Series languages, the data is always returned in ASCII format.

With all other languages, the format of the returned data is determined by the TDF (Trace Data Format) ([page 355](#)) command and, if TDF B (binary data format) has been selected, by the MDS command ([page 254](#)).

8566 and 8568 only: If the active marker has marker frequency count set to *On* when using the MKF? command, the marker frequency count value is returned to the controller.

## MKFC Marker Counter

### Syntax



### Description

Turns on or off the marker frequency counter. The resolution of the frequency marker counter is determined by the MKFCR command ([page 269](#)).

---


**NOTE**

The functions of the MKFC command are identical to the MC0 [*zero*] command ([page 252](#)) and MC1 [*one*] command ([page 253](#)).

---

## MKFCR Marker Counter Resolution

### Syntax



**NOTE** The marker counter can be specified in time units when operating in the time domain

### Description

Sets the resolution of the marker frequency counter. The marker counter value is always given either in Hertz or in seconds depending on whether the analyzer is operating in the frequency domain or the time domain.

**NOTE** On PSA analyzers, setting the marker frequency resolution will cause the Gate Time to change. The Gate Time is calculated using the following formula:

$$\text{Gate Time} = \frac{1}{\text{Resolution Value}}$$

For restrictions on the Gate Time, refer to the *PSA User's and Programmer's Reference, Volume 1*.

**NOTE** The functions of the MKFCR command are identical to the KS= command ([page 193](#)).

---

## MKMIN Marker Minimum

### Syntax



### Description

Moves the active marker to the minimum value detected.

---


**NOTE**

The functions of the MKMIN command are identical to the KSN command ([page 218](#)).

---

## MKN Marker Normal

### Syntax



Step Increment: by 1/10 of the frequency span.

### Description


The MKN command moves the active marker to the specified frequency. If the active marker has not been declared with MKACT, a normal marker is turned on and this active marker is assumed to be 1. If the active marker type is not currently normal (for example, it is delta or peak), the MKN command will change it to a normal marker.

#### NOTE

The functions of the MKN command are identical to the M2 [two] command ([page 247](#)).

## MKNOISE Marker Noise

### Syntax



### Description

Displays the average RMS noise density at the marker.

**NOTE** The functions of the MKNOISE command are identical to the KSM command ([page 215](#)) and the KSL command ([page 213](#)).

**NOTE** Some differences in marker noise may be seen between the legacy analyzers and the PSA Series analyzers due to the greater dynamic range of the PSA Series.

**NOTE** *All legacy analyzer languages:* If either the M3 command or the MKD command is executed with the marker noise function active (MKNOISE ON or KSM), the marker amplitude displayed and returned by the MKA? command ([page 261](#)) or the MA command ([page 251](#)) is the difference between the noise densities at the reference marker and at the delta marker position.

*8566, 8568 and 8590 Series only:* If either the M3 command or the MKD command is executed before marker noise has been activated (using the MKNOISE ON or KSM commands), the marker noise amplitude that is displayed on the screen is the difference between the carrier wave power and the noise density at the delta marker position. Regardless of the order in which Marker Noise and Delta Marker are activated, the marker amplitude displayed and returned by the MKA? command ([page 261](#)) or the MA command ([page 251](#)) is the difference between the noise densities at the reference marker and at the delta marker. That is, the value returned by MKA? and MA does not always agree with those displayed on the screen of the PSA Series analyzer. These returned values will only be correct as long as there has been no change in either the delta marker state or the marker noise state from the front



panel.

*8560 Series only:* If either the M3 command or the MKD command is executed before marker noise has been activated (using the MKNOISE ON or KSM commands), the marker noise amplitude that is displayed on the screen is the difference between the carrier wave power and the noise density at the delta marker position. The value returned by the MKA? or MA command is the difference between the carrier wave power and the noise density at the delta marker position. That is, the value returned by MKA? and MA will agree with that displayed on the screen of the PSA Series analyzer.

---

**NOTE**

The nominal noise bandwidth to RBW ratio for the RBW filters used in the ESA series analyzers is 1.128 for RBWs of 1 kHz and higher, and 1.055 for RBWs of less than 1 kHz.


The nominal ratio for PSA analyzers is 1.055 for all RBWs. The nominal ratio for the 8566 family of analyzers is 1.128 for RBWs 100 kHz and higher, and 1.114 for RBWs of 30kHz and lower.

---

---

## MKOFF Marker Off

### Syntax




### Description

Turns off either the active marker or all the markers. If the ALL parameter is omitted, only the active marker is turned off.

---

## MKP Marker Position

### Syntax




### Description

Places the active marker at the specified *X* co-ordinate.

## MKPK Marker Peak

### Syntax



### Description

Executing MKPK HI, or simply MKPK (no secondary keyword), positions the active marker at the highest signal detected. If an active marker is on the screen, the MKPK parameters move the marker as follows:

**HI** (highest) moves the active marker to the highest peak.

**NH** (next highest) moves the active marker to the next signal peak of lower amplitude.

**NR** (next right) moves the active marker to the next signal peak to the right of the current marker.

**NL** (next left) moves the active marker to the next signal peak to the left of the current marker.

---

**NOTE** The functions of the MKPK command (no secondary keyword) and the MKPK HI command are identical to the E1 commands: [\(page 165\)](#).


---

**NOTE** The functions of the MKPK NH command are similar to the KSK command [\(page 211\)](#), except that KSK does not take in to account the marker peak excursion or marker peak threshold values. For more details on marker peak excursion, see the MKPX command [\(page 278\)](#).

---

## MKPT Marker Threshold

### Syntax



### Description

The MKPT command sets the minimum amplitude level from which a peak on the trace can be detected.

#### NOTE

The default values and the range settings on Option 266 Programming Code Compatibility are different than on the legacy analyzers. The following table shows the differences.


**Table 4-20**

**Range Settings and Default Values with the MKPT Command**

Remote Language	Default Setting on Original Analyzer	Default Setting on ESA and PSA analyzers	Valid Range on Legacy Analyzer	Valid Range on ESA and PSA analyzers
8560E/EC Series analyzers	-130 dBm	-100 dBm	-200 dBm to 30 dBm	Ref Level to (Ref Level - (10 × Scale per Division))

## MKPX Marker Peak Excursion

### Syntax



Preset State: 6 dB.

Step Increment: by 1 dB.

### Description


Specifies the minimum signal excursion for the spectrum analyzer's internal peak identification routine.

The default value is 6 dB. In this case, any signal with an excursion of less than 6 dB on either side of the marker would not be identified.

Thus, if an MKPK NH command were to be executed on such a signal, the analyzer would not place a marker on this signal peak.

## MKREAD Marker Readout

### Syntax



### Description

Selects the type of active trace information displayed by the spectrum analyzer marker readout.

The MKREAD command can select the following types of active trace information:

<b>FRQ</b>	frequency
<b>SWT</b>	sweep time
<b>IST</b>	inverse sweep time
<b>PER</b>	period

**NOTE** The Inverse Sweep Time (IST) readout is only available when using a delta marker in zero span.

The results of the data depend on the MKREAD parameter and the frequency span, and whether the marker delta function is used.

**Table 4-21**

MKREAD Type	Non-Zero Span	Non-Zero Span Delta	Zero Span	Zero Span Delta
FRQ	Reads frequency	Reads delta frequency	N/A	N/A
SWT	Reads time since the start of sweep	Reads delta time between end points	Waveform measurements of detected modulation	Waveform measurements of detected modulation

Table 4-21

MKREAD Type	Non-Zero Span	Non-Zero Span Delta	Zero Span	Zero Span Delta
IST	N/A	N/A	N/A	Computes frequency corresponding to delta of markers. Performs $1/(T_1 - T_2)$
PER	Period of frequency	(Pulse measurement) delta time	N/A	N/A

---

**NOTE** FFT (Fast Fourier Transform) is not available in the Programming Code Compatibility Suite option.

---



---

## MKRL Marker to Reference Level

### Syntax



### Description

The MKRL command moves the active marker to the reference level.

---

#### NOTE

The functions of the MKRL command are identical to the E4 command ([page 168](#)).

---

## MKSP Marker to Span

### Syntax



### Description

The MKSP command operates only when the delta marker is On (see MKD ([page 265](#)) or M3 ([page 248](#))). When the delta marker is On and MKSP is executed, the delta marker and active marker determine the start and stop frequencies. The left marker specifies the start frequency, and the right marker specifies the stop frequency. If marker delta is Off, there is no operation.

---

**NOTE** If the active marker is not a delta marker, there is no change in its position.

---

---

**NOTE** The functions of the MKSP command are identical to the KSO command ([page 220](#)).

---

---

## MKSS Marker to Step Size

### Syntax



### Description

Sets the center-frequency step-size equal to the marker frequency. If the analyzer is in the delta mode, the step size is set to the frequency difference between the active and the delta marker.

---


#### NOTE

When the marker is a delta marker, the functions of the MKSS command are identical to the E3 command ([page 167](#)).

---

## MKT Marker Time

### Syntax



### Description


Places a marker at a position that corresponds to a specified point in time during the sweep.

**NOTE** The default unit of time is seconds ('S' or 'SC').

---

## MKTBL Marker Table

### Syntax




### Description

The MKTBL command turns the display of the marker table on or off.

---

## MKTRACE Marker Trace

### Syntax



---

**NOTE**

TRA corresponds to Trace 1, TRB corresponds to Trace 2, and TRC corresponds to Trace 3.

---

### Description

Moves the active marker to the corresponding position in Trace 1, Trace 2, or Trace 3.

---


**NOTE**

If the marker is moved to an inactive trace, the marker will move to the bottom of the screen on an ESA series analyzer, and to the top of the screen on a PSA analyzer.

---

## MKTRACK Marker Track

### Syntax



### Description

Moves the signal on which the active marker is located to the center of the spectrum analyzer display and keeps the signal peak at center screen.


To keep a drifting signal at center screen, place the active marker on the desired signal before turning on MKTRACK.

#### NOTE

The functions of the MKTRACK command are identical to the MT0 [zero] command ([page 291](#)) and the MT1 [one] command ([page 292](#)).

## MKTYPE Marker Type

### Syntax



### Description

Specifies the type of marker.

MKTYPE PSN allows the marker to be positioned horizontally in display units (default). Use the MKP and MKF commands to position the marker.

MKTYPE AMP allows the marker to be positioned according to amplitude. Use the MKA command to position the marker.

MKTYPE FIXED allows a marker to be placed at any fixed point on the display. Use the MKP, MKF, and MKA commands to position the marker.

#### NOTE


Marker type can only be set for an active marker. The marker type is reset to PSN when the marker is turned off (using the MKOFF command) or the instrument is preset.




## ML Mixer Level

### Syntax


#### 8560 Series Remote Language



#### 8566 and 8568 Remote Language



#### 8590 Series Remote Language



### Description

The ML command specifies the maximum signal level that is applied to the input mixer for a signal that is equal to or below the reference level. The effective mixer level is equal to the reference level minus the input

## Programming Commands

**ML Mixer Level**

attenuator setting.

---

**NOTE**

If an external amplifier gain value is set, the mixer level is determined using the following equation:

$$\text{Mixer Level} = \text{Ref. Level} - \text{Attenuation} + \text{Ext. Amplifier Gain}$$

The external amplifier gain is not preset by doing an IP command in case the analyzer is measuring a large signal. This is to protect the analyzer from damage from a large signal. For a helpful suggestion, see [Chapter 3](#), “Hints and Tips,” on page 83.

---

---

**NOTE**

The functions of the ML command are identical to the KS, command ([page 192](#)).

---

---

## MT0 [zero] Marker Track Off

### Syntax



### Description

The MT0 [zero] command disables the marker tracking mode.

---

#### NOTE

The functions of the MT0 [zero] command are identical to the MKTRACK OFF command ([page 287](#)).

---

## MT1 [one] Marker Track On

### Syntax



### Description

Moves the signal on which the active marker is located to the center of the spectrum analyzer display and keeps the signal peak at center screen.

To keep a drifting signal at center screen, place the active marker on the desired signal before issuing an MT1 [one] command.

---


#### NOTE

The functions of the MT1 command are identical to the MKTRACK ON command ([page 287](#)).

---

## MXMH Maximum Hold

### Syntax



---

**NOTE** TRA corresponds to Trace 1 and TRB corresponds to Trace 2.

---

### Description

Updates each trace element with the maximum level detected.

MXMH updates the specified trace (either Trace A or Trace B) with a new value from a detector only if the new value is larger than the previous trace data value.

---


**NOTE** The functions of the MXMH command are identical to the A2 command ([page 92](#)) and B2 command ([page 128](#)).

---

---

## O1 [*one*] Format - Display Units

### Syntax



### Description

The O1 [*one*] command transmits trace amplitude and position information as decimal values in display units.

---

## O2 [two] Format - Two 8-Bit Bytes

### Syntax




### Description

The O2 [two] command transmits trace amplitude and position information as two 8-bit binary numbers, or one instruction word.

## O3 [*three*] Format - Real Amplitude Units

### Syntax



### Description

The O3 [*three*] command transmits trace vertical axis information only, in measurement units of Hz, dBm, dB, volts or seconds.



---

## O4 [*four*] Format - One 8-Bit Byte

### Syntax




### Description

The O4 [*four*] command transmits trace amplitude information only as a binary number.

---

## OCCUP Percent Occupied Power Bandwidth

### Syntax




### Description

The OCCUP command is used to query the current value of the percent occupied power. This value is set by the DELMKBW ([page 157](#)) and the PWRBW command ([page 307](#)) command. The OCCUP command can also be used to set the percent occupied power.

---

## OL Output Learn String

### Syntax



### Description

The OL command transmits information to the controller that describes the state of the analyzer when the OL command is executed. This information is called the “learn string.” The learn string can be sent from the controller memory back to the analyzer to restore the analyzer to its original state.

---

**NOTE**

The OL command is not completely supported, due to differences between the PSA series, ESA series, and 8566/8568. This command is only supported on the PSA series of analyzers with firmware Rev. A.08.02 or later. It is not supported on ESA series analyzers

---

---

## OT Output Trace Annotations

### Syntax



### Description

The OT command sends 32 character-strings to the controller. Each character-string can be up to 64 characters long.

---

**NOTE**


The 'data invalid indicator' status report given in string 27 of the returned text is only supported on PSA.

---


# PEAKS Peaks

## Syntax


### 8560 Series Remote Language



### 8566 and 8568 Remote Language



### 8590 Series Remote Language



Prerequisite Commands: TS when using trace data


## Description

The PEAKS command sorts the signal peaks in the source trace by frequency or amplitude, and sends the sorted results to the destination trace.

---

## PKPOS Peak Position

### Syntax



---

**NOTE**

TRA corresponds to Trace 1, TRB corresponds to Trace 2, and TRC corresponds to Trace 3.

---


### Description

The PKPOS command returns the *X* co-ordinate value of the maximum peak in the specified trace.

# PLOT

## Plot

### Syntax



### Description

The PLOT command allows you transfer trace data, graticule and annotation information to a printer using a parallel port.

#### NOTE


The legacy analyzers transferred data directly to a plotter via the GPIB connection. The PLOT command now transfers data to a printer, and prints the entire screen. For instructions on connecting your analyzer to a printer, see the *PSA User's and Programmer's Reference, Volume 1*.

Although the PLOT command will read in plotter dimension values, these will be ignored.

---

## PP Preselector Peak

### Syntax



### Description

The PP command optimizes preselector tracking to peak the amplitude of a signal at the active marker. If a marker is not on the screen, PP places a marker at the highest signal level, and optimizes preselector tracking at that frequency.

---

#### NOTE

This command is only supported when the analyzer's maximum frequency limit is greater than 3 GHz. If the command is issued on an analyzer with a maximum frequency limit of 3 GHz or less, an error message will be generated stating that the command is not supported.


---



---

## PREAMPG External Preamplifier Gain

### Syntax



### Description

Subtracts a positive or negative preamplifier gain value from the displayed signal. The preamplifier gain is removed by entering a value of 0.

---

**NOTE**

An Instrument Preset (using the IP command [\(page 191\)](#)) does not reset the preamplifier gain to 0.


---

---

## PRINT

### Print

#### Syntax




#### Description

Transfers trace data, graticule and annotation of the analyzer screen directly to a printer via a parallel port.


# PWRBW Power Bandwidth

## Syntax


### 8560 Series Remote Language



### 8566 and 8568 Remote Language



### 8590 Series Remote Language



## Description

Computes the combined power of all signal responses in the specified trace, and returns the bandwidth of the specified percentage of total power. The number in the command is a percentage value, that is, it has a range of 0 to 100.

---

**NOTE** If the percent total power is 100%, the power bandwidth equals the frequency span.

---



---

**NOTE** On the 8566A/B analyzer, this command stops the trace. That is not the case with this Option 266 Programming Code Compatibility.

---

---

## **PWRUPTIME** **Power Up Time**

### **Syntax**

PWRUPTIME → ;


### **Description**

Returns the number of milliseconds that have elapsed since the spectrum analyzer was turned on.

---

## Q0 [zero] EMI Peak Detection

### Syntax



### Description

Sets the detector function to EMI detection. This is the same as Peak detection but uses CISPR related bandwidths.

---

#### NOTE

This command is only supported in PSA series analyzers. ESA series analyzers do not have quasi-peak detection functionality.


DET? will return EPK after execution of the Q0 command.

---

---

## Q1 [*one*] Quasi-Peak Detection

### Syntax



```
Q1 ;
```

### Description

Sets the detector function to Quasi-Peak detection. This is a fast-rise, slow-fall detector used to make CISPR compliant EMI measurements.

---

#### NOTE

This command is only supported in PSA series analyzers. ESA series analyzers do not have quasi-peak detection functionality.


---

DET? will return QPD after execution of the Q1 command.

---

## R1 [one] Illegal Command SRQ

### Syntax




### Description

The R1 [one] command deactivates all analyzer service requests (SRQs) except SRQ140, the illegal-command service request.

---

## R2 [*two*] End-of-Sweep SRQ

### Syntax



### Description

The R2 [*two*] command activates the end-of-sweep and illegal-command service requests.



---

## R3 [*three*] Hardware Broken SRQ

### Syntax



### Description

The R3 [*three*] command activates the hardware-broken and illegal-command service requests.

---

## R4 [four] Units-Key-Pressed SRQ

### Syntax



### Description

The R4 [four] command activate the units-key-pressed and illegal-command SRQs.

---

**NOTE**


PSA analyzers cannot replicate the units-key-pressed service request since no front panel interaction is supported.

---


# RB Resolution Bandwidth

## Syntax


### 8560 Series Remote Language



### 8566 and 8568 Remote Language



Preset State for all languages: 3 MHz, auto coupled.  
 Step Increment for all languages: In a 1, 3, 10 series.

**8590 Series Remote Language**

Preset State for all languages: 3 MHz, auto coupled.

Step Increment for all languages: In a 1, 3, 10 series.

**Description**


The RB command specifies the resolution bandwidth. Available bandwidths are 10 Hz, 30 Hz, 300 Hz, 1 kHz, 3kHz, 30 kHz, 100 kHz, 300 kHz, 1 MHz, and 3 MHz. The resolution bandwidths, video bandwidths, and sweep time are normally coupled. Executing RB decouples them. Execute CR ([page 152](#)) to reestablish coupling.

**NOTE**

Default values on PSA analyzers may vary from the legacy analyzers. Refer to the *PSA User's and Programmer's Reference, Volume 1* to find out any restrictions that may apply.

## RBR Resolution Bandwidth Ratio

### Syntax




### Description

This command is only supported in PSA Series analyzers. It sets the span to resolution bandwidth ratio. Allows you to set the Span/RBW ratio to 1/<value>, where <value> is sent by the user.

---

## RC Recall State

### Syntax



### Description


Recalls analyzer state data from the specified state register in the analyzer's memory.

Registers one through six are reserved for the user, and contain instrument states (such as front panel configuration) saved with the SAVES command ([page 329](#)) or the SV command ([page 348](#)).

- 
- NOTE** Recalling the Preset state (RC 7) is only supported in Option 266 Programming Code Compatibility Suite if **Save Reg 7 (Preset only)** is switched **On** in the Preferences menu ([see page 1-40](#)). Only the Preset state can be recalled. The previous state can not be recalled. With the **Save Reg 7 (Preset only)** turned **Off**, the measurement performance of the analyzer will be faster than when it is turned **On**.
- 
- NOTE** The functions of the RC command are identical to the RCLS command ([page 319](#)).
-

## RCLS Recall State

### Syntax



### Description

Recalls analyzer state data from the specified state register in the analyzer's memory.

Registers one through six are reserved for the user, and contain instrument states (such as front panel configuration) saved with the SAVES command ([page 329](#)) or the SV command ([page 348](#)).

---

**NOTE** The functions of the RCLS command are identical to the RC command ([page 318](#)).

---


**NOTE** Recalling the Preset state (RCLS 7) is only supported in Option 266 Programming Code Compatibility Suite if **Save Reg 7 (Preset only)** is switched **On** in the Preferences menu ([see page 1-40](#)). Only the Preset state can be recalled. The previous state can not be recalled.

---

---

## REV Revision

### Syntax



### Description

The REV command returns the firmware revision number.

---

**NOTE**

In PSA analyzers, this command returns the build date of Option 266 that you have installed in your analyzer. The date is returned in YYWW format where YY is the number of years since 1950, and WW is week number within that year (often referred to as “Work Week”).

As an example, if your Option 266 Programming Code Compatibility firmware was built on May 23 2004, the number 5421 would be returned. The year 2004 is 54 years after 1950, and May 23 is the 21st week of the year, hence 5421.


---




# RL Reference Level

## Syntax

### 8560 Series Remote Language



### 8566, 8568, and 8590 Series Remote Language



## Description

Specifies the amplitude level of the top graticule line on the display. This represents the reference level.

### CAUTION

Signal levels above +30 dBm will damage the spectrum analyzer. For a helpful suggestion on this subject, see [Chapter 3](#), “Hints and Tips,” on [page 83](#).

---

**NOTE**

The 8590 Series of analyzers have a maximum value of 60 dBm for the reference level. The range of reference levels for the PSA Series of analyzers depends on the other settings shown below:

- -170 dBm to +30 dBm with 0 dB reference level offset.
- -160 dBm to +40 dBm with 10 dB reference level offset.
- -180 dBm to +20 dBm with 10 dB external amplifier gain.
- -170 dBm to 0 dBm with preamp on (Option 1DS).

---

**NOTE**

If the display line is on, changing the reference level does not adjust the position of the display line.


---

---

## RMS Root Mean Square Value

Returns the root mean square value of the trace in measurement units.

### Syntax



Prerequisite Commands: TS when using trace data.

---

#### NOTE

TRA corresponds to Trace 1, TRB corresponds to Trace 2, and TRC corresponds to Trace 3.

---


### Description

Returns the RMS value of the trace in display units.


## ROFFSET Reference Level Offset

### Syntax

#### 8560 Series and 8590 Series Remote Language



#### 8566 Remote Language



### Description

Offsets all amplitude readouts without affecting the trace.

Once activated, the ROFFSET command displays the amplitude offset on the left side of the screen.

Entering ROFFSET 0 or presetting the spectrum analyzer eliminates an amplitude offset.


#### NOTE

The functions of the ROFFSET command are identical to the KSZ command ([page 227](#)).

---

## RQS Service Request Mask

### Syntax



### Description

Sets a bit mask for service requests.

---

#### NOTE

Some differences may be noticed in the value returned by the RQS query when compared with the value set. This is because Option 266 on ESA and PSA analyzers does not support the use of bit-1 of the status byte. Bit-1 of the status byte is always set to Off.

---

---

## S1[one] Continuous Sweep

### Syntax



### Description

The S1 command sets the spectrum analyzer to continuous sweep mode. In the continuous sweep mode, the spectrum analyzer takes its next sweep as soon as possible after the current sweep (as long as the trigger conditions are met). A sweep may temporarily be interrupted by data entries made over the remote interface.

---

**NOTE**

The functions of the command S1 are identical to the CONTS command ([page 149](#)).

---

---

## S2 [two] Single Sweep

### Syntax



### Description

The S2 command sets the analyzer to single sweep mode. Each subsequent time that the command S2 is sent, one sweep is started if the trigger conditions are met.

---

**NOTE**

The functions of the S2 command are similar to the SNGLS command ([page 338](#)).

---

---

## SADD Add Limit Line Segment

### Syntax



```
SADD ;
```

### Description


The SADD command is used to add a limit-line segment to the current limit line.



---

## **SAVES** **Save State**

### **Syntax**



### **Description**

Saves the current state of the spectrum analyzer in any of the registers one through six.

---

**NOTE**


The functions of the SAVES command are identical to the SV command ([page 348](#)).

---

---

## SDEL Delete Limit Line Segment

### Syntax



```
SDEL ;
```

### Description

The SDEL command deletes the limit-line segment specified with the SEDI command ([page 332](#)).

---

## SDON Terminate Limit Line Segment

### Syntax



### Description

The SDON command is used to terminate the SEDI command ([page 332](#)).

---

## SEDI Activate Limit Line Segment

### Syntax



### Description

The SEDI command activates the limit-line segment you identify by its segment number in the limit-line table.

---

**NOTE**


The maximum number of limit line points that can be specified on an ESA Series or a PSA Series analyzer is 200.

---

---

## SENDER Segment Entry for Frequency Limit Lines

### Syntax



### Description

The SENDER command is used to create a complete limit-line segment.

---

**NOTE**


Although the parameters SLOPE and FLAT are accepted in the command, they are ignored and have no effect.

---

---

## SER Serial Number

### Syntax




### Description

The SER command returns the analyzer serial number to the controller.

---

## **SETDATE** **Set Date**

### **Syntax**



### **Description**


The SETDATE command sets the date of the real-time clock of the spectrum analyzer. The date takes the form YYMMDD (Year, Month, Day)

---

## SETTIME

### Set Time

#### Syntax




#### Description

The SETTIME command sets the date of the real-time clock of the spectrum analyzer. The time takes the form HHMMSS (Hour, Minute, Second)



## SMOOTH Smooth Trace

### Syntax



Prerequisite Commands: TS when using trace data.

**NOTE** TRA corresponds to Trace 1, TRB corresponds to Trace 2, and TRC corresponds to Trace 3.

### Description

Smooths the trace according to the number of points specified for the running average.

Each point value is replaced with the average of the values (in measurement units) of the given number of points centered on it. Increasing the number of points increases smoothing at the cost of decreasing resolution. If the number of points is an even number, then the number of points is increased by one.

Smoothing decreases at the endpoints.

**NOTE** Some differences may be noticed between the smoothed trace in the legacy analyzers and the smoothed trace using the same signal in PSA analyzers.

---

## SNGLS Single Sweep

### Syntax

```
SNGLS → ;
```

### Description

Sets the spectrum analyzer to single-sweep mode. Each time TS (take sweep) is sent, one sweep taken as long as the trigger conditions are met.

---

**NOTE**


The functions of the SNGLS command are identical to the S2 command ([page 327](#)).

---

# SP Frequency Span


## Syntax

### 8560 Series Remote Language



Step Increment: 1, 2, 5, 10 sequence (up to the stop frequency of the spectrum analyzer)

**8566, 8568, and 8590 Series Remote Language**



Step Increment: 1, 2, 5, 10 sequence (up to the stop frequency of the spectrum analyzer)

**Description**

Changes the total displayed frequency range symmetrically about the center frequency.

If resolution and video bandwidths are coupled to the span width, the bandwidths change with the span width to provide a predetermined level of resolution and noise averaging. Likewise, the sweep time changes to maintain a calibrated display, if coupled. All of these functions are normally coupled, unless RB (page 315), VB (page 370), or ST (page 344) have been executed.


**NOTE**

Option 266 Programming Code Compatibility does not mimic the exact coupling behavior of the legacy analyzers. Refer to your *PSA User's and Programmer's Reference, Volume 1* for the values used.

---

## SRQ User-Defined SRQ

### Syntax



### Description

The SRQ command sends a service request to the controller when the SRQ operand fits the mask supplied with the RQS command.

---

**NOTE**


Option 266 Programming Code Compatibility does not support the setting of bit 1 (units-key-pressed) of the status byte. Bit-1 of the status byte is always set to Off

---


# SS Center Frequency Step Size

## Syntax


### 8560 Series Remote Language



### 8566 and 8568 Remote Language



### 8590 Series Remote Language




### Description

The SS command specifies center frequency step size.


# ST Sweep Time

## Syntax

### 8560 Series Remote Language




### 8566 and 8568 Remote Language





### 8590 Series Remote Language



### Description

The ST command specifies the time in which the analyzer sweeps the displayed frequency or time span.


#### NOTE

The OA option in the ST command behaves in the same manner as the ST? command in that it returns the current value to the controller. However, the OA option does not set the active function to Sweep Time.

---

## STB Status Byte Query

### Syntax




### Description

The STB command returns to the controller the decimal equivalent of the bits set in the status byte (see the RQS [\(page 325\)](#) and SRQ [\(page 341\)](#) commands). STB is equivalent to a serial poll. The RQS and associated bits are cleared in the same way that a serial poll would clear them.

---

## STDEV Standard Deviation of Trace Amplitudes

### Syntax



Prerequisite Commands: TS when using trace data

---

**NOTE**

TRA corresponds to Trace 1, TRB corresponds to Trace 2, and TRC corresponds to Trace 3.

---


### Description

Returns the standard deviation of the trace amplitude in display units.

---

## SV Save State

### Syntax



### Description

Saves the current state of the spectrum analyzer in any of the registers one through six.

---

**NOTE**


The functions of the SV command are identical to the SAVES command ([page 329](#)).

---

---

## T1 [one] Free Run Trigger

### Syntax



### Description

The T1 [one] command sets the analyzer sweep to free run trigger mode.

---

#### NOTE

The functions of the T1 [one] command are identical to the TM FREE command ([page 359](#)).

---

---

## T2 [*two*] Line Trigger

### Syntax



### Description

The T2 [*two*] command sets the analyzer sweep to line trigger mode.

---

**NOTE**

The functions of the T2 [*two*] command are identical to the TM LINE command ([page 359](#)).

---

---

## T3 [*three*] External Trigger

### Syntax



### Description

The T3 [*three*] command sets the analyzer sweep to external trigger mode.

---

#### NOTE

The functions of the T3 [*three*] command are identical to the TM EXT command ([page 359](#)).

---

---

## T4 [*four*] Video Trigger

### Syntax



### Description

The T4 [*four*] command sets the analyzer sweep to video trigger mode.

---

**NOTE**

The functions of the T4 [*four*] command are identical to the TM VID command ([page 359](#)).

---



---

## TA Trace A

### Syntax



### Description

Returns trace A amplitude values from the analyzer to the controller.

The display unit values are transferred in sequential order (from left to right) as seen on the screen. Display unit values can be transferred to the controller in any one of the four output formats as determined by the O1 [*one*] ([page 294](#)), O2 [*two*] ([page 295](#)), O3 [*three*] ([page 296](#)) and O4 [*four*] commands ([page 297](#)). The format of the returned data is also affected by the TDF (Trace Data Format) ([page 355](#)) command and, if TDF B (binary data format) has been selected, by the MDS command ([page 254](#)).

---

## TB Trace B

### Syntax




### Description

Returns trace B amplitude values from the analyzer to the controller.

The display unit values are transferred in sequential order (from left to right) as seen on the screen. Display unit values can be transferred to the controller in any one of the four output formats as determined by the O1 [*one*] ([page 294](#)), O2 [*two*] ([page 295](#)), O3 [*three*] ([page 296](#)) and O4 [*four*] commands ([page 297](#)). The format of the returned data is also affected by the TDF (Trace Data Format) ([page 355](#)) command and, if TDF B (binary data format) has been selected, by the MDS command ([page 254](#)).

## TDF Trace Data Format

### Syntax



### Description



Formats trace information for return to the controller.

The different trace data formats are as follows:

- Specifying M enables the 01 format and returns values in display units, from 0 to 1001.
- Specifying P enables the 03 format and returns absolute measurement values, such as dBm or Hz.
- Specifying A returns data as an A-block data field. The MDS command determines whether data comprises one or two 8-bit bytes. (See MDS [\(page 254\)](#))
- Specifying I returns data as an I-block data field. The MDS command determines whether data comprises one or two 8-bit bytes. (See MDS [\(page 254\)](#))
- Specifying B enables the 02 or 04 format. The MDS command determines whether data comprises one or two 8-bit bytes

# TH Threshold

## Syntax



## Description

The TH command blanks signal responses below the threshold level, similar to a base line clipper. The threshold level is nine major divisions below the reference level, unless otherwise specified. The UP and DN commands move the threshold 10 dB.


**NOTE**

The legacy analyzers all blank the display of everything below the threshold level, but this is not the case with PSA analyzers. Using the PEAKS (page 301) and MKPK (page 276) commands causes any values below the threshold level to be disregarded, even though the full trace will still be displayed. The 8560 Series supported the MKPT command in addition to the TH command. In PSA analyzers, both the TH and the MKPT commands will set the Marker Peak Threshold level; so if both commands are used in a single program, incompatibility issues may be experienced.

---

## **TIMEDATE** **Time Date**

### **Syntax**



### **Description**


Sets the date and time of the real-time clock of the spectrum analyzer. The number takes the form YYMMDDHHMMSS (Year, Month, Day, Hour, Minute, Second).

---

## TITLE

### Title

#### Syntax



#### Description


The TITLE command activates the screen title mode, enabling you to enter your own title for the screen. Valid string delimiters which must be used to start and terminate the title are shown below.

- !
- “
- \$
- %
- &
- ‘
- /
- :
- =
- \
- ~
- @

---

## TM Trigger Mode

### Syntax



### Description

Selects a trigger mode: free, line, video, or external.

---

#### NOTE


The functions of the TM command are identical to the T1 ([page 349](#)), T2 ([page 350](#)), T3 ([page 351](#)) and T4 ([page 352](#)) commands.

---

---

## TRA Trace Data Input and Output

### Syntax



### Description


The TRA command transfers Trace A amplitude values from the analyzer to the controller. The format depends on the trace data format selected. See the TDF command ([page 355](#)) for details on formatting.



---

## TRB Trace Data Input and Output

### Syntax




### Description

The TRB command transfers Trace B amplitude values from the analyzer to the controller. The format depends on the trace data format selected. See the TDF command ([page 355](#)) for details on formatting.

---

## TRC Trace Data Input and Output

### Syntax




### Description

The TRC command transfers Trace Amplitude values from the analyzer to the controller. The format depends on the trace data format selected. See the TDF command ([page 355](#)) for details on formatting.

## TRDSP Trace Display

### Syntax




### Description

The TRDSP command turns the display of the specified trace on or off.

---

## TRIGPOL Trigger Polarity

### Syntax



### Description

Selects the edge (positive or negative) of the trigger input that causes the trigger event. TRIGPOL is available in all trigger modes.

---

## TRPRST Trace Preset

### Syntax

TRPRST → ;

### Description

Sets the trace operations to their preset values.

---

#### NOTE


Option 266 Language Code Compatibility does not do the following stages of TRPRST as they are not supported:

- ANLGPLUS OFF
  - DISPOSE ONEOS
  - DISPOSE ONSWP
  - DISPOSE TRMATH
  - EM
-

---

## TRSTAT Trace State

### Syntax



### Description

The TRSTAT command returns trace states to the controller. Valid trace states are Clear-write, Off, View, Maximum Hold, and Blank.

**Table 4-22** Possible Trace States

Trace State Description	Trace State Data Returned
Clear-write	CLRW
View	VIEW
Blank	BLANK
Off	No data is returned
Maximum Hold	MXMH

---

## TS Take Sweep

### Syntax




### Description

Starts and completes one full sweep before the next command is executed.


A take sweep is required for each sweep in the single-sweep mode. TS prevents further input from the interface bus until the sweep is completed to allow synchronization with other instruments.

## VAVG Video Average

### Syntax



### 8566 and 8568 Remote Language



### Description

Enables the video-averaging function, which averages trace points to smooth the displayed trace. When queried, the VAVG command returns the average length.

---

**NOTE** The functions of the VAVG command are identical to the KSG command ([page 204](#)) and KSH command ([page 206](#)).

---

**NOTE** There are a few differences in the way video averaging works in Option 266 Programming Code Compatibility Suite compared to the legacy spectrum analyzers. See the following table for a summary of these differences.




**Table 4-23 Legacy Analyzers - Video Averaging Behavioral Differences**

<b>Condition</b>	<b>Legacy Spectrum Analyzers</b>	<b>Option 266 - Programming Code Compatibility Suite</b>
All conditions.	<i>8566 and 8568 only</i> - Original trace is displayed in Trace C.	Only displays the averaged trace. The averaged trace is displayed in Trace A.
Average Count value set to 0.	Cannot be set to 0.	Video averaging is turned off if the Averaging Count is set to 0.
Change in Average Count setting to a higher value.	<i>8566 and 8568 only</i> - Continues counting from where the previous value left off.	Resets the counter to zero and starts the measurement again.
Change in average counter setting to a lower value.	<i>8566 and 8568 only</i> - Updates the screen annotation with the lower averaging value.	If the new count value has not been reached, continues until the new lower count has been reached.  If the new, lower count value has already been reached, the analyzer will stop and wait until you take a new sweep.
Averaging turned on.	Sweep time remains unchanged.	Sweep time changes due to the selection of the sample detector.
Change in resolution bandwidth, video bandwidth, sweep time, reference level or attenuation.	<i>8566 and 8568 only</i> - In single sweep mode, resets counter to zero and starts the averaging again.	Continues the measurement without resetting the counter.
Change in center frequency or span.	In single sweep mode, resets counter to zero and starts the averaging again. <i>8566 and 8568 only</i> - Also reset the counter after changes in RBW, VBW, Sweep Time, Ref. Level and Attenuation.	In single sweep mode, resets counter to zero and starts the averaging again. Does not reset the counter after changes in RBW, VBW, Sweep Time, Ref. Level and Attenuation.


# VB Video Bandwidth

## Syntax


### 8560 Series Remote Language



### 8566 and 8568 Remote Language



### 8590 Series Remote Language



### Description

Specifies the video bandwidth, which is a post-detection, low-pass filter.

#### NOTE


Default values on the PSA analyzers may differ from the legacy analyzers. Refer to the *PSA User's and Programmer's Reference, Volume 1* for more details on the restrictions on the video bandwidth range.

When auto coupled, the video bandwidth is calculated as Resolution Bandwidth x Video Resolution Bandwidth Ratio. See the VBO command ([page 372](#)) for more details.

---

## VBO Video Bandwidth Coupling Offset

### Syntax




### Description

The VBO command specifies the relationship between the video and resolution bandwidths which is maintained when these bandwidths are coupled. The bandwidths are usually coupled unless the RB command ([page 315](#)) or VB command ([page 370](#)) have been executed.

- When 0 is selected, the ratio remains fixed at 1. That is, the resolution bandwidth and the video bandwidth are always equal
- When 1 is selected, the video bandwidth is one step higher than the resolution bandwidth. That is, the video bandwidth:resolution bandwidth ratio is three.
- When -1 is selected, the video bandwidth is one step lower than the resolution bandwidth. That is, the video bandwidth:resolution bandwidth ratio is 0.3.

## VBR Video Bandwidth to Resolution Bandwidth Ratio

### Syntax



### Description

The VBR command specifies the relationship between the video and resolution bandwidths that is maintained when these bandwidths are coupled.


#### NOTE

Some differences may be seen between the resolution bandwidth and video bandwidth settings when auto coupled on a PSA Series analyzer.

---

## VIEW View Trace

### Syntax



### Description

Displays Trace A, trace B, or trace C, and stops taking new data into the viewed trace.

---

**NOTE** The functions of the VIEW command are identical to the A3 (page 93), B3 (page 129) and KSj (page 210) commands.


---

**NOTE** TRA corresponds to Trace 1, TRB corresponds to Trace 2, and TRC corresponds to Trace 3.

---

## VTL Video Trigger Level

### Syntax




### Description

The VTL command sets the signal level that triggers a sweep.

## XCH Exchange

### Syntax



### Description

The XCH command exchanges the contents of the source and destination traces. The traces are analyzed and adjusted to fit the number of display points on the screen.

#### NOTE

The functions of the XCH TRA,TRB command are identical to the AXB (page 126) and EX (page 173) commands.

The functions of the XCH TRB,TRC command are identical to the BXC (page 135) and KSi (page 209) commands.



---

**5**

## **A Brief Introduction to the SCPI Language**

---

## SCPI Language Basics

This section is not intended to teach you everything about the SCPI (Standard Commands for Programmable Instruments) programming language. The SCPI Consortium or IEEE can provide that level of detailed information.

Topics covered in this chapter include:

- “Creating Valid Commands” on page 378
- “Command Keywords and Syntax” on page 378
- “Special Characters in Commands” on page 379
- “Parameters in Commands” on page 381
- “Putting Multiple Commands on the Same Line” on page 383

For more information refer to:

IEEE Standard 488.1-1987, *IEEE Standard Digital Interface for Programmable Instrumentation*. New York, NY, 1998.

IEEE Standard 488.2-1987, *IEEE Standard Codes, Formats, Protocols and Comment Commands for Use with ANSI/IEEE Std488.1-1987*. New York, NY, 1998.

### Command Keywords and Syntax

A typical command is made up of keywords set off by colons. The keywords are followed by parameters that can be followed by optional units.

Example: `SENSe:FREQuency:START 1.5 MHZ`

The instrument does not distinguish between upper and lower case letters. In the documentation, upper case letters indicate the short form of the keyword. The lower case letters, indicate the long form of the keyword. Either form may be used in the command.

Example: `Sens:Freq:Star 1.5 mhz`

is the same as `SENSE:FREQ:start 1.5 MHz`

---

#### NOTE

The command `SENS:FREQU:STAR` is not valid because `FREQU` is neither the short, nor the long form of the command. Only the short and long forms of the keywords are allowed in valid commands.

### Creating Valid Commands

Commands are not case sensitive and there are often many different ways of writing a particular command. These are examples of valid

commands for a given command syntax:

Command Syntax	Sample Valid Commands
[SENSe:]BANDwidth[:RESolution] <freq>	<p>The following sample commands are all identical. They will all cause the same result.</p> <ul style="list-style-type: none"> <li>• Sense:Band:Res 1700</li> <li>• BANDWIDTH:RESOLUTION 1.7e3</li> <li>• sens:band 1.7KHZ</li> <li>• SENS:band 1.7E3Hz</li> <li>• band 1.7kHz</li> <li>• bandwidth:RES 1.7e3Hz</li> </ul>
MEASure:SPECTrum [n] ?	<ul style="list-style-type: none"> <li>• MEAS:SPEC?</li> <li>• Meas:spec?</li> <li>• meas:spec3?</li> </ul> <p>The number 3 in the last meas example causes it to return different results than the commands above it. See the command description for more information.</p>
[ :SENSe ] :DETector [ :FUNCTion ] NEGative   POSitive   SAMPLe	<ul style="list-style-type: none"> <li>• DET:FUNC neg</li> <li>• Detector:Func Pos</li> </ul>
INITiate:CONTinuous ON OFF 1 0	<p>The sample commands below are identical.</p> <ul style="list-style-type: none"> <li>• INIT:CONT ON</li> <li>• init:continuous 1</li> </ul>

A Brief Introduction to the SCPI Language

### Special Characters in Commands

Special Character	Meaning	Example
	A vertical stroke between <b>parameters</b> indicates alternative choices. The effect of the command is different depending on which parameter is selected.	<p>Command:            TRIGger:SOURce            EXTernal   INTernal   LINE</p> <p>The choices are external, internal, and line.            Ex: TRIG:SOURCE INT</p> <p>is one possible command choice.</p>

Special Character	Meaning	Example
	A vertical stroke between <b>keywords</b> indicates identical effects exist for both keywords. The command functions the same for either keyword. Only one of these keywords is used at a time.	Command: SENSE: BANDwidth  BWIDth: OFFSet  Two identical commands are: Ex1: SENSE: BWIDth: OFFSET Ex2: SENSE: BAND: OFFSET
[ ]	keywords in square brackets are optional when composing the command. These implied keywords will be executed even if they are omitted.	Command: [SENSE:] BANDwidth[: RESolu tion]: AUTO  The following commands are all valid and have identical effects: Ex1: bandwidth: auto Ex2: band: resolution: auto Ex3: sense: bandwidth: auto
< >	Angle brackets around a word, or words, indicates they are not to be used literally in the command. They represent the needed item.	Command: SENS: FREQ <freq>  In this command example the word <freq> should be replaced by an actual frequency.  Ex: SENS: FREQ 9.7MHz.
{ }	Parameters in braces can optionally be used in the command either not at all, once, or several times.	Command: MEASure: BW <freq>{, level}  A valid command is: meas: BW 6 MHz, 3dB, 60dB

## Parameters in Commands

There are four basic types of parameters: booleans, keywords, variables and arbitrary block program data.

OFF|ON|0|1

(Boolean)

This is a two state boolean-type parameter. The numeric value 0 is equivalent to OFF. Any numeric value other than 0 is equivalent to ON. The numeric values of 0 or 1 are commonly used in the command instead of OFF or ON. Queries of the parameter always return a numeric value of 0 or 1.

keyword

The keywords that are allowed for a particular command are defined in the command syntax description.

Units

Numeric variables may include units. The valid units for a command depend on the variable type being used. See the following variable descriptions. The indicated default units will be used if no units are sent. Units can follow the numerical value with, or without, a space.

Variable

A variable can be entered in exponential format as well as standard numeric format. The appropriate range of the variable and its optional units are defined in the command description.

The following keywords may also be used in commands, but not all commands allow keyword variables.

- DEFault - resets the parameter to its default value.
- UP - increments the parameter.
- DOWN - decrements the parameter.
- MINimum - sets the parameter to the smallest possible value.
- MAXimum - sets the parameter to the largest possible value.

The numeric value for the function's MINimum, MAXimum, or DEFault can be queried by adding the keyword to the command in its query form. The keyword must be entered following the question mark.

Example query: SENSE:FREQ:CENTER? MAX

### Variable Parameters

<integer>

is an integer value with no units.

<real>

Is a floating point number with no units.

<freq>

<bandwidth> Is a positive rational number followed by optional units. The default unit is Hertz. Acceptable units include: Hz, kHz, MHz, GHz.

<time>

<seconds> Is a rational number followed by optional units. The default units are seconds. Acceptable units include: ks, s, ms, us, ns.

<voltage> Is a rational number followed by optional units. The default units are Volts. Acceptable units include: V, mV, uV, nV

<current> Is a rational number followed by optional units. The default units are Amperes. Acceptable units include: A, mA, uA, nA.

<power> Is a rational number followed by optional units. The default units are W. Acceptable units include: mAW, kW, W, mW, uW, nW, pW.

<ampl> Is a rational number followed by optional units. The default units are dBm. Acceptable units include: dBm, dBmV, dBuV.

<rel\_power>

<rel\_ampl> Is a positive rational number followed by optional units. The default units are dB. Acceptable units include: dB.

<percent> Is a rational number between 0 and 100. You can either use no units or use PCT.

<angle>

<degrees> Is a rational number followed by optional units. The default units are degrees. Acceptable units include: DEG, RAD.

<string> Is a series of alpha numeric characters.

<bit\_pattern> Specifies a series of bits rather than a numeric value. The bit series is the binary representation of a numeric value. There are no units.

Bit patterns are most often specified as hexadecimal numbers, though octal, binary or decimal numbers may also be used. In the SCPI language these numbers are specified as:

- Hexadecimal, #Hdddd or #hdddd where 'd' represents a hexadecimal digit 0 to 9 and 'a' to 'f'. So #h14 can be used instead of the decimal number 20.
- Octal, #Odddddd or #oddddddd where 'd' represents an octal digit 0 to 7. So #o24 can be used instead of the decimal number 20.
- Binary, #Bdddddddddddddd or #bdddddddddddddddd where 'd' represents a 1 or 0.

So #b10100 can be used instead of the decimal number 20.

### Block Program Data

Some parameters consist of a block of data. There are a few standard types of block data. Arbitrary blocks of program data can also be used.

<trace> Is an array of rational numbers corresponding to displayed trace data. See FORMat:DATA for information about available data formats.

A SCPI command often refers to a block of current trace data with a variable name such as: Trace1, TRACE2, or trace3, depending on which trace is being accessed.

<arbitrary block data> Consists of a block of data bytes. The first information sent in the block is an ASCII header beginning with #. The block is terminated with a semi-colon. The header can be used to determine how many bytes are in the data block. There are no units. (You will not get block data if your data type is ASCII, using FORMat:DATA ASCII command. Your data will be comma separated ASCII values.

Block data example: suppose the header is #512320.

- The first digit in the header (5) tells you how many additional digits/bytes there are in the header.
- The 12320 means 12 thousand, 3 hundred, 20 data bytes follow the header.
- Divide this number of bytes by your current data format (bytes/data point), either 8 (for real,64), or 4 (for real,32). For this example, if you're using real64 then there are 1540 points in the block.

### Putting Multiple Commands on the Same Line

Multiple commands can be written on the same line, reducing your code space requirement. To do this:

- Commands must be separated with a semicolon (;).
- If the commands are in different subsystems, the key word for the new subsystem must be preceded by a colon (:).
- If the commands are in the same subsystem, the full hierarchy of the command key words need not be included. The second command can start at the same key word level as the command that was just executed.

### SCPI Termination and Separator Syntax

All binary trace and response data is terminated with <NL><END>, as

SCPI Language Basics

defined in Section 8.5 of IEEE Standard 488.2-1992, *IEEE Standard Codes, Formats, Protocols and Common Commands for Use with ANSI/IEEE Std 488.1-1987*. New York, NY, 1992. (Although one intent of SCPI is to be interface independent, <END> is only defined for IEEE 488 operation.)

The following are some examples of good and bad commands. The examples are created from a theoretical instrument with the simple set of commands indicated below:

```

[:SENSe]
  :POWer
    [:RF]
      :ATTenuation 40dB

:TRIGger
  [:SEQuence]
  :EXTernal [1]
  :SLOPe
    POSitive

[:SENSe]
  :FREQuency
    :STARt
  :POWer
  [:RF]
  :MIXer
    :RANGe
  [:UPPer]
    
```

Bad Command	Good Command
PWR:ATT 40dB	POW:ATT 40dB
The short form of POWER is POW, not PWR.	
FREQ:STAR 30MHz;MIX:RANG -20dBm	FREQ:STAR 30MHz;POW:MIX:RANG -20dBm
The MIX:RANG command is in the same :SENSE subsystem as FREQ, but executing the FREQ command puts you back at the SENSE level. You must specify POW to get to the MIX:RANG command.	
FREQ:STAR 30MHz;POW:MIX RANG -20dBm	FREQ:STAR 30MHz;POW:MIX:RANG -20dBm
MIX and RANG require a colon to separate them.	
:POW:ATT 40dB;TRIG:FREQ:STAR 2.3GHz	:POW:ATT 40dB;:FREQ:STAR 2.3GHz
:FREQ:STAR is in the :SENSE subsystem, not the :TRIGGER subsystem.	
:POW:ATT?:FREQ:STAR?	:POW:ATT?;:FREQ:STAR?



Bad Command	Good Command
<b>:POW</b> and <b>FREQ</b> are within the same <b>:SENSE</b> subsystem, but they are two separate commands, so they should be separated with a semicolon, not a colon.	
<b>:POW:ATT -5dB ; :FREQ:STAR 10MHz</b>	<b>:POW:ATT 5dB ; :FREQ:STAR 10MHz</b>
Attenuation cannot be a negative value.	



## Numerics

8-bit bytes, 295, 297

## A

A minus B

into A, 114, 137

off, 136

plus display line, 115

A plus B

to A, 118, 199

A1, 91

A2, 92

A3, 93

A4, 94

absolute amplitude units, 121

AC coupling, 188

AC/DC, 186, 188

ACP, 95

ACPALPHA, 96

ACPALTCH, 97

ACPBRPER, 98

ACPBRWID, 99

ACPBW, 100

ACPCOMPUTE, 101

ACPFREQWT, 102

ACPLOWER, 103

ACPMAX, 104

ACPMEAS, 105

ACPMSTATE, 106

ACPPAR, 107

ACPPWRTX, 108

ACPRSLTS, 109

ACPSP, 110

ACPT, 111

ACPUPPER, 112

activate limit line segment, 332

activate marker, 262

active license key, 36

how to locate, 36

add limit line segment, 328

address

display, 156

adjacent channel power, 95

alpha weighting, 96

alternate channels, 97

auto, 107

bandwidth, 100

burst period, 98

burst width, 99

channel spacing, 110

compute, 101

frequency weighting, 102

lower, 103

manual, 107

maximum, 104

measure, 105

measurement results, 109

measurement state, 106

T weighting, 111

total power transmitted, 108

upper, 112

ADJALL, 113

Agilent Technologies URL, 2

alpha weighting, 96

AMB, 114

AMBPL, 115

amplitude

in dBm, 194

in dBmV, 196

in dBuv, 198

in volts, 200

amplitude marker, 261

amplitude units, 89, 121

analyzer command, 89

angle parameter (variables), 382

ANNOT, 117

annotation, 117

off, 221

on, 222

output trace, 300

APB, 118

arbitrary block data, 383

AT, 119

attenuation, 119

coupling, 138

AUNITS, 121

AUTO, 123

auto couple, 123, 125

AUTOCP, 125

average

video, 368

AXB, 126

## B

B minus display line, 131

B1, 127

B2, 128

B3, 129

B4, 130

bandwidth

channel power, 145

marker, 263

occupied power, 298

power, 307

resolution, 315

video, 370

video coupling offset, 372

bit\_pattern parameter

(variables), 382

BL, 131

BLANK, 132

blank

trace C, 212

blank trace, 94, 130, 132

block data

arbitrary, 383

identifying block size, 383

parsing output, 383

BML, 133

boolean parameter (commands),

381

BTC, 134

BXC, 135

bytes, 295, 297, 346

## C

C1, 136

C2, 137

CA, 138

CAL, 139

calibration, 139

carrier on power, 140

CARRON, 140

center frequency, 141

marker, 166, 264

step size, 342

CF, 141

CHANNEL, 142

channel power, 143, 144

bandwidth, 145

channel selection, 142

CHANPWR, 143

Chapter 6, "A Brief Introduction

to the SCPI Language," on

page 247, 46

character EOI, 89

characters, 88

Choose Option key, 35

CHP, 144

CHPWRBW, 145

CISPR, 309, 310

clear

average, 146

status byte, 148

write, 91, 127, 147

Clear Command Error Log, 45

CLRAVG, 146

CLRW, 147

CLS, 148

Cmd Error Log, 45

command

mnemonic, 88

terminators, 88

Command Err, 45

commands

boolean parameter, 381

keyword parameter, 381

multiple on a line, 383

parameters, 381

syntax, 378

termination, IEEE, 384

- units parameter, 381
- valid commands, 378
- variable parameter, 381
- variable parameter keywords, 381
- configure remote language, 41
- continuous sweep, 149, 326
- CONTS, 149
- correction factors on, 150
- CORREK, 150
- counter
  - marker, 268
- COUPLE, 151
- couple
  - attenuation, 138
  - auto, 123, 125
  - frequency
    - step size, 153
  - input, 151
  - resolution bandwidth, 152
  - sweep time, 154
  - video bandwidth, 155
- coupling
  - AC/DC, 186, 188
  - video bandwidth
    - offset, 372
- CR, 152
- CS, 153
- CT, 154
- current units, 89
- CV, 155
- D**
- DA, 156
- data
  - arbitrary blocks, 383
  - data byte, 89
  - data byte EOI, 89
  - data entry
    - hold, 185
  - data format
    - display units, 294
    - one 8-bit byte, 297
    - real amplitude units, 296
    - trace, 355
    - two 8-bit bytes, 295
  - date, 357
    - set, 335
  - DC coupling, 186
  - degree parameter (variables), 382
  - delay
    - sweep, 163
  - delete
    - current limit line, 238
    - limit line segment, 330
    - limit line table, 234
- deleting an
  - application/personality, 31
- delimit, 89
- DELMKBW, 157
- delta marker, 248, 265
  - occupied power bandwidth, 157
  - step size, 167
- delta value
  - limit line, 232
- DET, 158
- detection, 309, 310
  - mode, 158
  - negative peak, 201
  - normal, 195
  - positive peak, 197
  - sample, 203
- digit, 89
- display
  - address, 156
  - frequency, 176
  - limit line, 235
  - line, 160
  - line enable, 162
  - line off, 228
  - off, 205
  - on, 207
- display trace, 363
- DL, 160
- DLE, 162
- DLYSWP, 163
- documentation, 49
- DONE, 164
- dotted lines
  - optional path, 88
- E**
- E1, 165
- E2, 166
- E3, 167
- E4, 168
- EDITDONE, 169
- EDITLIML, 170
- elapsed time, 172
- EMC detection, 309, 310
- EMI peak detection, 309, 310
- enable
  - display line, 162
  - limit line testing, 240
- end-of-sweep SRQ, 312
- ERR, 171
- error, 171
- errors
  - clearing, 45
  - displaying, 45
  - storing, 45
- ET, 172
- EX, 173
- exchange traces, 376
  - A and B, 126, 173
  - B and C, 135, 209
- excursion
  - marker peak, 278
- extend analyzer reference level, 208
- external
  - preamplifier gain, 305
  - trigger, 225, 351
  - trigger mode, 359
- F**
- FA, 174
- fast preselector peak, 178
- fast preset, 223
- FB, 175
- FDSP, 176
- FOFFSET, 177
- format
  - display units, 294
  - one 8-bit byte, 297
  - real amplitude units, 296
  - trace data, 355
  - two 8-bit bytes, 295
- FPKA, 178
- free run trigger, 349
- free trigger mode, 359
- FREF, 179
- frequency
  - center, 141
  - display off, 176
  - limit line, 233
  - limit line segment entry, 333
  - marker, 267
  - marker readout, 279
  - offset, 177, 224
  - reference, 179
  - segment entry, 333
  - span, 339
  - start, 174
  - stop, 175
  - units, 89
- frequency parameter (variables), 382
- FS, 180
- full span, 180
- G**
- gain
  - external preamplifier, 305
- getting started, 26
- GRAT, 184
- graticule, 184
  - off, 217
  - on, 219

## H

hardware broken SRQ, 313  
hardware requirements, 28  
HD, 185  
hints, 84  
    compatibility, 84  
    instrument presets, 85  
    overloading, 85  
    SCPI language, 85  
    speed, 84  
    sweep times, 84  
    synchronization, 84, 85  
    time out, 84

## hold

data entry, 185  
maximum, 293  
mimimum, 259  
HP8560E/EC remote language, 41  
HP8561E/EC remote language, 41  
HP8562E/EC remote language, 41  
HP8563E/EC remote language, 41  
HP8564E/EC remote language, 41  
HP8565E/EC remote language, 41  
HP8566B remote language, 42  
HP8568B remote language, 42  
HP8590L remote language, 42  
HP8591E remote language, 42  
HP8592L remote language, 42  
HP8593E remote language, 42  
HP8594E remote language, 43  
HP8594EL remote language, 43  
HP8595E remote language, 43  
HP8596E remote language, 43

## I

I1, 186  
I2, 188  
ID, 190  
identify, 190  
IEEE command termination, 384  
IF adjustment, 113  
illegal command SRQ, 311  
impedance  
    units, 89  
input  
    attenuation, 119  
    coupling, 151  
input attenuation, 45  
Install Now key, 35  
Installing and Obtaining a license key, 35

installing measurement personalities, 31  
instrument preset, 191  
integer variable (variables), 381  
IP, 191

## K

keyword parameter (commands), 381  
KS,, 192  
KS=, 193  
KSA, 194  
KSA, 195  
KSB, 196  
KSb, 197  
KSC, 198  
KSc, 199  
KSD, 200  
KSD, 201  
KSE, 202  
KSe, 203  
KSG, 204  
Ksg, 205  
KSH, 206  
KSh, 207  
KSI, 208  
KSi, 209  
KSj, 210  
KSK, 211  
KSk, 212  
KSL, 213  
KSI, 214  
KSM, 215  
KSm, 217  
KSN, 218  
KSn, 219  
KSO, 220  
KSo, 221  
KSp, 222  
KST, 223  
KSV, 224  
KSx, 225  
KSy, 226  
KSZ, 227

## L

L0, 228  
level  
    mixer, 289  
    reference, 321  
    reference offset, 324  
    video trigger, 375  
LG, 230  
license key  
    obtaining and installing, 35  
LIMD, 232  
LIMF, 233

LIMIDEL, 234  
LIMIDESP, 235  
LIMIFAIL, 236  
LIMIFT, 237  
LIMPURGE, 238  
LIMIREL, 239  
limit  
    lower amplitude, 241  
    middle amplitude, 242  
    upper amplitude, 243  
limit line  
    activate segment, 332  
    add segment, 328  
    delete current, 238  
    delete segment, 330  
    delete table, 234  
    delta value, 232  
    display, 235  
    edit, 170  
    enable testing, 240  
    end of edit, 169  
    frequency, 237  
    frequency value, 233  
    relative, 239  
    segment activate, 332  
    segment entry for frequency  
        limit line, 333  
    segment terminate, 331  
    terminate segment, 331  
    testing, 240  
    time, 237  
limitations, 26  
    commands supported, 27  
    predefined functions, 27  
    user-defined functions, 27  
LIMITEST, 240  
limits failed, 236  
LIML, 241  
LIMM, 242  
LIMU, 243  
line trigger, 350  
line trigger mode, 359  
linear scale, 244  
LN, 244  
LO adjustment, 113  
LO and IF adjustment, 113  
loading an  
    application/personality, 31  
logarithmic scale, 230  
lower adjacent channel power, 103  
lower-limit amplitude, 241  
lsb length, 89  
LSPAN, 245

## M

M1, 246

- M2, 247
  - M3, 248
  - M4, 250
  - MA, 251
  - marker
    - activate, 262
    - amplitude, 261
    - amplitude output, 251
    - bandwidth, 263
    - center frequency, 166, 264
    - counter, 268
    - counter resolution, 269
    - delta, 248, 265
    - delta step size, 167
    - frequency, 267
    - frequency counter off, 252
    - frequency counter on, 253
    - frequency output, 258
    - minimum, 218, 270
    - next peak, 211
    - noise, 272
    - noise off, 213
    - noise on, 215
    - normal, 247, 271
    - occupied power bandwidth, 157
    - off, 246, 274
    - peak, 165, 276
    - peak excursion, 278
    - position, 275
    - readout, 279
    - reference level, 168
    - span, 220
    - step size, 167, 283
    - table, 285
    - threshold, 277
    - time, 284
    - to span, 282
    - trace, 286
    - track, 287
    - track off, 291
    - track on, 292
    - type, 288
    - zoom, 250
  - marker amplitude, 261
  - marker amplitude output, 251
  - marker bandwidth, 263
  - marker counter, 268
  - marker counter resolution, 193, 269
  - marker delta, 265
  - marker frequency, 267
  - marker frequency counter off, 252
  - marker frequency counter on, 253
  - marker frequency output, 258
  - marker minimum, 218, 270
  - marker noise, 272
  - marker noise off, 213
  - marker noise on, 215
  - marker normal, 247, 271
  - marker off, 246, 274
  - marker peak, 276
  - marker peak excursion, 278
  - marker position, 275
  - marker readout, 279
  - marker readout in frequency, 279
  - marker span, 220
  - marker step size, 283
  - marker table, 285
  - marker threshold, 277
  - marker time, 284
  - marker to center frequency, 166, 264
  - marker to next peak, 211
  - marker to reference level, 168, 281
  - marker to span, 282
  - marker trace, 286
  - marker track, 287
  - marker track off, 291
  - marker track on, 292
  - marker type, 288
  - max mixer level, 45
  - maximum adjacent channel power, 104
  - maximum hold, 92, 128, 293
  - MC0, 252
  - MC1, 253
  - MDS, 254
  - MEAN, 255
  - mean
    - power measurement, 256
    - trace, 255
  - mean power measurement, 256
  - MEANPWR, 256
  - MEASOFF, 257
  - measurement
    - data size, 254
    - off, 257
  - MF, 258
  - middle-amplitude
    - limit, 242
  - MINH, 259
  - minimum
    - hold, 259
    - marker, 218, 270
    - x position, 260
  - MINPOS, 260
  - missing options, 31
  - mixer level, 192, 289
  - MKA, 261
  - MKACT, 262
  - MKBW, 263
  - MKCF, 264
  - MKD, 265
  - MKF, 267
  - MKFC, 268
  - MKFCR, 269
  - MKMIN, 270
  - MKN, 271
  - MKNOISE, 272
  - MKOFF, 274
  - MKP, 275
  - MKPK, 276
  - MKPT, 277
  - MKPX, 278
  - MKREAD, 279
  - MKRL, 281
  - MKSP, 282
  - MKSS, 283
  - MKT, 284
  - MKTBL, 285
  - MKTRACE, 286
  - MKTRACK, 287
  - MKTYPE, 288
  - ML, 289
  - mnemonic
    - command, 88
  - msb length, 89
  - MT0, 291
  - MT1, 292
  - MXMH, 293
- ## N
- negative peak detection, 201
  - noise
    - marker, 272
    - marker off, 213
    - marker on, 215
    - measurement, 272
    - signal to noise ratio, 272
  - normal
    - detection, 195
    - marker, 247, 271
  - number, 89
- ## O
- O1, 294
  - O2, 295
  - O3, 296
  - O4, 297
  - OCCUP, 298
  - occupied power bandwidth, 298
  - occupied power bandwidth within
    - delta marker, 157
  - offset
    - frequency, 177, 224
    - reference level, 227, 324
    - video bandwidth
      - coupling, 372
  - OL, 299
  - Q1, 310

- options
    - loading/deleting, 31
    - options not in instrument
      - memory, 31
  - OT, 300
  - output data, identifying block size, 383
  - output learn string, 299
  - output termination, 89
  - output trace annotations, 300
- P**
- parameter (variables), 381
  - parameters (commands), 381
  - parameters, variable, 381
  - peak
    - excursion marker, 278
    - fast preselector, 178
    - marker, 165, 276
    - negative peak detection, 201
    - position, 302
    - preselector, 304
  - PEAKS, 301
  - percent occupied power
    - bandwidth, 298
  - percent parameter (variables), 382
  - personality options not in instrument, 31
  - phase parameter (variables), 382
  - PKPOS, 302
  - PLOT, 303
  - polarity
    - trigger, 364
  - position
    - minimum x, 260
  - positive peak detection, 197
  - power
    - bandwidth, 307
    - carrier on, 140
    - channel, 143, 144
    - percent occupied bandwidth, 298
    - up time, 308
  - power measurement
    - mean, 256
  - power parameter (variables), 382
  - power up time, 308
  - PP, 304
  - PREAMPG, 305
  - preamplifier
    - external gain, 305
  - preselector peak, 178, 304
  - preset
    - fast, 223
    - trace, 365
  - preset instrument, 191
  - previous span, 245
  - PRINT, 306
  - print, 306
  - programming
    - command parameters, 381
    - command syntax, 378
    - SCPI basics, 378
    - valid commands, 378
  - PWRBW, 307
  - PWRUPTIME, 308
- Q**
- quasi-peak detection, 309, 310
  - query
    - status byte, 346
- R**
- R1, 311
  - R2, 312
  - R3, 313
  - R4, 314
  - ratio
    - VBW/RBW, 373
  - RB, 152, 315
  - RBR, 317
  - RBW to span ratio, 317
  - RBW/VBW, 45
  - RC, 318
  - readout
    - marker, 279
  - recall last state, 318
  - recommended path, 88
  - reference
    - frequency, 179
  - reference level, 321
    - marker, 281
  - reference level marker, 168
  - reference level offset, 227, 324
  - register 7
    - saving, 45
  - relative limit lines, 239
  - relative power parameter (variables), 382
  - remote language
    - configuring, 41
    - HP8560E/EC, 41
    - HP8561E/EC, 41
    - HP8562E/EC, 41
    - HP8563E/EC, 41
    - HP8564E/EC, 41
    - HP8565E/EC, 41
    - HP8566B, 42
    - HP8568B, 42
    - HP8590L, 42
    - HP8591E, 42
    - HP8592L, 42
    - HP8593E, 42
    - HP8594E, 43
    - HP8594L, 43
    - HP8595E, 43
    - HP8596E, 43
    - SCPI, 41, 46
  - repeating syntax element, 88
  - reserved words, 88
  - resolution
    - marker counter, 193, 269
  - resolution bandwidth, 315, 317
    - coupling, 152
    - vide bandwidth ratio, 373
  - resolution bandwidth ratio, 317
  - results data, identifying block size, 383
  - returning or storing trace values, 360, 361, 362
  - REV, 320
  - revision, 320
  - RL, 321
  - RMS, 323
  - ROFFSET, 324
  - root mean square value, 323
  - RQS, 325
- S**
- S1, 326
  - S2, 327
  - SADD, 328
  - sample detection, 203
  - saving analyzer state, 299
  - saving register 7, 45
  - scale
    - linear, 244
    - logarithmic, 230
  - SCPI language, 41, 46
    - basic info, 378
    - command parameters, 381
    - command syntax, 378
    - keyword parameters, 381
    - valid commands, 378
  - screen title, 358
    - display, 358
  - SDEL, 330
  - SDON, 331
  - secondary keywords, 88
  - SEDI, 332
  - segment entry for frequency limit lines, 333
  - select frequency line, 237
  - select time limit line, 237
  - selection
    - channel, 142
  - SENDER, 333
  - SER, 334
  - serial number, 334
  - service request mask, 325

- set
    - date, 335
    - RF coupling to AC, 188
    - RF coupling to DC, 186
    - time, 336
  - SETDATE, 335
  - SETTIME, 336
  - setting the marker counter
    - resolution, 193
  - shipment
    - verification list, 49
  - signal-to-noise ratio, 272
  - single sweep, 327, 338
  - SMOOTH, 337
  - smooth trace, 337
  - SNGLS, 338
  - softkeys
    - Atten Offset, 45
    - Clear Command Error Log, 45
    - Cmd Error Log, 45
    - Command Err, 45
    - Config Remote Lang, 41
    - HP8560E/EC, 41
    - HP8561E/EC, 41
    - HP8562E/EC, 41
    - HP8563E/EC, 41
    - HP8564E/EC, 41
    - HP8565E/EC, 41
    - HP8566B, 42
    - HP8568B, 42
    - HP8590L, 42
    - HP8591E, 42
    - HP8592L, 42
    - HP8593E, 42
    - HP8594E, 43
    - HP8594L, 43
    - HP8595E, 43
    - HP8596E, 43
    - Language, 41
    - Limit RBW/VBW, 45
    - Save Reg 7, 45
    - SCPI, 41
  - SP, 339
  - span, 317, 339
    - frequency, 339
    - full, 180
    - marker, 220, 282
    - previous, 245
  - span zoom, 250
  - special, 88
    - numbers, 88
  - SPEEd | DRANge | DYNamicrange, 46
  - SRQ, 341
    - end-of-sweep, 312
    - hardware broken, 313
    - illegal command, 311
    - units-key-pressed, 314
    - user-defined, 341
  - SS, 342
  - ST, 344
  - standard deviation of trace
    - amplitudes, 347
  - start frequency, 174
  - state
    - recall, 318
    - trace, 366
  - status byte
    - clear, 148
    - query, 346
  - status byte query, 346
  - STB, 346
  - STDEV, 347
  - step size
    - center frequency, 342
    - marker, 283
  - stop frequency, 175
  - string parameter (variables), 382
  - sweep
    - continuous, 149, 326
    - delay, 163
    - single, 327, 338
    - take, 367
    - time, 344
  - sweep time
    - coupling, 154
  - syntax elements, 88
  - syntax for commands, 88
- T**
- T weighting
    - adjacent channel power, 111
  - T1, 349
  - T2, 350
  - T3, 351
  - T4, 352
  - table
    - marker, 285
  - take sweep, 367
  - TDF, 355
  - terminate limit line segment, 331
  - terminators
    - command, 88
  - TH, 356
  - threshold, 356
    - marker, 277
  - time, 357
    - elapsed, 172
    - marker, 284
    - set, 336
    - units, 89
  - time date, 357
  - time parameter (variables), 382
  - TIMEDATE, 357
  - tips, 84
    - compatibility, 84
    - instrument presets, 85
    - overloading, 85
    - SCPI language, 85
    - speed, 84
    - sweep times, 84
    - synchronization, 84, 85
    - time out, 84
  - TITLE, 358
  - title mode, 202
  - title, 358
  - TM, 359
  - TRA, 360
  - trace
    - A plus B to A, 199
    - blank, 94, 132
    - blank trace C, 212
    - data input, 360, 361, 362
    - data output, 360, 361, 362
    - display, 363
    - exchange, 376
    - exchange B and C, 209
    - mean, 255
    - output annotations, 300
    - preset, 365
    - returning values, 360, 361, 362
    - smooth, 337
    - standard deviation of
      - amplitudes, 347
    - state, 366
    - storing values, 360, 361, 362
    - transfer B to C, 214
    - view, 93, 129
    - view trace C, 210
  - trace B minus display line, 133
  - trace data format, 355, 383
  - trace data input, 360, 361, 362
  - trace data input and output, 360, 361, 362
  - trace data output, 360, 361, 362
  - trace marker, 286
  - trace mean, 255
  - trace preset, 365
  - track marker, 287
  - transfer traces
    - B to C, 134, 214
  - TRB, 361
  - TRC, 362
  - TRDSP, 363
  - trigger
    - external, 225, 351, 359
    - free, 359
    - free run, 349
    - line, 350, 359
    - mode, 359
    - polarity, 364



- TV, 359
- video, 226, 352, 359
- video level, 375
- triggering the spectrum analyzer, 359
- TRIGPOL, 364
- TRPRST, 365
- TRSTAT, 366
- TS, 367
- TV trigger mode, 359
- type marker, 288
  
- U**
- Uninstall Now, 36
- uninstalling measurement personalities, 31
- units, 89
- units parameter (commands), 381
- units-key-pressed SRQ, 314
- upper adjacent channel power, 112
- upper-limit amplitude, 243
- URL
  - Agilent Technologies, 2
  - firmware, 49
  - spectrum analyzer updates, 49
- user-defined SRQ, 341
  
- V**
- variable parameter (commands), 381
- variables
  - angle parameter, 382
  - bit\_data parameter, 382
  - degree parameter, 382
  - frequency parameter, 382
  - integer parameter, 381
  - parameters, 381
  - percent parameter, 382
  - phase parameter, 382
  - power parameter, 382
  - relative power parameter, 382
  - string parameter, 382
  - time parameter, 382
  - voltage parameter, 382
- VAVG, 368
- VB, 370
- VBO, 372
- VBR, 373
- VBW/RBW ratio, 373
- video average, 368
- video averaging
  - off, 206
  - on, 204
- video bandwidth, 370
  - coupling, 155
  - coupling offset, 372
  - resolution bandwidth ratio, 373
- video bandwidth to resolution bandwidth ratio, 373
- video trigger, 226, 352
  - level, 375
- video trigger level, 375
- video trigger mode, 359
- VIEW, 374
- view
  - mode, 93, 129
  - trace, 374
  - trace C, 210
  - view trace, 93, 129
  - view trace, trace
    - view, 374
  - voltage parameter (variables), 382
- VTL, 375
  
- W**
- website
  - firmware updates, 49
  
- X**
- XCH, 376
  
- Z**
- Q0, 309
- zoom marker, 250

