

INFORMATION ONLY – DOES NOT COMMUNICATE  
A MODIFICATION OR SAFETY CONDITION

**E4448A-09**

**S E R V I C E N O T E**

---

Supersedes:  
None

PSA Series Spectrum Analyzer E4440A, E4443A, E4445A,  
E4446A, E4448A

Serial Numbers: [0000A00000 / 9999Z99999]

**Please run the Frequency Response Adjustment (All Bands) after replacing the  
A10 Third Converter Assembly.**

**Parts Required:**

P/N	Description	Qty.
NONE		

**ADMINISTRATIVE INFORMATION**

SERVICE NOTE CLASSIFICATION: <b>INFORMATION ONLY</b>
AUTHOR: BNN      PRODUCT LINE: 12
ADDITIONAL INFORMATION:

© AGILENT TECHNOLOGIES, INC. 2004  
PRINTED IN U.S.A.



April 13, 2004

**Situation:**

The PSA may fail frequency the frequency response test after replacing the A10 Third Converter assembly.

**Solution/Action:**

Please run the Frequency Response Adjustment(All Bands) after replacing the A10 third converter.