11807F-01A

11807F Radio Test Software

Serial Numbers: 0000A00000 / 9999Z99999

Y2K update information.

 ${\tt HP}$ 11807 itself is Y2K compliant to ${\tt HP}$ standards and internally using 4-digit year data.

If platform OS, firmware and the hardware returns correct date and year data, 11807 works correctly. Possible malfunctions caused by OS, firmware and hardware which is not Y2K compliant are explained below.

Duplicate Service Notes: 11807C-03A

To Be Performed By: Customer

Situation:

HP 11807 run with following environment has defects related to the year 2000, if the platform OS, the firmware and the hardware is not Y2K compliant. Three types of defect are possible.

1. Incorrect TEST start date and time.

Environment: BASIC/WS, HT BASIC for DOS

When: BASIC/WS: When cycling the controller power after 0:00

of MAR/01/2000

HT BASIC for DOS:

After the time when the PC used with HP 11807

doesn't return correct date and time.

2. Incorrect TEST execution time.

Environment: IBASIC, HT BASIC for DOS

When: Executing the test at midnight DEC/31/1999 thru

JAN/01/2000.

3. Measurement never ends, hangs up.

Environment: IBASIC, HT BASIC for DOS

When: Executing the test at midnight DEC/31/1999 thru

JAN/01/2000.

OS, firmware, hardware environment which does not meet following conditions will cause the above malfunctions with HP 11807.

IBASIC HP 8920A: Firmware revisions A.18.03 or above are Y2K

compliant to HP standards.

IBASIC HP 8920B: Firmware revisions B.05.05 or above are Y2K

compliant to HP standards.

BASIC/WS: The controller must be HP 9000/362 or 382.

And BASIC OS revision 6.3 or above are Y2K

compliant to HP standards.

HT BASIC for DOS:

Depends on BIOS and DOS of the PC used with. Ask your vendor. In case of HP Vectra PC, see

the HP web site:

http://www.hp.com/vectra/year2000

Solution / Action:

Maintain the platform OS, the firmware, and the hardware in Y2K compliant.

Date	e: 01	Mar	1999										
***	****	****	****	*****	****	****	****	****	****	****	****	****	**
*													
*				IN	FORMA'	TION	ONLY						
*													
*													
*	AUTH	OR/EN	TITY:	SA/	3335								
			INFO:		_								
*			_										

For IBASIC users. See the 8920 service notes, 8920A-14 or 8920B-07 as well.