



Manual Changes

Manual Identification

Model Number: **OmniBER 719**
Date Printed: **9/00**
Part Number: **37719-90079**

Verification Manual

Manual Change Identification

Part Number: **37719-90083**



37719-90083

This supplement contains important information for correcting manual errors and for adapting the manual to instruments containing improvements made after the printing of the manual.

To use this supplement:

Make all ERRATA corrections.

Make all appropriate serial number related changes indicated in the table below.

Serial Prefix or Number	Make Manual Changes	Serial Prefix or Number	Make Manual Changes

*New Item

Note: Manual Change supplements are revised as often as necessary to keep manuals as current and accurate as possible. Agilent Technologies recommend that you periodically request the latest edition of the supplement. Free copies are available from all Agilent offices. When requesting copies, quote the manual identification information from your supplement or the model number and print date from the title page of the manual.

Date: 19 June 2001

Page: 1 of 2

Printed in UK

ERRATA

Page 1-11:

Modify the “**Electrostatic Discharge**” paragraph as follows:

Change: IEC 61000-4-3:1995 to IEC 61000-4-2:1995

***Page 3-2, Figure 3-1 Recommended Test Equipment:**

Table row ‘Optical Coupler’, table column ‘Recommended model’

Replace: ‘HP15477C’ with ‘15744C’

Page 3-118, Step 9:

Replace: “Check that the Synthesizer waveform on the Oscilloscope is between 364 and 546 mV pk_pk.”

with

“Check that the Synthesizer waveform on the Oscilloscope is between **60 and 100 mV pk_pk.**”