

## S E R V I C E N O T E

SUPERSEDES: None

**37724A SDH/PDH Test Set**

Serial Numbers: 0000U00000 / 9999U99999

**Performance Tests update****Calibration manual Part Numbers Required:**

<b>Part No.</b>	<b>INST Serial Number</b>
37724-90000	0000U00000 / 3403U00700*
37724-90019	3404U00700 / 9999U99999

\* If an instrument has been upgraded to Opt 005 then use 37724-90019.

**Situation:**

In April 1994 the 37724A capability was enhanced by adding Structured PDH (Opt 005). A new version of the Calibration Manual, which contains Performance Tests, was generated for instrument 3403U00700 / 9999U99999.

**Action:**

Stock both Calibration Manuals to cover instrument range.

DATE: May 1995

## ADMINISTRATIVE INFORMATION

SERVICE NOTE CLASSIFICATION:

**INFORMATION ONLY**

AUTHOR:	ENTITY:	ADDITIONAL INFORMATION:
ER	1400	