

**S E R V I C E N O T E**

**Agilent 5DX Systems  
Series 2L Performance Enhancement**

**Serial Numbers: USXXXX300/USXXXX466**

**An Improved Panel Clear Sensor is available for Series 2L systems.**

**To Be Performed By: Agilent-Qualified Personnel or Customer**

**Parts Required:**

P/N	Description	Qty.
N7200-68119	Kit, Panel Clear Upgrade	1

**Situation:**

The Panel Clear Sensor (Part Number E7200-61687) will occasionally give false indications when a panel contains cut outs or other unusual features.

**Solution/Action:**

A new Panel Clear Sensor Assembly (P/N N7200-60131) was developed for the Agilent 5DX Series 3 System that corrects these problems and is backward compatible to Series 2L Systems. The Panel Clear Sensor Replacement Kit, Part Number N7200-68119, is available as a customer installable upgrade for Series 2L 5DX Systems. The upgrade procedure is included in the kit.

**ADMINISTRATIVE INFORMATION**

SERVICE NOTE CLASSIFICATION:		
<b>MODIFICATION AVAILABLE</b>		
ACTION CATEGORY:	AGREEABLE TIME	X PERFORMANCE ENHANCEMENT [ ] SERVICE / RELIABILITY ENHANCEMENT
LOCATION CATEGORY:	X CUSTOMER INSTALLABLE [ ] ON-SITE [ ] SERVICE CENTER	AVAILABLE UNTIL: End of Support
AUTHOR: BRG    PRODUCT LINE: 5DX		
ADDITIONAL INFORMATION:		

© AGILENT TECHNOLOGIES, INC. 2001  
PRINTED IN U.S.A.



May 1, 2002