

Integra Series Firmware Information

The cart below summarizes the minimum firmware (FW) level required by the Integra (27xx) mainframe for use with various switching modules.

	2700	2750	2701
Initial FW released	A01	A01	A02
Current FW in release	B06	A08	A06
7700	A01	A01	A02
7701	B04	A03	A02
7702	A01	A01	A02
7703	A01	A01	A02
7705	A01	A01	A02
7706	A02	A01	A02
7707	B03	A01	A02
7708	B02	A01	A02
7709	B03	A01	A02
7710	B05	A04	A05
7711	B04	A03	A02
7712	B04	A03	A02

NOTE: Do not attempt a field upgrade from A0x to any Bxx level as it will render your instrument inoperative. To revise a model 2700 mainframe with Axx firmware to Bxx or higher, a hardware upgrade is required. Consult your local Keithley service center for instructions and pricing for this 2700 mainframe upgrade.

Last update: February 19, 2004