

This barrel adapter is terminated at both ends with 3-slot male triax connectors. Either end will connect to a standard 3-lug female triax connector.

When this adapter is connected to the Model 6430 Remote PreAmp, input/output LO is present on the outer shell of the adapter. Therefore, if a voltage is present on input/output LO, it will also be present on the shell of the triax adapter.

WARNING **To prevent injury from electric shock, DO NOT touch the adapter and triax cables while the instrument is turned on or is connected to an external source that is turned on.**



