

DECLARATION OF CONFORMITY
according to ISO/IEC Guide 22 and EN 45014:1998

Manufacturer's Name LeCroy Corporation
Manufacturer's Address 700 Chestnut Ridge Road
 Chestnut Ridge, NY 10977
 USA

herewith declare that

Product(s) Name: Current probe
Model Number(s): CP150

are in conformity with the provisions of the following EC directive(s), including the latest amendments,
and with national legislation implementing these directives:

73/23/EEC Low Voltage Directive
89/336/EEC EMC Directive


and that conformity with Council Directive 73/23/EEC is based on

EN 61010-1:1993	Safety requirements for electrical equipment for measurement control and laboratory use Part 1: General requirements
EN 61010-2-031:1994	Part 2-031: Particular requirements for hand-held probe assemblies for electrical measurement and test
EN 61010-2-032:1995	Part 2-032: Particular requirements for hand-held current clamps for electrical measurement and test

and that conformity with Council Directive 89/336/EEC is based on

EN 61326-1:1997+A1:1998	EMC requirements for electrical equipment for measurement control and laboratory use
Emissions:	EN 55011:1998 Class B Radiated Emissions
Immunity:	EN 61000-3-2:1995+A1:1998+A2:1998 Electrostatic Discharge EN 61000-3-3:1995 RF Radiated Electromagnetic Field

European Contact:



By: R. Kalyandrug
 Product Compliance manager
Place: LeCroy Corporation
 Chestnut Ridge, NY 10977

Your local LeCroy Sales Office or
LeCroy Europe GmbH
Waldhofer Str 104
D-69123 Heidelberg
Germany
Tel: (49) 6221 82700
Fax: (49) 6221 834655

Date: August 28, 2000