



100-Series
Resistance Standards
Instruction Manual



Copyright 2006 Ohm-Labs, Inc.

100-Series Resistance Standards Instruction Manual

1) General

Ohm-Labs' 100-Series Resistance Standards are designed as transportable or laboratory references for maintaining the ohm at levels between one ohm and ten megohms. Based on modern techniques of processing and construction, each standard is individually wound from selected resistance alloy and is carefully processed for low temperature coefficients of resistance and good long term stability. Hermetically sealed, these standards are highly immune from changes in barometric pressure and relative humidity. All models are supplied with a traceable report of calibration, including temperature coefficient data. This series is recommended for use at an ambient temperature near 23°C.

The 100-Series are offered in both decade and intermediate values.

2) Specifications

Model Number	Nominal Resistance	Tolerance in ppm	Rated Current	Typical Coefficients	Initial 12 mo. Stability
100	1 Ohm	<5	100 mA	Temperature $\alpha < 0.2 \text{ ppm} / ^\circ\text{C}$ $\beta < 0.5 \text{ ppm} / ^\circ\text{C}$	<5 ppm
101	10	<5	30		< 5
102	100	<5	10		< 5
103	1000	<5	3		< 5
104	10 K	<5	1	Voltage < 0.1 ppm / V Pressure < 0.1 ppm / kPa	< 5
105	100 K	<5	0.3		< 5
106	1 Meg	<5	0.1		< 5
107	10 Meg	<5	0.03		< 5
For non-decade values, specify: (1=100 series)+(multiplier)+(range)					
Examples	11.90	1.9	<5	Temperature, voltage & pressure coefficients, and stability, similar to above.	
	12.51	25	<5		
	142	400	<5		

Notes:

Tolerance is accuracy at time of manufacture

Temperature coefficient is at nominal 23°C +/-5°C.

Physical:

176 mm x 80 mm x 92 (6.9" x 3.15" x 3.625"); 1 kg (2 #)

Options available:

Installed 10K thermistor with mating 1 meter cable

Transit container, foam lined for protection during shipment (holds 4 standards)

Environmental Limits:

0-40°C, 0-95% RH, protect from shock or excessive vibration

3) Use

On receipt, inspect the standard for physical damage. If damaged, please immediately contact the carrier. We will assist with any damage claims and/or necessary repair.

Review the Report of Calibration accompanying the standard. The reported value is at 23 °C.

These standards must be used as four-terminal resistors to realize their stated accuracy. Make current and potential connections via the binding posts on the top of the standard. The standard is screened to identify these terminals.

Connection may be made with bare copper wire, silver plated solid copper wire, gold plated copper spade lugs or gold plated 4mm banana plugs. Wire may be passed through the hole in the binding post or wrapped around the post. Do not over tighten the binding posts; a snug finger tight pressure is adequate.

Allow 24 hours for the standard to acclimatize at ambient temperature (23 °C nominal).

For best measurement accuracy, do not exceed the current ratings of the standard. Although application of up to three times the rated current will not damage these standards, self-heating will change the resistance from the reported value.

Caution: Application of current in excess of three times the rated value may permanently shift the resistance of these standards.

4) Measured Value and Temperature Coefficients of Resistance

Each standard's Report of Calibration includes its measured value at 23 °C and its alpha and beta temperature coefficients of resistance. Barring damage, the temperature coefficients of resistance will not change over the life of the standard and do not need to be re-determined.

The alpha (α) coefficient is the change in resistance with temperature at 23 °C; the beta (β) coefficient is the curvature of this change. Within a temperature range around ambient (18 ° - 30 °C), the resistance of a standard may be accurately expressed as:

$$R_t = R_{23}[1 + \alpha(t-23) + \beta(t-23)^2]$$

Where:

R_t	=	Resistance at temperature 't'
R_{23}	=	Resistance at 23°C
t	=	Temperature of resistor

Each resistor is supplied with a table of corrections in parts per million (or micro-ohms per ohm) of the value at 23°C. As a visual aid, this data is also presented in a graph.

5) Maintenance and Repair

No maintenance is required. Repairs must be performed by the manufacturer.

6) Calibration

Periodically recertify the resistance of the standard at its recommended temperature. The calibration cycle will depend on the user's needs. We recommend annual calibration. Please return the standard to us at the below address (or send it to another qualified laboratory) for calibration.

7) Storage and Shipment

Do not expose the standard to temperatures above 40 °C. Protect from shock and extreme vibration. Handle as you would any other precision instrument. Shipment during cooler months is recommended.

8) Warrantee

These standards are warranted for five years from the date of shipment. Please see our Terms & Conditions for additional information.